

Bid Document

Bid Details	
Bid End Date/Time	20-07-2020 13:00:00
Bid Opening Date/Time	20-07-2020 13:30:00
Bid Life Cycle (From Publish Date)	90 (Days)
Bid Offer Validity (From End Date)	30 (Days)
Ministry/State Name	Ministry Of Labour And Employment
Department Name	Na
Organisation Name	Employees State Insurance Corporation (esic)
Office Name	Hyderabad
Total Quantity	10
Item Category	endotracheal or tracheotomy sets et tube with tapered shaped cuff with subglottice suction
MSE Exemption for Years of Experience and Turnover	No
Startup Exemption for Years of Experience and Turnover	No
Bid to RA enabled	No
Inspection Required	No

EMD Detail

Required	No
----------	----

ePBG Detail

Required	No
----------	----

Splitting

Bid splitting not applied.

Endotracheal Or Tracheotomy Sets Et Tube With Tapered Shaped Cuff With Subglottice Suction (10 meter)

Technical Specifications

[* As per GeM Category Specification](#)

Specification	Specification Name	Values	Bid Requirement (Allowed Values)
GENERAL	Purpose of ET tube with tapered shaped cuff with subglottice suction	To prevent microaspiration of fluid into the lungs	*
	Sterilized	YES	*
	Method of sterilization	ETO	*
	Shelf life (Years)	5	*
	The product should have shelf life of at least 2/3 rd of total shelf life at the time of supply	YES	*
DIMENSIONS	Internal diameter of ET tube (mm)	7	177.79986
	Over all length of ET tube (mm)	70	70.0
	Cuff resting diameter of ET tube (mm)	28	28.0
MATERIAL	Material of ET tube (Non-toxic Medical grade)	100% PVC	100% PVC
CONSTRUCTION	ET tube shall have tapered Shaped cuff with low pressure technology to reduce micro aspiration.	YES	*
	ET tube shall have hooded tip with murphy eye.	YES	*
	ET tube shall have one way cuff inflation valve and magill curve	yes	*
	ET tube shall enable continuous or intermittent secretion management through separate subglottic drainage	yes	*
	Cuff shape	Pear	Pear
	Pilot baloon	YES	*
	Kink resistant thermosensitive tube	YES	*

MARKING	I.D of the tube clearly marked on the pilot baloon and tube	YES	*
	Intubation depth marking	YES	*
PACKING	Packing	Blister pack	Blister pack
CERTIFICATIONS	Certification	-	*
	Product conformity Certificate is to be provided to the buyer when required	yes	*
	Tissue implantation testing certificate to be provided	yes	*
	Manufacturing Drug License	YES	*
	Drug License No	-	*
	Drug License Date	-	*

* Specifications highlighted in bold are the Golden Parameters.

* Bidders may note that In respect of non-golden Parameters, the specifications 'Values' chosen by Buyer will generally be preferred over 'Bid requirement (allowed Values) by the Buyer.

Additional Specification Documents

Consignees/Reporting Officer and Quantity

S.No.	Consignee/Reporting Officer	Address	Quantity	Delivery Days
1	Sailaja Vani Batchu	500038,Sanathanagar, Hyderabad	10	30

Bid Specific Additional Terms and Conditions

- Nominated Inspection Agency:** On behalf of the Buyer organization, any one of the following Inspection Agency would be conducting inspection of stores before acceptance:
Pre-dispatch Inspection at Seller Premises (applicable only if pre-dispatch inspection clause has been selected in ATC): pre-dispatch inspection at buyers place
Post Receipt Inspection at consignee site before acceptance of stores: post receipt inspection at buyers place

[This Bid is also governed by the General Terms and Conditions](#)

---Thank You---