RA Number: GEM/2020/RA/45097
Dated: 13-07-2020

RA Document

RA Details	
RA Technical End Date/Time	16-07-2020 23:00:00
RA Opening Date/Time	16-07-2020 23:30:00
RA Life Cycle (From Publish Date)	90 (Days)
RA Offer Validity (From End Date)	30 (Days)
Ministry/State Name	Ministry Of Labour And Employment
Department Name	Na
Organisation Name	Employees State Insurance Corporation (esic)
Office Name	Esi-pgimsr And Esic Medical College And Esic Hospital Joka
Total Quantity	80000
Item Category	Disposable Surgical Rubber Gloves - Prepowdered
MSE Exemption for Years of Experience and Turnover	No
Startup Exemption for Years of Experience and Turnover	No
Document required from seller	OEM Authorization Certificate *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer
Inspection Required	No

EMD Detail

Required	No

ePBG Detail

Required	No

Splitting

RA splitting not applied.

Disposable Surgical Rubber Gloves - Prepowdered (80000 pairs)

Technical Specifications

* As per GeM Category Specification

Specification	Specification Name	Values	Bid Requirement (Allowed Values)
GENERAL INFORMATION	Product description	Disposable Surgical Rubber Gloves conforming to IS:13422/1989 (RA 2003) -prepowdered	*
	Types	Type 1	*
	Single use	Yes	*
	Sterilized gloves	Yes	*
STANDARD	Conformity to standard	IS:13422/1992 (Reaffirmed 2003)	*
	Product is ISI Marked	Yes	*
	BIS Licence No (CM/L No)	6105044	*
DIMENSIONS	**Designation of Size of gloves as per Table 1 of IS:13422/1992(RA 2003)**	7	7
	Minimum length (mm)	245	245
	Width (mm)	76+/-6	76+/-6
	Minimum Thickness of Finger, Palm and Cuff (mm)	0.1	*
	The thickness of the cuff when measured using a gauge with a 5mm diameter foot shall not exceed 3mm	Yes	*
MATERIAL	**Material of the gloves**	Natural Rubber Latex Concentrate	Natural Rubber Latex Concentrate
	Surface treatment with lubricant or powder provided to facilitate donning of gloves as per Cl 5 decimel 1 of IS 13422/1992 (RA 2003)	Yes	*
	Lubricant or powder material used for surface treatment is Bio-absorbable, Non-toxic and Non-	Yes	*

	pyrogenic		
	Material composition of lubricant or powder used	-	*
TECHNICAL FEATURES	Cuf of the glove shall be resistant to tear	Yes	*
	Construction of the cuf shall fit closely without tendency to roll back or ruckle while in use	Yes	*
	Compliance to Cl 5 decimel 3 of IS:13422/1992 (RA 2003) in respect of Physical properties	Yes	*
	Compliance to Cl 5 decimel 5 of IS:13422/1992 (RA 2003) in respect of Air Tight test	Yes	*
	Compliance to Cl 5 decimel 4 of IS:13422/1992 (RA 2003) in respect of Sterility test	Yes	*
PACKING & MARKING	Packaging shall be provided to maintain sterility after sterilization during shipping and storage and to permit opening without contamination of the gloves as given in Cl 8 decimel 1 of IS:13422/1992 (RA 2003)	Yes	*
	Marking	As per Cl 8.2 of IS:13422/1992(RA 2003)	*
CERTIFICATIONS & REPORTS	Avaliability of Test report from Government / NABL / ILAC acredited Laboratory to prove conforming to the declared specification	Yes	*
	Test report Number and date	-	*
	Name and address of Test Lab	-	*
	Certification Available	EU-CE&US-FDA	*

	Submission of all the test reports to the buyer along with supplies on demand	Yes	*
SHELF LIFE / WARRANTY	Shelf life/warranty	-	*
	Minimum Shelf life at the time of supply	2/3rd of total shelf life	*

* Specifications highlighted in bold are the Golden Parameters.
* Bidders may note that In respect of non-golden Parameters, the specifications 'Values' chosen by Buyer will generally be preferred over 'Bid requirement (allowed Values) by the Buyer.

Additional Specification Documents

Consignee/Reporting Officer and Quantity

S.No.	Consignee/Reporting Officer	Address	Quantity	Delivery Days	
1	DEBAJIT DEB	700104,ESIC HOSPITAL JOKA, DIAMOND HARBOUR ROAD	80000	15	N/A

Bid Specific Additional Terms and Conditions

1. Bidder's offer is liable to be rejected if they don't upload any of the certificates / documents sought in the Bid document, ATC and Corrigendum if any.
2. IMPORTED PRODUCTS: In case of imported products, OEM or Authorized Seller of OEM should have a registered office in India to provide after sales service support in India. The certificate to this effect should be submitted.

This RA is also governed by the General Terms and Conditions

---Thank You---