

### Bid Document

Bid Details	
Bid End Date/Time	27-07-2020 17:00:00
Bid Opening Date/Time	27-07-2020 17:30:00
Bid Life Cycle (From Publish Date)	90 (Days)
Bid Offer Validity (From End Date)	30 (Days)
Ministry/State Name	Ministry Of Railways
Department Name	Na
Organisation Name	N/a
Office Name	North Western Railway
Total Quantity	3500
Item Category	hypodermic needle-Disposable, Sterile
MSE Exemption for Years of Experience and Turnover	No
Startup Exemption for Years of Experience and Turnover	No
Document required from seller	Certificate (Requested in ATC) *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer
Bid to RA enabled	No
Inspection Required	No

### EMD Detail

Required	No
----------	----

### ePBG Detail

Required	No
----------	----

### Splitting

Bid splitting not applied.

### Hypodermic Needle-Disposable, Sterile ( 3500 pieces )

#### Technical Specifications

[\\* As per GeM Category Specification](#)

Specification	Specification Name	Values	Bid Requirement (Allowed Values)
GENERAL	Material of Needle Tube (As per ISO 9626)	Stainless Steel (Medical Grade)	*
	<b>Needle Sheath</b>	With	With
	<b>Colouring of Sheath</b>	Pigmented	Pigmented, Unpigmented, NA
	Lubrication to Needle Tube	Yes	*
	Sterile Needle	Yes	*
	Method of sterilization	ETO	*
DIMENSIONS	<b>Nominal Length of Needle Tube</b>	25	25.0
	<b>Nominal Outside Diameter of Needle Tube (mm)</b>	0.63	0.63
STANDARDS	Conforming to Indian Standard	IS 10654 : latest	*
	Colour of Sheath (As per ISO 6009)	-	*
	Colour Coding (Accordance with ISO 6009)	-	*
PACKING & MARKING	<b>ISI Marking</b>	No	Yes, No
	CML Number	-	*
	Symbol for do not re-use as per ISO symbol number ISO 7000/1051	Yes	*
	Number of Needle Sealed in Primary Container	100	*
CERTIFICATION & LICENCES	Drug Licence no.	-	*
	Drug Licence date	-	*
	Submission of Test report on Form 39 or from central govt./NABL/ILAC accredited lab to provide conformity to specification at the time of supply	Yes	*

\* Specifications highlighted in bold are the Golden Parameters.

\* Bidders may note that In respect of non-golden Parameters, the specifications 'Values' chosen by Buyer will generally be preferred over 'Bid requirement ( allowed Values) by the Buyer.

#### Additional Specification Documents

#### Consignees/Reporting Officer and Quantity

S.No.	Consignee/Reporting Officer	Address	Quantity	Delivery Days
1	Sanjeev Malik	334001,DRM OFFICE, BIKANER	3500	15

#### Bid Specific Additional Terms and Conditions

- 1.Bidder's offer is liable to be rejected if they don't upload any of the certificates / documents sought in the Bid document, ATC and Corrigendum if any.
- 2.Material Test Certificate Should Be Sent Along with The Supply. The Material Will Be Checked by Buyer's Lab &

- the Results of the Lab will be the Sole Criteria for Acceptance of the Item.
- 3.The bidder is required to upload, along with the bid, all relevant certificates such as BIS licence, type test certificate, approval certificates and other certificates as prescribed in the Product Specification given in the bid document.
  - 4.To be eligible for award of contract, Bidder / OEM must possess following Certificates / Test Reports on the date of bid opening (to be uploaded with bid): Analysis Report from own test House or any NABL approved Lab.
  - 5.Scope of supply (Bid price to include all cost components) : Only supply of Goods

[This Bid is also governed by the General Terms and Conditions](#)

**---Thank You---**