

Bid Document

Bid Details	
Bid End Date/Time	20-07-2020 18:00:00
Bid Opening Date/Time	20-07-2020 18:30:00
Bid Life Cycle (From Publish Date)	90 (Days)
Bid Offer Validity (From End Date)	80 (Days)
Ministry/State Name	Gujarat
Department Name	Health & Family Welfare Department Gujarat
Organisation Name	Gujarat Medical Services Corporation Limited (gmscl)
Office Name	Gmscl
Total Quantity	2000
Item Category	Infrared Medical Thermometer For Measuring Body Temperature
Bidder Turnover (Last 3 Years)	49 Lakh (s)
OEM Average Turnover (Last 3 Years)	49 Lakh (s)
Experience Criteria	1 Year (s)
MSE Exemption for Years of Experience and Turnover	No
Startup Exemption for Years of Experience and Turnover	No
Document required from seller	Experience Criteria,Past Performance,Bidder Turnover,Certificate (Requested in ATC),OEM Authorization Certificate,OEM Annual Turnover *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer
Past Performance	10 %
Bid to RA enabled	Yes
Inspection Required	No

EMD Detail

Advisory Bank	State Bank of India
EMD Percentage(%)	3.00
EMD Amount	295484

ePBG Detail

Advisory Bank	State Bank of India
ePBG Percentage(%)	5.00
Duration of ePBG required (Months).	38

(a). EMD EXEMPTION: The bidder seeking EMD exemption, must submit the valid supporting document for the relevant category as per GeM GTC with the bid. Under MSE category, only manufacturers for goods and Service Providers for Services are eligible for exemption from EMD. Traders are excluded from the purview of this Policy.

(b). EMD & Performance security should be in favour of Beneficiary, wherever it is applicable.

Beneficiary:

Managing Director

Gujarat Medical Services Corporation Limited (GMSCL) , CH Rd, Sector 10B, Sector 17, Gandhinagar, Gujarat 382010
(Managing Director, Gujarat Medical Services Corporation Limited)

Splitting

Bid splitting not applied.

1. Experience Criteria: In respect of the filter applied for experience criteria, the Bidder or its OEM {themselves or through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU / Public Listed Company for number of years as indicated in the bid document before the bid opening date. Copies of relevant contracts to be submitted along with bid in support of having supplied some quantity during each of the year. In case of bunch bids, the category of primary product having highest value should meet this criterion.

2. OEM Turn Over Criteria: The minimum average annual financial turnover of the OEM of the offered product during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the OEM is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.

3. Past Performance: The Bidder or its OEM {themselves or through re-seller(s)} should have supplied same or similar Category Products for 10% of bid quantity, in at least one of the last three years before the bid opening date to any Central / State Govt Organization / PSU / Public Listed Company. Copies of relevant contracts (proving supply of cumulative order quantity in any one year) to be submitted along with bid in support of quantity supplied in the relevant year. In case of bunch bids, the category related to primary product having highest bid value should meet this criterion.

Infrared Medical Thermometer For Measuring Body Temperature (2000 pieces)

Technical Specifications

[* As per GeM Category Specification](#)

Specification	Specification Name	Values	Bid Requirement (Allowed Values)
GENERAL FEATURES	Product Description	Hand Held Digital Infrared medical thermometer for measuring body temperature	*
	Purpose	Infrared thermometer is a quick and easy-to-use method to measure body temperature	*
	Suitable for	All age group	*
PRODUCT INFORMATION	Type of Thermometer	Non Contact Infrared Thermometer	Non Contact Infrared Thermometer
	Body Material of IR Thermometer	-	*
	Color of Device	-	*

Ergonomic Design	Yes	*
Temperature display unit	Degree Celsius & Degree Fahrenheit	Degree Celsius & Degree Fahrenheit
Primary Measurement Mode	Body Temperature Mode	*
Measuring site for body temperature	Ear & Forehead	*
Lowest measurable temperature for body Mode	32°C (89.6°F) or Lower	*
Highest measurable temperature for body Mode	42°C (107.6°F) or Higher	*
Accuracy of measurement for body temperature in °C (±)	0.2	*
Accuracy of measurement for body temperature in °F (±)	0.2	*
Additional measurement modes available	Surface Temperature Mode, Room Temperature mode, NA	*
Lowest measurable temperature for surface mode	0°C (32 °F) or Lower	*
Highest measurable temperature for surface mode	60°C (140 °F) or Higher	*
Accuracy of measurement for surface temperature in °C (±)	0.5°C	*
Accuracy of measurement for surface temperature in °F (±)	0.5°F	*
Lowest measurable temperature for room temperature mode	0°C (32 °F) or Lower	*
Highest measurable temperature for room temperature mode	50°C (122 °F) or Higher	*
Sensor Type	Thermal Sensor	*
Display Resolution	0.1°C/0.1°F	*
Response Time	0.5	*
Measuring distance for non contact type IR thermometer	5 cm or better	*
Number of recent readings that can be stored	32	*
Built-in laser pointer	Yes	*
Laser sighting	Single point laser	*
Laser class	Class I	*
Weight of device with battery	-	*
Dimensions of device (mm x mm x mm)	-	*

DISPLAY	Display Type	LCD Display	*
	Backlit Display	Yes	Yes
	Display size	2	*
FEATURES	Features available	Interchangeable Celsius & Fahrenheit modes,Low battery indicator,Auto power on,Trigger/Button for operation,Auto power off when not in use,Display of previous reading on screen with current reading,Sound indication when the thermometer is ready to take or show the temperature,Audible alarm for fever,Different backlight colours to differentiate between the fever ranges,Silent mode,Auto hold function,Waterproof lens easy to clean and disinfect,Data transfer facility of other devices through USB/Bluetooth/WiFi/RS 232,Limit value for the adjustable alarm as programmable	Interchangeable Celsius & Fahrenheit modes, Low battery indicator, Auto power on, Trigger/Button for operation, Auto power off when not in use, Display of previous reading on screen with current reading, Sound indication when the thermometer is ready to take or show the temperature, Audible alarm for fever, Different backlight colours to differentiate between the fever ranges, Silent mode, Auto hold function, Waterproof lens easy to clean and disinfect, Data transfer facility of other devices through USB/Bluetooth/WiFi/RS 232, Limit value for the adjustable alarm as programmable
POWER SOURCE	Power source	Non Rechargeable Battery	Non Rechargeable Battery
	Battery chemistry	Li-ion	*
	Battery size, if Non Rechargeable Battery	9 Volt Rectangular Battery	*
	Number of batteries required for operation	1	*
	Number of batteries supplied (Inclusive in the scope of supply)	1	*
	Nominal voltage of each battery (V)	-	*
	Battery capacity (mAH)	-	*
ACCESSORIES	Accessories Provided	Carry Bag/case,User Manual in English,Probe cover,Data transfer software	*
ENVIRONMENTAL CONDITIONS	Minimum operating temperature	0	*
	Maximum operating temperature	50	*
	Maximum operating relative humidity (non condensing)	90	*
	Minimum storage temperature	0	*
	Maximum storage temperature	50	*
	Ingress Protection	IP 65	*
PACKING	The supplier/vendor shall ensure proper and adequate packing of IR thermometer	Yes	*

	to prevent damage in transporting, handling and storage		
CERTIFICATIONS & REPORTS	Availability of conformity certificate of IR thermomometer to prove conformity to the declared specifications from OEM	Yes	*
	Availability of calibration certificate for each device from OEM	Yes	*
	Product certifications	US-FDA	*
	Compliance to standards	IEC 61326-2 or Equivalent BIS, IEC 60825-1 or Equivalent BIS, IEC 61010-1 or Equivalent BIS, ROHS	*
	Manufacturer certifications	ISO:9001, ISO:13485	*
	Submission of all necessary certifications to the buyer along with supplies	Yes	*
TRAINING	The supplier/vendor shall provide free of cost training to the personnel in operation if required by the buyer	Yes	*
WARRANTY	Warranty	3	3 Or higher

* Specifications highlighted in bold are the Golden Parameters.

* Bidders may note that In respect of non-golden Parameters, the specifications 'Values' chosen by Buyer will generally be preferred over 'Bid requirement (allowed Values) by the Buyer.

Additional Specification Documents

Consignees/Reporting Officer and Quantity

S.No.	Consignee/Reporting Officer	Address	Quantity	Delivery Days
1	Prajapati Srushti Rushikumar	382010, Gujarat Medical Services Corporation Limited, Block No 14/1, Dr Jivraj Mehta Bhavan, Sector-10	2000	10

Bid Specific Additional Terms and Conditions

1. Bidder Turn Over Criteria: The minimum average annual financial turnover of the bidder during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the bidder is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.
2. Bidders can also submit the EMD with Account Payee Demand Draft in favour of Managing Director, Gujarat Medical Services Corporation limited, payable at Gandhinagar.
Bidder has to upload scanned copy / proof of the DD along with bid and has to ensure delivery of hardcopy to the Buyer within 5 days of Bid End date / Bid Opening date.
3. Bidders can also submit the EMD with Payment online through RTGS / internet banking in Beneficiary name Managing Director, Gujarat Medical Services Corporation limited, Account No. 32638352973 IFSC Code SBIN0008434 Bank Name State Bank of India Branch address NSC Branch, Gandhinagar..
Bidder to indicate bid number and name of bidding entity in the transaction details field at the time of on-line transfer. Bidder has to upload scanned copy / proof of the Online Payment Transfer along with bid.
4. **Upload Manufacturer authorization:** Wherever Authorised Distributors are submitting the bid, Manufacturers

- Authorisation Form (MAF)/Certificate with OEM details such as name, designation, address, e-mail Id and Phone No. required to be furnished along with the bid.
5. Escalation Matrix For Service Support : Bidder/OEM must provide Escalation Matrix of Telephone Numbers for Service Support.
 6. Experience Criteria: The Bidder or its OEM {themselves or through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU / Public Listed Company for 1 year before the bid opening date. Copies of relevant contracts to be submitted along with bid in support of having supplied some quantity during each of the year. In case of bunch bids, the primary product having highest value should meet this criterion.
 7. OPTION CLAUSE: The Purchaser reserves the right to increase or decrease the quantity to be ordered up to 25 percent at the time of placement of contract. The purchaser also reserves the right to increase the ordered quantity by up to 25% of the contracted quantity during the currency of the contract at the contracted rates. Bidders are bound to accept the orders accordingly.
 8. OEM Turn Over Criteria: The minimum average annual financial turnover of the OEM of the offered product during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the OEM is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria. In case of bunch bids, the OEM of CATEGORY RELATED TO primary product having highest bid value should meet this criterion.
 9. Over and above the normal Warranty terms as per GeM GTC, the successful bidder / OEM shall have to provide Comprehensive Warranty during the entire Standard warranty period as per contract. : The comprehensive warranty shall be covering the following scope Free replacement of parts (Upload an undertaking with the bid confirming compliance by the bidder if Bidder is taking onus of this compliance. In case OEM is taking onus of this compliance, OEM undertaking is to be uploaded along with Bidder undertaking)
 10. IMPORTED PRODUCTS: In case of imported products, OEM or Authorized Seller of OEM should have a registered office in India to provide after sales service support in India. The certificate to this effect should be submitted.
 11. Scope of supply (Bid price to include all cost components) : Supply Installation Testing Commissioning of Goods and Training of operators and providing Statutory Clearances required (if any)
 12. Successful Bidder can submit the Performance Security in the form of Account Payee Demand Draft also (besides PBG which is allowed as per GeM GTC). DD should be made in favour of Managing Director, Gujarat Medical Services Corporation limited payable at Gandhinagar. After award of contract, Successful Bidder can upload scanned copy of the DD in place of PBG and has to ensure delivery of hard copy to the original DD to the Buyer within 15 days of award of contract.
 13. Successful Bidder can submit the Performance Security in the form of Payment online through RTGS / internet banking also (besides PBG which is allowed as per GeM GTC). On-line payment shall be in Beneficiary name Managing Director, Gujarat Medical Services Corporation limited Account No. 32638352973 IFSC Code SBIN0008434 Bank Name State Bank of India Branch address NSC Branch, Gandhinagar.. Successful Bidder to indicate Contract number and name of Seller entity in the transaction details field at the time of on-line transfer. Bidder has to upload scanned copy / proof of the Online Payment Transfer in place of PBG within 15 days of award of contract.
 14. Timely Servicing / rectification of defects during warranty period: After having been notified of the defects / service requirement during warranty period, Seller has to complete the required Service / Rectification within 3 days time limit. If the Seller fails to complete service / rectification with defined time limit, a penalty of 0.5% of Unit Price of the product shall be charged as penalty for each week of delay from the seller. Seller can deposit the penalty with the Buyer directly else the Buyer shall have a right to recover all such penalty amount from the Performance Security (PBG). Cumulative Penalty cannot exceed more than 10% of the total contract value after which the Buyer shall have the right to get the service / rectification done from alternate sources at the risk and cost of the Seller besides forfeiture of PBG. Seller shall be liable to re-imburse the cost of such service / rectification to the Buyer.
 15. The bidder is required to upload, along with the bid, all relevant certificates such as BIS licence, type test certificate, approval certificates and other certificates as prescribed in the Product Specification given in the bid document.
 16. Warranty period of the supplied products shall be 3 years from the date of final acceptance of goods or after completion of installation, commissioning & testing of goods (if included in the scope of supply), at consignee location. OEM Warranty certificates must be submitted by Successful Bidder at the time of delivery of Goods. The seller should guarantee the rectification of goods in case of any break down during the guarantee period. Seller should have well established Installation, Commissioning, Training, Troubleshooting and Maintenance Service group in INDIA for attending the after sales service. Details of Service Centres near consignee destinations are to be uploaded along with the bid

[This Bid is also governed by the General Terms and Conditions](#)

---Thank You---