

### Bid Document

Bid Details	
Bid End Date/Time	27-07-2020 14:00:00
Bid Opening Date/Time	27-07-2020 14:30:00
Bid Life Cycle (From Publish Date)	90 (Days)
Bid Offer Validity (From End Date)	75 (Days)
Ministry/State Name	Gujarat
Department Name	Agriculture And Co-operation Department Gujarat
Organisation Name	Directorate Of Animal Husbandry
Office Name	Krishi Bhavan
Total Quantity	460
Item Category	NEEDLES AND SYRINGES DESTROYER
Bidder Turnover (Last 3 Years)	6 Lakh (s)
Experience Criteria	3 Year (s)
MSE Exemption for Years of Experience and Turnover	No
Startup Exemption for Years of Experience and Turnover	No
Document required from seller	Experience Criteria,Past Performance,Bidder Turnover *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer
Past Performance	50 %
Bid to RA enabled	Yes
Inspection Required	Yes
Inspection to be carried out by Buyers own empanelled agency	Yes
Type Of Inspection	Pre Dispatch
Name of the Empanelled Inspection Agency/ Authority	Board of Officers

#### EMD Detail

Required	No
----------	----

#### ePBG Detail

Required	No
----------	----

#### Splitting

Bid splitting not applied.

1. Experience Criteria: In respect of the filter applied for experience criteria, the Bidder or its OEM {themselves or through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU / Public Listed Company for number of years as indicated in the bid document before the bid opening date. Copies of relevant contracts to be submitted along with bid in support of having supplied some quantity during each of the year. In case of bunch bids, the category of primary product having highest value should meet this criterion.

2. Past Performance: The Bidder or its OEM {themselves or through re-seller(s)} should have supplied same or similar Category Products for 50% of bid quantity, in at least one of the last three years before the bid opening date to any Central / State Govt Organization / PSU / Public Listed Company. Copies of relevant contracts (proving supply of cumulative order quantity in any one year) to be submitted along with bid in support of quantity supplied in the relevant year. In case of bunch bids, the category related to primary product having highest bid value should meet this criterion.

## NEEDLES AND SYRINGES DESTROYER ( 460 pieces )

### Technical Specifications

[\\* As per GeM Category Specification](#)

Specification	Specification Name	Values	Bid Requirement (Allowed Values)
GENERAL	Purpose	To destroy the used injection needles by instant electromelting process and to cut syringe tips manually using the appliance	*
	Type	Portable and electrical	Portable and electrical
	Principle of operation	Electro-melting type	Electro-melting type
CONSTRUCTION	Suitable sliding tray shall be provided for collection of residue and destroyed syringe tips	yes	*
	Material of the Housing /enclosure	ABS Plastic	ABS Plastic
	Housing/enclosure shall be moulded type and shock proof	yes	*
	Provision to burn the needles and to cut the syringe tips shall be provided in the unit	yes	*
	Manual cutter provided in the equipment	Made of hardened blade of anti-corrossive SS material	*
	Transformer winding	Copper	Copper
FUNCTIONAL	Number of needles of 1mm dia and 80mm length that can be destroyed in continuous operation of 5 minutes	50	*
	Sizes of injection needles of all kinds which can be destroyed	Dia Ranging from 0.4mm to 1.6mm (26 SWG to 14 SWG) with length 12.5mm to 80mm	Dia Ranging from 0.4mm to 1.6mm (26 SWG to 14 SWG) with length 12.5mm to 80mm, Dia Ranging from 0.4mm to 1.0mm (26 SWG to 18 SWG) with length 12.5mm to 80mm
ELECTRICAL	Power supply	207V to 253V,50Hz AC	207V to 253V,50Hz AC

		supply	supply
	<b>Power ON/OFF switch with indicator shall be provided</b>	YES	YES
	<b>Rated power(Watts)</b>	60	60.0
	<b>Shall be provided with fuse and power cord of min 3m length and earthing point</b>	YES	YES
	Manual of usage instructions	To be supplied by the supplier	*
CERTIFICATIONS	Availability of Certification for the product to ISO-13485-2009 from Govt approve Lab	YES	*
	Comply in general to IS:302Pt-1/1979 to ensure safe and reliable operation in normal use	YES	*
	Occupational health and safety certification from approved bodies as per OHSAS 18001-2007	YES	*
	CE Certification from approved bodies	Yes	*
TESTS & REPORTS	Safety tests shall consist of marking,protection against electric shock,temperature rise ,IR,Leakage current ,HV,Stability and mechanical hazards,mechanical strength,constructional requirements,creepage distance and clearances,resistance to heat,fire and tracking as per IS:302Pt-1/79	NA	*
	Performance tests shall consist of capacity test as per clause 14 above ,power consumption test at maximum load and ability to cut syringe tips	YES	*

\* Specifications highlighted in bold are the Golden Parameters.

\* Bidders may note that In respect of non-golden Parameters, the specifications 'Values' chosen by Buyer will generally be preferred over 'Bid requirement ( allowed Values) by the Buyer.

#### Additional Specification Documents

#### Consignees/Reporting Officer and Quantity

S.No.	Consignee/Reporting Officer	Address	Quantity	Delivery Days
1	Patel Yashkumar	382330,GVK Emergency Management & Research Institute, Gujarat 108 Emergency Management Center, Naroda	460	15

## **Bid Specific Additional Terms and Conditions**

- 1.OPTION CLAUSE: The Purchaser reserves the right to increase or decrease the quantity to be ordered up to 25 percent at the time of placement of contract. The purchaser also reserves the right to increase the ordered quantity by up to 25% of the contracted quantity during the currency of the contract at the contracted rates. Bidders are bound to accept the orders accordingly.
- 2.Scope of supply (Bid price to include all cost components) : Only supply of Goods
- 3.After award of contract – Successful Bidder shall have to get advance sample approved from buyer before bulk manufacturing / starting bulk supplies. Successful Bidder shall submit 2 samples for Buyer's approval, within 5 days of award of contract. Buyer shall, as per contract specifications framework, either approve the advance sample or will provide complete list of modification required in the sample within 5 days of receipt of advance sample. Seller shall be required to ensure supply as per approved sample with modifications as communicated by Buyer. If there is delay from buyer side in approval of advance sample – the delivery period shall be refixed without LD for the period of delay in sample approval. In case, the sample is found to have major deviations / not conforming to the Contract specifications, the buyer at its discretion may call for fresh samples for approval before allowing bulk supplies or may terminate the contract after notifying the deviations to the seller. Unless otherwise provided in the contract, all samples required for test shall be supplied by the contractor free of cost. Where under the contract, the contractor is required to submit an advance sample, any expenses incurred by the contractor on or in connection with the production of stores in bulk, before the sample has been approved unconditionally shall be borne by the Seller and he shall not claim any compensation in the event of such sample being found unacceptable by the Buyer / Consignee.

[This Bid is also governed by the General Terms and Conditions](#)

**---Thank You---**