

Bid Document

Bid Details	
Bid End Date/Time	03-07-2020 16:00:00
Bid Opening Date/Time	2020-07-03 16:30:00
Bid Life Cycle (From Publish Date)	90 (Days)
Bid Offer Validity (From End Date)	30 (Days)
Ministry/State Name	Haryana
Department Name	Medical Education & Research Department Haryana
Organisation Name	Directorate Of Medical Education & Research
Office Name	Karnal
Total Quantity	10
Item Category	Intermittent Pneumatic Compression Device
Bidder Turnover (Last 3 Years)	7 Lakh (s)
OEM Average Turnover (Last 3 Years)	30 Lakh (s)
Experience Criteria	3 Year (s)
MSE Exemption for Years of Experience and Turnover	No
Startup Exemption for Years of Experience and Turnover	No
Document required from seller	Experience Criteria,Past Performance,Bidder Turnover,Certificate (Requested in ATC),OEM Authorization Certificate,OEM Annual Turnover *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer
Past Performance	80 %
Bid to RA enabled	No
Inspection Required	No

EMD Detail

Advisory Bank	State Bank of India
EMD Percentage(%)	2.00
EMD Amount	27000

ePBG Detail

Required	No
----------	----

(a). EMD EXEMPTION: The bidder seeking EMD exemption, must submit the valid supporting document for the relevant category as per GeM GTC with the bid. Under MSE category, only manufacturers for goods and Service Providers for Services are eligible for exemption from EMD. Traders are excluded from the purview of this Policy.

(b). EMD & Performance security should be in favour of Beneficiary, wherever it is applicable.

Beneficiary:

Accounts Officer

Kalpana Chawla Govt. Medical College, Karnal (Haryana) - 132001

(Ashok Kumar)

Splitting

Bid splitting not applied.

1. Experience Criteria: In respect of the filter applied for experience criteria, the Bidder or its OEM {themselves or through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU / Public Listed Company for number of years as indicated in the bid document before the bid opening date. Copies of relevant contracts to be submitted along with bid in support of having supplied some quantity during each of the year. In case of bunch bids, the category of primary product having highest value should meet this criterion.

2. OEM Turn Over Criteria: The minimum average annual financial turnover of the OEM of the offered product during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the OEM is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.

3. Past Performance: The Bidder or its OEM {themselves or through re-seller(s)} should have supplied same or similar Category Products for 80% of bid quantity, in at least one of the last three years before the bid opening date to any Central / State Govt Organization / PSU / Public Listed Company. Copies of relevant contracts (proving supply of cumulative order quantity in any one year) to be submitted along with bid in support of quantity supplied in the relevant year. In case of bunch bids, the category related to primary product having highest bid value should meet this criterion.

Intermittent Pneumatic Compression Device (10 pieces)**Technical Specifications**

* As per GeM Category Specification

Specification	Specification Name	Values	Bid Requirement (Allowed Values)
GENERAL FEATURES	Product Description	Intermittent Pneumatic Compression Device	*
	Clinical Purpose	To improve venous circulation and help prevent blood clots in the deep veins	*
	Portable, lightweight controller unit with handle	Yes	*
	Easy to use and easy to clean	Yes	*
PRODUCT TECHNICAL SPECIFICATIONS	Provision of Sequential, Gradient and circumferential pneumatic compression around the ankle, calf and then the thigh	Yes	*
	Venous refill detection technology	Yes	*
	Pressure adjustable Range	25 to 120 mm Hg	25 to 120 mm Hg
	Cycle Time (seconds)	60	*
	Approx Inflation time (seconds)	35	*
	Approx Deflation time (seconds)	32	*
	Graphical user interface of minimum 3 point 2 inch size in the controller unit	Yes	*
	Touch Screen Display	No	No
	Type of Screen	LCD	LCD
	Separate pressure display of both legs & indicating the Inflated Leg	Yes	*
	Material of controller unit should be compatible with most of hospital grade cleaning agents	Yes	*
	USB port in controller to make software updates easy	Yes	*

	Knobs for each cuff for setting and adjusting the pressure	Yes	*
	Provision of preset audible and visual alarms for leaks and maximum pressure	Yes	*
	Facility of automatic shutdown if pressure exceeds the maximum limit	Yes	*
	Facility of trouble shooting index in the device itself	Yes	*
	Ergonomic Handle and integrated bed mount with built in power cord storage	Yes	*
	Option of single leg usage or both if needed	Yes	*
	Locking feature that prevents accidental or intentional changing or prescribed settings	Yes	*
DIMENSIONS	Dimensions (L x B x H) in mm	-	*
	Item weight (Kg)	3.1	*
ACCESSORIES	Disposable Knee length sleeves of each size (No)	20	20
	Reusable Knee length sleeves of each size (No)	NA	NA
	Disposable Thigh length sleeves of each size (No)	20	20
	Reusable Thigh length sleeves of each size (No)	NA	NA
	Disposable Foot cuffs of each size (No)	10	10
	Reusable Foot cuffs each size (No)	NA	NA
	Suitable power adaptor/charger	Yes	*
	Standard accessories provided	-	*

GARMENTS/ SLEEVES SPECIFICATIONS	Requirement of Inflatable garment/sleeve for end use	Knee length	Knee length
	Type of inflatable garment/sleeve	Disposable	Disposable
	Material of garments should be non-woven and hydrophobic	Yes	*
	Skin friendly inner cotton lining in the garments	Yes	*
	Size of Knee length sleeve	L	L
	Size of Thigh length sleeve	L	L
	Size of Foot cuff	L	L
ELECTRICAL FEATURES	Input Power supply	220+/-10% V, 50/60 Hz, AC Single phase	*
	Power Consumption (Watt)	50	*
	Integrated Battery for power backup for uninterrupted compression	Yes	*
	Type of battery	Li-ion	Li-ion
	Battery backup time (Hrs)	4	4
PACKING MODE	The product is packed individually in a box with all its accessories in such a way that there will be no transit damage during transportation	Yes	*
ENVIRONMENTAL CONSIDERATIONS	Storage conditions	Temperature 5 deg C to 40 deg C, RH not more than 80%	*
	Work or usage environment	Temperature 5 deg C to 40 deg C, RH not more than 80%	*
CERTIFICATIONS & REPORTS	Submission of Test Report from Central Govt/NABL/ILAC accredited Lab to prove the conformity to declared specification at the time of supply	YES	*

	Should have safety certificate from a competent authority viz EU-CE/US-FDA/STQC/ERTL	Yes	*
	Certificate No	-	*
	Certificate Date	-	*
	Product certification	US-FDA	*
	Four digit number of notified body If product is EU-CE certified	-	*
	Product Certificate No	-	*
	Product Certificate Date	-	*
	Certificate issuing authority	-	*
	Certification, performance and safety standards specific to the device	IEC 60601-1:2012	*
	Submission of all the certifications and test reports to the buyer along with supplies on demand	Yes	*
INSTALLATION & TRAINING	Supplier to perform installation, safety and operation checks before handover	Yes	*
	Training of users in operation and basic maintenance shall be provided	Yes	*
WARRANTY & MAINTENANCE	On -site Warranty (Yrs)	3	3
	User technical and maintenance manual detailing complete maintaining schedule with routine maintenance should be provided	Yes	*
	Contact details of manufacturer, supplier and local service agent to be provided	Yes	*

* Specifications highlighted in bold are the Golden Parameters.

* Bidders may note that In respect of non-golden Parameters, the specifications 'Values' chosen by Buyer will generally be preferred over 'Bid requirement (allowed Values) by the Buyer.

Additional Specification Documents

Consignees/Reporting Officer and Quantity

S.No.	Consignee/Reporting Officer	Address	Quantity	Delivery Days
1	Ashok Kumar	132001, Kalpana Chawla Govt. Medical College, Karnal	10	30

Bid Specific Additional Terms and Conditions

1. Scope of supply (Bid price to include all cost components) : Supply Installation Testing and Commissioning of Goods
2. Bidder Turn Over Criteria: The minimum average annual financial turnover of the bidder during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the bidder is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.
3. OEM Turn Over Criteria: The minimum average annual financial turnover of the OEM of the offered product during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the OEM is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria. In case of bunch bids, the OEM of CATEGORY RELATED TO primary product having highest bid value should meet this criterion.
4. IMPORTED PRODUCTS: In case of imported products, OEM or Authorized Seller of OEM should have a registered office in India to provide after sales service support in India. The certificate to this effect should be submitted.
5. Availability of Service Centres: Bidder/OEM must have a Functional Service Centre in the State of each Consignee's Location in case of carry-in warranty. (Not applicable in case of goods having on-site warranty). If service center is not already there at the time of bidding, successful bidder / OEM shall have to establish one within 30 days of award of contract. Payment shall be released only after submission of documentary evidence of having Functional Service Centre.
6. ISO 9001: The bidder must have ISO 9001 certification.
7. Timely Servicing / rectification of defects during warranty period: After having been notified of the defects / service requirement during warranty period, Seller has to complete the required Service / Rectification within 7 days time limit. If the Seller fails to complete service / rectification with defined time limit, a penalty of 0.5% of Unit Price of the product shall be charged as penalty for each week of delay from the seller. Seller can deposit the penalty with the Buyer directly else the Buyer shall have a right to recover all such penalty amount from the Performance Security (PBG). Cumulative Penalty cannot exceed more than 10% of the total contract value after which the Buyer shall have the right to get the service / rectification done from alternate sources at the risk and cost of the Seller besides forfeiture of PBG. Seller shall be liable to re-imburse the cost of such service / rectification to the Buyer.

[This Bid is also governed by the General Terms and Conditions](#)

---Thank You---