

Bid Document

Bid Details	
Bid End Date/Time	13-07-2020 17:00:00
Bid Opening Date/Time	13-07-2020 17:30:00
Bid Life Cycle (From Publish Date)	90 (Days)
Bid Offer Validity (From End Date)	80 (Days)
Ministry/State Name	Pmo
Department Name	Department Of Atomic Energy
Organisation Name	Nuclear Power Corporation Of India Limited
Office Name	Kakrapar Gujarat Site
Total Quantity	10400
Item Category	Blood Glucose Test Strips (Glucostrips) with Glucometer
MSE Exemption for Years of Experience and Turnover	No
Startup Exemption for Years of Experience and Turnover	No
Document required from seller	Certificate (Requested in ATC) *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer
Bid to RA enabled	No
Inspection Required	No

EMD Detail

Required	No
----------	----

ePBG Detail

Required	No
----------	----

Splitting

Bid splitting not applied.

Blood Glucose Test Strips (Glucostrips) With Glucometer (10400 strip)

Technical Specifications

[* As per GeM Category Specification](#)

Specification	Specification Name	Values	Bid Requirement (Allowed Values)
GENERAL FEATURES	Product	Blood Glucose Test Strips (Glucostrips) with Glucometer	*
	Purpose	To Test Blood Glucose Level of an Individual	*
	One compatible glucometer should be provided free of cost on purchase of every	800	800 Or lower
GIUCOSTRIP SPECIFICATIONS	Enzyme used in the glucostrip	Glucose Oxidase	Glucose Oxidase
	Should be able to use fresh capillary whole blood samples	Yes	*
	Shelf life of glucostrips from the date of manufacture	18	18, 24, 30, 36 Or higher
	Glucostrips should have at least 3/4 th of the total shelf life at the time of supply to consignee	Yes	*
	Shelf life of glucostrips after first opening of the strip pack(Months)	6	6 Or higher
	Type of packing of glucostrips	Packed in an Air Tight Container	*
	Number of strips packed in one container	50	*
GLUCOMETER SPECIFICATIONS	Should be a hand held meter	Yes	*
	Minimum Measurement Range	30 to 600 mg/dl	*
	Accuracy	upto ± 20 % error allowed	*
	Test/Reading Time	≤ 10 sec	*

Blood Sample Size	≤ 1.5 micro liters	*
Display Type	LCD	*
Measuring Unit	mg/dl	*
Technology	Electrochemical	*
Reproducibility/Precision	± 5%	*
Free from Manual Coding	Yes	Yes, No
Battery operated	-	*
Battery should be replaceable without using any tools and should be available in the open market	Yes	*
Battert Life (Minimum number of tests)	500	*
Number of extra batteries provided with each glucometer	0	*
Minimum memory capacity of glucometer(tests)	50	*
Data transfer facility with help of USB cable	No	*
Minimum operating temperature range for both strips and glucometer	10 °C to 40 °C	*
Control solution of atleast 20 ml each for checking reliability of strips should be supplied free of cost	Yes	*
Warranty of Glucometer	1	*
Operating Manual of device to be provided	Yes	*
Holding Drug License from the appropriate Drug Authority	For Selling	*
Drung License Number	-	*
Drug License date	-	*
Availability of Test Report from Central Govt/NABL/ILAC accredited Lab to prove the conformity to the	Yes	*

	declared specification		
	Device and strips certified from National Institute of Biologicals (NIB), Govt of India	Yes	*
	Product Certification	ISO 15197 : Latest	*
	Four digit number of notified body If product is EU-CE certified	-	*
	Manufacturer Certification	ISO 13485 : Latest,ISO 9001 : Latest	*
	Submission of copy of licenses, test report and certifications to the buyer at the time of supply	Yes	*

* Specifications highlighted in bold are the Golden Parameters.

* Bidders may note that In respect of non-golden Parameters, the specifications 'Values' chosen by Buyer will generally be preferred over 'Bid requirement (allowed Values) by the Buyer.

Additional Specification Documents

Consignees/Reporting Officer and Quantity

S.No.	Consignee/Reporting Officer	Address	Quantity	Delivery Days
1	Kamleshkumar Gemalsinh Khengar	394651,KAPS PLANT SITE, PO ANUMALA, DISTRICT TAPI, GUJARAT	10400	30

Bid Specific Additional Terms and Conditions

1. Scope of supply (Bid price to include all cost components) : Only supply of Goods
2. Purchase preference to Micro and Small Enterprises (MSEs): Purchase preference will be given to MSEs as defined in Public Procurement Policy for Micro and Small Enterprises (MSEs) Order, 2012 dated 23.03.2012 issued by Ministry of Micro, Small and Medium Enterprises and its subsequent Orders/Notifications issued by concerned Ministry. If the bidder wants to avail the Purchase preference, the bidder must be the manufacturer of the offered product in case of bid for supply of goods. Traders are excluded from the purview of Public Procurement Policy for Micro and Small Enterprises. In respect of bid for Services, the bidder must be the Service provider of the offered Service. Relevant documentary evidence in this regard shall be uploaded along with the bid in respect of the offered product or service. If L-1 is not an MSE and MSE Seller (s) has/have quoted price within L-1+ 15% of margin of purchase preference /price band defined in relevant policy, such Seller shall be given opportunity to match L-1 price and contract will be awarded for percentage of 100% of total value.
3. The bidder is required to upload, along with the bid, all relevant certificates such as BIS licence, type test certificate, approval certificates and other certificates as prescribed in the Product Specification given in the bid document.

[This Bid is also governed by the General Terms and Conditions](#)

---Thank You---