

### Bid Document

Bid Details	
Bid End Date/Time	21-07-2020 19:00:00
Bid Opening Date/Time	21-07-2020 19:30:00
Bid Life Cycle (From Publish Date)	90 (Days)
Bid Offer Validity (From End Date)	30 (Days)
Ministry/State Name	Jammu & Kashmir
Department Name	Health And Family Welfare Department Jammu And Kashmir
Organisation Name	Government Medical College
Office Name	Govt Medical College Baramulla
Total Quantity	22
Item Category	Microscopes - Pathological and Research ( <b>PAC Only</b> )
Bidder Turnover (Last 3 Years)	11 Lakh (s)
OEM Average Turnover (Last 3 Years)	11 Lakh (s)
Experience Criteria	3 Year (s)
MSE Exemption for Years of Experience and Turnover	No
Startup Exemption for Years of Experience and Turnover	No
Document required from seller	Experience Criteria,Past Performance,Bidder Turnover,Certificate (Requested in ATC),OEM Authorization Certificate,OEM Annual Turnover *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer
Past Performance	50 %
Bid to RA enabled	No
Inspection Required	No

### EMD Detail

Required	No
----------	----

### ePBG Detail

Required	No
----------	----

### Splitting

Bid splitting not applied.

1. Experience Criteria: In respect of the filter applied for experience criteria, the Bidder or its OEM {themselves or through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU / Public Listed Company for number of years as indicated in the bid document before the bid opening date. Copies of relevant contracts to be submitted along with bid in support of having supplied some quantity during each of the year. In case of bunch bids, the category of primary product having highest value should meet this criterion.
2. OEM Turn Over Criteria: The minimum average annual financial turnover of the OEM of the offered product during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the OEM is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.
3. Past Performance: The Bidder or its OEM {themselves or through re-seller(s)} should have supplied same or similar Category Products for 50% of bid quantity, in at least one of the last three years before the bid opening date to any Central / State Govt Organization / PSU / Public Listed Company. Copies of relevant contracts (proving supply of cumulative order quantity in any one year) to be submitted along with bid in support of quantity supplied in the relevant year. In case of bunch bids, the category related to primary product having highest bid value should meet this criterion.

### Microscopes - Pathological And Research ( 22 pieces ) ( Under PAC )

Make	olympus
Model	cx23

### Technical Specifications

[\\* As per GeM Category Specification](#)

Specification	Specification Name	Values	Bid Requirement (Allowed Values)
CONFORMITY / TYPE / EYE PIECE - OBJECTIVE DETAILS	<b>Microscope Type</b>	pathological binocular non hinged type with built in light and with light intensity regulator	pathological binocular non hinged type with built in light and with light intensity regulator
	Conformity to Indian Standard	Pathological IS:4381 latest	*
	<b>ISI MARKED</b>	No	No

	C/ML No ( BIS License No)	NA	*
	<b>Eye piece Type</b>	Compensating	Compensating
	Binocular Eye pieces Confirming to the requirements of IS: 8275/1976 ( latest)	yes	*
	<b>Eye Piece with magnification</b>	Set of Two For Binocular 10x	Set of Two For Binocular 10x
	<b>Objective Type</b>	semiapochromatic	semiapochromatic
	<b>Objective Magnification</b>	100X	100X
	<b>Numerical Aperture of Objective</b>	1.25	1.25
	Type of lamp for Illumination	led	*
GENERIC	Plano Concave mirror attachment	NA	*
	Stage	Ractangular	*
	Size of Stage	76MM X 30MM	*
	Coarse and Fine Movement of stage	yes	*
	Co-axial focussing	Yes	*
	Micrometer Arrangement	Yes	*
ACCESSORIES	<b>Extra eyepieces (inclusive in the scope of supply)</b>	NA	NA
	Carrying case (inclusive in the scope of supply)	No	*
WARRANTY	<b>Warranty</b>	5	5 Or higher
CERTIFICATION	Availability of Test Report from Central Govt/NABL/ILAC accredited lab to prove conformity to specification	No	*
	Test Report No and Date	NA	*
	Name and Address of the Lab	NA	*

\* Specifications highlighted in bold are the Golden Parameters.

\* Bidders may note that In respect of non-golden Parameters, the specifications 'Values' chosen by Buyer will generally be preferred over 'Bid requirement ( allowed Values) by the Buyer.

#### Additional Specification Documents

**Consignees/Reporting Officer and Quantity**

S.No.	Consignee/Reporting Officer	Address	Quantity	Delivery Days
1	Abdul Hameed Kuthoo	193101, Govt Medical College Baramulla	22	15

**Bid Specific Additional Terms and Conditions**

1. Availability of Service Centres: Bidder/OEM must have a Functional Service Centre in the State of each Consignee's Location in case of carry-in warranty. (Not applicable in case of goods having on-site warranty). If service center is not already there at the time of bidding, successful bidder / OEM shall have to establish one within 30 days of award of contract. Payment shall be released only after submission of documentary evidence of having Functional Service Centre.
2. After award of contract – Successful Bidder shall have to get advance sample approved from buyer before bulk manufacturing / starting bulk supplies. Successful Bidder shall submit 1 samples for Buyer's approval, within 5 days of award of contract. Buyer shall, as per contract specifications framework, either approve the advance sample or will provide complete list of modification required in the sample within 5 days of receipt of advance sample. Seller shall be required to ensure supply as per approved sample with modifications as communicated by Buyer. If there is delay from buyer side in approval of advance sample – the delivery period shall be refixed without LD for the period of delay in sample approval. In case, the sample is found to have major deviations / not conforming to the Contract specifications, the buyer at its discretion may call for fresh samples for approval before allowing bulk supplies or may terminate the contract after notifying the deviations to the seller. Unless otherwise provided in the contract, all samples required for test shall be supplied by the contractor free of cost. Where under the contract, the contractor is required to submit an advance sample, any expenses incurred by the contractor on or in connection with the production of stores in bulk, before the sample has been approved unconditionally shall be borne by the Seller and he shall not claim any compensation in the event of such sample being found unacceptable by the Buyer / Consignee.
3. Bidder Turn Over Criteria: The minimum average annual financial turnover of the bidder during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the bidder is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.

[This Bid is also governed by the General Terms and Conditions](#)

**---Thank You---**