

Bid Document

Bid Details	
Bid End Date/Time	23-07-2020 17:00:00
Bid Opening Date/Time	23-07-2020 17:30:00
Bid Life Cycle (From Publish Date)	90 (Days)
Bid Offer Validity (From End Date)	65 (Days)
Ministry/State Name	Chandigarh
Department Name	Education Department Chandigarh
Organisation Name	Government Medical College And Hospital
Office Name	Sector 32, Chandigarh
Total Quantity	4000
Item Category	Digital Medical Thermometer
MSE Exemption for Years of Experience and Turnover	No
Startup Exemption for Years of Experience and Turnover	No
Bid to RA enabled	No
Inspection Required	No

EMD Detail

Required	No
----------	----

ePBG Detail

Required	No
----------	----

Splitting

Bid splitting not applied.

Digital Medical Thermometer (4000 pieces)

Technical Specifications

* As per GeM Category Specification

Specification	Specification Name	Values	Bid Requirement (Allowed Values)
GENERAL FEATURES	Product type	Electronic medical thermometer or Clinical digital thermometer	*
	Usage	Kids, infants, children and adults	*
	Clinical Purpose	For oral, rectal and armpit/axilla temperature measurement	*
PRODUCT INFORMATION	Type of display with minimum 3 digits	LED	LED
	Material	Plastic	*
	Type of tip	Flexible	Flexible
	Temperature measuring unit	Degree fahrenheit	*
	Temperature measurement range	89.60F to 107.60F	*
	Accuracy of measurement	+/- 0.2deg F	*
	No glass and no mercury present	Yes	*
	Safe to use and easy to read	Yes	*
	Water proof for ease of cleaning	Yes	*
	Should have clear instructions to use or preventive maintenance	Yes	*
	Pre-calibrated	Yes	*
	Display "Lo" if temperature is less than 32degC (89 decimel 6 deg F)	Yes	*
	Display "Hi" if temperature is more than 42degC (107degF)	Yes	*
	Rapid measurement of temperature for kids, infants, children and adults	Yes	*
	Beep sound when final steady temperature	Yes	*

	arrived during test		
	Auto saving of last reading before next test is performed	Yes	*
	Time for measurement of temperature, (Sec)	60	60.0
	Battery powered with low battery indicator	Yes	*
	Press type power button with beep for indication of start	Yes	*
	Auto off when not in use	For more than 1 minutes	*
	Weight (gm)	10	10
CERTIFICATIONS & REPORTS	Model approval certificate from Department of Legal Metrology (Weights and Measure), Ministry of Consumer Affairs, Government of India	Yes	*
	Certificate Number and Date from Department of Legal Metrology (Weights and Measure), Ministry of Consumer Affairs, Government of India	-	*
	Calibration certificate from concerned weights and measurement controller to be furnished at the time of supplies	Yes	*
	Product certification	EU-CE,US-FDA	*
	Product conformity certificate including Model approval certificate and EU-CE/US-FDA certificate to be furnished to the buyer on demand	Yes	*
WARRANTY AND MAINTENANCE	Warranty (Yrs)	1	1
	Clear instructions to use or preventive maintenance instructions to be provided	Yes	*

- * Specifications highlighted in bold are the Golden Parameters.
- * Bidders may note that In respect of non-golden Parameters, the specifications 'Values' chosen by Buyer will generally be preferred over 'Bid requirement (allowed Values) by the Buyer.

Additional Specification Documents

Consignees/Reporting Officer and Quantity

S.No.	Consignee/Reporting Officer	Address	Quantity	Delivery Days
1	Kul Bhushan	160031,Government Medical College & Hospital Sector 32, Chandigarh, 160031	4000	15

Bid Specific Additional Terms and Conditions

1. Scope of supply (Bid price to include all cost components) : Only supply of Goods

[This Bid is also governed by the General Terms and Conditions](#)

---Thank You---