

Bid Document

Bid Details	
Bid End Date/Time	20-07-2020 14:00:00
Bid Opening Date/Time	20-07-2020 14:30:00
Bid Life Cycle (From Publish Date)	90 (Days)
Bid Offer Validity (From End Date)	75 (Days)
Ministry/State Name	Gujarat
Department Name	Health & Family Welfare Department Gujarat
Organisation Name	Gujarat Medical Services Corporation Limited (gmscl)
Office Name	Gmscl
Total Quantity	90
Item Category	acoustic stethoscope
Experience Criteria	1 Year (s)
MSE Exemption for Years of Experience and Turnover	No
Startup Exemption for Years of Experience and Turnover	No
Document required from seller	Experience Criteria,Past Performance,Certificate (Requested in ATC),OEM Authorization Certificate *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer
Past Performance	10 %
Bid to RA enabled	No
Inspection Required	No

EMD Detail

Required	No
----------	----

ePBG Detail

Required	No
----------	----

Splitting

Bid splitting not applied.

1. Experience Criteria: In respect of the filter applied for experience criteria, the Bidder or its OEM {themselves or through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU / Public Listed Company for number of years as indicated in the bid document before the bid opening date. Copies of relevant contracts to be submitted along with bid in support of having supplied some quantity during each of the year. In case of bunch bids, the category of primary product having highest value should meet this criterion.

2. Past Performance: The Bidder or its OEM {themselves or through re-seller(s)} should have supplied same or similar Category Products for 10% of bid quantity, in at least one of the last three years before the bid opening date to any Central / State Govt Organization / PSU / Public Listed Company. Copies of relevant contracts (proving supply of cumulative order quantity in any one year) to be submitted along with bid in support of quantity supplied in the relevant year. In case of bunch bids, the category related to primary product having highest bid value should meet this criterion.

Acoustic Stethoscope (90 pieces)

Technical Specifications

[* As per GeM Category Specification](#)

Specification	Specification Name	Values	Bid Requirement (Allowed Values)
Warranty Scope	Warranty on Product (Year(s))	1	*
	Free Service Available during warranty	Yes	Yes
CONFIGURATION	Length of Y tubing (cm)	50	50.0 - 200.0 Or higher
	Recommended application	Physical Assessment and Diagnosis,Limited Physical Assessment and Blood Pressure,Cardiology,Physical Assessment, Diagnosis and cardiology	Physical Assessment and Diagnosis, Limited Physical Assessment and Blood Pressure, Cardiology, Physical Assessment, Diagnosis and cardiology
	Diameter of the Binaural tube	5 mm	5 mm
	Diaphragm type	Fixed	Fixed
	Chest piece material	Aluminium	Aluminium
	Material of snap on ring	Plastic	*
	Latex free product	Yes	Yes
	Noise Isolation earpiece	Yes	Yes
	Non chilling snap on	No	No

ring provided on chest piece		
Number of snap on ring provided	1	1
Binaural spring provided	No	No
Weight of the product (g)	500	500
Diameter of Chest piece(mm)	50	50
Number of spare diaphragm provided	1	1, 2 Or higher
Type of Stethoscope	Adult Stethoscopes	Adult Stethoscopes
Binaural tube material	Brass	Brass
Diameter of Chest piece;Diaphragm Side (mm)	25	25
Tunable Diaphragm	No	No
Finish of Binaural tube	Chrome plated over nickle conforming to Grade C of IS:1068-1958	Chrome plated over nickle conforming to Grade C of IS:1068-1958
Outer diameter of Y tubing (mm)	55	55
Type of Chest piece	Bell	Bell
Detectable Sound frequency range (Hz)	10	10
Ear pieces material	Bakelite	Bakelite
Number of ID tag provided	1	1
Material of the Spring	N/A	N/A
Over all dimension;LengthxHeightxWidth (mmxmmxmm)	-	*
Spare snap on ring Provided	No	No
Stethoscope design and manufacturing cofirms to IS 3391-1965 latest amendment	Yes	Yes
Material of Y tubing	Latex free PVC	*
Finish of chest piece	Chrome plated over nickle conforming to	Chrome plated over nickle conforming to

		Grade C of IS:1068-1958	Grade C of IS:1068-1958
	Diaphragm material	Celluloid	Celluloid
	Spare ear piece provided	No	No
	Number of Free Services, if provided	2	2
	Spare ID tag Provided	No	No
	Type of Y tubing	Single lumen	Single lumen
	Test Report Date	-	*
	ID Tag provided	No	No
	Number of spare Ear piece provided a set of two:	1	1
	Spare Diaphragm Provided	No	No
CERTIFICATION	Firm Certified to ISO 9001:2008 or latest	Yes	Yes
	ISI marked	Yes	Yes
	Other Certifications	-	*
	CE marked	Yes	Yes
	WHO-GMP certified	Yes	Yes
	Availability of Test Report from Central Govt./NABL/ILAC accredited lab to prove conformity to specification	No	No
	Test Report Number	-	*
	Name of the Lab	-	*
	Address of the Lab	-	*
PRODUCT DETAILS	Product code available on supplier website	Yes	Yes
	if Yes, Product code	-	*

* Specifications highlighted in bold are the Golden Parameters.

* Bidders may note that In respect of non-golden Parameters, the specifications 'Values' chosen by Buyer will generally be preferred over 'Bid requirement (allowed Values) by the Buyer.

Additional Specification Documents

Consignees/Reporting Officer and Quantity

S.No.	Consignee/Reporting Officer	Address	Quantity	Delivery Days
1	DR.Sureshbhai Punjabhai Rathod	362002,GMERS MEDICAL COLLEGE, NEAR MAJEVADI GATE, PADDOCK ROAD, JUNAGADH.	90	15

Bid Specific Additional Terms and Conditions

1. **OPTION CLAUSE:** The Purchaser reserves the right to increase or decrease the quantity to be ordered up to 25 percent at the time of placement of contract. The purchaser also reserves the right to increase the ordered quantity by up to 25% of the contracted quantity during the currency of the contract at the contracted rates. Bidders are bound to accept the orders accordingly.
2. **Experience Criteria:** The Bidder or its OEM {themselves or through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU / Public Listed Company for 1 year before the bid opening date. Copies of relevant contracts to be submitted along with bid in support of having supplied some quantity during each of the year. In case of bunch bids, the primary product having highest value should meet this criterion.
3. **IMPORTED PRODUCTS:** In case of imported products, OEM or Authorized Seller of OEM should have a registered office in India to provide after sales service support in India. The certificate to this effect should be submitted.
4. **Scope of supply (Bid price to include all cost components) :** Only supply of Goods
5. The bidder is required to upload, along with the bid, all relevant certificates such as BIS licence, type test certificate, approval certificates and other certificates as prescribed in the Product Specification given in the bid document.
6. **Escalation Matrix For Service Support :** Bidder/OEM must provide Escalation Matrix of Telephone Numbers for Service Support.
7. **Timely Servicing / rectification of defects during warranty period:** After having been notified of the defects / service requirement during warranty period, Seller has to complete the required Service / Rectification within 3 days time limit. If the Seller fails to complete service / rectification with defined time limit, a penalty of 0.5% of Unit Price of the product shall be charged as penalty for each week of delay from the seller. Seller can deposit the penalty with the Buyer directly else the Buyer shall have a right to recover all such penalty amount from the Performance Security (PBG). Cumulative Penalty cannot exceed more than 10% of the total contract value after which the Buyer shall have the right to get the service / rectification done from alternate sources at the risk and cost of the Seller besides forfeiture of PBG. Seller shall be liable to re-imburse the cost of such service / rectification to the Buyer.

[This Bid is also governed by the General Terms and Conditions](#)

---Thank You---