

### Bid Document

Bid Details	
Bid End Date/Time	10-07-2020 15:00:00
Bid Opening Date/Time	2020-07-10 15:30:00
Bid Life Cycle (From Publish Date)	90 (Days)
Bid Offer Validity (From End Date)	30 (Days)
Ministry/State Name	Haryana
Department Name	Medical Education & Research Department Haryana
Organisation Name	Directorate Of Medical Education & Research
Office Name	Karnal
Total Quantity	30
Item Category	acoustic stethoscope
OEM Average Turnover (Last 3 Years)	1 Lakh (s)
Experience Criteria	3 Year (s)
MSE Exemption for Years of Experience and Turnover	No
Startup Exemption for Years of Experience and Turnover	No
Document required from seller	Experience Criteria,Past Performance,Bidder Turnover,Certificate (Requested in ATC),OEM Authorization Certificate,OEM Annual Turnover *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer
Past Performance	80 %
Bid to RA enabled	No
Inspection Required	No

### EMD Detail

Required	No
----------	----

### ePBG Detail

Required	No
----------	----

## Splitting

Bid splitting not applied.

1. Experience Criteria: In respect of the filter applied for experience criteria, the Bidder or its OEM {themselves or through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU / Public Listed Company for number of years as indicated in the bid document before the bid opening date. Copies of relevant contracts to be submitted along with bid in support of having supplied some quantity during each of the year. In case of bunch bids, the category of primary product having highest value should meet this criterion.

2. OEM Turn Over Criteria: The minimum average annual financial turnover of the OEM of the offered product during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the OEM is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.

3. Past Performance: The Bidder or its OEM {themselves or through re-seller(s)} should have supplied same or similar Category Products for 80% of bid quantity, in at least one of the last three years before the bid opening date to any Central / State Govt Organization / PSU / Public Listed Company. Copies of relevant contracts (proving supply of cumulative order quantity in any one year) to be submitted along with bid in support of quantity supplied in the relevant year. In case of bunch bids, the category related to primary product having highest bid value should meet this criterion.

## Acoustic Stethoscope ( 30 pieces )

### Technical Specifications

[\\* As per GeM Category Specification](#)

Specification	Specification Name	Values	Bid Requirement (Allowed Values)
Warranty Scope	Warranty on Product (Year&#x28;s&#x29;)	3	*
	Free Service Available during warranty	Yes	Yes
CONFIGURATION	Length of Y tubing (cm)	98	98.0 - 200.0 Or higher
	Recommended application	Physical Assessment and Diagnosis,Limited Physical Assessment and Blood Pressure,Cardiology,Physical Assessment, Diagnosis	Physical Assessment and Diagnosis, Limited Physical Assessment and Blood Pressure, Cardiology, Physical Assessment, Diagnosis

	and cardiology	and cardiology
<b>Diameter of the Binaural tube</b>	5 mm	5 mm
<b>Diaphragm type</b>	Fixed	Fixed
<b>Chest piece material</b>	Aluminium	Aluminium
Material of snap on ring	Plastic	*
<b>Latex free product</b>	Yes	Yes
<b>Noise Isolation earpiece</b>	Yes	Yes
<b>Non chilling snap on ring provided on chest piece</b>	Yes	Yes
<b>Number of snap on ring provided</b>	1	1
<b>Binaural spring provided</b>	Yes	Yes
<b>Weight of the product (g)</b>	5	5
<b>Diameter of Chest piece(mm)</b>	100	100
<b>Number of spare diaphragm provided</b>	1	1, 2 Or higher
<b>Type of Stethoscope</b>	Adult Stethoscopes	Adult Stethoscopes
<b>Binaural tube material</b>	Drawn Brass confirming to IS 407-1961	Drawn Brass confirming to IS 407-1961
<b>Diameter of Chest piece;Diaphragm Side (mm)</b>	100	100
<b>Tunable Diaphragm</b>	Yes	Yes
<b>Finish of Binaural tube</b>	Chrome plated over nickle conforming to Grade C of IS:1068-1958	Chrome plated over nickle conforming to Grade C of IS:1068-1958
<b>Outer diameter of Y tubing (mm)</b>	100	100
<b>Type of Chest piece</b>	Bell	Bell
<b>Detectable Sound frequency range (Hz)</b>	1200	1200
<b>Ear pieces material</b>	Bakelite	Bakelite
<b>Number of ID tag provided</b>	1	1
<b>Material of the Spring</b>	N/A	N/A
Over all dimension;Len gthxHeightxWidth	-	*

	(mmxmmxmm)		
	<b>Spare snap on ring Provided</b>	Yes	Yes
	<b>Stethoscope design and manufacturing confirms to IS 3391-1965 latest amendment</b>	Yes	Yes
	Material of Y tubing	Latex free PVC	*
	<b>Finish of chest piece</b>	Chrome plated over nickle conforming to Grade C of IS:1068-1958	Chrome plated over nickle conforming to Grade C of IS:1068-1958
	<b>Diaphragm material</b>	Celluloid	Celluloid
	<b>Spare ear piece provided</b>	Yes	Yes
	<b>Number of Free Services, if provided</b>	2	2
	<b>Spare ID tag Provided</b>	Yes	Yes
	<b>Type of Y tubing</b>	Double lumen	Double lumen
	Test Report Date	-	*
	<b>ID Tag provided</b>	Yes	Yes
	<b>Number of spare Ear piece provided a set of two:</b>	1	1
	<b>Spare Diaphragm Provided</b>	Yes	Yes
CERTIFICATION	<b>Firm Certified to ISO 9001:2008 or latest</b>	Yes	Yes
	<b>ISI marked</b>	Yes	Yes
	Other Certifications	-	*
	<b>CE marked</b>	Yes	Yes
	<b>WHO-GMP certified</b>	Yes	Yes
PRODUCT DETAILS	<b>Product code available on supplier website</b>	Yes	Yes
	if Yes, Product code	-	*

\* Specifications highlighted in bold are the Golden Parameters.

\* Bidders may note that In respect of non-golden Parameters, the specifications 'Values' chosen by Buyer will generally be preferred over 'Bid requirement ( allowed Values) by the Buyer.

#### Additional Specification Documents

**Consignees/Reporting Officer and Quantity**

S.No.	Consignee/Reporting Officer	Address	Quantity	Delivery Days
1	Ashok Kumar	132001, Kalpana Chawla Govt. Medical College, Karnal	30	30

**Bid Specific Additional Terms and Conditions**

1. Scope of supply (Bid price to include all cost components) : Only supply of Goods
2. OEM Turn Over Criteria: The minimum average annual financial turnover of the OEM of the offered product during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the OEM is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria. In case of bunch bids, the OEM of CATEGORY RELATED TO primary product having highest bid value should meet this criterion.
3. IMPORTED PRODUCTS: In case of imported products, OEM or Authorized Seller of OEM should have a registered office in India to provide after sales service support in India. The certificate to this effect should be submitted.
4. Availability of Service Centres: Bidder/OEM must have a Functional Service Centre in the State of each Consignee's Location in case of carry-in warranty. (Not applicable in case of goods having on-site warranty). If service center is not already there at the time of bidding, successful bidder / OEM shall have to establish one within 30 days of award of contract. Payment shall be released only after submission of documentary evidence of having Functional Service Centre.
5. Timely Servicing / rectification of defects during warranty period: After having been notified of the defects / service requirement during warranty period, Seller has to complete the required Service / Rectification within 7 days time limit. If the Seller fails to complete service / rectification with defined time limit, a penalty of 0.5% of Unit Price of the product shall be charged as penalty for each week of delay from the seller. Seller can deposit the penalty with the Buyer directly else the Buyer shall have a right to recover all such penalty amount from the Performance Security (PBG). Cumulative Penalty cannot exceed more than 10% of the total contract value after which the Buyer shall have the right to get the service / rectification done from alternate sources at the risk and cost of the Seller besides forfeiture of PBG. Seller shall be liable to re-imburse the cost of such service / rectification to the Buyer.

[This Bid is also governed by the General Terms and Conditions](#)

**---Thank You---**