

Bid Document

Bid Details	
Bid End Date/Time	27-07-2020 18:00:00
Bid Opening Date/Time	27-07-2020 18:30:00
Bid Life Cycle (From Publish Date)	90 (Days)
Bid Offer Validity (From End Date)	75 (Days)
Ministry/State Name	Ministry Of Coal
Department Name	Materials Management
Organisation Name	Northern Coalfields Limited
Office Name	Singrauli
Total Quantity	1
Item Category	Arterial Blood Gas Analyser
MSE Exemption for Years of Experience and Turnover	No
Startup Exemption for Years of Experience and Turnover	No
Document required from seller	OEM Authorization Certificate *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer
Bid to RA enabled	No
Inspection Required	No
Estimated Bid Value	800000

EMD Detail

Advisory Bank	State Bank of India
EMD Percentage(%)	2.00
EMD Amount	16000

ePBG Detail

Advisory Bank	State Bank of India
ePBG Percentage(%)	10.00
Duration of ePBG required (Months).	62

(a). EMD EXEMPTION: The bidder seeking EMD exemption, must submit the valid supporting document for the relevant category as per GeM GTC with the bid. Under MSE category, only manufacturers for goods and Service Providers for Services are eligible for exemption from EMD. Traders are excluded from the purview of this Policy.

(b). EMD & Performance security should be in favour of Beneficiary, wherever it is applicable.

Beneficiary:

NCL, HQ, Singrauli

Singrauli, Materials Management, Northern Coalfields Limited, Ministry of Coal

(Ncl, Hq, Singrauli)

Splitting

Bid splitting not applied.

Arterial Blood Gas Analyser (1 pieces)**Technical Specifications**[* As per GeM Category Specification](#)

Specification	Specification Name	Values	Bid Requirement (Allowed Values)
Standards	Certification of the product	-	*
	Certification number and date	-	*
	Confirmity to Manufactures Certification	ISO 9001 & ISO 13485	*
Performance Parameters	Blood Gas Analyser (BGA) System	Blood gas analyser is an equipment for measuring and calculating complete Blood Gas Analysis	*
	Type of blood gas analyser	Bench top	Bench top
	Type of technology	Cartridge (multiple use)	Cartridge (multiple use)
	Parameters measured by blood gas analyser	pCO ₂ , pO ₂ , Na ⁺ , K ⁺ , pH and calculated parameters like HCO ₃ ⁻ , Base excess, TCO ₂	*
	Additional Parameters measurable by the blood gas analyser	Cl-,iCa,Glu,BUN,Hct,Creatinine,Lacate,Co-oximetry,Not available,Magnesium	Cl-, iCa, Glu, BUN, Hct, Creatinine, Lacate, Co-oximetry, Not available, Magnesium
	Cardiac Parameters measurable	Not available	Not available
	Neonatal Parameters	nBil	nBil
	Monitor type	TFT	TFT
	Type of printer	Inbuilt Printer	Inbuilt Printer
	Cartridge / reagent to be used can be stored mostly at room temperature between 2-30 Degree C, Automatic calibration (hint: seller shall inform the exact details to buyer at the time of making the supplies)	Yes	Yes
	"Shall have patient Bar Code facility, Ability to run a wide range of	Yes	Yes

	patient panels on one system"		
	Capability to transfer patient data to LIS/HIS	Yes	Yes
	Weight in kg	More than 15 kg	More than 15 kg
	Sample Volume in micro litre	135	*
	Power source	AC Mains + Rechargeable battery	AC Mains + Rechargeable battery
	Battery backup provided	Through UPS for 30 mins	Through UPS for 30 mins
	Shelf life of the reagent / catridges as applicable in months	6	6
	On-board shelf life of reagent in weeks	6	6
	Warranty (including electrode and membrane etc as applicable for the analyser offered) in years	5	5
	User/Technical/Maintenance manuals to be supplied in original in English in hard and soft copy	Yes	*
	Demonstration of equipment and training to be provided after completing supplies before acceptance	Yes	*
	The Principal Manufacturer must have direct Presence/approved service center In India	Yes	*
Reports	Availability of ISO certification to manufacturer	ISO 13485	*
	ISO certification date and number	-	*
	Copies of reports and certifications to be furnished to buyer on demand at time of supplies	Yes	*

* Specifications highlighted in bold are the Golden Parameters.

* Bidders may note that In respect of non-golden Parameters, the specifications 'Values' chosen by Buyer will generally be preferred over 'Bid requirement (allowed Values) by the Buyer.

Additional Specification Documents

Consignees/Reporting Officer and Quantity

S.No.	Consignee/Reporting Officer	Address	Quantity	Delivery Days
1	Sanjay Kumar Gupta	486890,NEHRU SATABDI CHIKITSHALYA, PO - JAYANT, DISTT - SINGRAULI M.P.	1	30

Bid Specific Additional Terms and Conditions

1.Scope of supply (Bid price to include all cost components) : Supply Installation Testing Commissioning of Goods and Training of operators and providing Statutory Clearances required (if any)

[This Bid is also governed by the General Terms and Conditions](#)

---Thank You---