

Bid Document

| Bid Details | |
|--|---|
| Bid End Date/Time | 27-07-2020 12:00:00 |
| Bid Opening Date/Time | 27-07-2020 12:30:00 |
| Bid Life Cycle (From Publish Date) | 90 (Days) |
| Bid Offer Validity (From End Date) | 30 (Days) |
| Ministry/State Name | Tripura |
| Department Name | Health And Family Welfare Department Tripura |
| Organisation Name | N/a |
| Office Name | Directorate Of Health Services, Agartala |
| Total Quantity | 1 |
| Item Category | Steam Sterlizer |
| Experience Criteria | 3 Year (s) |
| MSE Exemption for Years of Experience and Turnover | No |
| Startup Exemption for Years of Experience and Turnover | No |
| Document required from seller | Experience Criteria,Bidder Turnover,OEM Authorization Certificate *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer |
| Bid to RA enabled | No |
| Inspection Required | No |

EMD Detail

| | |
|----------|----|
| Required | No |
|----------|----|

ePBG Detail

| | |
|----------|----|
| Required | No |
|----------|----|

Splitting

Bid splitting not applied.

1. Experience Criteria: In respect of the filter applied for experience criteria, the Bidder or its OEM {themselves or through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU / Public Listed Company for number of years as indicated in the bid document before the bid opening date. Copies of relevant contracts to be submitted along with bid in support of having supplied some quantity during each of the year. In case of bunch bids, the category of primary product having highest value should meet this criterion.

Steam Sterilizer (1 pieces)

Technical Specifications

* As per GeM Category Specification

| Specification | Specification Name | Values | Bid Requirement (Allowed Values) |
|---------------|---|--|---|
| Construction | Type of Steriliser | VERTICAL CYLINDRICAL Steam Steriliser Pressure type, Steriliser Chamber size: 300 mm (Diameter) x 500 mm (Length), conforming to IS: 3829 (Part 3)/1985, reaffirmed 2001, with amendment 1 & 2, fitted with Silicon door gasket(expansion type).Electric load of the unit shall be 4 KW. | VERTICAL CYLINDRICAL Steam Steriliser Pressure type, Steriliser Chamber size: 300 mm (Diameter) x 500 mm (Length), conforming to IS: 3829 (Part 3)/1985, reaffirmed 2001, with amendment 1 & 2, fitted with Silicon door gasket(expansion type).Electric load of the unit shall be 4 KW. |
| | Steam supply to Steriliser | Direct Steam Operation from an external source | Direct Steam Operation from an external source |
| | Material: Steriliser parts; Chamber, Jacket, Door, Door Ring and shall be made of | SS 304 | SS 304 |
| | Type of control | Manual through indt. switch based | Manual through indt. switch based |
| CONSTRUCTION | Instruction plate on Steam Steriliser advising user to use soft water for steam generation for long service life of Steam Steriliser | YES | * |
| | Whether a compatible water softener/reversed osmosis based system to feed water to the steam generator unit is to be supplied with the steriliser: | No | No |
| | Provision of Lock on Steriliser Chamber door: | Manual Lock | Manual Lock |
| | Installation | The site ready for installation equipped with water supply, electric connection cabling and drainage outlet shall be provided by Buyer. The Seller shall be responsible to carry out for installation of the offered Steam Steriliser unit on turnkey basis i.e. Any civil work required for installation of the ordered equipment i.e. flooring/tilling/plastering/Paint and/or wood work shall be Seller's responsibility. The Seller shall also provide training to personnel of buyer. | * |
| | Type of Door: | Hinged type | Hinged type |

| | | | |
|---|---|---|---|
| | Display facility for Sterilisation Cycle parameters | Low water indicator, Chamber Temperature | Chamber Temperature, Low water indicator |
| | Print facility for record generation of dates, time, load, identification number and operating parameters (temperature and holding time etc): | Yes, port given to connect an external Printer Unit | Yes, port given to connect an external Printer Unit |
| | Warranty Period | 3 | 3, 5 Or higher |
| | No of Doors in Steriliser | 1 | 1 |
| | Whether Type Test Certificate from a Govt/NABL/ILAC approved testing laboratory is available (If Yes, to be furnished by Seller to Buyer on demand) for the offered product? | No | Yes, No |
| | Type of user friendly Interface (for microprocessor controlled Sterilizer) to specify parameters for running Sterilization cycle: | NA | NA |
| | Data Transfer facility through USB /RS232 port to computer/laptop on real time basis | No | Yes, No |
| | Air removal from Sterilisation chamber, after loading the steriliser, by on for more pressure/vacuums pulsing for proper steam penetration | YES | * |
| Integrated Shredder Facility | Integrated Shredder (for shredding the load in Sterilisation chamber) | Not available | Not available |
| | Material: Shredder parts; | SS 304 | SS 304 |
| | Power of Motor used for Shredding | NA, no integrated Shredder unit offered | NA, no integrated Shredder unit offered |
| Accessories (made of SS 316) for loading/unloading and handling of Load | Number of Rails: | 0 | 0, 2 Or higher |
| | Long Handle: | 1 | 1 |
| | Sterilisation Carriages | 1 | 1, 2 Or higher |
| | Pull out Trays/Tanks | 2 | 2 |
| | Floor loading carts | 0 | 0, 2 Or higher |
| | Transfer carriages: | 0 | 0, 1, 2 Or higher |

* Specifications highlighted in bold are the Golden Parameters.

* Bidders may note that In respect of non-golden Parameters, the specifications 'Values' chosen by Buyer will generally be preferred over 'Bid requirement (allowed Values) by the Buyer.

Additional Specification Documents

Consignees/Reporting Officer and Quantity

| S.No. | Consignee/Reporting Officer | Address | Quantity | Delivery Days |
|-------|-----------------------------|---------------------------|----------|---------------|
| 1 | SAMARPITA DUTTA | 799006,Kunjaban, Agartala | 1 | 15 |

Bid Specific Additional Terms and Conditions

- 1.Scope of supply (Bid price to include all cost components) : Supply Installation Testing Commissioning of Goods and Training of operators and providing Statutory Clearances required (if any)
- 2.IMPORTED PRODUCTS: In case of imported products, OEM or Authorized Seller of OEM should have a registered office in India to provide after sales service support in India. The certificate to this effect should be submitted.
- 3.Warranty period of the supplied products shall be 3 years from the date of final acceptance of goods or after completion of installation, commissioning & testing of goods (if included in the scope of supply), at consignee location. OEM Warranty certificates must be submitted by Successful Bidder at the time of delivery of Goods. The seller should guarantee the rectification of goods in case of any break down during the guarantee period. Seller should have well established Installation, Commissioning, Training, Troubleshooting and Maintenance Service group in INDIA for attending the after sales service. Details of Service Centres near consignee destinations are to be uploaded along with the bid
- 4.Dedicated /toll Free Telephone No. for Service Support : BIDDER/OEM must have Dedicated/toll Free Telephone No. for Service Support.
- 5.Availability of Service Centres: Bidder/OEM must have a Functional Service Centre in the State of each Consignee's Location in case of carry-in warranty. (Not applicable in case of goods having on-site warranty). If service center is not already there at the time of bidding, successful bidder / OEM shall have to establish one within 30 days of award of contract. Payment shall be released only after submission of documentary evidence of having Functional Service Centre.
- 6.ISO 9001: The bidder must have ISO 9001 certification.
- 7.Material Test Certificate Should Be Sent Along with The Supply. The Material Will Be Checked by Buyer's Lab & the Results of the Lab will be the Sole Criteria for Acceptance of the Item.
- 8.Bidder's offer is liable to be rejected if they don't upload any of the certificates / documents sought in the Bid document, ATC and Corrigendum if any.
- 9.Procurement under this bid is reserved for purchase from Micro and Small Enterprises from the State of Bid Inviting Authority whose credentials are validated online through Udyog Aadhaar for that product category. If the bidder wants to avail the reservation benefit, the bidder must be the manufacturer of the offered product in case of bid for supply of goods. Traders are excluded from the purview of Public Procurement Policy for Micro and Small Enterprises. In respect of bid for Services, the bidder must be the Service provider of the offered Service. Relevant documentary evidence in this regard shall be uploaded along with the bid in respect of the offered product or service.

[This Bid is also governed by the General Terms and Conditions](#)

---Thank You---