

Bid Document

Bid Details	
Bid End Date/Time	21-07-2020 14:00:00
Bid Opening Date/Time	21-07-2020 14:30:00
Bid Life Cycle (From Publish Date)	90 (Days)
Bid Offer Validity (From End Date)	30 (Days)
Ministry/State Name	Ministry Of Power
Department Name	Na
Organisation Name	N/a
Office Name	Bhakra Beas Management Board, Chandigarh
Total Quantity	450
Item Category	Disposable Surgical Rubber Gloves - Prepowdered
MSE Exemption for Years of Experience and Turnover	No
Startup Exemption for Years of Experience and Turnover	No
Bid to RA enabled	No
Inspection Required	No

EMD Detail

Required	No
----------	----

ePBG Detail

Required	No
----------	----

Splitting

Bid splitting not applied.

Disposable Surgical Rubber Gloves - Prepowdered (450 pairs)

Technical Specifications

[* As per GeM Category Specification](#)

Specification	Specification Name	Values	Bid Requirement (Allowed Values)
GENERAL INFORMATION	Product description	Disposable Surgical Rubber Gloves conforming to IS:13422/1989 (RA 2003)	*

		-prepowdered	
	Types	Type 1	*
	Single use	Yes	*
	Sterilized gloves	Yes	*
STANDARD	Conformity to standard	IS:13422/1992 (Reaffirmed 2003)	*
	Product is ISI Marked	Yes	*
	BIS Licence No (CM/L No)	6105044	*
DIMENSIONS	Designation of Size of gloves as per Table 1 of IS:13422/1992(RA 2003)	7.0 1/2	7.0 1/2
	Minimum length (mm)	245	245
	Width (mm)	89+/-6	89+/-6
	Minimum Thickness of Finger, Palm and Cuff (mm)	0.1	*
	The thickness of the cuff when measured using a gauge with a 5mm diameter foot shall not exceed 3mm	Yes	*
MATERIAL	Material of the gloves	Synthetic Latex or Rubber Solution	Synthetic Latex or Rubber Solution
	Surface treatment with lubricant or powder provided to facilitate donning of gloves as per Cl 5 decimal 1 of IS 13422/1992 (RA 2003)	Yes	*
	Lubricant or powder material used for surface treatment is Bio-absorbable, Non-toxic and Non-pyrogenic	Yes	*
	Material composition of lubricant or powder used	-	*
TECHNICAL FEATURES	Cuf of the glove shall be resistant to tear	Yes	*
	Construction of the cuff shall fit closely without tendency to roll back or ruckle while in use	Yes	*
	Compliance to Cl 5 decimal 3 of IS:13422/1992 (RA 2003) in respect of Physical properties	Yes	*
	Compliance to Cl 5 decimal 5 of IS:13422/1992 (RA 2003) in respect of Air Tight test	Yes	*
	Compliance to Cl 5 decimal 4 of IS:13422/1992 (RA 2003) in respect of Sterility test	Yes	*
PACKING & MARKING	Packaging shall be provided to maintain sterility after sterilization during shipping and storage and to permit	Yes	*

	opening without contamination of the gloves as given in Cl 8 decimal 1 of IS:13422/1992 (RA 2003)		
	Marking	As per Cl 8.2 of IS:13422/1992(RA 2003)	*
CERTIFICATIONS & REPORTS	Avaliability of Test report from Government / NABL / ILAC accredited Laboratory to prove conforming to the declared specification	Yes	*
	Test report Number and date	-	*
	Name and address of Test Lab	-	*
	Certification Available	ISO	*
	Submission of all the test reports to the buyer along with supplies on demand	Yes	*
SHELF LIFE / WARRANTY	Shelf life/warranty	-	*
	Minimum Shelf life at the time of supply	2/3rd of total shelf life	*

* Specifications highlighted in bold are the Golden Parameters.

* Bidders may note that In respect of non-golden Parameters, the specifications 'Values' chosen by Buyer will generally be preferred over 'Bid requirement (allowed Values) by the Buyer.

Additional Specification Documents

Consignees/Reporting Officer and Quantity

S.No.	Consignee/Reporting Officer	Address	Quantity	Delivery Days
1	Rakesh Kumar	175017,RE, Division, BBMB, Slapper, Distt. Mandi (HP)	450	10

Bid Specific Additional Terms and Conditions

1.Scope of supply (Bid price to include all cost components) : Only supply of Goods

[This Bid is also governed by the General Terms and Conditions](#)

---Thank You---