

Bid Document

| Bid Details | |
|---|---|
| Bid End Date/Time | 24-07-2020 15:00:00 |
| Bid Opening Date/Time | 24-07-2020 15:30:00 |
| Bid Life Cycle (From Publish Date) | 90 (Days) |
| Bid Offer Validity (From End Date) | 30 (Days) |
| Ministry/State Name | Ministry Of Agriculture And Farmers Welfare |
| Department Name | Department Of Agricultural Research And Education (dare) |
| Organisation Name | Central Agricultural University Imphal Manipur |
| Office Name | College Of Fisheries Tripura |
| Total Quantity | 1 |
| Item Category | Laboratory Stirrer |
| Experience Criteria | 1 Year (s) |
| MSE Exemption for Years Of Experience | Yes |
| Startup Exemption for Years Of Experience | Yes |
| Document required from seller | Experience Criteria,Past Performance,Certificate (Requested in ATC),OEM Authorization Certificate *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer |
| Past Performance | 80 % |
| Bid to RA enabled | No |
| Inspection Required | No |

EMD Detail

| | |
|----------|----|
| Required | No |
|----------|----|

ePBG Detail

| | |
|----------|----|
| Required | No |
|----------|----|

Splitting

Bid splitting not applied.

1. Experience Criteria: In respect of the filter applied for experience criteria, the Bidder or its OEM {themselves or through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU / Public Listed Company for number of years as indicated in the bid document before the bid opening date. Copies of relevant contracts to be submitted along with bid in support of having supplied some quantity during each of the year. In case of bunch bids, the category of primary product having highest value should meet this criterion.

2. Past Performance: The Bidder or its OEM {themselves or through re-seller(s)} should have supplied same or similar Category Products for 80% of bid quantity, in at least one of the last three years before the bid opening date to any Central / State Govt Organization / PSU / Public Listed Company. Copies of relevant contracts (proving supply of cumulative order quantity in any one year) to be submitted along with bid in support of quantity supplied in the relevant year. In case of bunch bids, the category related to primary product having highest bid value should meet this criterion.

Laboratory Stirrer (1 pieces)

Technical Specifications

[* As per GeM Category Specification](#)

| Specification | Specification Name | Values | Bid Requirement (Allowed Values) |
|------------------------|-------------------------|---|----------------------------------|
| Performance parameters | Application | Used for for stirring / mixing etc. of chemical, pharmaceutical, food and cosmetic products | * |
| | Type of base stand | U Type | U Type |
| | Type controller | Microprocessor | Microprocessor |
| | Type of Motor | DC | DC |
| | Motor rating in hp | 1/2 | 1/2 |
| | Geared | No | Yes, No |
| | Tacho feed back | Yes | Yes |
| | Speed control | Yes | Yes |
| | Speed range | 50-2000 | 50-2000 |
| | Wattage | 50 | * |
| | Soft start | Yes | * |
| | Stirring shaft length | 250 | 250 |
| | Stirring shaft diameter | 8 | 8 |
| | Stirring shaft material | SS | * |
| | Impeller type | Axial turbine | Axial turbine |
| | Impeller diameter | 38 | 38 |
| | Maximum stirring | 20 | 20 |

| | | | |
|---------|---|---------------------------------------|---------------------------------------|
| | capacity in litres | | |
| | Flame proof | No | Yes, No |
| | Type of display | LED | LED |
| | Timer range | 0 to 99 minutes | * |
| | Last parameter recall | Yes | Yes |
| | Auto restart | No | Yes, No |
| | chuck size in mm | 1.5 to 13 | 1.5 to 13 |
| | Motor protection in case of over load | Yes(LED Light flash ,auto stop) | * |
| | Power supply | 230 v plus minus 10% 50 hz | 230 v plus minus 10% 50 hz |
| | Weight of laboratory stir in kg | 2.4 | * |
| | Warranty in years | 1 | 1 |
| Reports | ISO certification of manufacturer | Yes | * |
| | ISO Certification No and date | - | * |
| | Availability of test report from govt/ NABL/ILAC accredited lab covering all parameters | No | Yes, No |
| | Test report number , Incase test reports not available , put NA | - | * |
| | Name of test lab , Incase test reports not available , put NA | - | * |
| | Address of test lab , Incase test reports not available , put NA | - | * |
| | Copies of all certifications and reports to be provided to buyer on demand at time of supplies | NA ,Incase test reports not available | NA ,Incase test reports not available |

* Specifications highlighted in bold are the Golden Parameters.

* Bidders may note that In respect of non-golden Parameters, the specifications 'Values' chosen by Buyer will generally be preferred over 'Bid requirement (allowed Values) by the Buyer.

Additional Specification Documents

Consignees/Reporting Officer and Quantity

| S.No. | Consignee/Reporting Officer | Address | Quantity | Delivery Days |
|-------|-----------------------------|---|----------|---------------|
| 1 | Sashadhar Chanda | 799210, College of Fisheries Central Agricultural University Lembucherra, Agartala - 799 210 Tripura, India | 1 | 30 |

Bid Specific Additional Terms and Conditions

1. IMPORTED PRODUCTS: In case of imported products, OEM or Authorized Seller of OEM should have a registered office in India to provide after sales service support in India. The certificate to this effect should be submitted.
2. Scope of supply (Bid price to include all cost components) : Supply Installation Testing Commissioning of Goods and Training of operators and providing Statutory Clearances required (if any)
3. Availability of Service Centres: Bidder/OEM must have a Functional Service Centre in the State of each Consignee's Location in case of carry-in warranty. (Not applicable in case of goods having on-site warranty). If service center is not already there at the time of bidding, successful bidder / OEM shall have to establish one within 30 days of award of contract. Payment shall be released only after submission of documentary evidence of having Functional Service Centre.
4. Dedicated /toll Free Telephone No. for Service Support : BIDDER/OEM must have Dedicated/toll Free Telephone No. for Service Support.
5. Escalation Matrix For Service Support : Bidder/OEM must provide Escalation Matrix of Telephone Numbers for Service Support.
6. Warranty period of the supplied products shall be 1 years from the date of final acceptance of goods or after completion of installation, commissioning & testing of goods (if included in the scope of supply), at consignee location. OEM Warranty certificates must be submitted by Successful Bidder at the time of delivery of Goods. The seller should guarantee the rectification of goods in case of any break down during the guarantee period. Seller should have well established Installation, Commissioning, Training, Troubleshooting and Maintenance Service group in INDIA for attending the after sales service. Details of Service Centres near consignee destinations are to be uploaded along with the bid

[This Bid is also governed by the General Terms and Conditions](#)

---Thank You---