

RA Document

RA Details	
RA Technical End Date/Time	27-07-2020 12:00:00
RA Opening Date/Time	27-07-2020 12:30:00
RA Life Cycle (From Publish Date)	90 (Days)
RA Offer Validity (From End Date)	80 (Days)
Ministry/State Name	Ministry Of Health And Family Welfare
Department Name	Department Of Health And Family Welfare
Organisation Name	All India Institute Of Medical Sciences (aiims)
Office Name	Raipur
Total Quantity	2
Item Category	Orthopedic Operating Table
Average Turnover (Last 3 Years)	1 Lakh (s)
OEM Average Turnover (Last 3 Years)	1 Lakh (s)
Experience Criteria	3 Year (s)
MSE Exemption for Years of Experience and Turnover	No
Startup Exemption for Years Of Experience and Turnover	Yes
Document required from seller	Experience Criteria,Past Performance,Bidder Turnover,OEM Authorization Certificate,OEM Annual Turnover *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer
Past Performance	50 %
Inspection Required	No

EMD Detail

Required	No
----------	----

ePBG Detail

Required	No
----------	----

Splitting

RA splitting not applied.

1. Experience Criteria: In respect of the filter applied for experience criteria, the Bidder or its OEM {themselves or through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU / Public Listed Company for number of years as indicated in the bid document before the bid opening date. Copies of relevant contracts to be submitted along with bid in support of having supplied some quantity during each of the year. In case of bunch bids, the category of primary product having highest value should meet this criterion.
2. OEM Turn Over Criteria: The minimum average annual financial turnover of the OEM of the offered product during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the OEM is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.
3. Past Performance: The Bidder or its OEM {themselves or through re-seller(s)} should have supplied same or similar Category Products for 50% of bid quantity, in at least one of the last three years before the bid opening date to any Central / State Govt Organization / PSU / Public Listed Company. Copies of relevant contracts (proving supply of cumulative order quantity in any one year) to be submitted along with bid in support of quantity supplied in the relevant year. In case of bunch bids, the category related to primary product having highest bid value should meet this criterion.

Orthopedic Operating Table (2 pieces)

Technical Specifications

[* As per GeM Category Specification](#)

Specification	Specification Name	Values	Bid Requirement (Allowed Values)
Standards	UMDNS Names	Tables,Operating-1396 1 , Operating Tables,Spinal-23732 , Operating table,multi- Specialty-25217, Operating Tabletops -25215	*
	Clinical Application in General for Orthopedic operating table	Orthopedic tables are specialized operating tables that provide an elevated surface for supporting the patient's body and include specific apparatus for supporting and/or	*

		<p>providing traction for the patient's limbs during orthopedic procedures. These tables stabilize the patient's position and provide the operating surgeon with optimal exposure of the surgical field. Applications for these tables include both upper-extremity procedures (e.g., shoulder surgery, nonoperative myelograms, spinal surgery) and lower-extremity/hip procedures (e.g., hip pinning, Ender nailing, intramedullary nailing of the femur, tibia, and fibula). Some orthopedic tables can also accommodate C-arm radiographic/fluoroscopic studies</p>	
	Conformity to Standards	-	*
	Conformity to standard for for electrical safety latest amendment	IEC- 60601-2-40 or BIS Equivalent	*
	Conformity to Manufacturing Standards	ISO 9001 & 13485 or ICMED 9001 & 13485	*
Table features	Type of OT table Actuator	Hydraulic	Hydraulic
	Type of mechanism for fuinctioning or controlling the table movements	Manual	Manual
	Remote controlled positions	NA (for manual operated)	NA (for manual operated)
	Number of section provided in Table top	upto 6 sections	*
	C-Arm Compatible	Yes	*
	Table top length, Range	1900 mm-2100 mm	1900 mm-2100 mm
	Table top width, Range	500 mm-580 mm	*
	Table top Height, Range	650 mm - 1150 mm	650 mm - 1150 mm

	Material	304 Grade SS	*
	Table top shall be X-ray compactable and stain Resistant	Yes	*
	Table Mattress Material	2 inch latex free fully radiolucent, detachable, impermeable to fluid and easily washable	*
	Mattress material shall be in parts as per table top	Yes	*
	Provision of X ray cassette channel	Yes	Yes
Table Positions	Trendelenburg Angle, Range	25° to 30°	*
	Reverse Trendelenburg Angle, Range	25° to 30°	*
	Lateral Tilt Angle (Left), Range	18° to 25°	*
	Lateral Tilt Angle (Right), Range	18° to 25°	*
	Head Section UP/Down Angle, Range	upto 60° to 90°	*
	Back Section UP/Down Angle, Range	upto 90° to 30°	*
	Horizontal Positioning, Range	0° to 90°	*
	Leg Section UP/Down, Range	20°-90°	*
	Flex	upto 230°	*
	Reflex	upto 100°	*
	Table top longitudinal sliding, Range	5" to 12"	5" to 12"
	Patient Weight Capacity	150-270 Kgs	*
	Kidney Bridge required	Yes	Yes
	Detachable Head, Leg & Pelvic Section	Yes	*
	Split leg section , detachable-1 pair	Yes	Yes
	Quick Switch for Zero position	Yes	*
	Battery backup of	Not provided	Not provided

	operating table in hrs		
	Battery backup provided	No	Yes, No
	Foot pump	Yes	Yes
	Both arm board should have up and down and rotation function	Yes	Yes
	Operating table should have foot brake for ease of use	Yes	*
Standard Accessories	Number of Anesthetic Screen with side clamp & lock	1	*
	Number of Arm rest with side clamp & lock	1pair	*
	Number of Shoulder support with side clamp & lock	1 pair	*
	Leg support or Lithothomy with side clamp & lock-1pair	Yes	Yes
	Extension bar-1pair	Yes	Yes
	Horizontal Attachments-1no	Yes	Yes
	Sacral Rest	Yes	*
	Lateral Intramedullary Nailing Support	Yes	*
	Retractable abductor bar	Yes	*
	Number of Foot Support	1	*
	Number of Bolster Pillow	1	*
	Remote-1 no	No	Yes, No
	IV pole and Bracket	Yes	*
Orthopedic Attachments	Orthopedic Attachments	Floor based	Floor based
	Foot plate mount	1	1
	Supports bars	4	4
	Radial setting clamps	2	2
	Countertraction post , including padding roll	2	*

	Side rail extension	2	2
	Transport cart	1	1
	Traction stirrup clamp with rotation	2	2
	Rotation and tilting clamp	2	2
	Foot plate for adult	2	2
	Traction boot pair	1	1
	Femur nailing - Counter traction post for femur	1	1
	Counter traction post for tibia	1	1
	Condylar fixation	1	1
	Guide bar	1	1
Power requirements	"Power input to be 220-240V AC, 50Hz as appropriate fitted with Indian plug "	Single phase 220-240 Volt, 50 Hz (AC Supply)	*
Reports	Warranty in years	3	3
	User/Technical/Maintenance manuals to be supplied in English in hard and soft copy	Yes	*
	Demonstration of equipment and training to be provided after completing supplies & Installation before acceptance	Yes	*
	The Principal Manufacturer must have direct Presence/approved service center In India	Yes	*
	Copies of reports and certifications to be furnished to buyer on demand at time of supplies	Yes	*
	Attach certificate from the principal that spare parts will be available for next 10 years	Yes	*
	QA, QC and other tests(6 months/ 1 year as per NABH/JCI)	Yes	*

* Specifications highlighted in bold are the Golden Parameters.

* Bidders may note that In respect of non-golden Parameters, the specifications 'Values' chosen by Buyer will generally be preferred over 'Bid requirement (allowed Values) by the Buyer.

Additional Specification Documents

Consignee/Reporting Officer and Quantity

S.No.	Consignee/Reporting Officer	Address	Quantity	Delivery Days	
1	Debashish Bhadra	492001,AIIMS Raipur Tatibandh	2	15	

Bid Specific Additional Terms and Conditions

1. Experience Certificate for the supply of the same to any Govt/ PSU/ any renowned private organisation along with Supply/ Purchase Order.
2. Make in india specific authorisation certificate needs to be enclosed.
3. Scope of supply (Bid price to include all cost components) : Supply Installation Testing and Commissioning of Goods
4. Escalation Matrix For Service Support : Bidder/OEM must provide Escalation Matrix of Telephone Numbers for Service Support.
5. Dedicated /toll Free Telephone No. for Service Support : BIDDER/OEM must have Dedicated/toll Free Telephone No. for Service Support.
6. Availability of Service Centres: Bidder/OEM must have a Functional Service Centre in the State of each Consignee's Location in case of carry-in warranty. (Not applicable in case of goods having on-site warranty). If service center is not already there at the time of bidding, successful bidder / OEM shall have to establish one within 30 days of award of contract. Payment shall be released only after submission of documentary evidence of having Functional Service Centre.
7. OPTION CLAUSE: The Purchaser reserves the right to increase or decrease the quantity to be ordered up to 25 percent at the time of placement of contract. The purchaser also reserves the right to increase the ordered quantity by up to 25% of the contracted quantity during the currency of the contract at the contracted rates. Bidders are bound to accept the orders accordingly.

[This RA is also governed by the General Terms and Conditions](#)

---Thank You---