

Bid Document

| Bid Details | |
|--|--|
| Bid End Date/Time | 20-07-2020 10:00:00 |
| Bid Opening Date/Time | 20-07-2020 10:30:00 |
| Bid Life Cycle (From Publish Date) | 90 (Days) |
| Bid Offer Validity (From End Date) | 50 (Days) |
| Ministry/State Name | Gujarat |
| Department Name | Health & Family Welfare Department Gujarat |
| Organisation Name | N/a |
| Office Name | General Hospital Jamkhambhalia |
| Total Quantity | 5 |
| Item Category | Medical Oxygen Gas Cylinders - Steel(Normalized) |
| MSE Exemption for Years of Experience and Turnover | No |
| Startup Exemption for Years of Experience and Turnover | No |
| Bid to RA enabled | No |
| Inspection Required | No |

EMD Detail

| | |
|----------|----|
| Required | No |
|----------|----|

ePBG Detail

| | |
|----------|----|
| Required | No |
|----------|----|

Splitting

Bid splitting not applied.

Medical Oxygen Gas Cylinders - Steel(Normalized) (5 pieces)

Technical Specifications

* As per GeM Category Specification

| Specification | Specification Name | Values | Bid Requirement (Allowed Values) |
|----------------------------------|---|--|----------------------------------|
| PRODUCT INFORMATION | Product description | Medical Oxygen Gas Cylinders - Steel (Normalized) | * |
| | Purpose | Filling of Compressed Medical Oxygen (I.P) gas for hospital purposes | * |
| CYLINDER TECHNICAL SPECIFICATION | Conformity to Standard | IS: 7285 (Part 1)/2004 (Latest) | * |
| | Product is ISI Marked | Yes | * |
| | BIS License No (CM/L No) | - | * |
| | Material of Cylinder body | Normalized Steel | * |
| | High pressure seamless Cylinder | Yes | * |
| | Colour coding of cylinders | Body - Black, Shoulder - White | * |
| | Cylinder Size | D | D |
| | Water Capacity of Cylinder (Litres) | 10 | * |
| | Hydrostatic test pressure (Kg/Cm2) | 250 | * |
| | Maximum working pressure at NTP (Normal Temperature and Pressure) (Kg/Cm2) | 150 | * |
| | Filling Pressure of Cylinder at NTP (Normal Temperature and Pressure)(Kg/Cm2) | 150+/-5% | * |
| | Outer diameter, approx, (mm) | 140 | * |
| | Minimum Wall thickness (mm) | 4 | * |
| | Minimum Cylinder Length (mm) | 830 | * |
| | Tare weight (Kg) | 14 | * |
| | Neck ring provided | No | Yes, No |
| | Foot ring provided | No | Yes, No |
| | Type of Valve fitted with the cylinder | Type B (Key operated) | Type B (Key operated) |

| | | | |
|------------------------|--|-----|---|
| STATUTORY REQUIREMENTS | Valve protection caps available | Yes | * |
| | ISI marked Valve conforming to IS 3224-2002 (Latest) fitted with the cylinder | Yes | * |
| | All fittings should be leak proof | Yes | * |
| | Cylinder conforms to requirements prescribed in Gas Cylinder Rules, 2016 of the Govt of India, with such modifications and amendments as may be ordered/issued from time to time by the Chief Controller of Explosives, Dept of Petroleum, Govt of India | Yes | * |
| | All cylinders must have filling permission or must have HST (Hydro Stretch Test) certification from test shop approved by Chief Controller of Explosives, Department of Petroleum, Government of India | Yes | * |
| | Testing of Cylinder as per standard protocol is the responsibility of Cylinders supplier | Yes | * |
| PACKING & MARKING | Each cylinder should be engraved the details viz (a) Serial No of cylinder/Identification, (b) Code of mfg, (c) Working pressure, (d) Water capacity, liters, (e) Test pressure in Kg/cm2, (f) Year of manufacture and testing, (g)Tare weight of cylinder | Yes | * |
| | All the cylinders should be properly painted as per their respective color code, Cylinder No, name of the supplier, name of the gas | Yes | * |

| | | | |
|---------------------------|---|-------------------|---|
| | cylinder and its contents should be clearly indicated on the neck of the cylinder | | |
| LICENSES & CERTIFICATIONS | Type approval certificate for the Cylinder to the relevant IS Standard from Competent authority to be made available to the buyer on demand | Yes | * |
| | Cylinder fitness certificate for each Cylinder is to be furnished to the buyer | Yes | * |
| | Validity of Cylinder safety certificate (years) | 5 | * |
| | Manufacturer facility certification available | ISO 9001 (Latest) | * |
| | Submission of all licenses, test reports and certificates to the buyer at the time of supply | Yes | * |

* Specifications highlighted in bold are the Golden Parameters.

* Bidders may note that In respect of non-golden Parameters, the specifications 'Values' chosen by Buyer will generally be preferred over 'Bid requirement (allowed Values) by the Buyer.

Additional Specification Documents

Consignees/Reporting Officer and Quantity

| S.No. | Consignee/Reporting Officer | Address | Quantity | Delivery Days |
|-------|-----------------------------|---|----------|---------------|
| 1 | Matani Harish Pravin | 361305,General Hospital, Jam-Khambhalia Near Reliance Circle, Highway Char Rasta, Jam-Khambhaliya | 5 | 15 |

Bid Specific Additional Terms and Conditions

1. Scope of supply (Bid price to include all cost components) : Only supply of Goods

[This Bid is also governed by the General Terms and Conditions](#)

---Thank You---