

### Bid Document

Bid Details	
Bid End Date/Time	21-07-2020 17:00:00
Bid Opening Date/Time	21-07-2020 17:30:00
Bid Life Cycle (From Publish Date)	90 (Days)
Bid Offer Validity (From End Date)	30 (Days)
Ministry/State Name	Ministry Of Defence
Department Name	Department Of Defence
Organisation Name	Indian Army
Office Name	*****
Total Quantity	1
Item Category	Canine Hydraulic Surgery and Examination Table
MSE Exemption for Years of Experience and Turnover	No
Startup Exemption for Years of Experience and Turnover	No
Bid to RA enabled	No
Inspection Required	No

#### EMD Detail

Required	No
----------	----

#### ePBG Detail

Required	No
----------	----

#### Splitting

Bid splitting not applied.

### Canine Hydraulic Surgery And Examination Table ( 1 pieces )

#### Technical Specifications

[\\* As per GeM Category Specification](#)

Specification	Specification Name	Values	Bid Requirement (Allowed Values)
Standards	Conformity to Standards	-	*
	Certification No and Date	-	*
	Conformity to standard for for electrical safety latest amendment	IEC- 60601- or BIS Equivalent	*
	Manufacturers Quality Standards	ISO 9001 & ISO 13485	*
Performance Parameters	Purpose	For surgeries and examinations in veterinary hospitals and animal care centers	*
	Material of table	Stainless Steel grade 304	*
	Availability of tilt-top mechanism	Yes	*
	The minimum width of the table (in inches)	78	*
	The minimum length of the table (in inches)	70	*
	The maximum height of the table (in inches)	26	*
	The minimum weight bearing capacity of the table(in Kg)	167	*
	Adjustability of the table height	Yes	*
	Furnishing with tie down and duplex electric output	Yes	*
	The system should be used to control the movement of the table	Hydraulic	*
	The voltage of the electric current required for the hydraulic system (in volts)	220	*
	The control of lowering and lifting of the table	Hydraulic system	*
	The table should be fitted with adjustable floor glides	Yes	*
	No of heavy duty swivel	4	*

	casters provided		
	The swivel caster should be provided with	Removable brakes	*
	Cleaning of Casters	Easy cleaning by pulling pins and sliding	*
	Control panel of the hydraulic system	Easily accessible and enclosed in waterproof casing	*
	The material of hydraulic lines	SS 304 grade	*
	Washability of the table	By hand or through pressure-washed for thorough cleaning	*
	Water Drainage system	Several drain holes provided throughout the table	*
	The table motor	Should be enclosed in watertight casing and should be enclosed in corrosion resistant receptacles and switches	*
Miscellaneous Parameters	<b>Comprehensive Warranty in years</b>	5	5
	User/Technical/Maintenance manuals to be supplied in English in hard and soft copy	Yes	*
	Demonstration of equipment and training to be provided after completing supplies & Installation	Yes	*
	Certificate of calibration	Inspection to be provided at time of supply of equipment	*
	List of equipment available for providing calibration and routine preventive maintenance check list as per manufacturers	Yes	*
	Log book maintenance	Log book with instructions for daily ,weekly,monthly and quarterly maintenance check list to be provided	*
	Authenticity of	The Principal	*

	Manufacturer	Manufacturer must have direct Presence/approved service center In India	
	Cost of Spares	List of important spare parts and accessories with their part no and the cost prevailing at time of supplies to be given to consignee/buyer	*
	Copies of reports and certifications to be furnished to buyer on demand	Yes	*
	After sales service	"On site after sales services to be provided"	*
	Certification to supply the spare parts	Certificate from the principal shall be provided to the buyer/consignee along with the equipment that spare parts will be available for next 10 years	*

\* Specifications highlighted in bold are the Golden Parameters.

\* Bidders may note that In respect of non-golden Parameters, the specifications 'Values' chosen by Buyer will generally be preferred over 'Bid requirement ( allowed Values) by the Buyer.

### Additional Specification Documents

### Consignees/Reporting Officer and Quantity

S.No.	Consignee/Reporting Officer	Address	Quantity	Delivery Days
1	*****	*****MEERUT	1	15

### Special terms and conditions for category Canine Hydraulic Surgery and Examination Table

1. Comprehensive warrantyComprehensive warranty shall include preventive maintenance including calibration as per technical/ service /operational manual of the manufacturer, service charges and spares,. During the warranty period commencing from date of the successful completion of warranty period, Service personnel shall visit each consignee site as recommended in the manufacturer's technical/ service /operational manual, at least once in six months. warranty shall not be including the consumables .Further there will be 98% uptime warranty during warranty period on 24 (hrs) X 7 (days) X 365 (days) basis, with

penalty, to extend warranty period by double the downtime period. Service centres Details of Service outlets in India to render services for equipment to be furnished to buyer/consignees with complete address ,telephone numbers, e mails etc at time of making the supplies .It shall be the responsibility of seller to ensure that authorized service centres are available to cater to the areas where supplies are made within reasonable distance from where the service calls can be handled .Details of toll free numbers for service call and online registration of service requests also to be provided buyer/consignee at the time of supplies. Source of supply It shall be responsibility of seller to provide Documents regarding source of equipments such as copy of Performa invoice or any other documents to establish that the products supplied are manufactured by OEM indicated and sourced from them. Packing and Marking Medical equipments being very delicate and sensitive packing for the goods should be strong and durable enough to withstand transit including transhipment (if any), rough handling, open storage etc. without any damage, deterioration etc. .The size, weights and volumes of the packing cases, remoteness of the final destination of the goods, availability or otherwise of transport and handling facilities at all points during transit up to final destination,. Quality of packing, the manner of marking within & outside the packages and provision of accompanying documentation shall take in to consideration the type of medical equipments being supplied. The accessories shall be suitably labelled and packed .Each of the package shall be marked on three sides with indelible paint of proper quality: indicating contract number and date , brief description of goods including quantity ,. Packing list reference number , country of origin of goods and any other relevant details. Spare Parts Seller shall provide materials, information etc. pertaining to spare parts manufactured and supplied by the OEM . It shall be ensured that the required spares are available for purchase at least for 10 years from date of supplies .In case due to any reasons the production of the spare parts is discontinued sufficient advance notice should be given to the buyer/consignee before such discontinuation to provide adequate time to purchase the required spare parts etc. Further, OEM and their service centres/dealers shall carry sufficient inventories to assure ex-stock supply of consumables and spares for the equipments so that the same are available. OEM or reseller shall always accord most favoured client status to the buyer/consignee and shall give the most competitive price for spares and consumables of its machines/equipments supplied. Installation, Training, Manuals Seller shall be responsible to carry out Installation & commissioning, Supervision and Demonstration of the goods. They shall provide required jigs and tools for assembly, minor civil works for the completion of the installation and Training of Consignee's representatives for operating and maintaining the equipment and Supplying required number of operation & maintenance manual for the goods. In case the category parameters are specifying any requirements regarding the installations , training and manuals the same shall also be applicable. Electrical safety checking Sellers are required to make sure that they furnish the list of equipments for carrying out routine and preventive maintenance to buyer/consignee. They should make sure to periodically check the electrical safety aspects as per BIS Safety Standards or equivalent .In case they do not have required equipment for such testing should ensure that the equipments checked for electrical safety compliance through labs with facilities for such checking during every preventive maintenance call. Software All software updates should be provided free of cost during warranty period.

## **Bid Specific Additional Terms and Conditions**

1. Scope of supply (Bid price to include all cost components) : Supply Installation Testing and Commissioning of Goods

[This Bid is also governed by the General Terms and Conditions](#)

**---Thank You---**