

3K04 Deliverable 1: Documentation

Group 33

Last Updated: 2025-10-26

Contents

List of Figures	iii
List of Tables	iii
1 Group Members	1
2 Abbreviations	2
2.1 General Abbreviations	2
2.2 Bradycardia Operating Abbreviations	2
3 Part 1	3
3.1 Introduction	3
3.2 Requirements	3
3.2.1 System Requirements	3
3.2.2 Programming Requirements	3
3.2.3 Hardware Requirements	4
3.2.4 DCM Requirements	4
3.3 Design	5
3.3.1 Simulink Design	5
3.3.1.1 Overall Design	5
3.3.1.2 Input Constant Variables	6
3.3.1.3 Monitored Input Variables	7
3.3.1.4 Stateflow Modules	8
3.3.1.5 AOO Stateflow Model	9
3.3.1.6 VOO Stateflow Model	10
3.3.1.7 VII Stateflow Model	11
3.3.1.8 AAI Stateflow Model	12
3.3.1.9 Hardware Hiding	13
3.3.2 DCM Design	14
3.3.2.1 Model	14
User Model	14
Pacing Model	14
Egram Model	14
3.3.2.2 View	14
Login View	15
Main View	15
3.3.2.3 Controller	15
Controller	15
Main	15
3.3.2.4 Views of DCM GUI	16
4 Part 2	18
4.1 Requirements Potential Changes	18
4.2 Design Decision Potential Changes	18
4.3 Module Description	19

4.3.1	Input Constant Variables Module	19
4.3.2	Stateflow Logic	19
4.3.3	Hardware Hiding	19
4.4	Testing	20
4.4.1	SimuLink Mode Testing	20
4.4.1.1	Testing of AOO	20
4.4.1.2	Testing of VOO	22
4.4.1.3	Testing of AAI	24
	No Natural Heart Rate	24
	Natural Heart Rate of 45 BPM	25
	Natural Heart Rate of 75 BPM	26
4.4.1.4	Testing of VVI	27
	No Natural Heart Rate	27
	Natural Heart Rate at 45 BPM	28
	Natural Heart Rate at 75 BPM	29
4.4.1.5	Hysteresis Testing	30
	Hysteresis Test 1 (60 BPM)	30
	Hysteresis Test 2 (50 BPM)	31
	Hysteresis Test 3 (40 BPM)	32
4.4.2	DCM Testing	33
4.4.2.1	Login and Registration	33
4.4.2.2	Parameter Input Validation	36
4.4.2.3	Mode Selection and Data Retrieval	39
4.5	GenAI Usage	40

List of Figures

1	Overall Simulink Mapping	5
2	Constant Input Variables	6
3	Monitored Input Variables	7
4	Stateflow Modules	8
5	AOO Stateflow Model	9
6	VOO Stateflow Model	10
7	VVI Stateflow Model	11
8	AAI Stateflow Model	12
9	Hardware Hiding of Model	13
10	Login Screen of DCM	16
11	Main Menu of DCM	16
12	Data Entry View	17
13	Switch in Device Detection from DCM	17
14	AOO Test	20
15	Close-up of AOO Pulse	21
16	VOO Test	22
17	Close-up of VOO Pulse	23
18	AAI Test No Heart Rate	24
19	AAI Test 45 BPM	25
20	AAI Test 75 BPM	26
21	VVI Test No Heart Rate	27
22	VVI Test 45 BPM	28
23	VVI Test 75 BPM	29
24	Hysteresis Test 1	30
25	Hysteresis Test 2	31
26	Hysteresis Test 3	32
27	Registration Test	33
28	Registration Result	34
29	Login Test	34
30	Login Result	35
31	DCM Parameter Page with Values Filled	36
32	DCM Parameter Error When Number Inputted is Out of Range	37
33	DCM Parameter Error When Non-Number is Inputted	37
34	DCM Parameter Error When URL is Larger Than LRL	38
35	Stored Parameter File	39
36	Stored Users File	39

List of Tables

1	Table of Group Members	1
2	Bradycardia Operating Abbreviations	2

1 Group Members

Table 1: Table of Group Members

Name	MacID	Student Number
Ryan Su	sur21	400507973
Cameron Lin	lin422	400535393
Braden McEachern	mceacb1	400527617
Damian Szydlowski	szydlowd	400512629
Menakan Thamilchelvan	thamilcm	400510755
Yash Panchal		
Said Dokmak		
Ishpreet Bal		

2 Abbreviations

2.1 General Abbreviations

BPM - Beats Per Minute

CCS - Cardiac Conduction System

DCM - Device Controller-Monitor

GPIO - General Purpose Input Output

GUI - Graphical User Interface

PWM - Pulse Width Modulation

2.2 Bradycardia Operating Abbreviations

Table 2: Bradycardia Operating Abbreviations

Category	Chambers Paced	Chambers Sensed	Response to Sensing	Rate Modulation
Letters	O-None	O-None	O-None	R-Rate Modulation
	A-Atrium	A-Atrium	T-Triggered	
	V-Ventricle	V-Ventricle	I-Inhibited	
	D-Dual	D-Dual	D-Tracked	

3 Part 1

3.1 Introduction

It is hard to understate the importance of the human heart. The heart is the core part of the cardiovascular system; supplying nutrients and oxygen to all the cells and removing carbon dioxide, especially to vital organs such as the brain, it is imperative for it to be working flawlessly and harmoniously at all times. Unfortunately, however, cardiovascular diseases are a leading cause of death globally, many of which are caused from complications with abnormal heart rhythms. A pacemaker is an implantable device capable of sending timed electrical impulses causing contractions at appropriate intervals. Understanding the operation and design of this life saving device will aid in developing more efficient and reliable cardiac assistive technology.

The purpose of this project is to design and implement a system that operates a cardiac pacemaker under specified modes. This project will be accomplished through an understanding of embedded systems and through engineering principles of software development.

The scope of this deliverable is to design and implement the embedded pacemaker software, driver software and user interface for the DCM while updating and maintaining documentation.

3.2 Requirements

- Overall system requirements (summarized from provided specification documents). It can be informal or semi-formal.
- Mode-specific requirements: AOO, VOO, AAI, VVI.

3.2.1 System Requirements

System Modes: The system shall implement the single chamber pacing modes AOO, VOO, VVI, and AAI. This will act as the foundational functionality of the pacemaker.

Hardware Hiding: Hardware interactions shall be abstracted in a hardware hiding layer such that logical control signals are mapped directly to hardware pins without direct pin references in the model. This will aid in readability and maintainability of the system.

Implementation: The pacing logic shall be implemented in Simulink and the DCM interface shall be implemented as a separate GUI that will communicate with the pacemaker model. This is to ensure modularity between the interface and embedded system.

3.2.2 Programming Requirements

Programmable Pulse Amplitude: The pacemaker shall generate atrial and ventricle signals with amplitudes that are configurable by the user. Adjustable amplitudes will allow for tuning of pacing strength.

Programmable Pulse Width: The pacemaker shall generate atrial and ventricle signals with pulse widths that are configurable by the user.

Programmable Rate Timing: The pacemaker shall have programmable LRL, lower rate limit, and URL, upper rate limits, that control the minimum and maximum pacing rates. This is to prevent bradycardia, abnormally low heart rhythms, and tachycardia, abnormally high heart rhythms.

Programmable Refractory Periods: Both the atrial and ventricle pulse modes shall both implement a refractory period during which sensed events are ignored to prevent double sensing of pulses or ringing.

Programmable Sensitivity: A programmable sensitivity, or threshold for event detection, shall be implemented to be adjustable by the DCM. This allows for adaptation to patients and noise.

Pacing Responses: Each pacing mode shall follow the response of its corresponding lettering:

O: Asynchronous pacing that ignores senses

I: Inhibited pacing from sensed pulses

T: Triggered pacing from sensed pulses

3.2.3 Hardware Requirements

Hardware Hiding: A layer of hardware abstraction shall be put in place to map digital logic to analog front-end hardware.

Front-End Enable: A signal shall enable or disable front-end sensing circuitry. This allows for controlled activation of ADCs and amplifiers to reduce noise and power consumption.

3.2.4 DCM Requirements

User Authentication: The DCM GUI shall provide user registration and login functionality supporting 10 different stored users. This provides access control for the programming interface.

Parameter Display and Editing: The DCM shall display and allow a user to edit pacemaker settings and parameters such as LRL, URL, Atrial/Ventricle Amplitude, Pulse Width, and Sensitivity.

Status Indicators: The DCM shall provide visible indicators of device connection status and communication loss. This is to ensure the user is aware of the systems states and faults.

3.3 Design

3.3.1 Simulink Design

3.3.1.1 Overall Design




Figure 1: Overall Simulink Mapping

The Pacemaker architecture can be split up into 4 main modules, input constant variables, input monitor variables, stateflow logic for modes, and hardware hiding. Figure 1 below shows the overarching workflow of the system:

3.3.1.2 Input Constant Variables




Figure 2: Constant Input Variables

In the above image, Figure 2, we find changeable variables relating to pacemaker operation. For general inputs, the changeable variables are:

- **Mode** - Refers to bradycardia operating modes, e.g AOO, VOO, AAI and VVI.
- **Low Rate Interval** - The number of generated pace pulses per minute, converted from a millisecond time period.
- **Hysteresis Pace** - When enabled, 1, a longer period is waited before pacing after sensing an event to prevent unwanted pacing pulses from ringing from an event.
- **Hysteresis Interval** - Specifies the time interval waited in the hysteresis mode in milliseconds.

The modifiable atrial variables are:


- **Pace Pulse Width** - Changes the width, length of time, of the pace pulse.
- **Pace Pulse Amplitude** - Changes the amplitude, voltage, of the pace pulse.
- **ARP (Atrial Refractory Period)** - The programmed time interval following an atrial event during which time atrial events shall not inhibit nor trigger pacing
- **Sense (Sensitivity)** - Determines the minimum value an atrial signal must be to be considered by the pacemaker.

The modifiable ventricular variables are:

- **Pace Pulse Width** - Changes the width, length of time, of the pace pulse.
- **Pace Pulse Amplitude** - Changes the amplitude, voltage, of the pace pulse.

- **VRP (Ventricle Refractory Period)** - The programmed time interval following an ventricle event during which time atrial events shall not inhibit nor trigger pacing.
- **Sense (Sensitivity)** - Determines the minimum value a ventricle signal must be to be considered by the pacemaker.

3.3.1.3 Monitored Input Variables



The monitored input variables can be seen in the above Figure 3. These are the atrial and ventricle detection variables. The pulses are sensed with through GPIO pins connecting to a board simulating heart conditions.

3.3.1.4 Stateflow Modules





Figure 4: Stateflow Modules


The stateflow diagram in Simulink is shown above. It shows the transition between each mode given the mode input shown in Figure 2. When switching between modes, a core state is returned to that resets variables to their nominal values.

3.3.1.5 AOO Stateflow Model



The above stateflow, Figure 5, shows the stateflow for the AOO mode. It sets values controlling discharge and charging of the capacitor to specified nominal values.

3.3.1.6 VOO Stateflow Model



Similar to the AOO stateflow, the above figure, Figure 6, shows the states of charging and discharging of the capacitor using specified nominal values.

3.3.1.7 VII Stateflow Model




Figure 7: VVI Stateflow Model

The above stateflow, Figure 7 shows the FSM for the VVI model. The initial state, pVRP, is the state that occurs right after the pacemaker delivers a ventricle pacing pulse. A refractory period occurs during this time, where the pacemaker ignores sensor inputs to prevent sensing of its own produced pulse or electrical ringing. After a certain amount of time, the pReady state is transitioned to where sensor inputs are allowed. This allows for the pacemaker to detect natural heart rhythms and correct heart rhythms accordingly, or solely deliver pacing pulses. The last state before going to the initial state is the discharging state where the pacing pulse is produced.

3.3.1.8 AAI Stateflow Model




Figure 8: AAI Stateflow Model

The above stateflow, Figure 8 shows the process for the AAI mode. The initial state, pARP, is the state occurs right after the pacemaker delivers an atrial pacing pulse. During this time, the pacemaker ignores sensing events to prevent it from sensing its own delivered pulse. The next state is transitioned to after a set time period, the refractory period, if no natural heartbeat is detected after a certain time interval, the discharging state is then transitioned to. However, if a natural heartbeat is detected, another refractory period occurs to prevent the pacemaker from sensing ringing. This state then transitions to the discharging state once an appropriate time has passed.

3.3.1.9 Hardware Hiding




Figure 9: Hardware Hiding of Model

The above image, Figure 9 shows abstraction of the GPIO pin functions. It connects logical output signals to the pins on the FRDM-K64F. This allows for better readability, thus making the code easier to maintain, debug and safer.

3.3.2 DCM Design

The design of the DCM has 3 separate sections; the **model, view and controller**.

3.3.2.1 Model

The model section handles the applications data and user logic. It does not interact directly with the interface but only manages data and enforces user authentication. Below is an explanation of the individual Python files and their functionality:

User Model

Purpose: To manage all user account information.

Functions:

- Loads and saves login information to or from the users.json file.
- Registers new users, checking for duplicates and the max user limit.
- Authenticates logins by verifying user passwords
- Provides a method to get the amount of registered users which is displayed by the DCM.

Pacing Model

Purpose: Manages pacing parameters for all users.

Functions:

- Loads and saves settings to or from pacing_settings.json
- The data is structured as a dictionary with settings as the key and value as the value.
- Saves a dictionary of parameters for a specific user and pacemaker mode
- Retrieves the saved parameters for a user and mode

Egram Model

Purpose: Manages the capture of electrogram data

Functions:

- Adds a single data point to an internal list
- Returns the list of captured data
- Clears the captured data to start a new session

3.3.2.2 View

The view section is what the user sees and interacts with. It is built with the customtkinter library for a more aesthetically pleasing interface. They do not have any control logic, only to display data and send user actions such as button clicks to the controller. Below are an explanation of each view:

Login View

Purpose: Provides a screen for users to login

Functions:

- Welcome class displays the login form and a register button.
- Register class displays the new user registration form and a back button.
- When a button is clicked, it collects text from the entry fields and calls a method on the controller that handles the login logic.
- The welcome view calls a function to display the current registered user count.

Main View

Purpose: Provides the main application screen after logging in

Functions:

- MainFrame is the main menu. It shows buttons for each pacing mode AOO, VOO, AAI and VVI as well as mock connection controls. It also displays the current connection status.
- DataEntry is the form for editing parameters. It displays all parameters and handles validation of data such as the input being in the correct range and is a numeric type before sending the data to the controller.

3.3.2.3 Controller

Controller

Purpose: The main application class that uses instances of the models and view frames.



Functions:

- It initializes itself and creates all view frames, UserModel and PacingModel. It then stores these in a dictionary.
- The navigation is handled by a show_frame function which requests a frame from the dictionary and uses tkinter's library function, tkraise(), to bring it into view.
- Event handling is separated into handling registration and saving settings. For registration it takes the username and password from the view, calls the UserModel to verify them, then shows a message to confirm or deny entry. The saving of settings takes mode and data from the DataEntry view and passes them to the PacingModel to be saved.

Main

Purpose: Its purpose and sole function is to create an instance of the controller and run it in the main loop, starting the customtkinter application.

3.3.2.4 Views of DCM GUI






Figure 12: Data Entry View




Figure 13: Switch in Device Detection from DCM

4 Part 2

4.1 Requirements Potential Changes

Pacing Modes: Additional pacing modes are to be developed such as AOOR, VOOR, AAIR, and VVIR.

Hardware Communication: The connection between hardware and GUI may be developed in the next deliverable. The functionality implemented in this deliverable will be used to interface with the board.

Parameter Changes: As additional pacing modes are to be developed, this in turn will bring additional parameters that will need to be added and configured.

Electrogram Functionality: The electrogram functionality will be used to store and display egram data provided by the simulating board.

4.2 Design Decision Potential Changes

Expansion of Libraries: As hardware communication will be implemented in the next iteration, hardware communication libraries such as those implementing serial communication will need to be explored.

Expansion of GUI: To accomodate for more pacing modes, the GUI will need to be modified. Furthermore, for improved quality of life features, more graphics and interactive features may be added.

DCM Architecture: The DCM system may need to be refactored to implement hardware communication.

Simulink Architecture: As 4 more additional modes are being added for a total of 8 pacing modes, the Simulink structure may need to be modified to integrate all new features more easily.

Data Verification: As data will be actively transfered from the DCM to the pacemaker hardware, data integrity further becomes more crucial to maintain. A system to verify parameters are correctly sent and saved will need to be implemented.

4.3 Module Description

Although briefly highlighted in the design section of this documentation. This section will go into more detail about the purpose of each module, key functionality, variables, and how each module interacts with one another.

As shown in Figure 1, there are 3 primary modules operating the pacemaker; **input constant variables, stateflow logic for modes, and hardware hiding.**

4.3.1 Input Constant Variables Module

This module is highlighted in the design section. It goes through all the state variables that are changeable parameters of the pacemaker. These inputs include general inputs such as the pacing mode, low rate interval, and hysteresis settings, as well as parameters for atrial and ventricle pacing. These parameters are then used in the stateflow logic module to complete pacing to user specifications.

4.3.2 Stateflow Logic

This module has many submodules which are also covered in the design section. The overall state machine controlling mode selection and resetting of variables to prevent unwanted pacing behaviour is shown in Figure 4. This module converts input parameters into raw data that can be further converted to electrical signals. This module receives data from the input constant variables module as well as from the monitored input variables. This module then feeds into hardware hiding.

4.3.3 Hardware Hiding

Hardware hiding is also briefly covered in the design section. The purpose of this module is to convert the signals from the stateflow logic module into outputted electrical signals through pins. As shown in Figure 9, pins are mapped to certain electrical signals such as atrial outputs, ventricle outputs, front end signals, and general pin configurations such as grounding pins. This is designed to ensure coding and modules are easier to debug with higher level identification and function of each GPIO.

4.4 Testing

4.4.1 SimuLink Mode Testing

4.4.1.1 Testing of AOO

Purpose: The purpose of this test is to test basic AOO functionality.

Input Conditions: General inputs of mode = 0, AOO, and hysteresis = 0, off, and standard atrial inputs.

Expected Output: Consistent, evenly spaced pulses in the output with disregard for natural heart beats.

Actual Output: Output of testing is exactly that of expected, as shown below in Figure 10.

Result: Pass




Figure 11 below shows a zoomed in view of one of the pulses in the AOO test above, Figure 10.




Figure 15: Close-up of AOO Pulse

4.4.1.2 Testing of VOO

Purpose: The purpose of this test is to test basic VOO functionality.

Input Conditions: General inputs of mode = 1, VOO, and hysteresis = 0, off, and standard ventricle inputs.

Expected Output: Consistent, evenly spaced pulses in the output with disregard for natural heart beats.

Actual Output: Output of testing is exactly that of expected, as shown below in Figure 12.

Result: Pass




Figure 16: VOO Test

Figure 13 below shows a zoomed in view of one of the pulses in the VOO test above, Figure 12




Figure 17: Close-up of VOO Pulse

4.4.1.3 Testing of AAI

No Natural Heart Rate


Purpose: The purpose of this test is to test basic AAI functionality with no natural heart rhythms.

Input Conditions: General inputs of mode = 2, AAI, and hysteresis = 0, off, and standard artial inputs.

Expected Output: Consistent, evenly spaced pulses in the output.

Actual Output: Output of testing is exactly that of expected, as shown below in Figure 13.

Result: Pass



Natural Heart Rate of 45 BPM


Purpose: The purpose of this test is to test basic AAI functionality at a low heart rate.

Input Conditions: General inputs of mode = 2, AAI, and hysteresis = 0, off, standard artial inputs, and monitored atrial pulses at 45 BPM.

Expected Output: The pacemaker should pulse after the refractory period expires. An output of blue pulses right before the natural red pulses is expected.

Actual Output: Output of testing is exactly that of expected, as shown below in Figure 15.

Result: Pass



Natural Heart Rate of 75 BPM


Purpose: The purpose of this test is to test basic AAI functionality at a nominal heart rate.

Input Conditions: General inputs of mode = 2, AAI, and hysteresis = 0, off, standard artial inputs, and monitored atrial pulses at 75 BPM.

Expected Output: As a nominal heart reate is being inputted, the pacemaker should not be delivering pacing pulses as the simulated heart rate is nominal and healthy.

Actual Output: Output of testing is exactly that of expected, as shown below in Figure 16.

Result: Pass



4.4.1.4 Testing of VVI

No Natural Heart Rate


Purpose: The purpose of this test is to test basic VVI functionality with no inputted heart rate.

Input Conditions: General inputs of mode = 3, VVI, and hysteresis = 0, off, standard ventricle inputs, and monitored ventricle pulses.

Expected Output: Consistent, evenly spaced pulses in the output.

Actual Output: Output of testing is exactly that of expected, as shown below in Figure 17.

Result: Pass



Natural Heart Rate at 45 BPM


Purpose: The purpose of this test is to test basic VVI functionality with no inputted heart rate.

Input Conditions: General inputs of mode = 3, VVI, and hysteresis = 0, off, standard ventricle inputs, and monitored ventricle pulses.

Expected Output: The pacemaker should pulse after the refractory period expires. An output of blue pulses right before the natural red pulses is expected.

Actual Output: Output of testing is exactly that of expected, as shown below in Figure 18.

Result: Pass



Natural Heart Rate at 75 BPM


Purpose: The purpose of this test is to test basic VVI functionality with no inputted heart rate.

Input Conditions: General inputs of mode = 3, VVI, and hysteresis = 0, off, standard ventricle inputs, and monitored ventricle pulses.

Expected Output: As a nominal heart rate is being inputted, the pacemaker should not be delivering pacing pulses as the simulated heart rate is nominal and healthy.

Actual Output: Output of testing is exactly that of expected, as shown below in Figure 18.

Result: Pass



4.4.1.5 Hysteresis Testing

Hysteresis Test 1 (60 BPM)


Purpose: The purpose of this test is to test basic hysteresis mode functionality.

Input Conditions: General inputs of mode = 2, AAI, and hysteresis = 1, on, standard atrium inputs, and monitored atrial pulses at 50 BPM.

Expected Output: No output of the pacemaker.

Actual Output: Output of testing is exactly that of expected, as shown below in Figure 20.

Result: Pass



Hysteresis Test 2 (50 BPM)


Purpose: The purpose of this test is to test basic hysteresis mode functionality.

Input Conditions: General inputs of mode = 2, AAI, and hysteresis = 1, on, standard atrium inputs, and monitored atrial pulses at 50 BPM.

Expected Output: No output of the pacemaker.

Actual Output: Output of testing is exactly that of expected, as shown below in Figure 21.

Result: Pass



Hysteresis Test 3 (40 BPM)


Purpose: The purpose of this test is to test basic hysteresis mode functionality.

Input Conditions: General inputs of mode = 2, AAI, and hysteresis = 1, on, standard atrium inputs, and monitored atrial pulses at 50 BPM.

Expected Output: Delayed output signals from the pacemaker.

Actual Output: Output of testing is exactly that of expected, as shown below in Figure 22.

Result: Pass



4.4.2 DCM Testing

4.4.2.1 Login and Registration

Purpose: The purpose of this test is to verify correct storage of newly registered user data and allow login.

Input Conditions: A random username and password. This will be used again in the login screen to access the DCM.

Expected Output: A window should pop up notifying the user an account has been registered. The DCM controls should be accessible after the user logs in.

Actual Output: Dialogue is shown and the file is updated to include new user data. The user is then brought to the

Result: Pass




Figure 27: Registration Test




Figure 28: Registration Result




Figure 29: Login Test




Figure 30: Login Result

4.4.2.2 Parameter Input Validation

Purpose: To enforce numeric types within an allowed range and to ensure the upper rate interval is greater than lower rate interval.


Input Conditions: Entering a non-numeric, an out of range number and an upper rate interval that is greater than lower rate interval in parameter settings.

Expected Output: Invalid input dialogue is shown and parameter changes are not saved.

Actual Output:

Result: Pass

The below figures show the general parameter page with some values inputted as well as the results of putting invalid values into parameter page.



The screenshot shows a web application window titled "DCM - Device Control Monitor" with a subtitle "Editing Parameters for: ADO". The page contains a list of parameters with their current values and ranges:

Parameter	Value	Range
Lower Rate Limit (bpm)	123	120 - 170
Upper Rate Limit (bpm)	123	120 - 170
Atrial Amplitude (mV)	2	0.5 - 7.0
Atrial Pulse Width (ms)	1	0.5 - 5.0
Ventricular Amplitude (mV)		0.5 - 7.0
Ventricular Pulse Width (ms)		0.5 - 5.0
VPP (V)		1.0 - 9.0
AMP (V)		1.0 - 9.0

At the bottom of the page, there are two buttons: "Back to Device Selection" and "Save Parameters".

Figure 31: DCM Parameter Page with Values Filled




Figure 32: DCM Parameter Error When Number Inputted is Out of Range




Figure 33: DCM Parameter Error When Non-Number is Inputted




Figure 34: DCM Parameter Error When URL is Larger Than LRL

4.4.2.3 Mode Selection and Data Retrieval

Purpose: To test data storage, ensuring proper saving of user data

Input Conditions: Registering account, logging in, and saving parameters.

Expected Output: User data is now found in the associated JSON files.

Actual Output: Registered user data and parameters are found in their respective JSON files.

Result: Pass

This test was done in conjunction to previous tests except with a different user registered. A user "asd" with password "asd" was used for faster log ins. The following images are of the JSON files and the saved parameters from the previous test.

```
3K04-Project > DCM > models > {} pacing_settings.json > ...
1  {
2    "asd": {
3      "A00": {
4        "Lower Rate Limit": "123",
5        "Upper Rate Limit": "123",
6        "Atrial Amplitude": "2",
7        "Atrial Pulse Width": "1.9"
8      }
9    }
10 }
```

Figure 35: Stored Parameter File

```
3K04-Project > DCM > models > {} users.json > ...
1  {
2    "asd": "asd"
3  }
```

Figure 36: Stored Users File

4.5 GenAI Usage

We used a Generative AI assistant to support development of the DCM. It provided starter boilerplate for a Tkinter app with a welcome screen, registration and login, JSON storage capped at ten users, which we then adapted and tested. We also used it to clarify Python functions and libraries such as Tkinter, JSON, etc. and to troubleshoot installing tkinter. We had AI to clarify comments within the code as well. All design decisions, requirements, and validation were done by our team, and we reviewed and verified all AI outputs before inclusion.