

Turbidity sensor setup/calibration:

- We need to tell the sensor what 0 NTU and 100 NTU look like
- Be careful with glassware
- Wipe bottles "clear"
- Fill open container with tap water
- "Experiment" → "Calibrate" → "turbidity"
- Problems?
 - Wipe inside of sensor
 - Recalibrate