



**Astellas BQT Assay Report**  
**Test Article Report**

---

**Assay Details**

---

**User Information**  
User Name: harding  
Computer Name: DESKTOP-RFHI5SO  
Logon Server: \\DESKTOP-RFHI5SO  
User Domain: DESKTOP-RFHI5SO  
Astellas BQT Infectivity PLA Script Version 0.1  
JMP Version 18.1.0

---

Analyst Signature/Date

---

Approver Signature/Date

Astellas BQT Infectivity Files

First Data File	Second Data File
18OCT2024_Plate01_KL-S3	18OCT2024_Plate01_KL-S4

50% L01-240910\_1 & Reference Standard Data

Group	Sampling	N Rows	Vg/mL	Accepted	Std		MOI	Log10 MOI	Log10 Vg/mL	Jackknife z	Residual	Std Dev(Residual	Outlier Within Group	Externally Outlier	
				Droplets	Dev(Vg/mL)	CV(Vg/mL)					Vg/mL	Vg/mL)		Studentized Residuals	Between Group
50% L01-240910_1	13	2	1.797e+10	19656.5	359645621.8	2.0013493063	1.6e+1	1.2041199827	10.254551883	3.508	2016984707.3	975150051.65	Ok	1.528	Ok
50% L01-240910_1	14	2	1.130e+10	20631	235053015.22	2.0794349286	8e+0	0.903089987	10.053220492	4.319	2850624619.9	1279054942.5	Outlier	2.970	Ok
50% L01-240910_1	15	2	5.246e+9	20620.5	325698869.5	6.2090657061	4e+0	0.6020599913	9.7197899954	2.437	750892487.00	483150314.72	Ok	1.512	Ok
50% L01-240910_1	16	2	2.863e+9	20059	133535135.16	4.6637199832	2e+0	0.3010299957	9.4568630835	2.669	368152542.00	200351201.56	Ok	1.872	Ok
50% L01-240910_1	17	2	1.555e+10	19717.5	489935793.6	3.1497867731	1.6e+1	1.2041199827	10.191858014	0.233	-398602596.8	975150051.65	Ok	-0.139	Ok
50% L01-240910_1	18	2	8.920e+9	19954.5	145520323.79	1.6314329748	8e+0	0.903089987	9.9503544163	0.309	466713638.63	1279054942.5	Ok	0.527	Ok
50% L01-240910_1	19	2	4.446e+9	20536.5	22592759.573	0.5081526529	4e+0	0.6020599913	9.6479750835	0.071	-48587418.25	483150314.72	Ok	-0.226	Ok
50% L01-240910_1	20	2	2.457e+9	20746.5	3583547.517	0.1458313452	2e+0	0.3010299957	9.3904622856	0.115	-37799377.35	200351201.56	Ok	0.061	Ok
50% L01-240910_1	21	2	1.433e+10	18854	691328511.13	4.8227318281	1.6e+1	1.2041199827	10.156391355	1.421	-1618382111	975150051.65	Ok	-1.061	Ok
50% L01-240910_1	22	2	7.986e+9	19410.5	272371628.48	3.410460901	8e+0	0.903089987	9.9023487925	1.261	-466713638.6	1279054942.5	Ok	-0.586	Ok
50% L01-240910_1	23	2	3.792e+9	19610	17206271.137	0.4537112028	4e+0	0.6020599913	9.5789072595	1.863	-702305068.8	483150314.72	Ok	-2.010	Ok
50% L01-240910_1	24	2	2.165e+9	19152	13728570.717	0.6341816577	2e+0	0.3010299957	9.3354116483	1.726	-330353164.6	200351201.56	Ok	-1.373	Ok
Ref.Std (L01-240910)	1	2	3.467e+10	18613.5	587669438.42	1.6949040913	1.6e+1	1.2041199827	10.539987978	1.175	-735884400.0	975150051.65	Ok	-0.903	Ok
Ref.Std (L01-240910)	2	2	1.622e+10	19311.5	463230896.97	2.8555810473	8e+0	0.903089987	10.210103028	2.499	-2504692995	1279054942.5	Ok	-0.843	Ok
Ref.Std (L01-240910)	3	2	7.657e+9	19803.5	231970536.3	3.0296105469	4e+0	0.6020599913	9.8840460225	3.606	-657906493.8	483150314.72	Ok	-0.782	Ok
Ref.Std (L01-240910)	4	2	3.761e+9	19680.5	94431706.818	2.5108464532	2e+0	0.3010299957	9.5752976845	0.366	32829494.625	200351201.56	Ok	-0.354	Ok
Ref.Std (L01-240910)	5	2	3.614e+10	19200	2238782751.2	6.1939800559	1.6e+1	1.2041199827	10.55804215	0.356	735884400.00	975150051.65	Ok	-0.415	Ok
Ref.Std (L01-240910)	6	2	1.892e+10	19002.5	282808622.15	1.4950841497	8e+0	0.903089987	10.276827009	0.083	189258117.50	1279054942.5	Ok	0.729	Ok
Ref.Std (L01-240910)	7	2	8.450e+9	18863	81272179.859	0.96178488	4e+0	0.6020599913	9.9268639631	0.243	135457229.00	483150314.72	Ok	0.211	Ok
Ref.Std (L01-240910)	8	2	3.695e+9	18885.5	10054551.433	0.2720908438	2e+0	0.3010299957	9.5676487726	1.157	-32829494.63	200351201.56	Ok	-0.548	Ok
Ref.Std (L01-240910)	9	2	4.060e+10	18942	2841771707.6	7.0001149946	1.6e+1	1.2041199827	10.608484012	4.985	5187462150.0	975150051.65	Outlier	0.923	Ok
Ref.Std (L01-240910)	10	2	2.104e+10	20346	443443222.92	2.1074118715	8e+0	0.903089987	10.3230886	1.823	2315434877.0	1279054942.5	Ok	1.985	Ok
Ref.Std (L01-240910)	11	2	8.837e+9	20066	112974810.7	1.2784101859	4e+0	0.6020599913	9.9463114	1.397	522449264.75	483150314.72	Ok	0.663	Ok
Ref.Std (L01-240910)	12	2	3.968e+9	20399	132165288.52	3.3311430456	2e+0	0.3010299957	9.5985241259	5.157	239444102.63	200351201.56	Within Analytical Error	0.224	Ok

Within Group Jackknife z Outlier Limit (≥): 4  
Between Group Externally Studentized Residuals Outlier Limit (≥): 4  
Accepted Droplets: Invalid wells with < 10,000 accepted droplets excluded from all calculations


50% L01-240910\_1 Test Sample & Reference Standard Summary Statistics

Group	MOI	N Rows	Mean(Vg/mL)	Std Dev(Vg/mL)
Ref.Std (L01-240910)	2e+0	3	3.81e+9	1.42e+8
Ref.Std (L01-240910)	4e+0	3	8.31e+9	6.02e+8
Ref.Std (L01-240910)	8e+0	3	1.9e+10	2.42e+9
Ref.Std (L01-240910)	1.6e+1	2	3.5e+10	1.04e+9
50% L01-240910_1	2e+0	3	2.5e+9	3.51e+8
50% L01-240910_1	4e+0	3	4.49e+9	7.28e+8
50% L01-240910_1	8e+0	2	8.45e+9	6.6e+8
50% L01-240910_1	1.6e+1	3	1.6e+10	1.85e+9

50% L01-240910\_1 Model Selection

Model	Parallelism	Linearity	Validity		Selected Model
	Slope Ratio	Ratio	R2	RMSE Evaluation	
Model 2, Low Standard and Test Doses Excluded	0.869	3.430	0.979	0.047 Parallel and Linear	Model 2, Low Standard and Test Doses Excluded
Model 1, All Doses	0.822	1.105	0.987	0.045 Parallel and Linear	
Model 3, High Standard and Test Doses Excluded	0.769	1.427	0.981	0.046 Parallel and Linear	
Model 4, Standard Low Dose Only Excluded	0.849	0.157	0.986	0.048 Parallel and Linear	
Model 6, Test Low Dose Only Excluded	0.841	3.152	0.985	0.044 Parallel and Linear	
Model 8, Standard Low Dose and Test High Dose Excluded	0.835	3.136	0.987	0.050 Parallel and Linear	
Model 7, Test High Dose Only Excluded	0.808	2.668	0.988	0.046 Parallel and Linear	
Model 9, Standard High Dose and Test Low Dose Excluded	0.801	0.704	0.979	0.044 Parallel and Linear	
Model 5, Standard High Dose Only Excluded	0.783	1.343	0.983	0.045 Parallel and Linear	

50% L01-240910\_1 Graphs



50% L01-240910\_1 Validity Report

Validity Criteria	LSL	USL	Validity Results	Assay Validity	Overall Validity
Dose Response Test	.	0.05	0.000	Passed Validity Criteria	Assay is Valid
Reference Standard Curve Depth	2720000000	.	31600673338	Passed Validity Criteria	
% Relative Infectivity Delta (Constrained - Unconstrained)	.	15	0.074	Passed Validity Criteria	
Parallelism Slope Ratio	0.7	1.4	0.869	Passed Validity Criteria	
Linearity Ratio	.	26.3	3.430	Passed Validity Criteria	
Unconstrained EC50 Standard	0.04	61.8	4.760	Passed Validity Criteria	
Number of Wells that Failed Accepted Droplets (<10000)	.	5	0.000	Passed Validity Criteria	

50% L01-240910\_1 Relative Infectivity and Infectious Particle Ratio

		RI	Reference	Relative Infectivity	Assay RI	Assay RI	Upper	Lower	CI Range as %					
EC50 Ref	EC50 Test	Uncorrected	Reference CF	Stability CF	Reportable Result	Upper 95%	Lower 95%	Spec Limit	Spec Limit	CI Range	of Tolerance	CI Range % of Tolerance	Check	OOS Validity
5.26	11.16	47.1	0	0	47.1	51.7	42.6	150	50	9.1	9.1	Bioassay Results are Reportable Assay is Valid and OOS		

Unconstrained RI	Constrained RI	Relative Infectivity Delta
47.2	47.1	0.1

Infectious Particle Ratio	Infectious Particle Ratio Lower Limit	Infectious Particle Ratio Upper Limit
0.6	0.3	1.0

150% L01-240910\_1 & Reference Standard Data

Group	Sampling	N Rows	Vg/mL	Accepted		Std		MOI	Log10 MOI	Log10 Vg/mL	Jackknife z	Residual	Std Dev(Residual	Outlier Within Group	Externally Outlier	
				Droplets	Dev(Vg/mL)	CV(Vg/mL)	Vg/mL					Vg/mL)	Studentized Residuals		Between Group	
150% L01-240910_1	25	2	4.333e+10	19449.5	1415326191.4	3.2663094426	1.6e+1	1.2041199827	10.636799217	2.752	498233642.50	975150051.65	Ok		0.240	Ok
150% L01-240910_1	26	2	2.527e+10	19744	642816777.6	2.5434988531	8e+0	0.903089987	10.402655657	13.343	2495354001.8	1279054942.5	Within Analytical Error		1.422	Ok
150% L01-240910_1	27	2	1.199e+10	19708.5	370403067.43	3.0902382741	4e+0	0.6020599913	10.078682607	7.469	816961441.50	483150314.72	Within Analytical Error		-0.354	Ok
150% L01-240910_1	28	2	7.037e+9	20511	40166095.497	0.5707730578	2e+0	0.3010299957	9.8473961516	0.935	39120598.125	200351201.56	Ok		1.012	Ok
150% L01-240910_1	29	2	4.278e+10	19064	380998669.04	0.8906886483	1.6e+1	1.2041199827	10.631197541	0.129	-57075807.50	975150051.65	Ok		0.049	Ok
150% L01-240910_1	30	2	2.291e+10	19624.5	1437202301	6.2733021375	8e+0	0.903089987	10.360021699	0.565	132241517.25	1279054942.5	Ok		0.008	Ok
150% L01-240910_1	31	2	1.109e+10	20716	889172320.19	8.0163895138	4e+0	0.6020599913	10.045007123	1.003	-77338485.00	483150314.72	Ok		-1.514	Ok
150% L01-240910_1	32	2	6.959e+9	19792	77086896.492	1.1077457191	2e+0	0.3010299957	9.842540481	0.511	-39120598.13	200351201.56	Ok		0.826	Ok
150% L01-240910_1	33	2	4.239e+10	18382	931314530.59	2.1969287211	1.6e+1	1.2041199827	10.627280412	1.685	-441157835.0	975150051.65	Ok		-0.085	Ok
150% L01-240910_1	34	2	2.265e+10	19918.5	228694495.38	1.0098965802	8e+0	0.903089987	10.35497881	0.865	-132241517.3	1279054942.5	Ok		-0.150	Ok
150% L01-240910_1	35	2	1.125e+10	20499	114153032.32	1.0149997396	4e+0	0.6020599913	10.051021522	0.463	77338485.000	483150314.72	Ok		-1.289	Ok
150% L01-240910_1	36	2	6.474e+9	19781.5	69066883.248	1.0668379646	2e+0	0.3010299957	9.8111713961	9.472	-524038123.9	200351201.56	Outlier		-0.317	Ok
Ref.Std (L01-240910)	1	2	3.467e+10	18613.5	587669438.42	1.6949040913	1.6e+1	1.2041199827	10.539987978	1.175	-735884400.0	975150051.65	Ok		-1.266	Ok
Ref.Std (L01-240910)	2	2	1.622e+10	19311.5	463230896.97	2.8555810473	8e+0	0.903089987	10.210103028	2.499	-2504692995	1279054942.5	Ok		-1.178	Ok
Ref.Std (L01-240910)	3	2	7.657e+9	19803.5	231970536.3	3.0296105469	4e+0	0.6020599913	9.8840460225	3.606	-657906493.8	483150314.72	Ok		-1.090	Ok
Ref.Std (L01-240910)	4	2	3.761e+9	19680.5	94431706.818	2.5108464532	2e+0	0.3010299957	9.5752976845	0.366	32829494.625	200351201.56	Ok		-0.488	Ok
Ref.Std (L01-240910)	5	2	3.614e+10	19200	2238782751.2	6.1939800559	1.6e+1	1.2041199827	10.55804215	0.356	735884400.00	975150051.65	Ok		-0.572	Ok
Ref.Std (L01-240910)	6	2	1.892e+10	19002.5	282808622.15	1.4950841497	8e+0	0.903089987	10.276827009	0.083	189258117.50	1279054942.5	Ok		1.014	Ok
Ref.Std (L01-240910)	7	2	8.450e+9	18863	81272179.859	0.96178488	4e+0	0.6020599913	9.9268639631	0.243	135457229.00	483150314.72	Ok		0.290	Ok
Ref.Std (L01-240910)	8	2	3.695e+9	18885.5	10054551.433	0.2720908438	2e+0	0.3010299957	9.5676487726	1.157	-32829494.63	200351201.56	Ok		-0.757	Ok
Ref.Std (L01-240910)	9	2	4.060e+10	18942	2841771707.6	7.0001149946	1.6e+1	1.2041199827	10.608484012	4.985	5187462150.0	975150051.65	Outlier		1.267	Ok
Ref.Std (L01-240910)	10	2	2.104e+10	20346	443443222.92	2.1074118715	8e+0	0.903089987	10.3230886	1.823	2315434877.0	1279054942.5	Ok		3.053	Ok
Ref.Std (L01-240910)	11	2	8.837e+9	20066	112974810.7	1.2784101859	4e+0	0.6020599913	9.9463114	1.397	522449264.75	483150314.72	Ok		0.921	Ok
Ref.Std (L01-240910)	12	2	3.968e+9	20399	132165288.52	3.3311430456	2e+0	0.3010299957	9.5985241259	5.157	239444102.63	200351201.56	Within Analytical Error		0.308	Ok

Within Group Jackknife z Outlier Limit (≥): 4  
Between Group Externally Studentized Residuals Outlier Limit (≥): 4  
Accepted Droplets: Invalid wells with < 10,000 accepted droplets excluded from all calculations


150% L01-240910\_1 Test Sample & Reference Standard Summary Statistics

Group	MOI	N Rows	Std	
			Mean(Vg/mL)	Dev(Vg/mL)
Ref.Std (L01-240910)	2e+0	3	3.81e+9	1.42e+8
Ref.Std (L01-240910)	4e+0	3	8.31e+9	6.02e+8
Ref.Std (L01-240910)	8e+0	3	1.9e+10	2.42e+9
Ref.Std (L01-240910)	1.6e+1	2	3.5e+10	1.04e+9
150% L01-240910_1	2e+0	2	7e+9	5.53e+7
150% L01-240910_1	4e+0	3	1.1e+10	4.78e+8
150% L01-240910_1	8e+0	3	2.4e+10	1.45e+9
150% L01-240910_1	1.6e+1	3	4.3e+10	4.72e+8

150% L01-240910\_1 Model Selection

Model	Parallelism		Linearity		Validity	
	Slope Ratio	Ratio	R2	RMSE Evaluation	Selected Model	
Model 2, Low Standard and Test Doses Excluded	0.903	5.183	0.987	0.034 Parallel and Linear	Model 2, Low Standard and Test Doses Excluded	
Model 1, All Doses	0.823	0.482	0.992	0.033 Parallel and Linear		
Model 3, High Standard and Test Doses Excluded	0.776	4.452	0.987	0.035 Parallel and Linear		
Model 6, Test Low Dose Only Excluded	0.874	3.985	0.993	0.032 Parallel and Linear		
Model 4, Standard Low Dose Only Excluded	0.850	0.942	0.988	0.035 Parallel and Linear		
Model 8, Standard Low Dose and Test High Dose Excluded	0.842	1.866	0.980	0.038 Parallel and Linear		
Model 9, Standard High Dose and Test Low Dose Excluded	0.832	2.539	0.994	0.030 Parallel and Linear		
Model 7, Test High Dose Only Excluded	0.816	1.216	0.989	0.036 Parallel and Linear		
Model 5, Standard High Dose Only Excluded	0.783	2.392	0.993	0.032 Parallel and Linear		

150% L01-240910\_1 Graphs



150% L01-240910\_1 Validity Report

Validity Criteria	Validity		Overall Validity
	LSL	USL	
Dose Response Test	-	0.05	0.000 Passed Validity Criteria
Reference Standard Curve Depth	2720000000	-	31600673338 Passed Validity Criteria
% Relative Infectivity Delta (Constrained - Unconstrained)	-	15	0.246 Passed Validity Criteria
Parallelism Slope Ratio	0.7	1.4	0.903 Passed Validity Criteria
Linearity Ratio	-	26.3	5.183 Passed Validity Criteria
Unconstrained EC50 Standard	0.04	61.8	4.760 Passed Validity Criteria
Number of Wells that Failed Accepted Droplets (<10000)	-	5	0.000 Passed Validity Criteria

150% L01-240910\_1 Relative Infectivity and Infectious Particle Ratio

		RI		Reference	Relative Infectivity	Assay RI	Assay RI	Upper	Lower		CI Range as %		
EC50 Ref	EC50 Test	Uncorrected	Reference CF	Stability CF	Reportable Result	Upper 95%	Lower 95%	Spec Limit	Spec Limit	CI Range	of Tolerance	CI Range % of Tolerance	Check OOS Validity
8.77	6.83	128.4	0	0	128.4	136.4	121.1	150	50	15.2	15.2	Bioassay Results are Reportable Assay is Valid and Within Limits	
		Relative											
Unconstrained RI	Constrained RI	Infectivity Delta											
128.2	128.4	0.2											
Infectious Particle Ratio	Infectious Particle Ratio Lower Limit	Infectious Particle Ratio Upper Limit											
1.7	0.3	1.0											



100% L01-240910 & Reference Standard Data

Group	Sampling	N Rows	Vg/mL	Accepted		Std		MOI	Log10 MOI	Log10 Vg/mL	Jackknife z	Residual	Std Dev(Residual	Outlier Within Group	Externally Outlier	
				Droplets	Dev(Vg/mL)	CV(Vg/mL)						Vg/mL	Vg/mL)		Studentized Residuals	Between Group
100% L01-240910	49	2	3.584e+10	18158.5	513669225.3	1.4330867287		1.6e+1	1.2041199827	10.554411073	1.078	-529564818.8	975150051.65	Ok	-0.941	Ok
100% L01-240910	50	2	1.644e+10	18226.5	20350893.08	0.1237600509		8e+0	0.903089987	10.216002993	2.888	-2146416014	1279054942.5	Ok	-0.964	Ok
100% L01-240910	51	2	7.608e+9	19725	256471113.81	3.3709586709		4e+0	0.6020599913	9.8812850299	2.601	-651703948.3	483150314.72	Ok	-0.965	Ok
100% L01-240910	52	2	3.716e+9	17735.5	289685040.08	7.7950422058		2e+0	0.3010299957	9.570107597	0.049	-6072311.750	200351201.56	Ok	-0.287	Ok
100% L01-240910	53	2	3.690e+10	19450	1728233892.7	4.6832199996		1.6e+1	1.2041199827	10.567057958	0.413	529564818.75	975150051.65	Ok	-0.500	Ok
100% L01-240910	54	2	1.888e+10	18863.5	258789690.42	1.3710651781		8e+0	0.903089987	10.27588887	0.151	284836422.00	1279054942.5	Ok	0.854	Ok
100% L01-240910	55	2	8.320e+9	18746	56525600.254	0.6793927511		4e+0	0.6020599913	9.9201242741	0.102	60059737.750	483150314.72	Ok	0.198	Ok
100% L01-240910	56	2	3.593e+9	12685.5	.	.		2e+0	0.3010299957	9.5554301725	1.940	-129568938.5	200351201.56	Ok	-0.766	Ok
100% L01-240910	57	2	4.094e+10	18744	1870104954.7	4.5675232426		1.6e+1	1.2041199827	10.612185215	6.103	4570404048.8	975150051.65	Outlier	1.034	Ok
100% L01-240910	58	2	2.045e+10	20813.5	573367617.57	2.8035032798		8e+0	0.903089987	10.310732093	1.624	1861579591.5	1279054942.5	Ok	2.097	Ok
100% L01-240910	59	2	8.852e+9	19702.5	137595150.53	1.5544659683		4e+0	0.6020599913	9.9470219092	1.763	591644210.50	483150314.72	Ok	1.014	Ok
100% L01-240910	60	2	3.858e+9	19416	144262472.04	3.7393203403		2e+0	0.3010299957	9.586360698	2.330	135641250.25	200351201.56	Ok	0.232	Ok
Ref.Std (L01-240910)	1	2	3.467e+10	18613.5	587669438.42	1.6949040913		1.6e+1	1.2041199827	10.539987978	1.175	-735884400.0	975150051.65	Ok	-1.175	Ok
Ref.Std (L01-240910)	2	2	1.622e+10	19311.5	463230896.97	2.8555810473		8e+0	0.903089987	10.210103028	2.499	-2504692995	1279054942.5	Ok	-1.094	Ok
Ref.Std (L01-240910)	3	2	7.657e+9	19803.5	231970536.3	3.0296105469		4e+0	0.6020599913	9.8840460225	3.606	-657906493.8	483150314.72	Ok	-1.013	Ok
Ref.Std (L01-240910)	4	2	3.761e+9	19680.5	94431706.818	2.5108464532		2e+0	0.3010299957	9.5752976845	0.366	32829494.625	200351201.56	Ok	-0.455	Ok
Ref.Std (L01-240910)	5	2	3.614e+10	19200	2238782751.2	6.1939800559		1.6e+1	1.2041199827	10.55804215	0.356	735884400.00	975150051.65	Ok	-0.533	Ok
Ref.Std (L01-240910)	6	2	1.892e+10	19002.5	282808622.15	1.4950841497		8e+0	0.903089987	10.276827009	0.083	189258117.50	1279054942.5	Ok	0.944	Ok
Ref.Std (L01-240910)	7	2	8.450e+9	18863	81272179.859	0.96178488		4e+0	0.6020599913	9.9268639631	0.243	135457229.00	483150314.72	Ok	0.270	Ok
Ref.Std (L01-240910)	8	2	3.695e+9	18885.5	10054551.433	0.2720908438		2e+0	0.3010299957	9.5676487726	1.157	-32829494.63	200351201.56	Ok	-0.706	Ok
Ref.Std (L01-240910)	9	2	4.060e+10	18942	2841771707.6	7.0001149946		1.6e+1	1.2041199827	10.608484012	4.985	5187462150.0	975150051.65	Outlier	1.183	Ok
Ref.Std (L01-240910)	10	2	2.104e+10	20346	443443222.92	2.1074118715		8e+0	0.903089987	10.3230886	1.823	2315434877.0	1279054942.5	Ok	2.755	Ok
Ref.Std (L01-240910)	11	2	8.837e+9	20066	112974810.7	1.2784101859		4e+0	0.6020599913	9.9463114	1.397	522449264.75	483150314.72	Ok	0.857	Ok
Ref.Std (L01-240910)	12	2	3.968e+9	20399	132165288.52	3.3311430456		2e+0	0.3010299957	9.5985241259	5.157	239444102.63	200351201.56	Within Analytical Error	0.288	Ok

Within Group Jackknife z Outlier Limit (≥): 4  
Between Group Externally Studentized Residuals Outlier Limit (≥): 4  
Accepted Droplets: Invalid wells with < 10,000 accepted droplets excluded from all calculations


100% L01-240910 Test Sample & Reference Standard Summary Statistics

Group	MOI	N Rows	Std	
			Mean(Vg/mL)	Dev(Vg/mL)
Ref.Std (L01-240910)	2e+0	3	3.81e+9	1.42e+8
Ref.Std (L01-240910)	4e+0	3	8.31e+9	6.02e+8
Ref.Std (L01-240910)	8e+0	3	1.9e+10	2.42e+9
Ref.Std (L01-240910)	1.6e+1	2	3.5e+10	1.04e+9
100% L01-240910	2e+0	3	3.72e+9	1.33e+8
100% L01-240910	4e+0	3	8.26e+9	6.24e+8
100% L01-240910	8e+0	3	1.9e+10	2.02e+9
100% L01-240910	1.6e+1	2	3.6e+10	7.49e+8

100% L01-240910 Model Selection

Model	Parallelism		Linearity		Validity	
	Slope Ratio	Ratio	R2	RMSE Evaluation	Selected Model	
Model 3, High Standard and Test Doses Excluded	1.011	0.658	0.989	0.034 Parallel and Linear	Model 3, High Standard and Test Doses Excluded	
Model 1, All Doses	1.019	3.047	0.992	0.035 Parallel and Linear		
Model 2, Low Standard and Test Doses Excluded	1.021	4.978	0.981	0.040 Parallel and Linear		
Model 6, Test Low Dose Only Excluded	0.989	3.757	0.990	0.037 Parallel and Linear		
Model 5, Standard High Dose Only Excluded	0.970	2.144	0.991	0.034 Parallel and Linear		
Model 4, Standard Low Dose Only Excluded	1.052	3.664	0.990	0.037 Parallel and Linear		
Model 9, Standard High Dose and Test Low Dose Excluded	0.941	2.400	0.989	0.037 Parallel and Linear		
Model 7, Test High Dose Only Excluded	1.062	2.580	0.991	0.035 Parallel and Linear		
Model 8, Standard Low Dose and Test High Dose Excluded	1.097	3.519	0.989	0.038 Parallel and Linear		

100% L01-240910 Graphs



100% L01-240910 Validity Report

Validity Criteria	LSL		USL		Validity		Overall Validity
	Dose Response Test				Results Assay Validity	Assay Validity	
Reference Standard Curve Depth	2720000000	.	0.05	31600673338	0.000 Passed Validity Criteria	Assay is Valid	
% Relative Infectivity Delta (Constrained - Unconstrained)		.	15	0.000	Passed Validity Criteria		
Parallelism Slope Ratio	0.7		1.4	1.011	Passed Validity Criteria		
Linearity Ratio	.		26.3	0.658	Passed Validity Criteria		
Unconstrained EC50 Standard	0.04		61.8	4.760	Passed Validity Criteria		
Number of Wells that Failed Accepted Droplets (<10000)	.		5	0.000	Passed Validity Criteria		

100% L01-240910 Relative Infectivity and Infectious Particle Ratio

		RI		Reference	Relative Infectivity	Assay RI	Assay RI	Upper	Lower		CI Range as %		
EC50 Ref	EC50 Test	Uncorrected	Reference CF	Stability CF	Reportable Result	Upper 95%	Lower 95%	Spec Limit	Spec Limit	CI Range	of Tolerance	CI Range % of Tolerance Check	OOS Validity
3.98	4.02	99.0	0	0	99.0	103.8	94.4	150	50	9.3	9.3	Bioassay Results are Reportable Assay is Valid and Within Limits	
		Relative											
Unconstrained RI	Constrained RI	Infectivity Delta											
99.0	99.0	0.0											
Infectious Particle Ratio	Infectious Particle Ratio	Infectious Particle Ratio											
1.3	0.3	1.0											

50% L01-240910\_2 & Reference Standard Data

Group	Sampling	N Rows	Vg/mL	Accepted		Std		MOI	Log10 MOI	Log10 Vg/mL	Jackknife z	Residual	Std Dev(Residual		Externally Outlier	
				Droplets	Dev(Vg/mL)	CV(Vg/mL)						Vg/mL	Vg/mL	Outlier Within Group	Studentized Residuals	Between Group
50% L01-240910_2	61	2	1.807e+10	18507	67065519.327	0.371135865	1.6e+1	1.2041199827	10.256966367	3.436	1852530364.8	975150051.65	Ok		0.865	Ok
50% L01-240910_2	62	2	1.102e+10	19613.5	190120617.32	1.7258574881	8e+0	0.903089987	10.042024285	2.575	1694588623.5	1279054942.5	Ok		2.435	Ok
50% L01-240910_2	63	2	5.159e+9	19546	199264473.2	3.8622715539	4e+0	0.6020599913	9.7125870699	2.102	690231131.75	483150314.72	Ok		1.087	Ok
50% L01-240910_2	64	2	2.761e+9	19386.5	142914635.79	5.1765314071	2e+0	0.3010299957	9.4410378536	2.485	318069076.60	200351201.56	Ok		1.368	Ok
50% L01-240910_2	65	2	1.586e+10	19528	426059891.54	2.6857978728	1.6e+1	1.2041199827	10.200397327	0.224	-354373015.3	975150051.65	Ok		-0.434	Ok
50% L01-240910_2	66	2	9.172e+9	19640	381089296.09	4.1548290727	8e+0	0.903089987	9.9624735888	0.098	-149218902.8	1279054942.5	Ok		0.455	Ok
50% L01-240910_2	67	2	4.472e+9	20154	205743731.17	4.6005207441	4e+0	0.6020599913	9.6505196182	0.005	3158760.500	483150314.72	Ok		-0.242	Ok
50% L01-240910_2	68	2	2.420e+9	19772	116728454.59	4.8244854223	2e+0	0.3010299957	9.3837257373	0.080	-23248787.30	200351201.56	Ok		0.007	Ok
50% L01-240910_2	69	2	1.472e+10	19282.5	392975759.31	2.6697350163	1.6e+1	1.2041199827	10.167897604	1.440	-1498157350	975150051.65	Ok		-1.213	Ok
50% L01-240910_2	70	2	7.776e+9	19753.5	311276312.65	4.0030128902	8e+0	0.903089987	9.8907590846	1.778	-1545369721	1279054942.5	Ok		-1.077	Ok
50% L01-240910_2	71	2	3.776e+9	19467	89472972.022	2.3697464095	4e+0	0.6020599913	9.5769899896	2.141	-693389892.3	483150314.72	Ok		-1.955	Ok
50% L01-240910_2	72	2	2.148e+9	19516	136835293.16	6.3705683843	2e+0	0.3010299957	9.332019945	1.832	-294820289.3	200351201.56	Ok		-1.208	Ok
Ref.Std (L01-240910)	1	2	3.467e+10	18613.5	587669438.42	1.6949040913	1.6e+1	1.2041199827	10.539987978	1.175	-735884400.0	975150051.65	Ok		-0.821	Ok
Ref.Std (L01-240910)	2	2	1.622e+10	19311.5	463230896.97	2.8555810473	8e+0	0.903089987	10.210103028	2.499	-2504692995	1279054942.5	Ok		-0.767	Ok
Ref.Std (L01-240910)	3	2	7.657e+9	19803.5	231970536.3	3.0296105469	4e+0	0.6020599913	9.8840460225	3.606	-657906493.8	483150314.72	Ok		-0.711	Ok
Ref.Std (L01-240910)	4	2	3.761e+9	19680.5	94431706.818	2.5108464532	2e+0	0.3010299957	9.5752976845	0.366	32829494.625	200351201.56	Ok		-0.323	Ok
Ref.Std (L01-240910)	5	2	3.614e+10	19200	2238782751.2	6.1939800559	1.6e+1	1.2041199827	10.55804215	0.356	735884400.00	975150051.65	Ok		-0.379	Ok
Ref.Std (L01-240910)	6	2	1.892e+10	19002.5	282808622.15	1.4950841497	8e+0	0.903089987	10.276827009	0.083	189258117.50	1279054942.5	Ok		0.664	Ok
Ref.Std (L01-240910)	7	2	8.450e+9	18863	81272179.859	0.96178488	4e+0	0.6020599913	9.9268639631	0.243	135457229.00	483150314.72	Ok		0.193	Ok
Ref.Std (L01-240910)	8	2	3.695e+9	18885.5	10054551.433	0.2720908438	2e+0	0.3010299957	9.5676487726	1.157	-32829494.63	200351201.56	Ok		-0.499	Ok
Ref.Std (L01-240910)	9	2	4.060e+10	18942	2841771707.6	7.0001149946	1.6e+1	1.2041199827	10.608484012	4.985	5187462150.0	975150051.65	Outlier		0.842	Ok
Ref.Std (L01-240910)	10	2	2.104e+10	20346	443443222.92	2.1074118715	8e+0	0.903089987	10.3230886	1.823	2315434877.0	1279054942.5	Ok		1.770	Ok
Ref.Std (L01-240910)	11	2	8.837e+9	20066	112974810.7	1.2784101859	4e+0	0.6020599913	9.9463114	1.397	522449264.75	483150314.72	Ok		0.604	Ok
Ref.Std (L01-240910)	12	2	3.968e+9	20399	132165288.52	3.3311430456	2e+0	0.3010299957	9.5985241259	5.157	239444102.63	200351201.56	Within Analytical Error		0.205	Ok

Within Group Jackknife z Outlier Limit (≥): 4  
Between Group Externally Studentized Residuals Outlier Limit (≥): 4  
Accepted Droplets: Invalid wells with < 10,000 accepted droplets excluded from all calculations


50% L01-240910\_2 Test Sample & Reference Standard Summary Statistics

Group	MOI	N Rows	Std	
			Mean(Vg/mL)	Dev(Vg/mL)
Ref.Std (L01-240910)	2e+0	3	3.81e+9	1.42e+8
Ref.Std (L01-240910)	4e+0	3	8.31e+9	6.02e+8
Ref.Std (L01-240910)	8e+0	3	1.9e+10	2.42e+9
Ref.Std (L01-240910)	1.6e+1	2	3.5e+10	1.04e+9
50% L01-240910_2	2e+0	3	2.44e+9	3.07e+8
50% L01-240910_2	4e+0	3	4.47e+9	6.92e+8
50% L01-240910_2	8e+0	3	9.32e+9	1.63e+9
50% L01-240910_2	1.6e+1	3	1.6e+10	1.7e+9

50% L01-240910\_2 Model Selection

Model	Parallelism		Linearity		Validity	
	Slope Ratio	Ratio	R2	RMSE Evaluation	Selected Model	
Model 2, Low Standard and Test Doses Excluded	0.884	6.084	0.971	0.054 Parallel and Linear	Model 2, Low Standard and Test Doses Excluded	
Model 1, All Doses	0.849	2.251	0.984	0.049 Parallel and Linear		
Model 3, High Standard and Test Doses Excluded	0.840	2.744	0.976	0.052 Parallel and Linear		
Model 8, Standard Low Dose and Test High Dose Excluded	0.912	0.584	0.983	0.055 Parallel and Linear		
Model 7, Test High Dose Only Excluded	0.883	1.844	0.984	0.050 Parallel and Linear		
Model 4, Standard Low Dose Only Excluded	0.877	2.164	0.982	0.053 Parallel and Linear		
Model 6, Test Low Dose Only Excluded	0.856	4.405	0.979	0.050 Parallel and Linear		
Model 9, Standard High Dose and Test Low Dose Excluded	0.815	3.758	0.970	0.051 Parallel and Linear		
Model 5, Standard High Dose Only Excluded	0.808	0.704	0.979	0.050 Parallel and Linear		

50% L01-240910\_2 Graphs





50% L01-240910\_2 Validity Report

Validity Criteria	LSL	USL	Validity Results	Assay Validity	Overall Validity
Dose Response Test	.	0.05	0.000	Passed Validity Criteria	Assay is Valid
Reference Standard Curve Depth	2720000000	.	31600673338	Passed Validity Criteria	
% Relative Infectivity Delta (Constrained - Unconstrained)	.	15	0.191	Passed Validity Criteria	
Parallelism Slope Ratio	0.7	1.4	0.884	Passed Validity Criteria	
Linearity Ratio	.	26.3	6.084	Passed Validity Criteria	
Unconstrained EC50 Standard	0.04	61.8	4.760	Passed Validity Criteria	
Number of Wells that Failed Accepted Droplets (< 10000)	.	5	0.000	Passed Validity Criteria	

50% L01-240910\_2 Relative Infectivity and Infectious Particle Ratio

EC50 Ref	EC50 Test	RI Uncorrected	Reference CF	Reference Stability CF	Relative Infectivity Reportable Result	Assay RI Upper 95%	Assay RI Lower 95%	Upper Spec Limit	Lower Spec Limit	CI Range	CI Range as % of Tolerance	CI Range % of Tolerance Check	OOS Validity
5.27	10.73	49.1	0	0	49.1	54.0	44.3	150	50	9.7	9.7	Bioassay Results are Reportable	Assay is Valid and OOS

Unconstrained RI	Constrained RI	Relative Infectivity Delta
49.3	49.1	0.2

Infectious Particle Ratio	Infectious Particle Ratio Lower Limit	Infectious Particle Ratio Upper Limit
0.7	0.3	1.0

150% L01-240910\_2 & Reference Standard Data

Group	Sampling	N Rows	Vg/mL	Accepted Droplets	Std Dev(Vg/mL)	CV(Vg/mL)	MOI	Log10 MOI	Log10 Vg/mL	Jackknife z	Residual Vg/mL	Std Dev(Residual Vg/mL)	Outlier Within Group	Externally Studentized Residuals	Outlier Between Group
150% L01-240910_2	73	2	4.346e+10	20242.5	164290989.13	0.3780442877	1.6e+1	1.2041199827	10.638071064	0.720	1326603.750	975150051.65	Ok		0.291 Ok
150% L01-240910_2	74	2	2.554e+10	19859.5	172756060.68	0.6764341882	8e+0	0.903089987	10.407207743	18.374	2516037139.5	1279054942.5	Outlier		1.586 Ok
150% L01-240910_2	75	2	1.238e+10	20354	532896929.11	4.304486152	4e+0	0.6020599913	10.092721903	8.443	954160308.75	483150314.72	Outlier		0.031 Ok
150% L01-240910_2	76	2	6.956e+9	20033.5	95618993.904	1.3746652481	2e+0	0.3010299957	9.8423472161	0.242	71649780.750	200351201.56	Ok		0.722 Ok
150% L01-240910_2	77	2	4.315e+10	20507.5	567354794.46	1.3147474299	1.6e+1	1.2041199827	10.635012399	161.86	-303666686.3	975150051.65	Within Analytical Error		0.180 Ok
150% L01-240910_2	78	2	2.312e+10	20364	1420489654.8	6.1439823415	8e+0	0.903089987	10.363988116	0.602	96825714.000	1279054942.5	Ok		0.102 Ok
150% L01-240910_2	79	2	1.135e+10	21188.5	621413242.94	5.4769542447	4e+0	0.6020599913	10.054841391	0.966	-79911804.00	483150314.72	Ok		-1.293 Ok
150% L01-240910_2	80	2	7.162e+9	19824.5	294757875.4	4.1154812625	2e+0	0.3010299957	9.8550447888	1.400	278020400.25	200351201.56	Ok		1.224 Ok
150% L01-240910_2	81	2	4.346e+10	19857	175599334.22	0.4040901983	1.6e+1	1.2041199827	10.638044549	0.695	-1326603.750	975150051.65	Ok		0.290 Ok
150% L01-240910_2	82	2	2.293e+10	20353	694780318.29	3.0304863146	8e+0	0.903089987	10.36033518	0.820	-96825714.00	1279054942.5	Ok		-0.020 Ok
150% L01-240910_2	83	2	1.151e+10	19997.5	194500111.81	1.690454602	4e+0	0.6020599913	10.060916343	0.489	79911804.000	483150314.72	Ok		-1.064 Ok
150% L01-240910_2	84	2	6.534e+9	19613	156104914.68	2.3889408814	2e+0	0.3010299957	9.8152111738	3.594	-349670181.0	200351201.56	Ok		-0.272 Ok
Ref.Std (L01-240910)	1	2	3.467e+10	18613.5	587669438.42	1.6949040913	1.6e+1	1.2041199827	10.539987978	1.175	-735884400.0	975150051.65	Ok		-1.342 Ok
Ref.Std (L01-240910)	2	2	1.622e+10	19311.5	463230896.97	2.8555810473	8e+0	0.903089987	10.210103028	2.499	-2504692995	1279054942.5	Ok		-1.248 Ok
Ref.Std (L01-240910)	3	2	7.657e+9	19803.5	231970536.3	3.0296105469	4e+0	0.6020599913	9.8840460225	3.606	-657906493.8	483150314.72	Ok		-1.153 Ok
Ref.Std (L01-240910)	4	2	3.761e+9	19680.5	94431706.818	2.5108464532	2e+0	0.3010299957	9.5752976845	0.366	32829494.625	200351201.56	Ok		-0.513 Ok
Ref.Std (L01-240910)	5	2	3.614e+10	19200	2238782751.2	6.1939800559	1.6e+1	1.2041199827	10.55804215	0.356	735884400.00	975150051.65	Ok		-0.602 Ok
Ref.Std (L01-240910)	6	2	1.892e+10	19002.5	282808622.15	1.4950841497	8e+0	0.903089987	10.276827009	0.083	189258117.50	1279054942.5	Ok		1.072 Ok
Ref.Std (L01-240910)	7	2	8.450e+9	18863	81272179.859	0.96178488	4e+0	0.6020599913	9.9268639631	0.243	135457229.00	483150314.72	Ok		0.305 Ok
Ref.Std (L01-240910)	8	2	3.695e+9	18885.5	10054551.433	0.2720908438	2e+0	0.3010299957	9.5676487726	1.157	-32829494.63	200351201.56	Ok		-0.799 Ok
Ref.Std (L01-240910)	9	2	4.060e+10	18942	2841771707.6	7.0001149946	1.6e+1	1.2041199827	10.608484012	4.985	5187462150.0	975150051.65	Outlier		1.334 Ok
Ref.Std (L01-240910)	10	2	2.104e+10	20346	443443222.92	2.1074118715	8e+0	0.903089987	10.3230886	1.823	2315434877.0	1279054942.5	Ok		3.375 Ok
Ref.Std (L01-240910)	11	2	8.837e+9	20066	112974810.7	1.2784101859	4e+0	0.6020599913	9.9463114	1.397	522449264.75	483150314.72	Ok		0.972 Ok
Ref.Std (L01-240910)	12	2	3.968e+9	20399	132165288.52	3.3311430456	2e+0	0.3010299957	9.5985241259	5.157	239444102.63	200351201.56	Within Analytical Error		0.324 Ok

Within Group Jackknife z Outlier Limit (≥): 4  
Between Group Externally Studentized Residuals Outlier Limit (≥): 4  
Accepted Droplets: Invalid wells with < 10,000 accepted droplets excluded from all calculations


150% L01-240910\_2 Test Sample & Reference Standard Summary Statistics

Group	MOI	N Rows	Mean(Vg/mL)	Std Dev(Vg/mL)
Ref.Std (L01-240910)	2e+0	3	3.81e+9	1.42e+8
Ref.Std (L01-240910)	4e+0	3	8.31e+9	6.02e+8
Ref.Std (L01-240910)	8e+0	3	1.9e+10	2.42e+9
Ref.Std (L01-240910)	1.6e+1	2	3.5e+10	1.04e+9
150% L01-240910_2	2e+0	3	6.88e+9	3.2e+8
150% L01-240910_2	4e+0	2	1.1e+10	1.13e+8
150% L01-240910_2	8e+0	2	2.3e+10	1.37e+8
150% L01-240910_2	1.6e+1	3	4.3e+10	1.75e+8

150% L01-240910\_2 Model Selection

Model	Parallelism Slope Ratio	Linearity Ratio	R2	Validity RMSE Evaluation	Selected Model
Model 2, Low Standard and Test Doses Excluded	0.909	4.414	0.989	0.032 Parallel and Linear	Model 2, Low Standard and Test Doses Excluded
Model 1, All Doses	0.820	0.283	0.994	0.031 Parallel and Linear	
Model 3, High Standard and Test Doses Excluded	0.754	3.494	0.989	0.032 Parallel and Linear	
Model 6, Test Low Dose Only Excluded	0.880	3.484	0.994	0.030 Parallel and Linear	
Model 4, Standard Low Dose Only Excluded	0.847	1.444	0.990	0.033 Parallel and Linear	
Model 9, Standard High Dose and Test Low Dose Excluded	0.837	0.583	0.995	0.028 Parallel and Linear	
Model 8, Standard Low Dose and Test High Dose Excluded	0.818	0.569	0.985	0.036 Parallel and Linear	
Model 7, Test High Dose Only Excluded	0.792	1.988	0.991	0.034 Parallel and Linear	
Model 5, Standard High Dose Only Excluded	0.781	3.014	0.994	0.029 Parallel and Linear	

150% L01-240910\_2 Graphs





BQT Infectivity\_18Nov2024-14-14-03

## 150% L01-240910\_2 Validity Report

Validity Criteria	LSL	USL	Validity Results	Assay Validity	Overall Validity
Dose Response Test	.	0.05	0.000	Passed Validity Criteria	Assay is Valid
Reference Standard Curve Depth	2720000000	.	31600673338	Passed Validity Criteria	
% Relative Infectivity Delta (Constrained - Unconstrained)	.	15	0.299	Passed Validity Criteria	
Parallelism Slope Ratio	0.7	1.4	0.909	Passed Validity Criteria	
Linearity Ratio	.	26.3	4.414	Passed Validity Criteria	
Unconstrained EC50 Standard	0.04	61.8	4.760	Passed Validity Criteria	
Number of Wells that Failed Accepted Droplets (<10000)	.	5	0.000	Passed Validity Criteria	

### 150% L01-240910\_2 Relative Infectivity and Infectious Particle Ratio

EC50 Ref		EC50 Test	RI		Reference	Relative Infectivity	Assay RI	Assay RI	Upper	Lower	CI Range as %			
8.94		7.05	Uncorrected	Reference CF	Stability CF	Reportable Result	Upper 95%	Lower 95%	Spec Limit	Spec Limit	CI Range	of Tolerance	CI Range % of Tolerance	Check OOS Validity
			126.7	0	0	126.7	135.5	118.9	150	50	16.6	16.6	Bioassay Results are Reportable Assay is Valid and Within Limits	
				Relative										
Unconstrained RI		Constrained RI		Infectivity Delta										
127.0		126.7		0.3										
Infectious Particle Ratio		Infectious Particle Ratio Lower Limit		Infectious Particle Ratio Upper Limit										
1.7		0.3		1.0										

### Relative Infectivity All Samples

Sample Name	EC50 Standard	EC50 Test	Infectious			
			Ratio	Reportable RI	RI Lower 95	RI Upper 95
50% L01-240910_1	5.2600554481	11.157346029	0.6	47.1	42.6	51.7
150% L01-240910_1	8.7673145259	6.8278698132	1.7	128.4	121.1	136.4
200% L01-240910	8.8323386383	5.8607726426	2.0	150.7	141.5	160.9
100% L01-240910	3.9797170781	4.0203862953	1.3	99.0	94.4	103.8
50% L01-240910_2	5.2698045757	7.034794737	0.7	49.1	44.3	54.0
150% L01-240910_2	8.9351490447	7.0504906686	1.7	126.7	118.9	135.5

Sample Name	Overall		Reportable
	Validity	OOS	
50% L01-240910_1	Assay is Valid	OOS	Reportable
150% L01-240910_1	Assay is Valid	Within Limits	Reportable
200% L01-240910	Assay is Valid	OOS	Reportable
100% L01-240910	Assay is Valid	Within Limits	Reportable
50% L01-240910_2	Assay is Valid	OOS	Reportable
150% L01-240910_2	Assay is Valid	Within Limits	Reportable

## Astellas BQT Infectivity Bioassay Materials and Reference Standard Report

### Assay Details

Assay	Date Assay		Bioassay	Analyst	Signal	Method	Instrument ID	Instrument internal no.	Bioassay preparation (date_operator)	Bioassay review (date_reviewer)
	Site:	Initiated:								
Astellas BQT Infectivity			PLA	BQT Test Run 1	Vg/mL	KT430				

Notes  
Assay Range Check

## Materials

	Reagents	Material	Source	Catalog#	Lot#	Expiration Date
1						
2						
3						
4						
5						
6						
7						
8						

### Reference Details

Reference		Expiry/		RS Correction	RS Stability	
Reference/Control	Standard (RS)	Description	Lot#	Reevaluation	Factor	Correction Factor
1	Ref.Std	Test	Test	Test	0	0



[illegible]

BQT Infectivity\_18Nov2024-14-14-03

Well	Sample		Sample		Sample		Sample		Conc(copies/ µL)		Status	Experiment	SampleType	TargetType	Supermix	DyeName(s)	Accepted		
	description 1	description 2	description 3	description 4	Target				Droplets	Positives							Negatives		
B03	8	RS	REP3		BDNF	1423.709229	OK	DQ	Unknown	Unknown	ddPCR Supermix for Probes (No dUTP)	FAM					21479	15075	6404
A03	16	RS	REP3		BDNF	2840.367188	OK	DQ	Unknown	Unknown	ddPCR Supermix for Probes (No dUTP)	FAM					19714	17951	1763
E10	NTC				BDNF	No Call	CHECK	DQ	Unknown	Unknown	ddPCR Supermix for Probes (no dUTP)	FAM					19126	0	19126
E11	NTC				BDNF	No Call	CHECK	DQ	Unknown	Unknown	ddPCR Supermix for Probes (no dUTP)	FAM					20944	0	20944
E12	NTC				BDNF	No Call	CHECK	DQ	Unknown	Unknown	ddPCR Supermix for Probes (no dUTP)	FAM					18778	0	18778
F10	PC				BDNF	1280.408691	OK	DQ	Unknown	Unknown	ddPCR Supermix for Probes (no dUTP)	FAM					18318	12149	6169
F11	PC				BDNF	1218.153198	OK	DQ	Unknown	Unknown	ddPCR Supermix for Probes (no dUTP)	FAM					19227	12400	6827
F12	PC				BDNF	1202.61377	OK	DQ	Unknown	Unknown	ddPCR Supermix for Probes (no dUTP)	FAM					17535	11226	6309

Well	Sample		Sample		Sample		Sample		Conc(copies/ µL)		Status	Experiment	SampleType	TargetType	Supermix	DyeName(s)	Accepted		
	description 1	description 2	description 3	description 4	Target				Droplets	Positives							Negatives		
D01	2	RS	REP1		BDNF	127.5908127	OK	DQ	Unknown	Unknown	ddPCR Supermix for Probes (No dUTP)	FAM					19041	1957	17084
C01	4	RS	REP1		BDNF	260.693512	OK	DQ	Unknown	Unknown	ddPCR Supermix for Probes (No dUTP)	FAM					18807	3738	15069
B01	8	RS	REP1		BDNF	551.6500854	OK	DQ	Unknown	Unknown	ddPCR Supermix for Probes (No dUTP)	FAM					19425	7271	12154
A01	16	RS	REP1		BDNF	1169.609009	OK	DQ	Unknown	Unknown	ddPCR Supermix for Probes (No dUTP)	FAM					16931	10666	6265
D10	2	200	REP1		BDNF	348.1827393	OK	DQ	Unknown	Unknown	ddPCR Supermix for Probes (No dUTP)	FAM					20431	5234	15197
C10	4	200	REP1		BDNF	450.1703796	OK	DQ	Unknown	Unknown	ddPCR Supermix for Probes (No dUTP)	FAM					20145	6405	13740
B10	8	200	REP1		BDNF	985.618042	OK	DQ	Unknown	Unknown	ddPCR Supermix for Probes (No dUTP)	FAM					19405	11009	8396
A10	16	200	REP1		BDNF	1667.197388	OK	DQ	Unknown	Unknown	ddPCR Supermix for Probes (No dUTP)	FAM					19306	14626	4680
H07	2	150.2	REP1		BDNF	234.1138306	OK	DQ	Unknown	Unknown	ddPCR Supermix for Probes (No dUTP)	FAM					20427	3686	16741
G07	4	150.2	REP1		BDNF	425.228363	OK	DQ	Unknown	Unknown	ddPCR Supermix for Probes (No dUTP)	FAM					19787	6002	13785
F07	8	150.2	REP1		BDNF	847.2356567	OK	DQ	Unknown	Unknown	ddPCR Supermix for Probes (No dUTP)	FAM					20344	10443	9901
E07	16	150.2	REP1		BDNF	1452.476807	OK	DQ	Unknown	Unknown	ddPCR Supermix for Probes (No dUTP)	FAM					19756	14008	5748
D07	2	150	REP1		BDNF	235.5180359	OK	DQ	Unknown	Unknown	ddPCR Supermix for Probes (No dUTP)	FAM					20813	3776	17037
C07	4	150	REP1		BDNF	408.2714844	OK	DQ	Unknown	Unknown	ddPCR Supermix for Probes (No dUTP)	FAM					20299	5952	14347
B07	8	150	REP1		BDNF	857.5824585	OK	DQ	Unknown	Unknown	ddPCR Supermix for Probes (No dUTP)	FAM					19821	10259	9562
A07	16	150	REP1		BDNF	1411.008789	OK	DQ	Unknown	Unknown	ddPCR Supermix for Probes (No dUTP)	FAM					19095	13340	5755
H01	2	100.2	REP1		BDNF	130.7037048	OK	DQ	Unknown	Unknown	ddPCR Supermix for Probes (No dUTP)	FAM					17071	1795	15276
G01	4	100.2	REP1		BDNF	259.6535645	OK	DQ	Unknown	Unknown	ddPCR Supermix for Probes (No dUTP)	FAM					18435	3651	14784
F01	8	100.2	REP1		BDNF	548.6073608	OK	DQ	Unknown	Unknown	ddPCR Supermix for Probes (No dUTP)	FAM					17108	6376	10732
E01	16	100.2	REP1		BDNF	1206.892456	OK	DQ	Unknown	Unknown	ddPCR Supermix for Probes (No dUTP)	FAM					18313	11748	6565
H04	2	50.2	REP1		BDNF	95.39581299	OK	DQ	Unknown	Unknown	ddPCR Supermix for Probes (No dUTP)	FAM					19811	1543	18268
G04	4	50.2	REP1		BDNF	176.6719055	OK	DQ	Unknown	Unknown	ddPCR Supermix for Probes (No dUTP)	FAM					18976	2646	16330
F04	8	50.2	REP1		BDNF	362.7191162	OK	DQ	Unknown	Unknown	ddPCR Supermix for Probes (No dUTP)	FAM					19656	5215	14441
E04	16	50.2	REP1		BDNF	600.7639771	OK	DQ	Unknown	Unknown	ddPCR Supermix for Probes (No dUTP)	FAM					18915	7564	11351
D04	2	50	REP1		BDNF	98.58995819	OK	DQ	Unknown	Unknown	ddPCR Supermix for Probes (No dUTP)	FAM					19369	1557	17812
C04	4	50	REP1		BDNF	182.5280457	OK	DQ	Unknown	Unknown	ddPCR Supermix for Probes (No dUTP)	FAM					20047	2881	17166
B04	8	50	REP1		BDNF	382.3301392	OK	DQ	Unknown	Unknown	ddPCR Supermix for Probes (No dUTP)	FAM					20367	5651	14716
A04	16	50	REP1		BDNF	590.5283203	OK	DQ	Unknown	Unknown	ddPCR Supermix for Probes (No dUTP)	FAM					19478	7687	11791
D02	2	RS	REP2		BDNF	122.939415	OK	DQ	Unknown	Unknown	ddPCR Supermix for Probes (No dUTP)	FAM					17909	1777	16132
C02	4	RS	REP2		BDNF	279.7557678	OK	DQ	Unknown	Unknown	ddPCR Supermix for Probes (No dUTP)	FAM					19392	4104	15288
B02	8	RS	REP2		BDNF	623.8641357	OK	DQ	Unknown	Unknown	ddPCR Supermix for Probes (No dUTP)	FAM					18906	7781	11125
A02	16	RS	REP2		BDNF	1152.047852	OK	DQ	Unknown	Unknown	ddPCR Supermix for Probes (No dUTP)	FAM					19079	11913	7166
D11	2	200	REP2		BDNF	277.4240723	OK	DQ	Unknown	Unknown	ddPCR Supermix for Probes (No dUTP)	FAM					19822	4164	15658
C11	4	200	REP2		BDNF	471.7400513	OK	DQ	Unknown	Unknown	ddPCR Supermix for Probes (No dUTP)	FAM					21018	6943	14075
B11	8	200	REP2		BDNF	902.1112671	OK	DQ	Unknown	Unknown	ddPCR Supermix for Probes (No dUTP)	FAM					20704	11087	9617
A11	16	200	REP2		BDNF	1909.904663	OK	DQ	Unknown	Unknown	ddPCR Supermix for Probes (No dUTP)	FAM					18943	15207	3736
H08	2	150.2	REP2		BDNF	245.6865997	OK	DQ	Unknown	Unknown	ddPCR Supermix for Probes (No dUTP)	FAM					20624	3887	16737
G08	4	150.2	REP2		BDNF	392.8456421	OK	DQ	Unknown	Unknown	ddPCR Supermix for Probes (No dUTP)	FAM					21121	5996	15125
F08	8	150.2	REP2		BDNF	804.1484375	OK	DQ	Unknown	Unknown	ddPCR Supermix for Probes (No dUTP)	FAM					19757	9783	9974
E08	16	150.2	REP2		BDNF	1451.810669	OK	DQ	Unknown	Unknown	ddPCR Supermix for Probes (No dUTP)	FAM					20370	14440	5930
D08	2	150	REP2		BDNF	230.1463165	OK	DQ	Unknown	Unknown	ddPCR Supermix for Probes (No dUTP)	FAM					20008	3555	16453
C08	4	150	REP2		BDNF	390.6889954	OK	DQ	Unknown	Unknown	ddPCR Supermix for Probes (No dUTP)	FAM					20745	5862	14883
B08	8	150	REP2		BDNF	797.5358887	OK	DQ	Unknown	Unknown	ddPCR Supermix for Probes (No dUTP)	FAM					19989	9841	10148
A08	16	150	REP2		BDNF	1434.838257	OK	DQ	Unknown	Unknown	ddPCR Supermix for Probes (No dUTP)	FAM					19225	13547	5678
H02	2	100.2	REP2		BDNF	No Call	CHECK	DQ	Unknown	Unknown	ddPCR Supermix for Probes (No dUTP)	FAM					5936	0	5936
G02	4	100.2	REP2		BDNF	276.0016174	OK	DQ	Unknown	Unknown	ddPCR Supermix for Probes (No dUTP)	FAM					18148	3795	14353
F02	8	100.2	REP2		BDNF	635.269165	OK	DQ	Unknown	Unknown	ddPCR Supermix for Probes (No dUTP)	FAM					19061	7953	11108
E02	16	100.2	REP2		BDNF	1189.354614	OK	DQ	Unknown	Unknown	ddPCR Supermix for Probes (No dUTP)	FAM					19625	12484	7141
H05	2	50.2	REP2		BDNF	77.89870453	OK	DQ	Unknown	Unknown	ddPCR Supermix for Probes (No dUTP)	FAM					20072	1286	18786
G05	4	50.2	REP2		BDNF	144.2233582	OK	DQ	Unknown	Unknown	ddPCR Supermix for Probes (No dUTP)	FAM					20932	2415	18517
F05	8	50.2	REP2		BDNF	296.7576904	OK	DQ	Unknown	Unknown	ddPCR Supermix for Probes (No dUTP)	FAM					19987	4456	15531
E05	16	50.2	REP2		BDNF	518.7389526	OK	DQ	Unknown	Unknown	ddPCR Supermix for Probes (No dUTP)	FAM					20134	7179	12955
D05	2	50	REP2		BDNF	81.8263092	OK	DQ	Unknown	Unknown	ddPCR Supermix for Probes (No dUTP)	FAM					20688	1390	19298
C05	4	50	REP2		BDNF	148.734436	OK	DQ	Unknown	Unknown	ddPCR Supermix for Probes (No dUTP)	FAM					19889	2362	17527