

OpenBedside

An open-source gateway between bedside medical devices and the hospital's EHR. Infusion pumps first. Built so it can be extended to ventilators and other devices.

Not a medical device. Not for clinical use. See [DISCLAIMER.md](#).

Why I built it

My reason is a selfish one, and I would rather say so up front.

I am building **AI MedAgent**, a medication reasoning engine (aimedagent.net; the public demonstration runs on a simulated patient). A reasoning engine is only as good as what it can see, and it needs two feeds: the EHR and the bedside devices.

The EHR half has a door now. Federal certification rules require certified EHR technology to offer a standardized FHIR API (ONC, 45 CFR 170.315(g)(10)). Access is not free and not always quick, but the door exists and it looks much the same from one EHR to the next.

The device half has no door. Every infusion pump that reports to an EHR does it through its manufacturer's own gateway, in its own protocol, and each one is closed. IV infusions are one of the most important feeds a medication reasoning engine can have, because that is where the drug actually enters the patient, and they are the hardest data to get.

So I am doing the selfish thing in the most useful way I know: opening the device half for everyone, me included. OpenBedside is a **free, open-source device gateway. No fees, no royalties, no permission needed.** It starts with IV infusion pumps because that is the feed I need most, and it is built to take ventilators and other devices next.

OpenBedside contains no AI MedAgent code and no clinical reasoning. It is plumbing. It sends standard HL7 to any hospital system that wants it. AI MedAgent is one possible consumer of that data, and not a required one.

What it is

Every infusion pump that connects to an EHR today does it through its manufacturer's own gateway: years of work, sold with the pump, closed, and each one different. OpenBedside is one gateway, open to everyone. The parts are the same for every device:

- **the internal logic:** a common data model, state tracking, and events worked out from what the device reports;
- **store-and-forward:** every message saved to disk before it is sent, delivered in order, retried until the hospital acknowledges it;

- **EHR connectivity:** standard IHE Devices HL7 v2 messages over MLLP, the way hospital interface engines already expect them;
- **device registration and patient association:** nothing is sent until a device is registered, and each device's patient comes only from sources the hospital allows (the device itself, the ADT feed for its bed, or a person), with conflicts withheld;
- **a web page:** a live view of every device, connection, event and message, plus the Devices and Patients pages for registration and association.

The only part anyone writes for a new device, whether a manufacturer, a hospital integration team or a researcher, is a **device adapter**: a short piece of Python that translates that device's protocol into OpenBedside's data model. Every pump speaks differently, and they all carry the same facts. The adapter is the translation.

Open source, for anyone

Nothing here is proprietary. OpenBedside is released under the [Apache License 2.0](#). Anyone may download it, use it, test it, modify it, build it into their own product, and extend it to other devices. Improvements are welcome. See [CONTRIBUTING.md](#).

It was designed from public sources: the IHE Devices Technical Framework, ISO/IEEE 11073 nomenclature as published, FDA guidance, and experience. No vendor's protocol or code is in it. The worked example uses a fictional pump with an invented protocol.

Start here

You want to	Read
install it and see it run in 15 minutes	START-HERE-install-on-your-Mac.md (Linux is the same; Windows notes in the user guide)
run it, read the status page, change settings	docs/USER-GUIDE.md
connect your own device	docs/VENDOR-INTEGRATION-GUIDE.md
know exactly what the EHR receives	docs/HL7-INTERFACE.md
understand the design	docs/ARCHITECTURE.md and docs/DECISIONS.md
read it on paper or pass it on	docs/pdf/ : every guide as a PDF, and the OpenBedside Handbook with all of them in one file

The short version

```
python3 -m venv ~/.venvs/openbedside
source ~/.venvs/openbedside/bin/activate
pip install -e ".[dev]"          # from inside this folder
openbedside demo                # gateway + pump simulator + stand-in EHR, one window
# then open http://127.0.0.1:8080
```

Needs Python 3.11 or newer and nothing else. No database server, no web framework, no third-party packages at run time.

What it does in 0.2

Piece	Status
Data model: device, modules, channels, primary and secondary sources	working
Pump simulator: two modules, piggyback, rate change, KVO	working
Events worked out from observations: start, stop, complete, rate change, KVO, secondary to primary	working, each labelled <i>derived</i>
Store-and-forward (SQLite): ordered, retried, survives restarts	working
EHR outbound: IHE PCD-01 0RU^R01 over MLLP, application ACK required	working
Web page: Activity, Devices (registration), Patients (association), sign-in	working
Device registry: vendor + device id, EHR identifier (serial, asset tag, EUI-64), location, allowed patient sources	working
ADT inbound (A01, A02, A03, A04, A08, A11, A12, A13, A40): census of identifiers and locations	working
Patient association: device-reported, manual, ADT by bed; conflicts withheld; mid-infusion changes need confirmation	working
PID always sent (Unknown when unassociated); PV1 with location and visit	working
Adapter kit: base class, worked example, checker, starter-folder generator	working
Orders inbound: IHE PCD-03 RGV^015	simulator only , see the disclaimer
PCD-10 infusion event messages	planned

Piece	Status
Ventilators	registration accepts them; data model planned (the vendor guide explains what changes)
Encryption (MLLP, web page, database) and access audit log	not yet: synthetic data only

Commands

<code>openbedside demo</code>	everything in one window
<code>openbedside run -c config.toml</code>	the gateway
<code>openbedside ehr-sim</code>	a stand-in hospital interface engine
<code>openbedside check -c config.toml</code>	check an adapter's output before a hospital sees it
<code>openbedside new-adapter NAME</code>	write a starter adapter folder for a device
<code>openbedside fake-pump</code>	the fictional ACME pump used by the worked example
<code>openbedside send-order ...</code>	send a test order to the simulator
<code>openbedside send-adt ...</code>	send a test ADT message (synthetic identifiers only)
<code>openbedside set-password</code>	set the admin password for the Devices and Patients pages

Daniel C. Pettus. Apache License 2.0.