

Analytical and Bioanalytical Chemistry

Electronic Supplementary Material

**Isothermal circular strand displacement polymerization of DNA and microRNA
in digital microfluidic devices**

Maria Chiara Giuffrida, Laura Maria Zanolli, Roberta D'Agata, Alessia Finotti, Roberto Gambari,
Giuseppe Spoto




Fig. S1 Average referenced fluorescence intensity produced after the droplet ICSDP detection of a 100 pM **RRSoy FM** solution compared with the average referenced fluorescence intensity produced by the unrelated **RRSoy CTR** sequence at the same concentrations (two-tailed t-test, level 95%, $\nu = 8$ $p < 0.0001$). Error bars represent the confidence interval of the mean at the 95% level




Fig. S2 A,B. RT-qPCR of the external miR-210 control. Decreasing concentration of miR-210 (400 nM, 40 nM, 4 nM, 0.4 nM, 0.04 nM) were cDNA converted and PCR amplified. Panel **A** shows the reaction curves (the highest concentrations on the left), in panel **B** the titration curve (the highest concentrations on the right). **C.** RT-qPCR amplification curves of samples obtained from K562 cells after 96 hours after transfection with 270 nM pre-miR-210 (**270 96h**) (duplicated curves of a representative experiment). The amplification product correspond to the arrow indicated in panel **B**. **D.** Average referenced fluorescence intensity produced after droplet ICSDP detection of miR-210 from samples obtained from K562 cells after 96 hours after transfection with 270 nM pre-miR-210 (**270 96h**), compared with the average referenced fluorescence intensity produced by control samples **C1 96h** (two-tailed t-test, level 95%, $v = 8$ $p < 0.0001$). Error bars represent the confidence interval of the mean at the 95% level