
BISCHOFF CHROMATOGRAPHY
GERMANY | INDIA


ProntoSIL

CHIRAL


High performance liquid chromatography (HPLC) is becoming more widely used as a technique for the direct separation of chiral compounds and after contemplating an advisability of Chiral technologies, Bischoff Chromatography which is leading column manufactures whereas with in short period of time span known for its technique, quality and performance in Germany, Europe. We are launching our new product range **PRONTOSIL CHIRAL COLUMNS with Chiral column Screening facility.** Whereas ProntoSIL Chiral columns it can be used for separation of many enantiomers.




ProntoSil Chiral PAD
(Chiralpak AD, Lux Amylose-1)




ProntoSil Chiral POD
(Chiralcel OD, Lux Cellulose-1)




ProntoSil Chiral PAS
(Chiralpak AS)



ProntoSil Chiral POG
(Chiralcel OJ, Lux Cellulose-3)



ProntoSil Chiral PIA
(Chiralpak IA)



ProntoSil Chiral PIC
(Chiralpak IC)

PRONTOSIL CHIRAL COLUMN PHASE Discription

PRONTOSIL CHIRAL PHASES	Chiral sector
PRONTOSIL CHIRAL PAD	Amylose tris (3,5-dimethylphenylcarbamat)
PRONTOSIL CHIRAL POD	Cellulose tris (3,5-dimethylphenylcarbamate)
PRONTOSIL CHIRAL PAS	Amylose tris(S)-a-methylbenzylcarbamate)
PRONTOSIL CHIRAL POJ	Cellulose tris (4-methylbenzoate)
PRONTOSIL CHIRAL PIA	Amylose tris (3,5-dimethylphenylcarbamate)
PRONTOSIL CHIRAL PIC	Cellulose tris (3,5-dichlorophenylcarbamate)

PHASE COMPARISION:

ProntoSIL	Daicel	Phenominex	YMC
PRONTOSIL CHIRAL PAD-10	Chiral Pak AD	Lux Amylose-1	CHIRAL ART Amylose-C
PRONTOSIL CHIRAL POD-10	Chiral Pak OD	Lux Cellulose-1	CHIRAL ART Cellulose-C
PRONTOSIL CHIRAL PAS-10	Chiral Pak AS	Amylose tris(S)-a-methylbenzylcarbamate)	
PRONTOSIL CHIRAL POJ-10	Chiral Pak OJ	Lux Cellulose-3	
PRONTOSIL CHIRAL PIA-10	Chiral Pak IA	Lux i-Amylose-1	
PRONTOSIL CHIRAL PIC-10	Chiral Pak IC	Lux i-Cellulose-5	
PRONTOSIL CHIRAL PAD-5	Chiral Pak AD-H	Lux Amylose-1	CHIRAL ART Amylose-C
PRONTOSIL CHIRAL POD-5	Chiral Pak OD-H	Lux Cellulose-1	CHIRAL ART Cellulose-C
PRONTOSIL CHIRAL PAS-5	Chiral Pak AS-H		
PRONTOSIL CHIRAL POJ-5	Chiral Pak OJ-H	Lux Cellulose-3	
PRONTOSIL CHIRAL PIA-5	Chiral Pak IA-H	Lux i-Amylose-1	Amylose-SA
PRONTOSIL CHIRAL PIC-5	Chiral pak IC-H	Lux i-Cellulose-5	Cellulose-SC

Chiral Column care:

(Normal phase CHIRAL column)

- System must be free from aqueous contamination.
- Flush the system without column with isopropyl alcohol then by using respective mobile phase after this procedure connect the respective column and set parameters as per methods requirement.
- Samples should preferably be dissolved in the mobile phase.
- The mobile phase and the sample solution should be filtered through a membrane filter of approximately 0.5µm
- Porosity to ensure that there is no precipitate before using.

CHIRAL COLUMN STORAGE:

(Normal phase CHIRAL column)

- For column storage, remove the acidic or basic additives by flushing the column with several column volumes of the same mobile phase, but without the additive.
- Columns can be stored with ends capped in the additive-free mobile phase, or the shipping solvent, at room temperature.
- Operating these columns in accordance with the guidelines outlined here will result in a long column life.