



Inspection Report

Emergent Biodefense Operations Lansing L L C
3500 N Martin Luther King Blvd
Lansing, MI 48909

Customer ID: 197

Certificate: 34-R-0027

Site: 001

EMERGENT BIODEFENSE

Type: ROUTINE INSPECTION

Date: 04-JAN-2016

2.38(f)(1)

MISCELLANEOUS.

On 11-9-15, a technician working in the biocontainment area found a dead guinea pig on the floor of the autoclave. The technician had checked and loaded a used cage rack into the autoclave and ran it about one hour previous to finding the dead guinea pig upon opening the autoclave. The guinea pig was presumed to have been killed by the cycling of the autoclave. Ensure that animal racks are empty and/or no loose guinea pigs are in the autoclave prior to running it. The facility has adjusted its cage and autoclave checking procedures to help ensure that this will not happen again.

Correct by: Corrected by the time of inspection.

Exit interview conducted with the Institutional Official, the Attending Veterinarian, Facility Manager and APHIS representative Kurt Hammel.

Prepared By: KURT HAMMEL, D V M USDA, APHIS, Animal Care

Date:
29-JAN-2016

Title: VETERINARY MEDICAL OFFICER 1023

Received by Title: 70151520000070765189

Date:
29-JAN-2016



Species Inspected

Cust No	Cert No	Site	Site Name	Inspection
197	34-R-0027	001	EMERGENT BIODEFENSE	04-JAN-16

Count	Scientific Name	Common Name
004065	<i>Cavia porcellus</i>	DOMESTIC GUINEA PIG
004065	Total	