

santesoft fft imaging \*

### Description:

In Santesoft Sante FFT Imaging versions 1.4.1 and prior once a user opens a malicious DCM file on affected FFT Imaging installations , a local attacker could perform an out-of-bounds write, which could allow for arbitrary code execution.

### Main problems:

787 "Out-of-bounds Write"