

Sample Result

Sample Properties

Type	Control
Method	260813-UTAR SB2

	Sample Result Name	Measurement Date	Calculation Date	Wavelength Drift Status
1.	QC 8ppm	13/8/2026 3:59:56 PM	13/8/2026 4:02:58 PM	Success
2.	QC 8ppm	13/8/2026 2:55:14 PM	13/8/2026 4:02:58 PM	Success
3.	QC 8ppm	13/8/2026 2:19:36 PM	13/8/2026 4:02:58 PM	Success
4.	QC 8ppm	13/8/2026 1:42:19 PM	13/8/2026 4:02:58 PM	Success

Sample Inputs

	VOLUME	DILUTION	WEIGHT
1.	1	1	1
2.	1	1	1
3.	1	1	1
4.	1	1	1

Results in Concentration

		Ce 418.660	Ce 413.765	La 408.672	La 379.478	Nd 401.225	Nd 417.731	Fe 259.941	Fe 238.204	Fe 239.562
		ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
		Conc	Conc	Conc	Conc	Conc	Conc	Conc	Conc	Conc
1.	QC 8ppm									
	1	104 %	104 %	107 %	108 %	106 %	104 %	93.5 %	97.7 %	96.3 %
	2	105 %	105 %	107 %	108 %	106 %	104 %	93.3 %	96.8 %	95.7 %
	3	105 %	105 %	108 %	109 %	106 %	104 %	92.8 %	96.7 %	95.2 %
	Min. Calib.	0.001239	0.009240	0.006764	0.001134	0.002983	0.0141	0.0749	0.0452	0.0460
	Min. Accept.	90.0 %	90.0 %	90.0 %	90.0 %	90.0 %	90.0 %	90.0 %	90.0 %	90.0 %
	Reported	104 %	104 %	108 %	108 %	106 %	104 %	93.2 %	97.1 %	95.7 %
	Max. Accept.	110 %	110 %	110 %	110 %	110 %	110 %	110 %	110 %	110 %
	Max. Calib.	30.0	30.0	30.0	30.0	30.0	30.0	30.0	30.0	30.0
	Mean	8.35	8.36	8.60	8.66	8.48	8.31	7.46	7.76	7.66
	SD	0.0278	0.0194	0.0484	0.0379	0.0321	0.0156	0.0307	0.0429	0.0444
	RSD	0.333 %	0.232 %	0.563 %	0.438 %	0.378 %	0.187 %	0.412 %	0.553 %	0.580 %
2.	QC 8ppm									
	1	103 %	104 %	107 %	108 %	106 %	103 %	101 %	104 %	104 %
	2	103 %	103 %	107 %	108 %	106 %	103 %	103 %	105 %	105 %
	3	104 %	104 %	108 %	109 %	107 %	104 %	102 %	105 %	104 %
	Min. Calib.	0.001239	0.009240	0.006764	0.001134	0.002983	0.0141	0.0749	0.0452	0.0460
	Min. Accept.	90.0 %	90.0 %	90.0 %	90.0 %	90.0 %	90.0 %	90.0 %	90.0 %	90.0 %
	Reported	103 %	104 %	108 %	108 %	106 %	104 %	102 %	105 %	104 %
	Max. Accept.	110 %	110 %	110 %	110 %	110 %	110 %	110 %	110 %	110 %
	Max. Calib.	30.0	30.0	30.0	30.0	30.0	30.0	30.0	30.0	30.0
	Mean	8.28	8.29	8.61	8.65	8.49	8.28	8.16	8.38	8.35
	SD	0.0354	0.0292	0.0500	0.0425	0.0281	0.0247	0.0657	0.0403	0.0657
	RSD	0.427 %	0.353 %	0.580 %	0.491 %	0.331 %	0.298 %	0.806 %	0.481 %	0.786 %
3.	QC 8ppm									
	1	103 %	103 %	106 %	107 %	105 %	103 %	104 %	106 %	105 %
	2	103 %	104 %	106 %	107 %	105 %	104 %	103 %	105 %	105 %
	3	105 %	105 %	107 %	107 %	105 %	105 %	105 %	107 %	107 %
	Min. Calib.	0.001239	0.009240	0.006764	0.001134	0.002983	0.0141	0.0749	0.0452	0.0460


			Ce 418.660	Ce 413.765	La 408.672	La 379.478	Nd 401.225	Nd 417.731	Fe 259.941	Fe 238.204	Fe 239.562
			ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
			Conc	Conc	Conc	Conc	Conc	Conc	Conc	Conc	Conc
3.	QC 8ppm										
		Min. Accept.	90.0 %	90.0 %	90.0 %	90.0 %	90.0 %	90.0 %	90.0 %	90.0 %	90.0 %
		Reported	104 %	104 %	107 %	107 %	105 %	104 %	104 %	106 %	105 %
		Max. Accept.	110 %	110 %	110 %	110 %	110 %	110 %	110 %	110 %	110 %
		Max. Calib.	30.0	30.0	30.0	30.0	30.0	30.0	30.0	30.0	30.0
		Mean	8.30	8.32	8.53	8.56	8.41	8.32	8.31	8.46	8.44
		SD	0.0923	0.0753	0.0399	0.0326	0.0201	0.0754	0.0787	0.0806	0.0895
		RSD	1.11 %	0.904 %	0.468 %	0.381 %	0.239 %	0.907 %	0.948 %	0.953 %	1.06 %
4.	QC 8ppm										
		1	107 %	107 %	107 %	107 %	107 %	108 %	109 %	109 %	109 %
		2	105 %	105 %	105 %	106 %	104 %	105 %	107 %	107 %	107 %
		3	105 %	105 %	107 %	107 %	105 %	106 %	104 %	105 %	105 %
		Min. Calib.	0.001239	0.009240	0.006764	0.001134	0.002983	0.0141	0.0749	0.0452	0.0460
		Min. Accept.	90.0 %	90.0 %	90.0 %	90.0 %	90.0 %	90.0 %	90.0 %	90.0 %	90.0 %
		Reported	105 %	106 %	106 %	106 %	106 %	106 %	107 %	107 %	107 %
		Max. Accept.	110 %	110 %	110 %	110 %	110 %	110 %	110 %	110 %	110 %
		Max. Calib.	30.0	30.0	30.0	30.0	30.0	30.0	30.0	30.0	30.0
		Mean	8.44	8.45	8.50	8.50	8.45	8.52	8.54	8.58	8.57
		SD	0.120	0.106	0.0706	0.0529	0.111	0.139	0.175	0.144	0.155
		RSD	1.42 %	1.25 %	0.830 %	0.622 %	1.31 %	1.63 %	2.04 %	1.68 %	1.81 %

Regression

Regression Properties


Method	260813-UTAR SB2
Version	6
Regression Date	13/8/2026 4:02:58 PM

Regression Graph of Ce 418.660




Model	-	Polynomial
Coeff. A0	ppm	-3.4517E-001
Coeff. A1	ppm/cps	1.0872E-005
LOD	ppm	0.001239
BEC	ppm	0.790
Std. Error	ppm	0.464
Regr. Error	-	None
Min. Correlation	-	0.996000
Correlation	-	0.999024
Concentration	ppm	0.001239 to 30.0

Regression Graph of Ce 413.765




Model	-	Polynomial
Coeff. A0	ppm	-3.2450E-001
Coeff. A1	ppm/cps	1.2973E-005
LOD	ppm	0.009240
BEC	ppm	0.871
Std. Error	ppm	0.484
Regr. Error	-	None
Min. Correlation	-	0.996000
Correlation	-	0.998935
Concentration	ppm	0.009240 to 30.0

Regression Graph of La 408.672




Model	-	Polynomial
Coeff. A0	ppm	-3.9970E-001
Coeff. A1	ppm/cps	2.0482E-006
LOD	ppm	0.006764
BEC	ppm	0.275
Std. Error	ppm	0.433
Regr. Error	-	None
Min. Correlation	-	0.996000
Correlation	-	0.999147
Concentration	ppm	0.006764 to 30.0

Regression Graph of La 379.478




Model	-	Polynomial
Coeff. A0	ppm	-3.7001E-001
Coeff. A1	ppm/cps	2.3021E-006
LOD	ppm	0.001134
BEC	ppm	0.243
Std. Error	ppm	0.433
Regr. Error	-	None
Min. Correlation	-	0.996000
Correlation	-	0.999149
Concentration	ppm	0.001134 to 30.0

Regression Graph of Nd 401.225




Model	-	Polynomial
Coeff. A0	ppm	-1.2166E-001
Coeff. A1	ppm/cps	5.9354E-006
LOD	ppm	0.002983
BEC	ppm	0.511
Std. Error	ppm	0.183
Regr. Error	-	None
Min. Correlation	-	0.996000
Correlation	-	0.999848
Concentration	ppm	0.002983 to 30.0

Regression Graph of Nd 417.731




Model	-	Polynomial
Coeff. A0	ppm	-8.7124E-002
Coeff. A1	ppm/cps	2.4816E-005
LOD	ppm	0.0141
BEC	ppm	1.73
Std. Error	ppm	0.170
Regr. Error	-	None
Min. Correlation	-	0.996000
Correlation	-	0.999868
Concentration	ppm	0.0141 to 30.0

Regression Graph of Fe 259.941




Model	-	Polynomial
Coeff. A0	ppm	-1.5900E+000
Coeff. A1	ppm/cps	5.4848E-006
LOD	ppm	0.0749
BEC	ppm	2.04
Std. Error	ppm	0.879
Regr. Error	-	None
Min. Correlation	-	0.996000
Correlation	-	0.996445
Concentration	ppm	0.0749 to 30.0

Regression Graph of Fe 238.204



Model	-	Polynomial
Coeff. A0	ppm	-1.4346E+000
Coeff. A1	ppm/cps	3.8957E-006
LOD	ppm	0.0452
BEC	ppm	2.00
Std. Error	ppm	0.913
Regr. Error	-	None
Min. Correlation	-	0.996000
Correlation	-	0.996210
Concentration	ppm	0.0452 to 30.0

Regression Graph of Fe 239.562



Model	-	Polynomial
Coeff. A0	ppm	-1.4265E+000
Coeff. A1	ppm/cps	5.2568E-006
LOD	ppm	0.0460
BEC	ppm	2.01
Std. Error	ppm	0.903
Regr. Error	-	None
Min. Correlation	-	0.996000
Correlation	-	0.996293
Concentration	ppm	0.0460 to 30.0