

Sample Result

Sample Properties

Type Control
Method 260901-UTAR SB3

Sample Result Name Measurement Date Calculation Date Wavelength Drift Status

1. QC 10ppm 1/9/2026 9:12:01 AM 1/9/2026 10:20:00 AM Success
2. QC 10ppm 1/9/2026 9:44:37 AM 1/9/2026 10:20:00 AM Success
3. QC 10ppm 1/9/2026 10:18:52 AM 1/9/2026 10:20:00 AM Success

Results in Concentration

Cu 324.754 Cu 327.396 Zn 213.856 Mn 257.610 Mn 259.373 Cr 267.716
ppm ppm ppm ppm ppm ppm
Conc Conc Conc Conc Conc Conc

1. QC 10ppm

1 97.8 % 98.4 % 102.6 % 99.3 % 99.0 % 96.7 %
2 98.1 % 98.7 % 102.9 % 99.6 % 99.3 % 97.0 %
3 98.4 % 99.0 % 103.2 % 99.9 % 99.6 % 97.3 %
Min. Accept. 90.0 % 90.0 % 90.0 % 90.0 % 90.0 % 90.0 %
Reported 97.5 % 98.1 % 102.3 % 99.0 % 98.7 % 96.4 %
Max. Accept. 110 % 110 % 110 % 110 % 110 % 110 %

2. QC 10ppm

1 101.5 % 101.1 % 103.8 % 101.4 % 101.2 % 100.1 %
2 101.8 % 101.4 % 104.1 % 101.7 % 101.5 % 100.4 %
3 102.1 % 101.7 % 104.4 % 102.0 % 101.8 % 100.7 %
Min. Accept. 90.0 % 90.0 % 90.0 % 90.0 % 90.0 % 90.0 %
Reported 101.2 % 100.8 % 103.5 % 101.1 % 100.9 % 99.8 %
Max. Accept. 110 % 110 % 110 % 110 % 110 % 110 %

3. QC 10ppm

1 104.3 % 103.9 % 106.5 % 104.7 % 104.2 % 102.4 %
2 104.6 % 104.2 % 106.8 % 105.0 % 104.5 % 102.7 %
3 104.9 % 104.5 % 107.1 % 105.3 % 104.8 % 103.0 %
Min. Accept. 90.0 % 90.0 % 90.0 % 90.0 % 90.0 % 90.0 %
Reported 104.0 % 103.6 % 106.2 % 104.4 % 103.9 % 102.1 %
Max. Accept. 110 % 110 % 110 % 110 % 110 % 110 %