

Sample Result

Sample Properties

Type Control

Method 260814-UTAR SB3

Sample Result Name	Measurement Date	Calculation Date	Wavelength	Drift Status
--------------------	------------------	------------------	------------	--------------

1. QC 5ppm	14/8/2026 1:01:47 PM	14/8/2026 1:01:47 PM		Success
2. QC 5ppm	14/8/2026 2:14:08 AM	14/8/2026 1:01:47 PM		Success
3. QC 5ppm	14/8/2026 5:05:37 PM	14/8/2026 1:01:47 PM		Success

Sample Inputs

VOLUME	DILUTION	WEIGHT
--------	----------	--------

1. 1	1	1
2. 1	1	1
3. 1	1	1

Results in Concentration

Ce 418.660 La 408.672 Pr 422.293 Nd 401.225 Sm 359.260

ppm ppm ppm ppm ppm

Conc Conc Conc Conc Conc

1. QC 5ppm

1	100.6 %	96.8 %	94.2 %	103.3 %	99.2 %
2	102.6 %	98.6 %	94.8 %	102.8 %	98.3 %
3	103.1 %	98.8 %	94.0 %	102.8 %	98.3 %
Min. Calib.	0.001100	0.006200	0.004500	0.002900	0.008100
Min. Accept.	90.0 %	90.0 %	90.0 %	90.0 %	90.0 %
Reported	102 %	98 %	95 %	103 %	99 %
Max. Accept.	110 %	110 %	110 %	110 %	110 %
Max. Calib.	30.0	30.0	30.0	30.0	30.0
Mean	7.76	8.42	7.62	7.96	7.93
SD	0.0582	0.0470	0.0161	0.0743	0.0275
RSD	1.852 %	0.250 %	0.657 %	1.294 %	1.782 %

2. QC 5ppm

1	101.6 %	97.1 %	93.7 %	103.5 %	99.8 %
2	103.5 %	99.1 %	96.1 %	102.6 %	98.9 %
3	103.0 %	97.0 %	94.6 %	103.5 %	99.6 %
Min. Calib.	0.001100	0.006200	0.004500	0.002900	0.008100
Min. Accept.	90.0 %	90.0 %	90.0 %	90.0 %	90.0 %
Reported	102 %	98 %	95 %	103 %	99 %
Max. Accept.	110 %	110 %	110 %	110 %	110 %
Max. Calib.	30.0	30.0	30.0	30.0	30.0
Mean	8.32	7.59	8.26	8.14	7.79
SD	0.0747	0.0478	0.1396	0.1063	0.0407
RSD	0.716 %	1.560 %	0.206 %	1.661 %	1.630 %

3. QC 5ppm

1	101.7 %	96.7 %	96.2 %	103.2 %	99.7 %
2	101.1 %	98.0 %	96.2 %	103.4 %	97.9 %
3	100.9 %	98.7 %	95.1 %	103.7 %	98.8 %
Min. Calib.	0.001100	0.006200	0.004500	0.002900	0.008100
Min. Accept.	90.0 %	90.0 %	90.0 %	90.0 %	90.0 %
Reported	102 %	98 %	95 %	103 %	99 %
Max. Accept.	110 %	110 %	110 %	110 %	110 %
Max. Calib.	30.0	30.0	30.0	30.0	30.0
Mean	8.20	7.93	8.70	7.67	8.09
SD	0.1158	0.1306	0.0314	0.0324	0.1053
RSD	1.233 %	0.831 %	1.232 %	0.989 %	0.578 %

Regression

Regression Properties

Method 260814-UTAR SB3

Version 4

Regression Date 14/8/2026 1:01:47 PM

Regression Graph of Ce 418.660

-5 0 5 10 15 20 25 30

Calc. Conc. /ppm Intensity /cps

Method:260814-UTAR SB3 - Regression:Ce 418.660 Model - Polynomial

Coeff. A0 ppm -9.2121E-01

Coeff. A1 ppm/cps 2.8330E-05

LOD ppm 0.001100

BEC ppm 1.73

Std. Error ppm 0.203

Regr. Error - None

Min. Correlation - 0.996000

Correlation - 0.999451

Concentration ppm 0.001100 to 30.0

Regression Graph of La 408.672

-5 0 5 10 15 20 25 30

Calc. Conc. /ppm Intensity /cps

Method:260814-UTAR SB3 - Regression:La 408.672 Model - Polynomial

Coeff. A0 ppm -5.3579E-01

Coeff. A1 ppm/cps 2.3289E-05

LOD ppm 0.006200

BEC ppm 0.916

Std. Error ppm 0.364

Regr. Error - None

Min. Correlation - 0.996000

Correlation - 0.996617

Concentration ppm 0.006200 to 30.0

Regression Graph of Pr 422.293

-5 0 5 10 15 20 25 30

Calc. Conc. /ppm Intensity /cps

Method:260814-UTAR SB3 - Regression:Pr 422.293 Model - Polynomial

Coeff. A0 ppm -1.9937E+00

Coeff. A1 ppm/cps 2.1941E-05

LOD ppm 0.004500

BEC ppm 1.83

Std. Error ppm 0.975

Regr. Error - None

Min. Correlation - 0.996000

Correlation - 0.998972

Concentration ppm 0.004500 to 30.0

Regression Graph of Nd 401.225

-5 0 5 10 15 20 25 30

Calc. Conc. /ppm Intensity /cps

Method:260814-UTAR SB3 - Regression:Nd 401.225 Model - Polynomial

Coeff. A0 ppm -1.0100E+00

Coeff. A1 ppm/cps 4.0859E-06

LOD ppm 0.002900

BEC ppm 1.6

Std. Error ppm 0.858

Regr. Error - None

Min. Correlation - 0.996000

Correlation - 0.997980

Concentration ppm 0.002900 to 30.0

Regression Graph of Sm 359.260

-5 0 5 10 15 20 25 30

Calc. Conc. /ppm Intensity /cps

Method:260814-UTAR SB3 - Regression:Sm 359.260 Model - Polynomial

Coeff. A0 ppm -1.6121E+00

Coeff. A1 ppm/cps 1.1844E-05

LOD ppm 0.008100

BEC ppm 0.488

Std. Error ppm 0.958

Regr. Error - None

Min. Correlation - 0.996000

Correlation - 0.999597

Concentration ppm 0.008100 to 30.0