

Sample Result

Sample Properties

Type Control

Method 260820-UTAR SB4

Sample Result Name	Measurement Date	Calculation Date	Wavelength	Drift Status
1. QC 10ppm	20/8/2026 1:38:20 PM	20/8/2026 1:38:20 PM		Success
2. QC 10ppm	20/8/2026 1:07:59 PM	20/8/2026 1:38:20 PM		Success
3. QC 10ppm	20/8/2026 3:15:03 AM	20/8/2026 1:38:20 PM		Success
4. QC 10ppm	20/8/2026 5:05:05 PM	20/8/2026 1:38:20 PM		Success

Sample Inputs

VOLUME DILUTION WEIGHT

- 1 1 1
- 1 1 1
- 1 1 1
- 1 1 1

Results in Concentration

Fe 259.941 Mn 257.610 Cu 324.754 Zn 213.856 Ni 231.604

ppm ppm ppm ppm ppm

Conc Conc Conc Conc Conc

1. QC 10ppm

1 97.9 % 102.4 % 104.1 % 94.9 % 100.5 %
2 98.3 % 100.0 % 104.1 % 96.3 % 101.4 %
3 98.3 % 101.8 % 104.6 % 96.6 % 99.7 %
Min. Calib. 0.074900 0.018000 0.005500 0.009200 0.013400
Min. Accept. 90.0 % 90.0 % 90.0 % 90.0 % 90.0 %
Reported 97 % 101 % 104 % 96 % 100 %
Max. Accept. 110 % 110 % 110 % 110 % 110 %
Max. Calib. 30.0 30.0 30.0 30.0 30.0
Mean 8.31 7.95 8.58 8.04 7.80
SD 0.0190 0.0129 0.0875 0.0924 0.0110
RSD 1.445 % 0.212 % 0.228 % 0.160 % 0.728 %

2. QC 10ppm

1 97.0 % 100.3 % 104.0 % 96.1 % 100.7 %
2 98.1 % 101.2 % 103.2 % 95.9 % 99.7 %
3 95.8 % 101.5 % 103.6 % 95.7 % 101.1 %
Min. Calib. 0.074900 0.018000 0.005500 0.009200 0.013400
Min. Accept. 90.0 % 90.0 % 90.0 % 90.0 % 90.0 %
Reported 97 % 101 % 104 % 96 % 100 %
Max. Accept. 110 % 110 % 110 % 110 % 110 %
Max. Calib. 30.0 30.0 30.0 30.0 30.0
Mean 7.57 8.28 8.28 7.57 8.37
SD 0.1221 0.0253 0.0368 0.0851 0.0296
RSD 0.449 % 0.979 % 1.762 % 0.243 % 1.635 %

3. QC 10ppm

1 98.1 % 99.8 % 104.5 % 96.1 % 98.5 %
2 95.8 % 101.8 % 103.2 % 95.7 % 99.9 %
3 98.1 % 102.2 % 103.0 % 94.5 % 99.7 %
Min. Calib. 0.074900 0.018000 0.005500 0.009200 0.013400
Min. Accept. 90.0 % 90.0 % 90.0 % 90.0 % 90.0 %
Reported 97 % 101 % 104 % 96 % 100 %
Max. Accept. 110 % 110 % 110 % 110 % 110 %
Max. Calib. 30.0 30.0 30.0 30.0 30.0
Mean 8.61 8.44 7.84 8.34 8.38
SD 0.1197 0.1027 0.0781 0.0366 0.0405
RSD 0.211 % 1.498 % 0.216 % 0.696 % 0.195 %

4. QC 10ppm

1	96.9 %	102.3 %	104.1 %	94.7 %	100.0 %
2	98.1 %	99.7 %	102.7 %	97.1 %	99.7 %
3	98.3 %	101.2 %	104.2 %	94.6 %	98.7 %
Min. Calib.	0.074900	0.018000	0.005500	0.009200	0.013400
Min. Accept.	90.0 %	90.0 %	90.0 %	90.0 %	90.0 %
Reported	97 %	101 %	104 %	96 %	100 %
Max. Accept.	110 %	110 %	110 %	110 %	110 %
Max. Calib.	30.0	30.0	30.0	30.0	30.0
Mean	8.29	8.18	7.88	7.81	8.30
SD	0.0540	0.0472	0.0283	0.1004	0.0740
RSD	1.865 %	1.878 %	0.118 %	1.280 %	1.170 %

Regression

Regression Properties

Method 260820-UTAR SB4

Version 7

Regression Date 20/8/2026 1:38:20 PM

Regression Graph of Fe 259.941

-5 0 5 10 15 20 25 30

Calc. Conc. /ppm Intensity /cps

Method:260820-UTAR SB4 - Regression:Fe 259.941 Model - Polynomial

Coeff. A0 ppm -1.8050E+00

Coeff. A1 ppm/cps 1.6591E-05

LOD ppm 0.074900

BEC ppm 1.31

Std. Error ppm 0.219

Regr. Error - None

Min. Correlation - 0.996000

Correlation - 0.997361

Concentration ppm 0.074900 to 30.0

Regression Graph of Mn 257.610

-5 0 5 10 15 20 25 30

Calc. Conc. /ppm Intensity /cps

Method:260820-UTAR SB4 - Regression:Mn 257.610 Model - Polynomial

Coeff. A0 ppm -1.8659E+00

Coeff. A1 ppm/cps 8.0842E-06

LOD ppm 0.018000

BEC ppm 0.784

Std. Error ppm 0.494

Regr. Error - None

Min. Correlation - 0.996000

Correlation - 0.998119

Concentration ppm 0.018000 to 30.0

Regression Graph of Cu 324.754

-5 0 5 10 15 20 25 30

Calc. Conc. /ppm Intensity /cps

Method:260820-UTAR SB4 - Regression:Cu 324.754 Model - Polynomial

Coeff. A0 ppm -1.4101E+00

Coeff. A1 ppm/cps 2.9532E-05

LOD ppm 0.005500

BEC ppm 2.04

Std. Error ppm 0.576

Regr. Error - None

Min. Correlation - 0.996000

Correlation - 0.998605

Concentration ppm 0.005500 to 30.0

Regression Graph of Zn 213.856

-5 0 5 10 15 20 25 30

Calc. Conc. /ppm Intensity /cps

Method:260820-UTAR SB4 - Regression:Zn 213.856 Model - Polynomial

Coeff. A0 ppm -9.1853E-01

Coeff. A1 ppm/cps 2.8021E-05

LOD ppm 0.009200

BEC ppm 0.349

Std. Error ppm 0.89

Regr. Error - None

Min. Correlation - 0.996000

Correlation - 0.997031

Concentration ppm 0.009200 to 30.0

Regression Graph of Ni 231.604

-5 0 5 10 15 20 25 30

Calc. Conc. /ppm Intensity /cps

Method:260820-UTAR SB4 - Regression:Ni 231.604 Model - Polynomial

Coeff. A0 ppm -2.6506E-01

Coeff. A1 ppm/cps 2.2530E-05

LOD ppm 0.013400

BEC ppm 0.473

Std. Error ppm 0.354

Regr. Error - None

Min. Correlation - 0.996000

Correlation - 0.996821

Concentration ppm 0.013400 to 30.0