

Sample Result

Sample Properties

Type Control

Method 260828-UTAR SB2

Sample Result Name	Measurement Date	Calculation Date	Wavelength Drift Status
--------------------	------------------	------------------	-------------------------

1. QC 8ppm	28/8/2026 3:13:43 PM	28/8/2026 3:13:43 PM	Success
2. QC 8ppm	28/8/2026 5:31:16 AM	28/8/2026 3:13:43 PM	Success

Sample Inputs

VOLUME	DILUTION	WEIGHT
--------	----------	--------

1. 1	1	1
2. 1	1	1

Results in Concentration

Ce 413.765 La 379.478 Fe 238.204

ppm ppm ppm

Conc Conc Conc

1. QC 8ppm

1	104.8 %	108.8 %	93.3 %
2	104.5 %	109.8 %	94.4 %
3	105.3 %	109.7 %	94.2 %
Min. Calib.	0.009240	0.001134	0.045200
Min. Accept.	90.0 %	90.0 %	90.0 %
Reported	106 %	109 %	94 %
Max. Accept.	110 %	110 %	110 %
Max. Calib.	30.0	30.0	30.0
Mean	8.01	7.51	7.59
SD	0.1336	0.1366	0.0864
RSD	1.686 %	1.207 %	0.381 %

2. QC 8ppm

1	104.9 %	108.4 %	95.2 %
2	106.9 %	110.1 %	95.2 %
3	105.1 %	108.2 %	92.8 %
Min. Calib.	0.009240	0.001134	0.045200
Min. Accept.	90.0 %	90.0 %	90.0 %
Reported	106 %	109 %	94 %
Max. Accept.	110 %	110 %	110 %
Max. Calib.	30.0	30.0	30.0
Mean	8.44	8.56	7.99
SD	0.0969	0.0316	0.1402
RSD	1.743 %	1.955 %	1.640 %

Regression

Regression Properties

Method 260828-UTAR SB2

Version 6

Regression Date 28/8/2026 3:13:43 PM

Regression Graph of Ce 413.765

-5 0 5 10 15 20 25 30

Calc. Conc. /ppm Intensity /cps

Method:260828-UTAR SB2 - Regression:Ce 413.765 Model - Polynomial

Coeff. A0 ppm -2.8124E-01

Coeff. A1 ppm/cps 1.7188E-06

LOD ppm 0.009240

BEC ppm 1.87

Std. Error ppm 0.399

Regr. Error - None

Min. Correlation - 0.996000

Correlation - 0.999630

Concentration ppm 0.009240 to 30.0

Regression Graph of La 379.478

-5 0 5 10 15 20 25 30

Calc. Conc. /ppm Intensity /cps

Method:260828-UTAR SB2 - Regression:La 379.478 Model - Polynomial

Coeff. A0 ppm -4.3564E-01

Coeff. A1 ppm/cps 2.6058E-05

LOD ppm 0.001134

BEC ppm 2.05

Std. Error ppm 0.34

Regr. Error - None

Min. Correlation - 0.996000

Correlation - 0.999071

Concentration ppm 0.001134 to 30.0

Regression Graph of Fe 238.204

-5 0 5 10 15 20 25 30

Calc. Conc. /ppm Intensity /cps

Method:260828-UTAR SB2 - Regression:Fe 238.204 Model - Polynomial

Coeff. A0 ppm -1.7892E+00

Coeff. A1 ppm/cps 2.6293E-05

LOD ppm 0.045200

BEC ppm 2.16

Std. Error ppm 0.3

Regr. Error - None

Min. Correlation - 0.996000

Correlation - 0.999185

Concentration ppm 0.045200 to 30.0