

Gradient-Sensitive Functional Descriptors for Interferometric Surface Metrology: Complementing Amplitude-Averaging ISO Parameters

Abstract

This paper presents a complete interferometric pipeline for optical evaluation of roundness and surface profiles using a self-built Michelson interferometer. The method is validated on two certified JENOPTIK standards: the roundness standard FN 111 (24,000 interferograms) and the magnification standard FN 101 with a certified Flick flat (100,000 interferograms), each measured in five 400° sweeps. External validation against 104 tactile reference measurements yields an E_n score of 0.61 for the FN 111 and 0.3% agreement for the FN 101, demonstrating metrological consistency of the optical reconstruction. In addition to conventional profile roughness parameters (ISO 21920-2, formerly ISO 4287) and roundness metrics ($RONt$, $RONq$), the reconstructed nanometre profiles are analysed using functional-analytic descriptors including L^p , Sobolev $W^{1,2}$, and bounded-variation (BV) norms. Controlled defect-injection experiments and analysis of the certified Flick flat show that gradient-based descriptors provide complementary sensitivity to localised geometric features that may remain below the detection threshold of amplitude-averaging parameters. For a 100 nm synthetic spike, BV density increased by 18% and the $W^{1,2}$ seminorm by 622%, while R_a showed no detectable response at the reconstruction resolution. The BV descriptor also detected the certified Flick flat in a sliding-window analysis where R_a remained insensitive. These results suggest that functional descriptors can complement standard profile and roundness parameters, providing additional information on localised geometric variations without requiring changes to the measurement hardware. All reported uncertainties are within-

run repeatability estimates from five sweeps of a single continuous rotation and should be interpreted as lower bounds; an indicative type-B uncertainty budget is provided.

Keywords: interferometry, roundness, surface profile, functional norms, bounded variation, Sobolev seminorm, uncertainty

1. Introduction

Surface topography is one of the most important properties of technical parts. The accuracy of form and roughness influences contact conditions, wear and lifetime. Roundness, understood as deviation from a perfect circle, remains a central requirement in precision manufacturing; Zhu et al. emphasised that precise measurements are needed both for form tolerances and for tribological surface texture assessment [1]. At nanometric scale, such measurements require high-precision optical systems [2]. The foundational reference for surface metrology terminology and parameter definitions is the handbook of Whitehouse [3].

Optical dimensional metrology has a long connection with interference [4]. Modern optical surface metrology relies on diffraction, polarisation and interference [5, 6]. Diffraction and scattering have been applied to roughness measurement in several distinct ways: Brodmann introduced a scattered-light roughness instrument for production environments [7] and subsequently extended the scattered-light principle to combined roughness, form and waviness measurement [8]; Fan and Huynh analysed light scattering from flat periodic rough surfaces to estimate roughness parameters [9]; Hertzsch et al. demonstrated model-based scatterometry for microtopographic analysis of turned surfaces [10]; and Kucharski and Zdunek reported a low-cost scatterometry setup achieving nanometric precision in fast roughness measurement [11]. Interference remains central for ultra-precise surface-form measurement, from the early interference-based length comparisons of Benoît [12] to modern large-scale precision optics programmes [13]. De Groot systematised the principles of phase-shifting and coherence-scanning interference microscopy for topography evaluation [14, 15],

25 and Chatterjee described phase-evaluation interferometric testing of reflective
26 surfaces [16].

27 Several interferometric instruments have been proposed for high-accuracy
28 surface and roundness measurement. Polack et al. described a Michelson-based
29 system for surface-shape determination built around a high-resolution stitching
30 interferometer developed by Thomasset et al. [17, 18]. Fan et al. presented
31 a roundness measuring system for microspheres based on two small Michel-
32 son interferometers [19]. Kuhnel et al. proposed an interferometric system for
33 roundness and cylindricity measurement of ring gauges with 0.1 nm resolution
34 over a large measurement range [20]. Commercial instruments can be highly ac-
35 curate, but they are often costly and may not expose the full acquisition chain
36 needed for experimental method development.

37 The numerical analysis of fringe patterns has its own methodological lineage,
38 which frames the phase-decoding route used here. Takeda et al. introduced the
39 Fourier-transform method of fringe-pattern analysis, in which a spatial carrier
40 allows the phase to be recovered from a single interferogram without temporal
41 phase shifting [21]. Kemaio extended this concept with the two-dimensional win-
42 dowed Fourier transform, improving robustness to local fringe-quality variations
43 at the cost of window-parameter selection [22]. More recently, Feng et al. applied
44 deep learning to single-frame fringe analysis, training convolutional networks to
45 estimate the intermediate quantities of the arctangent phase calculation and
46 reporting accuracy approaching multi-frame phase shifting [23]. Galaktionov
47 and Toporovsky proposed a multi-stage fringe-map reconstruction algorithm
48 for fibre-end surface analysis in non-phase-shifting interferometry, showing that
49 geometric fringe processing remains competitive where phase-shifting hardware
50 is unavailable [24]. The radial excess-fraction method used in the present work
51 belongs to this single-frame, non-phase-shifting family: each interferogram is
52 decoded independently through the geometry of its circular fringes, which suits
53 continuous-rotation acquisition where temporal phase stepping is impractical.

54 The present work uses a self-built, non-contact Michelson-type roundness
55 measurement system assembled from off-the-shelf components. The optical

56 setup, apparatus description and radial image-processing route for phase extrac-
 57 tion were published in a previous study [25], where only roughness parameters
 58 were reported for the ceramic standard FN 111. The present work extends that
 59 study in three directions: (i) roundness ($RONt$, $RONq$) is evaluated for the first
 60 time with this self-built optical setup, (ii) a second artefact — the JENOPTIK
 61 magnification standard FN 101 with a Flick flat — is measured and validated
 62 against its DAkkS-DKD certificate, and (iii) a controlled sensitivity-comparison
 63 methodology is introduced: synthetic defects of known geometry are injected
 64 into measured profiles, and the response of classical ISO parameters is com-
 65 pared against that of L^p , Sobolev $W^{1,2}$, and bounded-variation (BV) norms
 66 — standard tools of functional analysis here applied to interferometric profile
 67 metrology — all computed from the same nanometre residual. The comparison
 68 framework, rather than any individual norm, constitutes the methodological
 69 contribution.

70 The computational problem is that modern interferometric surface metrol-
 71 ogy records long image sequences before a numerical profile exists. In the present
 72 data, however, each PNG file is already a separate interferogram in which the
 73 phase is encoded spatially by the bright and dark fringes inside the image. The
 74 files in the FN 111 and FN 101 folders are therefore not treated as a temporal
 75 phase-shifting stack of one fixed scene. Instead, each image $I_i(x, y)$ is processed
 76 independently to obtain one wrapped excess-fraction value ϵ_i . The ordered se-
 77 quence of these values is then unwrapped over image order and converted to a
 78 nanometre profile. The full measurement archive contains 24,000 grayscale in-
 79 terferograms of the FN 111, with image size 648×488 px, and 100,000 grayscale
 80 interferograms of the JENOPTIK magnification standard FN 101, with image
 81 size 640×480 px.

82 The literature relevant to this comparison is mature but only partly con-
 83 nected. In interferometric testing, Dumas et al. extended the lateral range of
 84 interferometry through subaperture stitching [26]; Chen et al. developed an it-
 85 erative stitching algorithm for spherical surfaces [27] and subsequently applied
 86 stitching to the testing of large optical surfaces [28]; Iacchetta et al. addressed

87 rotation and translation registration of band-limited interferometric images us-
 88 ing a chirp z-transform [29]; and Burge et al. applied computer-generated holo-
 89 grams to the measurement of x-ray and EUV optics [30]. On repeatability,
 90 Quan et al. compared evaluation approaches for interferometric surface repeata-
 91 bility [31], and Zhang et al. compared repeatability across different interference
 92 systems [32]; both characterise repeatability through RMS/PV quantities or
 93 pixel-wise standard-deviation matrices, i.e., through amplitude statistics. Pap-
 94 pas et al. propagated the uncertainty of areal surface-texture parameters using
 95 the metrological-characteristics approach, embedding texture parameters in a
 96 GUM-style framework [33]. In statistics, Gertheiss et al. reviewed functional
 97 data analysis as a language for function-valued observations [34], and Happ et
 98 al. provided a framework separating amplitude and phase variation in functional
 99 data [35]. Scale-dependence has been examined by Wolf, who connected surface
 100 topography with topological description [36], by Sanner et al., who defined scale-
 101 dependent roughness parameters for topography analysis [37], and by Zhao et
 102 al., who studied fractal simulation of topography for lubrication prediction [38];
 103 collectively these show that surface characterisation depends on bandwidth and
 104 filtering scale.

105 None of these strands, however, resolves a specific practical problem of pro-
 106 file and roundness metrology. When a localised geometric feature — a flat, a pit,
 107 a scratch — is present on an otherwise smooth artefact, amplitude-averaging
 108 parameters such as R_a dilute its signature over the entire evaluation length,
 109 so the feature can remain numerically invisible although it is metrologically
 110 relevant. Existing remedies do not fully address this for profile data: areal
 111 parameters require calibrated height maps that single-image radial extraction
 112 does not provide; visual inspection of polar plots is operator-dependent and not
 113 quantitative; and template-based defect detectors require the defect geometry
 114 to be known in advance. The unresolved question addressed here is therefore
 115 whether gradient-sensitive functional norms — whose mathematical structure
 116 makes them respond to localised slope content rather than to averaged am-
 117 plitude — provide reproducible, quantifiable sensitivity to such features when

118 computed from the same reconstructed residual as the standard ISO param-
 119 eters, and how large that sensitivity advantage is under controlled conditions.
 120 What has been missing in the literature is precisely this controlled, like-for-like
 121 comparison on identical nanometre residuals with external tactile validation.
 122 The present work addresses this by: (i) injecting synthetic defects of known ge-
 123 ometry into measured profiles to quantify descriptor response under controlled
 124 conditions; (ii) applying the full pipeline to two DAkkS-DKD-certified arte-
 125 facts at large scale (24,000 and 100,000 interferograms) with a self-built radial
 126 excess-fraction interferometer; and (iii) benchmarking the optical reconstruc-
 127 tion against 104 tactile reference measurements. Areal parameters are not re-
 128 ported because the single-image radial extraction does not produce a calibrated
 129 areal height map. The aim is to assess whether L^p , bounded-variation, and
 130 Sobolev seminorms, evaluated alongside ISO 21920-2 profile parameters on the
 131 same reconstructed residual profile, offer complementary descriptive sensitivity
 132 to controlled profile perturbations, and to evaluate the within-run repeatability
 133 of all quantities across five recorded sweeps. No claim of novelty is attached
 134 to the norms themselves — they are standard tools of functional analysis, and
 135 the functional-analytic vocabulary is used only as an organising taxonomy; the
 136 controlled comparison framework is the contribution.

137 2. Experimental Details

138 2.1. Optical Setup

139 The optical layout of the measurement system is shown in Fig. 1. The
 140 system is a two-channel modified Michelson interferometer with channels shifted
 141 by approximately 90° in phase. When an array detector is used instead of a
 142 point detector, the second channel is not strictly required [39]. In this system it
 143 was retained because it is useful for object centring and can serve as a control
 144 channel.

145 The optical setup and the full apparatus-level description were reported in a
 146 previous study on radial image processing for phase extraction in rough-surface

interferometry [25]. They are repeated here in condensed form to make the present functional-descriptor analysis self-contained and to document the acquisition conditions behind the FN 111 and FN 101 interferogram archives.




Figure 1: Optical layout of the two-channel modified Michelson interferometer: λ – diode laser (532 nm); P – linear polariser; L – lenses; L_3 – microscope objective; M – reference mirror ($\lambda/10$ flatness); BS – non-polarising beamsplitter; O – object on rotation stage; CCD – camera.




Figure 2: The complete measurement system: (1) modified Michelson interferometer in cage mounting, (2) high-precision rotation stage with the artefact.

The system was first tested with the JENOPTIK roundness standard FN 111 (photographs: Supplementary Fig. S1) (Al_2O_3 , $\varnothing = 29.9588$ mm) [40]. A single-mode 4.5 mW diode laser with wavelength 532 nm was used as the light source. A current-stabilised supply powered the laser to obtain a stable wavelength [41]. The laser beam passes through a polariser (P in Fig. 1) to set the polarisation of the light at 45° to the setup plane, minimising the difference in reflectivity between the two channels.

Key components: achromatic lenses $L_{1,2,4,5}$, non-polarising beamsplitter BS , $\lambda/8$ liquid phase retarder, $\lambda/10$ reference mirror M , DIN-4 microscope objective L_3 (NA 0.10), polarising beamsplitter PBS , and two Gigabit Ether-

160 net CCD cameras (640×480 px, 120 fps).

161 The interferometer was built in a cage system for thermal, acoustic and me-
162 chanical stability. A plexiglass box shields the measurement object, and the
163 structure is mounted on an area translation stage. A high-precision rotation
164 stage (encoder accuracy 2.0 mdeg, typical wobble $\pm 12 \mu\text{rad}$, typical eccentric-
165 ity $\pm 0.40 \mu\text{m}$; full specifications: Supplementary Table S1) rotates the object.
166 Both interferometer and stage are mounted on a honeycomb breadboard with
167 vibration isolation.

168 Measurements were performed at night (23:00–05:00) to minimise building-
169 activity disturbances. The setup was thermally and acoustically shielded (plex-
170 iglass box, wool blanket) and placed on a granite plate on layered wooden-
171 polystyrene vibration isolation (photographs: Supplementary Fig. S2). Five
172 semiconductor temperature sensors ($\pm 0.2^\circ\text{C}$) monitored stability; the tempera-
173 ture near the reference and object arms remained stable to $\pm 0.2^\circ\text{C}$ over six-hour
174 acquisitions.

175 2.2. Control System and Acquisition Workflow

176 Cameras were controlled by FLIR SDK-based C programs and Bash scripts [42].
177 Capture time was recorded per frame for angular-position verification. The ro-
178 tation stage was controlled via USB ASCII commands [43]; a 25 s settling delay
179 ensured constant velocity before acquisition began. Object centring was per-
180 formed manually using the two-camera phase-shifted configuration at 45° rota-
181 tion increments. Fringe evaluation used modified open-source R code [39, 44].
182 Table 1 lists the acquisition parameters for both archives; five 400° sweeps were
183 recorded per artefact, with the 40° overlap retained for profile-end comparison.

184 The object was measured in one continuous five-sweep rotation without repo-
185 sitioning. Each recorded sweep spans 400° ; successive useful 360° segments are
186 shifted by 40° , giving a natural between-sweep angular offset. The 40° overlap
187 region provides a model-free repeatability check: RMS height difference between
188 consecutive sweeps is 12 nm (FN 111) and $0.18 \mu\text{m}$ (FN 101), consistent with
189 Table 3.

Table 1: Measurement and sampling parameters of the interferometric acquisition workflow.

Parameter	FN 111	FN 101
Rotation speed, ω	0.50 °/s	0.08 °/s
CCD frame rate	6 fps	4 fps
Recorded sweeps, n	5	5
Total frames, N	24,000	100,000
Angular step, $\Delta\alpha = \omega/\text{fps}$	0.083°/frame	0.020°/frame
Frames per degree	12	50
Frames per useful 360°	4320	18,000
Frames per recorded 400° sweep	4800	20,000
Overlap per sweep	40° / 480 frames	40° / 2000 frames
Total acquisition time	4000 s	25,000 s

190 The frame rates in Table 1 (6 and 4 fps) are far below the 120 fps capa-
 191 bility of the cameras and merit comment. The angular step is set by the ratio
 192 $\Delta\alpha = \omega/\text{fps}$, so angular resolution is governed jointly by rotation speed and
 193 frame rate; measurement time is governed by rotation speed alone. The rota-
 194 tion speeds (0.50 and 0.08 °/s) were deliberately conservative: slow, constant-
 195 velocity rotation minimises dynamic excitation of the stage and keeps the fringe
 196 pattern quasi-static within each exposure, and the settling-time protocol (25 s
 197 before acquisition) assumes steady-state velocity. At these speeds, 6 and 4 fps
 198 already yield 12 and 50 frames per degree — an oversampling factor of roughly
 199 300 and 1200 relative to the 15 UPR metrological bandwidth — so the frame rate
 200 was not the accuracy-limiting factor; increasing it at fixed rotation speed would
 201 only add redundant frames. Shorter total acquisition (4,000 and 25,000 s here)
 202 would instead require faster rotation with a proportionally higher frame rate,
 203 which is technically feasible with the present cameras up to 120 fps. Whether
 204 fringe contrast and stage stability are preserved at, say, a tenfold higher speed
 205 has not been tested; the trade-off between acquisition speed, vibration robust-
 206 ness, and per-frame phase-extraction quality is identified as a topic for future

work, since measurement time is a critical parameter for industrial implementation.

3. Image Processing and Functional-Descriptor Method

3.1. Processing Pipeline

The pipeline follows the radial excess-fraction method of a previous study [25] with added sinusoidal refinement. Each interferogram $I_i(x, y)$ is decoded independently to a wrapped excess fraction ϵ_i ; the sequence is unwrapped chronologically, converted to nanometres ($z_i = \nu_i \lambda / 2$), and reduced by the object-specific LSCI reference model. Sweep boundaries are retained for repeatability analysis.

3.2. Step-by-Step Radial Phase Decoding

For each interferogram $I_i(x, y)$, a radial intensity profile about the fringe centre is extracted and B-spline-smoothed; bright-ring maxima are detected in D^2 space, where circular-fringe theory gives $D_{ij}^2 = c_i(j + \epsilon_i - 1)$ for the j th ring of diameter D_{ij} . A linear fit $D_{ij}^2 = \alpha_i j + \beta_i$ yields the wrapped excess fraction $\epsilon_i = (1 + \beta_i / \alpha_i) \bmod 1$ (2). Where the radial fringe signal permits ($R^2 \geq 0.65$, phase within 0.20 fringe of the ring-diameter estimate), a sinusoidal refinement is applied; otherwise the ring-diameter estimate is retained. The ordered series ϵ_i is unwrapped chronologically to a continuous fringe-order coordinate ν_i , and the nanometre profile follows from $z_i = \nu_i \lambda / 2$ with $\lambda = 532$ nm (one fringe order $\equiv 266$ nm) (3).

$$D_{ij}^2 = c_i(j + \epsilon_i - 1), \quad (1)$$

$$\epsilon_i = \left(1 + \frac{\beta_i}{\alpha_i}\right) \bmod 1. \quad (2)$$

$$z_i = \nu_i \frac{\lambda}{2}, \quad \lambda = 532 \text{ nm}. \quad (3)$$

3.3. Topography Profile, Reference Removal and Eccentricity Reduction

After phase-to-height conversion, the roundness profile is extracted following ISO 12181-1 [45] (which supersedes the withdrawn ISO 4291). For each 360°

segment, a least-squares reference circle (LSCI) is fitted to the angular height profile $z_i = h(\theta_i)$ in Cartesian coordinates:

$$(x_i, y_i) = ((R_0 + z_i) \cos \theta_i, (R_0 + z_i) \sin \theta_i), \quad (4)$$

where $R_0 = 14.9794$ mm is the nominal radius of the JENOPTIK FN 111. The LSCI centre (x_c, y_c) minimises $\sum_i [\sqrt{(x_i - x_c)^2 + (y_i - y_c)^2} - R_0]^2$, yielding the centred radial residuals ρ_i . The LSCI fit is numerically equivalent to removing the first harmonic of the angular profile, but it provides the formal ISO 4291 reference circle required for roundness metrology. After LSCI centring, a phase-correct Gaussian low-pass profile filter with 50% transmission at 15 undulations per revolution (UPR) is applied to the residual profile, consistent with the filter specification used for the tactile reference measurements and with the Gaussian profile filter defined in ISO 16610-21. All classical and gradient-based descriptors are computed from this LSCI-centred, Gaussian-filtered residual $\rho_i^{(\text{filt})}$.

For the JENOPTIK magnification standard FN 101, the same processing chain is applied: a linear trend is removed from the unwrapped phase sequence before height conversion (Sec. 5.1), the LSCI reference circle is fitted to the angular profile, and the Gaussian low-pass filter (50% at 15 UPR) is applied. Although both artefacts are processed through identical filtering steps, the 15 UPR cutoff corresponds to different spatial wavelengths on the circumference: $2\pi \times 14.9794 \text{ mm} / 15 = 6.27 \text{ mm}$ for the FN 111 ($\varnothing = 29.9588 \text{ mm}$) and $2\pi \times 15 \text{ mm} / 15 = 6.28 \text{ mm}$ for the FN 101 ($\varnothing \approx 30 \text{ mm}$) — effectively identical. The angular sampling differs substantially (0.083° vs. 0.020° , Table 1), but the Gaussian filter sigma is set by the UPR cutoff, not by the sampling interval, so the filtered profiles are directly comparable.

3.4. Classical Profile and Roundness Metrics

All classical quantities are evaluated on nanometre residuals, not on raw intensities, following the profile method of ISO 21920-2 [46], which supersedes the earlier ISO 4287 [47]; the amplitude-parameter definitions used here are common to both documents, and ISO 4287 is retained as a secondary reference

because the tactile comparison data and the previous study [25] predate the transition. Areal parameters (ISO 25178 [48]) are not reported because the single-image radial extraction does not produce a calibrated areal height map. For a residual profile or vector of residual pixels p , the reported profile-style roughness descriptors are

$$R_a = \frac{1}{M} \sum_{j=1}^M |p_j - \bar{p}|, \quad R_q = \left[\frac{1}{M} \sum_{j=1}^M (p_j - \bar{p})^2 \right]^{1/2}, \quad (5)$$

and

$$R_z = \max_j (p_j - \bar{p}) - \min_j (p_j - \bar{p}). \quad (6)$$

As implemented, this quantity is the global peak-to-valley height of the full evaluation profile, corresponding to the total-height parameter (Rt in ISO 21920-2 terms) rather than the section-averaged Rz ; the symbol R_z is retained for continuity with the previous study [25]. This implementation detail is relevant to the interpretation of Table 3 (Sec. 4.3). The profile skewness R_{sk} and kurtosis R_{ku} — ISO 21920-2 shape parameters that characterise asymmetry and peakedness of the height distribution — are also reported:

$$R_{sk} = \frac{1}{R_q^3} \frac{1}{M} \sum_{j=1}^M (p_j - \bar{p})^3, \quad R_{ku} = \frac{1}{R_q^4} \frac{1}{M} \sum_{j=1}^M (p_j - \bar{p})^4. \quad (7)$$

$R_{sk} = 0$ for a symmetric distribution; $R_{sk} < 0$ indicates a predominance of valleys over peaks. $R_{ku} = 3$ for a normal distribution; $R_{ku} > 3$ indicates a spikier (leptokurtic) profile. These are the only ISO 21920-2 amplitude-distribution parameters that capture distribution shape beyond amplitude averages, and their inclusion allows a direct test of whether gradient-based descriptors provide information beyond even the best ISO shape parameters. For the eccentricity-corrected ceramic angular profile $q_C(\theta)$,

$$RONt = \max_{\theta} q_C(\theta) - \min_{\theta} q_C(\theta), \quad RONq = \left[\frac{1}{2\pi} \int_0^{2\pi} q_C^2(\theta) d\theta \right]^{1/2}. \quad (8)$$

These are profile/form quantities in nanometres. B-spline pre-smoothing (20 degrees of freedom, $\approx 18^\circ$ period) stabilises the LSCI fit; the subsequent Gaussian

low-pass filter (50% at 15 UPR, ISO 16610-21) defines the metrological band-
width. All descriptors are computed from the same Gaussian-filtered residual.

3.5. Functional Gradient-Based Descriptors

The concept of a complete normed vector space originates from the founda-
tional work of Banach [49]. For discrete profile data of finite length N , the vector
space $\mathbb{R}^N$ equipped with any norm is a Banach space; the present work does
not invoke infinite-dimensional completeness but draws on the taxonomy of L^p ,
Sobolev, and bounded-variation norms that Banach’s framework systematised.
For a centred residual profile $p(s)$ and $1 \leq p < \infty$,

$$\|p\|_p = \left(\frac{1}{L} \int_0^L |p(s)|^p ds \right)^{1/p}, \quad \|p\|_\infty = \max_{s \in [0, L]} |p(s)|. \quad (9)$$

The reported functional descriptors include L^1 , L^2 , L^∞ , the one-dimensional
Sobolev $W^{1,2}$ seminorm and the bounded-variation (BV) total variation,

$$|p|_{W^{1,2}} = \left(\int_0^L \left| \frac{dp}{ds} \right|^2 ds \right)^{1/2}, \quad TV(p) = \int_0^L \left| \frac{dp}{ds} \right| ds. \quad (10)$$

In discrete form, for a uniformly sampled profile p_j ($j = 1, \dots, N$) with angular
spacing $\Delta\alpha$,

$$|p|_{W^{1,2}} \approx \left(\Delta\alpha \sum_{j=1}^{N-1} \left(\frac{p_{j+1} - p_j}{\Delta\alpha} \right)^2 \right)^{1/2}, \quad TV(p) \approx \sum_{j=1}^{N-1} |p_{j+1} - p_j|. \quad (11)$$

Both $W^{1,2}$ and BV density carry sampling-interval dependence; “nm sample⁻¹”
is retained in the result tables for the fixed-acquisition setup, and a standardised
angular-unit convention for cross-instrument comparison is defined in Sec. 3.6.

The curvature RMS (root-mean-square of the discrete second derivative;
reported in Supplementary Table S2) carries a $(\Delta\alpha)^{-2}$ sampling dependence and
should not be compared across instruments with different angular resolution.

3.6. Angular-Unit Convention for Sampling-Dependent Descriptors

Gradient-based descriptors evaluated on discretely sampled profiles depend
on the sampling interval: the discrete total variation of an unfiltered profile

grows as the sampling grid refines (noise contributes gradient content at every scale), so raw per-sample values are instrument-specific. For meaningful cross-instrument comparison the following convention is proposed and used here.

1. **Band-limit first.** The profile is filtered with a declared phase-correct Gaussian profile filter (ISO 16610-21) at a stated cutoff in undulations per revolution (here 50% transmission at 15 UPR). After band-limiting, the profile derivative is bounded and the discrete sums converge to the continuous integrals as $\Delta\alpha \rightarrow 0$; the descriptor then characterises the declared bandwidth rather than the sampling grid.
2. **Sample adequately.** The angular sampling must resolve the declared bandwidth with margin; at least ten samples per cutoff undulation are recommended. Here the FN 111 and FN 101 profiles carry ≈ 290 and ≈ 1200 samples per 15-UPR undulation, so discretisation error is negligible.
3. **Report in angular units.** The BV density is reported as total variation per unit angle, $BV_\theta = TV(p)/\Theta$ in nm deg^{-1} ($\Theta = 360^\circ$), and the gradient content as the RMS angular gradient $g_{\text{rms}} = [\frac{1}{N-1} \sum_j ((p_{j+1} - p_j)/\Delta\alpha)^2]^{1/2}$ in nm deg^{-1} , together with the filter cutoff.

For the present data, the per-sample values in Table 3 convert as follows. FN 111 ($\Delta\alpha = 0.083^\circ$): BV density $0.263 \text{ nm sample}^{-1} \rightarrow 3.2 \text{ nm deg}^{-1}$; RMS gradient $0.310 \text{ nm sample}^{-1} \rightarrow 3.7 \text{ nm deg}^{-1}$. FN 101 ($\Delta\alpha = 0.020^\circ$): BV density $6.58 \text{ nm sample}^{-1} \rightarrow 329 \text{ nm deg}^{-1}$; RMS gradient $8.76 \text{ nm sample}^{-1} \rightarrow 438 \text{ nm deg}^{-1}$, all at the 15-UPR cutoff. Because the conversion is a fixed multiplicative factor for a given acquisition, relative uncertainties and relative defect responses are identical in either unit system. Adoption of gradient descriptors in reference software for roundness evaluation would require exactly this pair of declarations — filter cutoff and angular units — which parallels the established practice of declaring the filter for *RONt*.

3.7. Uncertainty Evaluation

Each dataset contains five 400° sweeps recorded in a single continuous rotation without stopping or repositioning. The five useful 360° segments are ex-

334 tracted from this continuous sequence; because each recorded sweep spans 400° ,
335 successive useful revolutions are shifted by 40° on the object. These segments
336 are not strictly independent — they share the same continuous-run conditions
337 (laser warm-up state, stage-bearing temperature, ambient drift). The standard
338 deviation across sweeps therefore reflects within-run repeatability rather than
339 between-run reproducibility. For a descriptor X evaluated on each sweep, the
340 mean value is $\bar{X}$ and the within-run standard uncertainty is estimated as

$$u_{\text{within}}(X) = \frac{s_X}{\sqrt{n}}, \quad n = 5, \quad (12)$$

341 where s_X is the sample standard deviation across the five sweeps. The reported
342 expanded uncertainty is

$$U_{95}(X) = t_{0.975, n-1} u_{\text{within}}(X). \quad (13)$$

343 With $n = 5$, $t_{0.975, 4} = 2.776$; the χ^2_4 -based 95% CI for the standard deviation
344 spans $[0.6s_X, 2.9s_X]$, so U_{95} is indicative rather than definitive. Because the five
345 segments share continuous-run conditions (laser state, stage temperature, am-
346 bient drift), the $1/\sqrt{n}$ factor in u_{within} may underestimate the true uncertainty
347 of the mean; all reported uncertainties are *lower bounds*. An indicative type-B
348 budget for the principal systematic terms is given in Sec. 3.8; a complete GUM-
349 compliant budget [50] with experimentally characterised input distributions, and
350 a between-run reproducibility study with independent setups, thermal cycles,
351 and physical re-centring of the artefact, remain future work (Sec. 5.6). With
352 $n = 5$ sweeps and the observed within-run variability, the minimum detectable
353 relative difference between two descriptors (80% power, $\alpha = 0.05$, two-sided
354 paired t -test) is approximately 30% for the FN 111 and 5% for the FN 101, con-
355 sistent with the exploratory character of the descriptor comparison. A per-sweep
356 trend of FN 111 $RONt$ values gives a slope of $8.4 \pm 4.1 \text{ nm sweep}^{-1}$ ($p = 0.12$),
357 consistent with zero drift, though the sweep-to-sweep scatter cautions that U_{95}
358 may underestimate long-term reproducibility.

359 3.8. Indicative Type-B Uncertainty Budget

360 Although a complete GUM budget requires between-run data that the present
 361 single-rotation acquisition cannot supply, the principal type-B contributions can
 362 be bounded from component specifications and pipeline properties. Table 2
 363 summarises the estimates for the FN 111 *RONt* evaluation; the same relative
 364 arguments apply to the FN 101.

365 *Laser wavelength stability.* The height scale $z = \nu\lambda/2$ inherits the relative
 366 wavelength error: a ± 1 nm tolerance on the 532 nm diode module (rectangu-
 367 lar distribution) gives a relative standard uncertainty of 0.11%, i.e. 0.3 nm on
 368 *RONt* = 278 nm (and 46 nm on the FN 101 *RONt* = 42.07 μm , within its
 369 ± 0.48 μm repeatability interval). Wavelength error acts as a pure scale factor
 370 and cannot create or mask localised features.

371 *Refractive index of air.* The in-air wavelength varies with temperature and
 372 pressure by $\sim 10^{-6}$ per 1°C or 3 hPa; with the observed $\pm 0.2^\circ\text{C}$ stability the
 373 contribution is below 10^{-5} relative, i.e. < 0.01 nm. Negligible.

374 *Abbe error and stage wobble.* The specified axis wobble (± 12 μrad) enters
 375 through two mechanisms: a cosine-type radial error $R\theta_w^2/2 \approx 1$ pm, and a lateral
 376 beam walk $R\theta_w \approx 0.18$ μm on the circumference, equivalent to an angular
 377 mislocation of 0.0007° , which maps through the local band-limited gradient
 378 (≈ 3.2 nm deg^{-1} , Sec. 3.6) to ≈ 0.002 nm. Both are negligible after LSCI
 379 first-harmonic removal.

380 *Camera nonlinearity and EFM localisation.* Intensity nonlinearity perturbs
 381 the radial fringe-peak positions; its effect is bounded empirically by the per-
 382 frame excess-fraction uncertainty of ± 0.05 fringe (13 nm). The 15-UPR band
 383 limit averages ≈ 290 frames per resolved undulation, reducing the per-point
 384 contribution to ≈ 0.8 nm and the peak-to-valley contribution to ≈ 1.1 nm.

385 *Form removal.* The LSCI reference (first-harmonic removal) is exact only
 386 in the limaçon approximation; the second-order term $a^2/(2R)$ for residual cen-
 387 tring eccentricity a leaks into the second harmonic of the residual. Manual
 388 centring leaves a of the order of tens of micrometres (the shared first-harmonic

389 model of Sec. 5.2.1, with its 63 nm/0.18% per-sweep amplitude variation, im-
390 plies $a \approx 35 \mu\text{m}$); for $a = 15\text{--}35 \mu\text{m}$ the term evaluates to 7–41 nm. Because
391 this contribution is common to all five sweeps, it does not appear in the type-A
392 scatter and must be carried as a bounded systematic term. Its magnitude is
393 consistent with the 34 nm difference between the optical *RONt* (278 ± 22 nm)
394 and the certified tactile value (244 ± 52 nm), a difference that the $E_n = 0.61$
395 comparison shows to be within the combined uncertainties.

Table 2: Indicative type-B uncertainty budget for the FN 111 *RONt* evaluation. Random contributions are standard uncertainties combined in quadrature with the type-A term; the form-removal term is a bounded systematic (bias) contribution common to all sweeps and is reported separately.

Source	Basis	Contribution to <i>RONt</i>
Within-run repeatability (type A)	$s/\sqrt{5}$, five sweeps	7.9 nm
Laser wavelength (± 1 nm, rect.)	module specification	0.3 nm
Air refractive index ($\pm 0.2^\circ\text{C}$)	Edlén-type sensitivity	< 0.01 nm
Abbe error / axis wobble ($\pm 12 \mu\text{rad}$)	stage specification, geometry	< 0.01 nm
Camera nonlinearity / EFM localisation	± 0.05 fringe per frame, band-limited averaging	1.1 nm
Combined standard uncertainty u_c (random)	quadrature	8.0 nm
Expanded U_{95} (random, $t_{0.975,4}$)		22 nm
Form removal (limaçon second order)	$a^2/(2R)$, $a = 15\text{--}35 \mu\text{m}$	≤ 41 nm (systematic bound)

396 The quadrature combination is dominated by the type-A term, so the re-
397 ported U_{95} values are essentially unchanged by the enumerated random type-B
398 terms; the budget nevertheless shows that the leading unquantified risks are the
399 form-removal systematic and the untested between-run effects, not the enumer-
400 ated instrument terms.

401 3.9. Computational Implementation

402 The reconstruction workflow (`scripts/reconstruct_radial_profile_sequence.`
403 `py`) decodes ϵ_i from each interferogram, unwraps the sequence, converts to
404 nanometres, applies LSCI fitting and Gaussian low-pass filtering (50% at 15 UPR,
405 ISO 16610-21), and exports CSV/JSON metrics with sweep-repeatability uncer-

406 tainty. Vector PDF figures document each processing stage. The pipeline is in-
407 voked as `PYTHONPATH=src.venv/bin/pythonscripts/reconstruct_radial_profile_`
408 `sequence.py` with dataset-specific parameters (full command in the data repos-
409 itory). Per-image excess-fraction decoding and unwrapping require approxi-
410 mately 0.5 ms per frame (measured on an Intel Core i7-9700K, single-threaded
411 Python 3.9), so the full 124,000-frame archive processes in under two minutes
412 on consumer hardware.

413 4. Results

414 The results follow the updated radial sequence route only. Each raw im-
415 age is first decoded to a wrapped excess fraction from its radial fringe pat-
416 tern. The ordered excess-fraction series is then unwrapped in chronological
417 order, converted to nanometres, and corrected for eccentricity. Finally, classi-
418 cal profile/roundness descriptors and functional gradient-based descriptors are
419 computed from the same nanometre residual profiles.

420 4.1. Radial Phase Decoding and Nanometre Profile Reconstruction

421 The FN 111 folder contained 24,000 interferograms and the steel folder con-
422 tained 100,000 interferograms. The reported run used full-resolution images, the
423 central radial aperture of 240 px, B-spline-smoothed peak detection, continuity-
424 aware EFM peak-subset tracking and an EFM prominence threshold of 0.04 of
425 the radial fringe-signal range. All five recorded sweeps were processed.

426 Figs. S3-S6 (single-image EFM and sequence processing steps) are provided
427 in the Supplementary Materials.

428 4.2. Roughness, Roundness and Functional Descriptors

429 Table 3 gives the numerical comparison after LSCI reference circle and Gaus-
430 sian low-pass filtering (50% at 15 UPR). Both artefacts are processed through
431 the same filtering chain. Each descriptor is reported as the full-sequence value
432 together with the 95% expanded type-A repeatability uncertainty estimated

433 from the five sweeps. Per-sweep values for all descriptors are provided in Sup-
434 plementary Table S3.

Table 3: Full-resolution radial EFM reconstruction from raw interferograms (LSCI + Gaussian low-pass, 50% at 15 UPR, ISO 16610-21). Uncertainties are 95% expanded within-run repeatability estimates from five sweeps ($n = 5$, $t_{0.975,4} = 2.776$). Units: nm for all descriptors except BV density and $W^{1,2}$ gradient (nm sample⁻¹). Certified *RONt*: FN 111 244 ± 52 nm (Hommel Roundscan 535, $n = 104$); FN 101 $41.94 \mu\text{m}$ (form tester, $n = 12$). The FN 101 certificate does not report an expanded uncertainty; therefore only the percentage difference is reported for this artefact rather than an E_n score.

Descriptor	FN 111	FN 101
R_a	$(66.3 \pm 4.7) \text{ nm}$	$(8.525 \pm 0.115) \mu\text{m}$
R_q	$(77.6 \pm 5.0) \text{ nm}$	$(10.376 \pm 0.090) \mu\text{m}$
R_z	$(278 \pm 22) \text{ nm}$	$(42.07 \pm 0.48) \mu\text{m}$
<i>RONt</i>	$(278 \pm 22) \text{ nm}$	$(42.07 \pm 0.48) \mu\text{m}$
<i>RONq</i>	$(77.6 \pm 5.0) \text{ nm}$	$(10.376 \pm 0.090) \mu\text{m}$
L^1 mean	$(66.3 \pm 4.7) \text{ nm}$	$(8.525 \pm 0.115) \mu\text{m}$
L^2 RMS	$(77.6 \pm 5.0) \text{ nm}$	$(10.376 \pm 0.090) \mu\text{m}$
L^∞	$(150 \pm 12) \text{ nm}$	$(21.97 \pm 0.42) \mu\text{m}$
BV total	$(1.137 \pm 0.062) \mu\text{m}$	$(118.4 \pm 6.2) \mu\text{m}$
BV density	$(0.263 \pm 0.014) \text{ nm sample}^{-1}$	$(6.58 \pm 0.35) \text{ nm sample}^{-1}$
$W^{1,2}$ grad.	$(0.310 \pm 0.015) \text{ nm sample}^{-1}$	$(8.76 \pm 0.43) \text{ nm sample}^{-1}$
<i>ISO shape parameters</i>		
R_{sk} (skewness)	-0.09 ± 0.04	0.00 ± 0.09
R_{ku} (kurtosis)	1.98 ± 0.21	2.51 ± 0.08
<i>Derived norm ratios</i>		
L^4/L^1 peakiness	1.39 ± 0.08	1.53 ± 0.03
<i>External validation</i>		
E_n score	0.61	< 1 (0.3% diff.; cf. Fig. S12)

435 4.3. Coincidence of R_z with *RONt* and R_q with *RONq*

436 The identical values of R_z and *RONt* (and of R_q and *RONq*) for both arte-
437 facts in Table 3 are expected by construction, not coincidental. In this pipeline
438 both descriptor families are evaluated on the same LSCI-centred, Gaussian-
439 filtered closed residual profile (Sec. 3.3; harmonic decomposition of the residual:

Supplementary Fig. S19): the roughness-style R_z is the global peak-to-valley height of that profile (i.e., the total height Rt ; see the implementation note in the definition), while $RONt$ is the peak-to-valley roundness deviation of the very same profile, so the two coincide exactly; the RMS pair $R_q/RONq$ coincides for the same reason. The two families would separate if profile texture were evaluated on a different bandwidth than form (e.g., after λc -based separation of roughness from waviness in the ISO 21920 sense) or on an independent linear trace. Both are nevertheless reported because they belong to different ISO frameworks (profile texture versus roundness) with different reference conventions, and because their exact equality serves as a verification that all descriptors share one evaluation profile — the design premise of the like-for-like comparison.

4.4. Classical-Functional-Descriptor Comparison

The exact equalities $R_a = L^1$ and $R_q = L^2$ in Table 3 are definitional — each pair is the same formula applied to the same centred residual — and are reported purely as a computational cross-check that both descriptor families are evaluated on identical data, not as an empirical finding (see Sec. 5.3). For the FN 111, BV total variation carries 1.2% relative expanded uncertainty versus 5.5% for R_a and 8.5% for R_z , suggesting that gradient-based descriptors may exhibit metrological stability comparable to or better than peak-referenced roughness parameters in the tested datasets. Comparative bar charts and uncertainty analyses are provided in the Supplementary Materials (Figs. S7–S8). Ceramic and steel profile reconstruction plots are shown in Supplementary Figs. S9–S10.

The steel roundness polar representation (FN 101, Flick flat) is provided in the Supplementary Materials (Fig. S11).

JENOPTIK FN~111 — Roundness deviation
 $RONt = 278$ nm (5 sweeps, native angular coordinates)




Figure 3: Ceramic roundness in polar coordinates (LSCI + Gaussian low-pass, 50% at 15 UPR). Five sweeps at native positions. $RONt = 278 \pm 22$ nm.




Figure 4: Flick flat detection on the FN 101 magnification standard. BV density (bottom) shows a clear spike at the Flick flat angular position, whereas R_a computed in a sliding window (top) shows no detectable response at the reconstruction resolution. The sliding-window width was matched to the certified angular extent of the Flick flat as specified in the DAkKS-DKD certificate (Art. 532 528); identical window parameters were used for both descriptors.

464 The Flick flat produces a clear response in BV density while showing no
 465 detectable response in a sliding-window R_a analysis at the reconstruction reso-
 466 lution (Fig. 4), illustrating the potential complementary sensitivity of gradient-
 467 based descriptors to localised geometric features.

468 5. Discussion

469 5.1. Confirmed Results

470 The analysis rests on a key metrological distinction: an interferogram is
 471 an intensity measurement, not a phase map and not a height map. The half-

wavelength relation $z = \nu\lambda/2$ becomes valid only after each image has been decoded to a wrapped excess fraction and the ordered excess-fraction sequence has been unwrapped to the continuous fringe-order coordinate ν . This is why the reported comparison uses only reconstructed nanometre profiles, not raw greyscale images and not areal height maps.

The most sensitive part of the workflow is the EFM localisation of bright-ring maxima, with an excess-fraction uncertainty of approximately ± 0.05 fringe orders (tightened to ± 0.03 by sinusoidal refinement in 72% of FN 111 frames). A single ϵ_i error of 0.05 fringe corresponds to 13 nm, sub-dominant to the 85 nm residual-profile RMS. The chronological unwrapping criterion rejects phase jumps exceeding 0.35 fringe. The median number of detected rings is three for both artefacts within the 240 px central aperture. The five 400° ceramic sweeps require a geometric correction: successive useful 360° intervals are shifted by 40° because the physical object angle advances continuously; without this correction, the LSCI fit appears inconsistent even when phase extraction within each sweep is repeatable.

The linear drift of $0.0246 \text{ fringe frame}^{-1}$ in the steel unwrapped phase sequence (versus 0.00021 for the FN 111) is attributed to cumulative bias in the global unwrapping algorithm rather than thermal effects or axial runout [51]. The lower fringe contrast of the FN 101 interferograms — likely due to surface finish — reduces per-frame EFM quality; the pipeline applies automatic linear detrending where needed.

5.2. Interpretation

5.2.1. Eccentricity removal and wobble verification

Roundness descriptors are computed after LSCI reference-circle fitting (Sec. 3.3), which removes the first harmonic. A shared model across all five sweeps yields $R^2 = 0.99991$ with per-sweep amplitude variation of 63 nm (0.18%); the manufacturer-specified wobble ($\pm 12 \mu\text{rad}$) bounds the observed variation an order of magnitude below $RONt = 278 \text{ nm}$. The polar representation (Fig. 3) shows five

501 overlaid sweeps repeatable to ± 50 nm, illustrating that the residual is domi-
 502 nated by low-order form deviation rather than random roughness.

503 5.3. Comparison with ISO Parameters

504 A central question of this work is whether gradient-based functional descrip-
 505 tors provide additional descriptive sensitivity alongside classical ISO parameters
 506 when applied to the same reconstructed profiles. To answer this rigorously, every
 507 descriptor was computed on the same reconstructed nanometre residual profiles.

508 **Mathematical equivalences.** The L^1 norm is defined as $\frac{1}{N} \sum |p_i - \bar{p}|$,
 509 which is identical to the definition of R_a . The L^2 norm is $\sqrt{\frac{1}{N} \sum (p_i - \bar{p})^2}$,
 510 identical to R_q . Computing both on the same centred residual yields *exact*
 511 *numerical equality*: $R_a - L^1 = 0$ and $R_q - L^2 = 0$ to machine precision for
 512 both artefacts (Table 3). These equalities are not empirical findings; they are
 513 definitional consequences. Labelling the arithmetic mean of absolute deviations
 514 as L^1 rather than R_a provides no new metrological information and should not
 515 be presented as a methodological advance.

516 It is emphasised that neither the L^p norms nor the Sobolev and BV semi-
 517 norms are presented as novel mathematical constructs; they are standard tools
 518 whose definitions are restated for completeness. Their value in the present con-
 519 text is threefold. First, the functional-analytic taxonomy places ISO parameters
 520 at specific positions (L^1 , L^2 , L^∞) within a larger space of possible descriptors,
 521 revealing that descriptors outside the ISO set — BV density, $W^{1,2}$ seminorm
 522 — respond differently to the same perturbations. Second, the synthetic-defect
 523 injection methodology provides a controlled, like-for-like comparison framework
 524 that can be applied to any descriptor computed from the same residual profile;
 525 the framework itself, rather than any individual norm, constitutes the method-
 526 ological contribution. Third, the Flick flat on the certified FN 101 standard pro-
 527 vides a concrete existence proof: a real, DAKS-DKD-certified artefact whose
 528 defining geometric feature is detected unambiguously by BV density while re-
 529 maining below the detection threshold of sliding-window R_a (Fig. 4).

530 **Extrema: R_z versus L^∞ .** $R_z/L^\infty \approx 2$ for both artefacts, consistent with

531 symmetric positive and negative excursions. L^∞ captures the single largest
 532 one-sided excursion; the ratio serves as a simple symmetry indicator.

533 **Variation-based descriptors.** A controlled sensitivity analysis (Supple-
 534 mentary Table S2) examined five synthetic defect types injected into the FN 111
 535 residual profile. Amplitude-averaging descriptors (R_a , R_q , R_z) showed no de-
 536 tectable change at the reconstruction resolution — a single 100 nm spike changed
 537 R_a by less than 0.05%, below the per-frame phase-extraction precision — while
 538 BV density increased by +18% and the $W^{1,2}$ gradient seminorm by +622% for
 539 the same spike (Fig. 5). ISO shape parameters responded to asymmetry (R_{sk}
 540 changed by −1104% for a step) but not to symmetric localised features. R_{sk}
 541 and BV density appear complementary in the tested scenarios: R_{sk} responds to
 542 asymmetry, BV density responds to localised gradients.

543 **Translation of synthetic responses to realistic defects.** The injected
 544 defects are idealised (isolated spike, uniform scratch, pure sinusoid), whereas
 545 real manufacturing defects have complex, random morphologies. The synthetic
 546 responses nevertheless translate to other backgrounds through a simple scal-
 547 ing argument. A localised defect adds a total-variation increment ΔTV that
 548 depends only on the defect geometry ($\Delta TV \approx 2A$ for an isolated spike of ampli-
 549 tude A), so the relative full-profile BV-density response is $\Delta TV / TV_{\text{background}}$.
 550 On the FN 111 background ($TV = 1.137 \mu\text{m}$) a 100 nm spike yields +18%;
 551 on the FN 101 background ($TV = 118.4 \mu\text{m}$) the identical spike would yield
 552 only +0.17%, below the within-run repeatability. Detectability of a given abso-
 553 lute defect in the full-profile descriptor therefore scales inversely with the back-
 554 ground gradient content, which quantifies the reviewer-relevant observation that
 555 the FN 101 defect-to-background ratio is roughly 500 times less favourable than
 556 typical smooth-surface applications. Two consequences follow for industrial in-
 557 spection. First, full-profile BV density is best suited to smooth, finished surfaces
 558 — bearing raceways, optical components, precision seals — whose residual am-
 559 plitudes (tens of nanometres) are comparable to the FN 111 case, and where the
 560 synthetic archetypes map onto realistic features (spike $\leftrightarrow$ pit or inclusion edge,
 561 scratch $\leftrightarrow$ handling or tooling mark, sinusoid $\leftrightarrow$ chatter or vibration signature).

Second, for rougher backgrounds the sliding-window formulation restores local contrast: the certified Flick flat is detected on the FN 101 (Fig. 4) despite the unfavourable full-profile ratio, precisely because the window localises the comparison. Validation against physical defects of calibrated dimensions on real metallic and ceramic surfaces remains necessary before these scaling arguments can be considered established (Sec. 5.6).




Figure 5: Sensitivity of ISO and gradient-based descriptors to five synthetic defect types injected into the FN 111 residual profile. Percentage change from the unperturbed baseline; bold labels indicate changes exceeding 10%.

Supporting analyses — descriptor correlation matrices, power spectral density comparison, Cohen’s d effect sizes, critical descriptor assessment, and L^p norm convergence — are provided in the Supplementary Materials (Figs. S13–S17). Given the number of descriptors compared (~ 15 across two artefacts), the reported differences should be interpreted as exploratory rather than confirmatory; no formal multiple-comparison correction has been applied.

Practical considerations. A practical suggestion for interferometric profile and roundness metrology is to retain R_a , R_q , R_z , R_{sk} , $RONt$ and $RONq$ as the primary engineering layer compliant with ISO 21920-2, and to consider

577 supplementing them with BV density (in angular units, nm deg^{-1} , Sec. 3.6) as
 578 a gradient-sensitive functional-analysis complement. The $W^{1,2}$ gradient semi-
 579 norm is strongly correlated with BV density in the present datasets (within the
 580 FN 111: Pearson $r = 0.95$, $n = 5$ sweeps; within the FN 101: $r = 0.93$, $n = 5$).
 581 The pooled $r = 0.94$ (95% CI $[0.72, 0.99]$, $n = 10$) is reported for descriptive
 582 purposes only — the sweeps are not independent observations and this pooled
 583 correlation should be interpreted with caution. This strong correlation indicates
 584 that, as *repeatability metrics across sweeps*, BV density and the $W^{1,2}$ seminorm
 585 carry largely redundant information, and routine reporting of both is not war-
 586 ranted. They are not interchangeable in their response to isolated features,
 587 however: for the 100 nm synthetic spike the $W^{1,2}$ seminorm responded with
 588 +622% versus +18% for BV density, because the quadratic gradient weighting
 589 amplifies a single steep event far more strongly than the linear total-variation
 590 measure. A practical reading of the descriptor space in these data is therefore:
 591 (i) R_a/R_q (equivalently L^1/L^2) carry amplitude information; (ii) R_{sk} carries an
 592 independent asymmetry axis; (iii) BV density carries the localised-gradient axis
 593 and is recommended as the primary gradient descriptor; (iv) the $W^{1,2}$ semi-
 594 norm adds value mainly as a high-sensitivity alarm for isolated steep features
 595 rather than as an independent information axis; and (v) the L^4/L^1 peakiness
 596 ratio provides a supplementary distribution-shape axis with a direct physical in-
 597 terpretation as the amplification of dominant profile excursions relative to R_a .
 598 These additional descriptors are computed from the same residual profile with-
 599 out additional hardware or changes to the EFM pipeline. The cross-instrument
 600 validation suggests that gradient-based descriptors are applicable to any 1D
 601 profile, whether from interferometry or tactile roundness measurement.

602 **Practical value and the known-feature caveat.** It must be acknowl-
 603 edged that the Flick flat is a certified, deliberately manufactured feature whose
 604 presence and position were known in advance: Fig. 4 is an existence proof under
 605 ground-truth conditions, not a blind detection study. The practical argument
 606 for a gradient-sensitive layer is therefore not that it revealed an unknown defect
 607 here. It is that it converts what an operator might notice in a visual polar-

608 plot inspection into a quantitative, thresholdable, and automatable indicator
609 computed from data already acquired for ISO reporting — with no operator
610 in the loop, a defined angular position estimate, and a repeatability statement.
611 Whether this indicator reveals defects that routine quality-control reporting
612 would otherwise miss has not been demonstrated; a blind-detection protocol on
613 artefacts with independently characterised defects is outlined in Sec. 5.6 as the
614 decisive next test of practical value.

615 **Planned versus exploratory comparisons.** The synthetic-defect re-
616 sponses (Supplementary Table S2, Fig. 5) and the Flick flat detection (Fig. 4)
617 were planned comparisons motivated by the study’s objective of assessing gradient-
618 based descriptor sensitivity. The correlation matrices, power spectral density
619 comparison, Cohen’s d effect sizes, L^p norm convergence analysis, and the cross-
620 descriptor stability comparison (Supplementary Figs. S13–S17) are exploratory
621 and are intended to characterise the descriptor space rather than to test specific
622 hypotheses. This distinction should be borne in mind when interpreting the
623 reported associations. The present study is a proof-of-concept demonstration;
624 generalisability to other artefact types, materials, and measurement instruments
625 remains to be established.

626 5.4. Limitations

627 Several limitations should be acknowledged. First, the uncertainty analysis
628 is restricted to within-run Type A repeatability from five sweeps in a single con-
629 tinuous rotation, complemented by the indicative type-B budget of Sec. 3.8. The
630 small $n = 5$ limits statistical power (relative differences below $\sim 30\%$ cannot be
631 resolved for the FN 111), and between-run reproducibility (independent setups,
632 thermal cycles, physical re-centring of the artefact) has not been assessed; the
633 reported expanded uncertainties should be interpreted as lower bounds, and the
634 form-removal systematic (bounded at ≤ 41 nm) is carried outside the quadra-
635 ture combination.

636 Second, the validation is restricted to two artefacts from a single manufac-
637 turer (JENOPTIK), both of ~ 30 mm diameter. The two artefacts do differ

638 substantially in material (Al_2O_3 ceramic versus steel) and in residual amplitude
 639 (two orders of magnitude in R_a), which provides some breadth, but generalis-
 640 ability to other diameters, materials such as optical glass or tungsten carbide,
 641 and other surface finishes has not been tested. Third, the synthetic-defect anal-
 642 ysis uses idealised mathematical defects (isolated spike, uniform scratch, pure
 643 sinusoid) injected into the FN 111 ceramic profile only; the response to real
 644 manufacturing defects with complex morphologies may differ, and the scaling
 645 analysis of Sec. 5.3 shows quantitatively that full-profile detection sensitivity
 646 degrades in proportion to background gradient content (the FN 101 ratio is
 647 ~ 500 times less favourable). Fourth, BV density and the $W^{1,2}$ seminorm carry
 648 intrinsic sampling-interval dependence; the angular-unit convention of Sec. 3.6
 649 standardises reporting, but no inter-laboratory evidence yet exists that the con-
 650 vention transfers across instruments. Fifth, the Flick flat result (Supplemen-
 651 tary Fig. S18) is demonstrated on a single artefact with a known, certified
 652 macroscopic feature; blind-detection performance and performance for subtler,
 653 micrometre-scale defects have not been characterised.

654 *5.5. Metrological Implications*

655 The results carry several implications for profile and roundness metrology
 656 practice. The finding that BV density responds to localised gradients that am-
 657 plitude averaging may not capture suggests — subject to further validation —
 658 a potential role for gradient-based descriptors in applications where isolated
 659 defects are of concern; possible examples include quality control of bearing sur-
 660 faces, optical components, or precision seals. Because these descriptors are
 661 computed from the same residual profile used for ISO 21920-2 parameters, their
 662 adoption does not require changes to existing measurement acquisition proto-
 663 cols.

664 The ISO 21920-2 framework (with ISO 4287 as its predecessor) remains the
 665 appropriate baseline for surface texture specification and conformity assessment.
 666 The gradient-based descriptors discussed here are proposed as optional supple-
 667 mentary indicators, not as replacements for standardised parameters. Their

668 metrological utility would be strengthened by inter-laboratory comparisons, by
 669 evaluation on a wider range of artefact types, and by adoption of the angular-
 670 unit reporting convention of Sec. 3.6. Until such steps are taken, BV density and
 671 related descriptors should be regarded as research-grade tools whose diagnostic
 672 value has been observed under controlled laboratory conditions on two specific
 673 artefacts.

674 *5.6. Future Work*

675 The present study suggests several directions for further investigation. (i) A
 676 between-run reproducibility study with independent setups, distinct thermal
 677 cycles, and physical re-centring of the artefact would establish whether the
 678 gradient-based descriptors retain their relative stability under realistic labora-
 679 tory conditions, and would supply the input distributions needed to upgrade the
 680 indicative type-B budget of Sec. 3.8 to a complete GUM-compliant evaluation.
 681 (ii) Extending the validation to artefacts with different diameters, materials
 682 (e.g., optical glass, tungsten carbide), and surface finishes would test the gener-
 683 alisability of the findings beyond the two ~ 30 mm JENOPTIK artefacts studied
 684 here. (iii) A systematic characterisation of BV density response to real manufac-
 685 turing defects (scratches, pits, inclusions) of calibrated dimensions, culminating
 686 in a blind-detection protocol in which the analyst does not know defect presence
 687 or position, would bridge the gap between the present synthetic-defect analy-
 688 sis and industrial inspection and would provide the decisive test of practical
 689 value. (iv) An inter-laboratory comparison using a common artefact set and a
 690 standardised processing pipeline would provide the reproducibility data needed
 691 to assess whether BV density could eventually be considered for standardisa-
 692 tion. (v) Empirical verification of the angular-unit convention of Sec. 3.6 across
 693 instruments with different angular resolutions would support the potential in-
 694 clusion of gradient descriptors in reference software for roundness evaluation.
 695 (vi) A systematic study of the speed-accuracy trade-off (rotation speed, frame
 696 rate, fringe contrast) would establish the shortest acquisition compatible with
 697 the reported uncertainty, a prerequisite for industrial adoption.

698 Data Availability

699 The analysis code, processed results, and all figures are available at <https://github.com/dawidkucharski/banach-profile-metrology> (release v1.1.0).
700
701 The raw interferometric PNG archives (~ 32 GB) are available from the corre-
702 sponding author upon reasonable request.

703 References

- 704 [1] L. Zhu, C. Ni, Z. Yang, C. Liu, Investigations of micro-textured surface
705 generation mechanism and tribological properties in ultrasonic vibration-
706 assisted milling of Ti-6Al-4V, Precision Engineering 57 (November 2018)
707 (2019) 229–243. doi:10.1016/j.precisioneng.2019.04.010.
708 URL <https://doi.org/10.1016/j.precisioneng.2019.04.010https://linkinghub.elsevier.com/retrieve/pii/S0141635918307384>
709
- 710 [2] R. Leach, Optical Measurement of Surface Topography, Springer Berlin
711 Heidelberg, Berlin, Heidelberg, 2011. doi:10.1007/978-3-642-12012-1.
712 URL <http://link.springer.com/10.1007/978-3-642-12012-1>
- 713 [3] D. J. Whitehouse, Handbook of Surface and Nanometrology, 2nd Edition,
714 CRC Press, 2011.
- 715 [4] A. Michelson, Light-Waves and their Application to Metrology, Nature
716 49 (1255) (1893) 56–60. doi:10.1038/049056a0.
717 URL <https://doi.org/10.1038/049056a0http://10.0.4.14/049056a0http://www.nature.com/articles/049056a0>
718
- 719 [5] F. A. Jenkins, H. E. White, Fundamentals of Optics, International student
720 edition, McGraw-Hill, 1976.
721 URL <https://books.google.pl/books?id=dCdRAAAAMAAJhttps://books.google.pl/books?id=4SPCJ0wtFSUC>
722
- 723 [6] M. Born, E. Wolf, E. Hecht, Principles of Optics: Electromagnetic Theory
724 of Propagation, Interference and Diffraction of Light, Physics Today 53 (10)

- (2000) 77–78. doi:10.1063/1.1325200.
 URL <http://physicstoday.scitation.org/doi/10.1063/1.1325200>
- [7] R. Brodmann, Optisches Rauheitsmessgeraet fuer die Fertigung, Feinwerk-
 technik und Messtechnik 91 (2) (1983) 63–66.
- [8] R. Brodmann, Roughness form and waviness measurement by means of
 light-scattering, Precision Engineering 8 (4) (1986) 221–226. doi:10.1016/
 0141-6359(86)90063-2.
- [9] Y. Y. Fan, V. M. Huynh, An investigation of light scattering from flat peri-
 odic rough surfaces for surface roughness estimation, Precision Engineering
 16 (3) (1994) 205–211. doi:10.1016/0141-6359(94)90126-0.
- [10] A. Hertzsch, K. Kröger, H. Truckenbrodt, Microtopographic analysis of
 turned surfaces by model-based scatterometry, Precision Engineering 26 (3)
 (2002) 306–313. doi:10.1016/S0141-6359(02)00116-2.
- [11] D. Kucharski, H. Zdunek, A low-cost, simple optical setup for a fast
 scatterometry surface roughness measurements with nanometric precision,
 Bulletin of the Polish Academy of Sciences: Technical Sciences 68 (48)
 (2020) 485–490. doi:10.24425/bpasts.2020.133364.
 URL http://journals.pan.pl/Content/116520/PDF/10_485-490_01475_Bpast.No.68-3_30.06.20_K3A_TeX.pdf
- [12] R. Benoît, Application des phénomènes d’interférence à des déterminations
 métrologiques, Journal de Physique Théorique et Appliquée 7 (1) (1898)
 57–68. doi:10.1051/jphystap:01898007005700.
 URL <http://www.edpsciences.org/10.1051/jphystap:01898007005700%0Apapers3://publication/doi/10.1051/jphystap:01898007005700http://www.edpsciences.org/10.1051/jphystap:01898007005700>
- [13] G. Xue-fei, X. Sheng-nian, Y. Ye-fei, B. Shan, B. Xing, C. Zhou-jian,
 C. Ge-rui, D. Peng, G. Tian-shu, G. Wei, H. Shuang-lin, L. Yu-long,

- 753 L. Ying, L. Zi-ren, S. Ming-xue, S. Bao-san, T. Wen-lin, Y. Pin, X. Peng,
754 Z. Yun-long, Z. Hai-peng, L. Yun-kau, Laser Interferometric Gravita-
755 tional Wave Detection in Space and Structure Formation in the Early
756 Universe, Chinese Astronomy and Astrophysics 39 (4) (2015) 411–446.
757 doi:10.1016/j.chinastron.2015.10.001.
758 URL [http://www.sciencedirect.com/science/article/pii/](http://www.sciencedirect.com/science/article/pii/S0275106215000843)
759 [S0275106215000843https://linkinghub.elsevier.com/retrieve/](https://linkinghub.elsevier.com/retrieve/pii/S0275106215000843)
760 [pii/S0275106215000843](https://linkinghub.elsevier.com/retrieve/pii/S0275106215000843)
- 761 [14] P. de Groot, Phase Shifting Interferometry, in: R. K. Leach (Ed.), Optical
762 Measurement of Surface Topography, Springer Berlin Heidelberg, Berlin,
763 Heidelberg, 2011, pp. 167–186. doi:10.1007/978-3-642-12012-1_8.
764 URL http://link.springer.com/10.1007/978-3-642-12012-1_8
- 765 [15] P. de Groot, Principles of interference microscopy for the measurement
766 of surface topography, Advances in Optics and Photonics 7 (2015) 1–65.
767 doi:10.1364/AOP.7.000001.
- 768 [16] S. Chatterjee, Phase-Shifting Laser Interferometry for Measurement of Sur-
769 face Form Error, in: P.K.Gupta, R.Khare (Eds.), Springer Proceedings in
770 Physics, Vol. 160 of Springer Proceedings in Physics, 2015, pp. 209–222.
771 doi:10.1007/978-81-322-2000-8_10.
772 URL http://link.springer.com/10.1007/978-81-322-2000-8_10
- 773 [17] F. Polack, M. Thomasset, S. Brochet, D. Denetiere, Surface shape deter-
774 mination with a stitching Michelson interferometer and accuracy evalua-
775 tion, Review of Scientific Instruments 90 (2). doi:10.1063/1.5061930.
- 776 [18] M. Thomasset, M. Idir, F. Polack, M. Bray, J. J. Servant, A new phase-shift
777 microscope designed for high accuracy stitching interferometry, Nuclear
778 Instruments and Methods in Physics Research, Section A: Accelerators,
779 Spectrometers, Detectors and Associated Equipment 710 (2013) 7–12. doi:
780 10.1016/j.nima.2012.10.123.

- [19] K.-C. Fan, N. Wang, Z.-W. Wang, H. Zhang, Development of a roundness measuring system for microspheres, *Measurement Science and Technology* 25 (6) (2014) 064009. doi:10.1088/0957-0233/25/6/064009.
URL <http://stacks.iop.org/0957-0233/25/i=6/a=064009?key=crossref.73300f1601df9e917d2e7830f09f45e7>
- [20] M. Kühnel, V. Ullmann, U. Gerhardt, E. Manske, Automated setup for non-tactile high-precision measurements of roundness and cylindricity using two laser interferometers, *Measurement Science and Technology* 23 (7) (2012) 74016. doi:10.1088/0957-0233/23/7/074016.
URL <http://stacks.iop.org/0957-0233/23/i=7/a=074016?key=crossref.8429e4ab7d2d23702c0a05dba94b50d5http://stacks.iop.org/0957-0233/23/i=7/a=074016>
- [21] M. Takeda, H. Ina, S. Kobayashi, Fourier-transform method of fringe-pattern analysis for computer-based topography and interferometry, *Journal of the Optical Society of America* 72 (1) (1982) 156–160. doi:10.1364/JOSA.72.000156.
- [22] Q. Kemao, Two-dimensional windowed Fourier transform for fringe pattern analysis: Principles, applications and implementations, *Optics and Lasers in Engineering* 45 (2) (2007) 304–317. doi:10.1016/j.optlaseng.2005.10.012.
- [23] S. Feng, Q. Chen, G. Gu, T. Tao, L. Zhang, Y. Hu, W. Yin, C. Zuo, Fringe pattern analysis using deep learning, *Advanced Photonics* 1 (2) (2019) 025001. doi:10.1117/1.AP.1.2.025001.
- [24] I. Galaktionov, V. Toporovsky, A Multi-Stage Algorithm of Fringe Map Reconstruction for Fiber-End Surface Analysis and Non-Phase-Shifting Interferometry, *Applied System Innovation* 9 (2) (2026) 31. doi:10.3390/asi9020031.
- [25] D. Kucharski, M. Wiczorowski, Radial image processing for phase ex-

- traction in rough-surface interferometry, Measurement (2025) 117102doi:
10.1016/j.measurement.2025.117102.
- [26] P. Dumas, J. Fleig, G. Forbes, P. E. Murphy, Extending the range of
interferometry though subaperture stitching, Vol. 10314, 2003, p. 134 –
137, cited by: 9. doi:10.1117/12.2284035.
URL <https://www.scopus.com/inward/record.uri?eid=2-s2.0-85016807739&doi=10.1117%2f12.2284035&partnerID=40&md5=8aa04fbb9453cddff6da95df376ffdb6>
- [27] S. Chen, S. Li, Y. Dai, Z. Zheng, Iterative algorithm for subaperture
stitching test with spherical interferometers, Journal of the Optical Society
of America A: Optics and Image Science, and Vision 23 (5) (2006) 1219 –
1226, cited by: 48. doi:10.1364/JOSAA.23.001219.
URL <https://www.scopus.com/inward/record.uri?eid=2-s2.0-33745276550&doi=10.1364%2fJOSAA.23.001219&partnerID=40&md5=4d72b8095c89e81b1ba058a414b7b0c2>
- [28] S. Chen, S. Li, Y. Dai, Z. Zheng, Testing of large optical surfaces with
subaperture stitching, Applied Optics 46 (17) (2007) 3504 – 3509, cited
by: 44. doi:10.1364/AO.46.003504.
URL <https://www.scopus.com/inward/record.uri?eid=2-s2.0-34547536867&doi=10.1364%2fAO.46.003504&partnerID=40&md5=9e100905a834eebb31d582fc8884de99>
- [29] A. S. Iacchetta, J. R. Fienup, D. T. Leisawitz, Rotation and transla-
tion registration of bandlimited interferometric images using a chirp
z-transform, Vol. 9907, 2016, cited by: 3. doi:10.1117/12.2232178.
URL <https://www.scopus.com/inward/record.uri?eid=2-s2.0-85006847872&doi=10.1117%2f12.2232178&partnerID=40&md5=25831551d0ae177fef89c421998e31d4>
- [30] J. Burge, S. Ament, J. Beverage, C. Muccianti, C. Zhao, Applications of
computer-generated holograms for measuring x-ray and euv optics, Vol.

838 13620, 2025, cited by: 0. doi:10.1117/12.3067011.
839 URL [https://www.scopus.com/inward/record.uri?eid=2-s2.](https://www.scopus.com/inward/record.uri?eid=2-s2.0-105024341829&doi=10.1117%2f12.3067011&partnerID=40&md5=92bd5b2fbfd00333a45113efd7a5be87)
840 [0-105024341829&doi=10.1117%2f12.3067011&partnerID=40&md5=](https://www.scopus.com/inward/record.uri?eid=2-s2.0-105024341829&doi=10.1117%2f12.3067011&partnerID=40&md5=92bd5b2fbfd00333a45113efd7a5be87)
841 [92bd5b2fbfd00333a45113efd7a5be87](https://www.scopus.com/inward/record.uri?eid=2-s2.0-105024341829&doi=10.1117%2f12.3067011&partnerID=40&md5=92bd5b2fbfd00333a45113efd7a5be87)

842 [31] H. Quan, H. Xi, W. Fan, Evaluating surface repeatability for interfero-
843 metric measurement: A comparative study, Vol. 9684, 2016, cited by: 1.
844 doi:10.1117/12.2243301.
845 URL [https://www.scopus.com/inward/record.uri?eid=2-s2.](https://www.scopus.com/inward/record.uri?eid=2-s2.0-85002262299&doi=10.1117%2f12.2243301&partnerID=40&md5=98a061c8f8cdb6f89be75a953cd88b3f)
846 [0-85002262299&doi=10.1117%2f12.2243301&partnerID=40&md5=](https://www.scopus.com/inward/record.uri?eid=2-s2.0-85002262299&doi=10.1117%2f12.2243301&partnerID=40&md5=98a061c8f8cdb6f89be75a953cd88b3f)
847 [98a061c8f8cdb6f89be75a953cd88b3f](https://www.scopus.com/inward/record.uri?eid=2-s2.0-85002262299&doi=10.1117%2f12.2243301&partnerID=40&md5=98a061c8f8cdb6f89be75a953cd88b3f)

848 [32] L. Zhang, B. Zhang, S. Han, S. Tang, C. Zhuang, Repeatability
849 study on different interference systems, Vol. 10847, 2018, cited by: 0.
850 doi:10.1117/12.2505406.
851 URL [https://www.scopus.com/inward/record.uri?eid=2-s2.](https://www.scopus.com/inward/record.uri?eid=2-s2.0-85059070480&doi=10.1117%2f12.2505406&partnerID=40&md5=d2fe726c4d532c3f34340ac03143d343)
852 [0-85059070480&doi=10.1117%2f12.2505406&partnerID=40&md5=](https://www.scopus.com/inward/record.uri?eid=2-s2.0-85059070480&doi=10.1117%2f12.2505406&partnerID=40&md5=d2fe726c4d532c3f34340ac03143d343)
853 [d2fe726c4d532c3f34340ac03143d343](https://www.scopus.com/inward/record.uri?eid=2-s2.0-85059070480&doi=10.1117%2f12.2505406&partnerID=40&md5=d2fe726c4d532c3f34340ac03143d343)

854 [33] A. Pappas, L. Newton, A. Thompson, H. Hooshmand, R. Leach, Un-
855 certainty propagation of field areal surface texture parameters using the
856 metrological characteristics approach, 2023, p. 419 – 422, cited by: 3.
857 URL [https://www.scopus.com/inward/record.](https://www.scopus.com/inward/record.uri?eid=2-s2.0-85165481981&partnerID=40&md5=3c2a4bd283d7a91350fa5adce5f6941b)
858 [uri?eid=2-s2.0-85165481981&partnerID=40&md5=](https://www.scopus.com/inward/record.uri?eid=2-s2.0-85165481981&partnerID=40&md5=3c2a4bd283d7a91350fa5adce5f6941b)
859 [3c2a4bd283d7a91350fa5adce5f6941b](https://www.scopus.com/inward/record.uri?eid=2-s2.0-85165481981&partnerID=40&md5=3c2a4bd283d7a91350fa5adce5f6941b)

860 [34] J. Gertheiss, D. Rügamer, B. X. W. Liew, S. Greven, Functional data
861 analysis: An introduction and recent developments, Biometrical Journal
862 66 (7), cited by: 32; All Open Access, Hybrid Gold Open Access.
863 doi:10.1002/bimj.202300363.
864 URL [https://www.scopus.com/inward/record.uri?eid=2-s2.](https://www.scopus.com/inward/record.uri?eid=2-s2.0-85205152017&doi=10.1002%2fbimj.202300363&partnerID=40&md5=8d9e42e9cd2f23350c0bf0b477169043)
865 [0-85205152017&doi=10.1002%2fbimj.202300363&partnerID=40&md5=](https://www.scopus.com/inward/record.uri?eid=2-s2.0-85205152017&doi=10.1002%2fbimj.202300363&partnerID=40&md5=8d9e42e9cd2f23350c0bf0b477169043)
866 [8d9e42e9cd2f23350c0bf0b477169043](https://www.scopus.com/inward/record.uri?eid=2-s2.0-85205152017&doi=10.1002%2fbimj.202300363&partnerID=40&md5=8d9e42e9cd2f23350c0bf0b477169043)

- [35] C. Happ, F. Scheipl, A.-A. Gabriel, S. Greven, A general framework for multivariate functional principal component analysis of amplitude and phase variation, *Stat* 8 (1), cited by: 12; All Open Access, Green Open Access. doi:10.1002/sta4.220.
URL <https://www.scopus.com/inward/record.uri?eid=2-s2.0-85064480543&doi=10.1002%2fst4.220&partnerID=40&md5=e8d4b76faff42103fe5637961836125d>
- [36] G. W. Wolf, Surfaces - topography and topology, *Surface Topography: Metrology and Properties* 8 (1), cited by: 14; All Open Access, Hybrid Gold Open Access. doi:10.1088/2051-672X/ab70e8.
URL <https://www.scopus.com/inward/record.uri?eid=2-s2.0-85084666906&doi=10.1088%2f2051-672X%2fab70e8&partnerID=40&md5=006779646e49c4fc727d508bc63eabfa>
- [37] A. Sanner, W. G. Nöhring, L. A. Thimons, T. D. Jacobs, L. Pastewka, Scale-dependent roughness parameters for topography analysis, *Applied Surface Science Advances* 7, cited by: 28; All Open Access, Gold Open Access, Green Open Access. doi:10.1016/j.apsadv.2021.100190.
URL <https://www.scopus.com/inward/record.uri?eid=2-s2.0-85119021594&doi=10.1016%2fj.apsadv.2021.100190&partnerID=40&md5=7ad03ad2f796c56f6bec60b5febac199>
- [38] J. Zhao, Z. Wang, L. Zhengminqing, R. Zhu, Fractal simulation of surface topography and prediction of its lubrication characteristics, *Surface Topography: Metrology and Properties* 9 (4), cited by: 17. doi:10.1088/2051-672X/ac3f26.
URL <https://www.scopus.com/inward/record.uri?eid=2-s2.0-85122423109&doi=10.1088%2f2051-672X%2fac3f26&partnerID=40&md5=7eb253c5518d5356ead64c22878553ba>
- [39] F. Meijer, D. Kucharski, E. Stachowska, Determination of the phase in the center of a circular two-beam interference pattern to determine the

896 displacement of a rough surface, *Optical Engineering* 57 (10) (2018) 1.
897 doi:10.1117/1.OE.57.10.104101.
898 URL [https://www.spiedigitallibrary.org/journals/
899 optical-engineering/volume-57/issue-10/104101/
900 Determination-of-the-phase-in-the-center-of-a-circular/10.
901 1117/1.OE.57.10.104101.full](https://www.spiedigitallibrary.org/journals/optical-engineering/volume-57/issue-10/104101/Determination-of-the-phase-in-the-center-of-a-circular/10.1117/1.OE.57.10.104101.full)

902 [40] JENOPTIK Industrial Metrology, Unequivocal measurement of form and
903 positional tolerances using innovative measuring systems, JENOPTIK In-
904 dustrial Metrology Germany GmbH, 2018.

905 [41] Thorlabs Inc., CPS532 Collimated Laser Diode Module, 532 nm, 4.5 mW,
906 Round Beam, Ø11 mm Housing (2019).
907 URL [https://www.thorlabs.com/thorproduct.cfm?partnumber=
908 CPS532](https://www.thorlabs.com/thorproduct.cfm?partnumber=CPS532)

909 [42] FLIR Systems, Software Development Kit (2019).
910 URL [https://www.flir.com/browse/camera-cores-amp-components/
911 software-development-kits/](https://www.flir.com/browse/camera-cores-amp-components/software-development-kits/)

912 [43] Newport Corporation, ESP301 Controller GUI Manual, Newport Corpora-
913 tion, 2017.
914 URL [https://www.newport.com/medias/sys_master/images/images/
915 h7e/hfe/9044140982302/ESP301-Controller-GUI-Manual.pdf](https://www.newport.com/medias/sys_master/images/images/h7e/hfe/9044140982302/ESP301-Controller-GUI-Manual.pdf)

916 [44] R. Ihaka, R. Gentleman, R: A Language for Data Analysis and Graphics,
917 *Journal of Computational and Graphical Statistics* 5 (3) (1996) 299–314.
918 doi:10.1080/10618600.1996.10474713.
919 URL [http://www.tandfonline.com/doi/abs/10.1080/10618600.1996.
920 10474713](http://www.tandfonline.com/doi/abs/10.1080/10618600.1996.10474713)

921 [45] ISO 12181-1:2011, Geometrical product specifications (GPS) — Roundness
922 — Part 1: Vocabulary and parameters of roundness, Tech. rep., Interna-
923 tional Organization for Standardization (2011).

- 924 [46] ISO 21920-2:2021, Geometrical product specifications (GPS) — Surface
 925 texture: Profile — Part 2: Terms, definitions and surface texture param-
 926 eters, Tech. rep., International Organization for Standardization, Geneva
 927 (2021).
- 928 [47] ISO 4287:1997, Geometrical Product Specifications (GPS) — Surface tex-
 929 ture: Profile method — Terms, definitions and surface texture parameters,
 930 Tech. rep., International Organization for Standardization (1997).
- 931 [48] ISO 25178-2:2012, Geometrical Product Specifications (GPS) — Surface
 932 texture: Areal — Part 2: Terms, definitions and surface texture paramet-
 933 ers, Tech. rep., International Organization for Standardization (2012).
- 934 [49] S. Banach, Théorie des opérations linéaires, Monografie Matematyczne,
 935 Warsaw, 1932.
- 936 [50] JCGM 100:2008, Evaluation of measurement data — Guide to the expres-
 937 sion of uncertainty in measurement (GUM), Tech. rep., Joint Committee
 938 for Guides in Metrology (2008).
- 939 [51] Newport Corporation, Precision Motion Control Catalog, Newport Corpo-
 940 ration, 2018.