

Gradient-Sensitive Functional Descriptors for Interferometric Surface Metrology: Complementing Amplitude-Averaging ISO Parameters

Abstract

This paper presents a complete interferometric pipeline for optical evaluation of roundness and surface profiles using a self-built Michelson interferometer. The method is validated on two certified JENOPTIK standards: the roundness standard FN 111 (24,000 interferograms) and the magnification standard FN 101 with a certified Flick flat (100,000 interferograms), each measured in five 400° sweeps. External validation against 104 tactile reference measurements yields an E_n score of 0.61 for the FN 111 and 0.3% agreement for the FN 101, demonstrating metrological consistency of the optical reconstruction. In addition to conventional profile roughness parameters (ISO 21920-2, formerly ISO 4287) and roundness metrics ($RONt$, $RONq$), the reconstructed nanometre profiles are analysed using functional-analytic descriptors including L^p , Sobolev $W^{1,2}$, and bounded-variation (BV) norms. Controlled defect-injection experiments and analysis of the certified Flick flat show that gradient-based descriptors provide complementary sensitivity to localised geometric features that may remain below the detection threshold of amplitude-averaging parameters. For a 100 nm synthetic spike, BV density increased by 18% and the $W^{1,2}$ seminorm by 622%, while R_a showed no detectable response at the reconstruction resolution. The BV descriptor also detected the certified Flick flat in a sliding-window analysis where R_a remained insensitive. These results suggest that functional descriptors can complement standard profile and roundness parameters, providing additional information on localised geometric variations without requiring changes to the measurement hardware. All reported uncertainties are within-

run repeatability estimates from five sweeps of a single continuous rotation and should be interpreted as lower bounds; an indicative type-B uncertainty budget is provided.

Keywords: interferometry, roundness, surface profile, functional norms, bounded variation, Sobolev seminorm, uncertainty

1. Introduction

Surface topography is one of the most important properties of technical parts. The accuracy of form and roughness influences contact conditions, wear and lifetime. Roundness, understood as deviation from a perfect circle, remains a central requirement in precision manufacturing; Zhu et al. emphasised that precise measurements are needed both for form tolerances and for tribological surface texture assessment [1]. At nanometric scale, such measurements require high-precision optical systems [2]. The foundational reference for surface metrology terminology and parameter definitions is the handbook of Whitehouse [3].

Optical dimensional metrology has a long connection with interference [4]. Modern optical surface metrology relies on diffraction, polarisation and interference [5, 6]. Diffraction and scattering have been applied to roughness measurement in several distinct ways: Brodmann introduced a scattered-light roughness instrument for production environments [7] and subsequently extended the scattered-light principle to combined roughness, form and waviness measurement [8]; Fan and Huynh analysed light scattering from flat periodic rough surfaces to estimate roughness parameters [9]; Hertzsch et al. demonstrated model-based scatterometry for microtopographic analysis of turned surfaces [10]; and Kucharski and Zdunek reported a low-cost scatterometry setup achieving nanometric precision in fast roughness measurement [11]. Interference remains central for ultra-precise surface-form measurement, from the early interference-based length comparisons of Benoît [12] to modern large-scale precision optics programmes [13]. De Groot systematised the principles of phase-shifting and coherence-scanning interference microscopy for topography evaluation [14, 15],

25 and Chatterjee described phase-evaluation interferometric testing of reflective
26 surfaces [16].

27 Several interferometric instruments have been proposed for high-accuracy
28 surface and roundness measurement. Polack et al. described a Michelson-based
29 system for surface-shape determination built around a high-resolution stitching
30 interferometer developed by Thomasset et al. [17, 18]. Fan et al. presented
31 a roundness measuring system for microspheres based on two small Michel-
32 son interferometers [19]. Kuhnel et al. proposed an interferometric system for
33 roundness and cylindricity measurement of ring gauges with 0.1 nm resolution
34 over a large measurement range [20]. Commercial instruments can be highly ac-
35 curate, but they are often costly and may not expose the full acquisition chain
36 needed for experimental method development.

37 The numerical analysis of fringe patterns has its own methodological lineage,
38 which frames the phase-decoding route used here. Takeda et al. introduced the
39 Fourier-transform method of fringe-pattern analysis, in which a spatial carrier
40 allows the phase to be recovered from a single interferogram without temporal
41 phase shifting [21]. Kemaio extended this concept with the two-dimensional win-
42 dowed Fourier transform, improving robustness to local fringe-quality variations
43 at the cost of window-parameter selection [22]. More recently, Feng et al. applied
44 deep learning to single-frame fringe analysis, training convolutional networks to
45 estimate the intermediate quantities of the arctangent phase calculation and
46 reporting accuracy approaching multi-frame phase shifting [23]. Galaktionov
47 and Toporovsky proposed a multi-stage fringe-map reconstruction algorithm
48 for fibre-end surface analysis in non-phase-shifting interferometry, showing that
49 geometric fringe processing remains competitive where phase-shifting hardware
50 is unavailable [24]. The radial excess-fraction method used in the present work
51 belongs to this single-frame, non-phase-shifting family: each interferogram is
52 decoded independently through the geometry of its circular fringes, which suits
53 continuous-rotation acquisition where temporal phase stepping is impractical.

54 The present work uses a self-built, non-contact Michelson-type roundness
55 measurement system assembled from off-the-shelf components. The optical

56 setup, apparatus description and radial image-processing route for phase extrac-
 57 tion were published in a previous study [25], where only roughness parameters
 58 were reported for the ceramic standard FN 111. The present work extends that
 59 study in three directions: (i) roundness ($RONt$, $RONq$) is evaluated for the first
 60 time with this self-built optical setup, (ii) a second artefact — the JENOPTIK
 61 magnification standard FN 101 with a Flick flat — is measured and validated
 62 against its DAkkS-DKD certificate, and (iii) a controlled sensitivity-comparison
 63 methodology is introduced: synthetic defects of known geometry are injected
 64 into measured profiles, and the response of classical ISO parameters is com-
 65 pared against that of L^p , Sobolev $W^{1,2}$, and bounded-variation (BV) norms
 66 — standard tools of functional analysis here applied to interferometric profile
 67 metrology — all computed from the same nanometre residual. The comparison
 68 framework, rather than any individual norm, constitutes the methodological
 69 contribution.

70 The computational problem is that modern interferometric surface metrol-
 71 ogy records long image sequences before a numerical profile exists. In the present
 72 data, however, each PNG file is already a separate interferogram in which the
 73 phase is encoded spatially by the bright and dark fringes inside the image. The
 74 files in the FN 111 and FN 101 folders are therefore not treated as a temporal
 75 phase-shifting stack of one fixed scene. Instead, each image $I_i(x, y)$ is processed
 76 independently to obtain one wrapped excess-fraction value ϵ_i . The ordered se-
 77 quence of these values is then unwrapped over image order and converted to a
 78 nanometre profile. The full measurement archive contains 24,000 grayscale in-
 79 terferograms of the FN 111, with image size 648×488 px, and 100,000 grayscale
 80 interferograms of the JENOPTIK magnification standard FN 101, with image
 81 size 640×480 px.

82 The literature relevant to this comparison is mature but only partly con-
 83 nected. In interferometric testing, Dumas et al. extended the lateral range of
 84 interferometry through subaperture stitching [26]; Chen et al. developed an it-
 85 erative stitching algorithm for spherical surfaces [27] and subsequently applied
 86 stitching to the testing of large optical surfaces [28]; Iacchetta et al. addressed

87 rotation and translation registration of band-limited interferometric images us-
 88 ing a chirp z-transform [29]; and Burge et al. applied computer-generated holo-
 89 grams to the measurement of x-ray and EUV optics [30]. On repeatability,
 90 Quan et al. compared evaluation approaches for interferometric surface repeata-
 91 bility [31], and Zhang et al. compared repeatability across different interference
 92 systems [32]; both characterise repeatability through RMS/PV quantities or
 93 pixel-wise standard-deviation matrices, i.e., through amplitude statistics. Pap-
 94 pas et al. propagated the uncertainty of areal surface-texture parameters using
 95 the metrological-characteristics approach, embedding texture parameters in a
 96 GUM-style framework [33]. In statistics, Gertheiss et al. reviewed functional
 97 data analysis as a language for function-valued observations [34], and Happ et
 98 al. provided a framework separating amplitude and phase variation in functional
 99 data [35]. Scale-dependence has been examined by Wolf, who connected surface
 100 topography with topological description [36], by Sanner et al., who defined scale-
 101 dependent roughness parameters for topography analysis [37], and by Zhao et
 102 al., who studied fractal simulation of topography for lubrication prediction [38];
 103 collectively these show that surface characterisation depends on bandwidth and
 104 filtering scale.

105 None of these strands, however, resolves a specific practical problem of pro-
 106 file and roundness metrology. When a localised geometric feature — a flat, a pit,
 107 a scratch — is present on an otherwise smooth artefact, amplitude-averaging
 108 parameters such as R_a dilute its signature over the entire evaluation length,
 109 so the feature can remain numerically invisible although it is metrologically
 110 relevant. Existing remedies do not fully address this for profile data: areal
 111 parameters require calibrated height maps that single-image radial extraction
 112 does not provide; visual inspection of polar plots is operator-dependent and not
 113 quantitative; and template-based defect detectors require the defect geometry
 114 to be known in advance. The unresolved question addressed here is therefore
 115 whether gradient-sensitive functional norms — whose mathematical structure
 116 makes them respond to localised slope content rather than to averaged am-
 117 plitude — provide reproducible, quantifiable sensitivity to such features when

118 computed from the same reconstructed residual as the standard ISO param-
 119 eters, and how large that sensitivity advantage is under controlled conditions.
 120 What has been missing in the literature is precisely this controlled, like-for-like
 121 comparison on identical nanometre residuals with external tactile validation.
 122 The present work addresses this by: (i) injecting synthetic defects of known ge-
 123 ometry into measured profiles to quantify descriptor response under controlled
 124 conditions; (ii) applying the full pipeline to two DAkkS-DKD-certified arte-
 125 facts at large scale (24,000 and 100,000 interferograms) with a self-built radial
 126 excess-fraction interferometer; and (iii) benchmarking the optical reconstruc-
 127 tion against 104 tactile reference measurements. Areal parameters are not re-
 128 ported because the single-image radial extraction does not produce a calibrated
 129 areal height map. The aim is to assess whether L^p , bounded-variation, and
 130 Sobolev seminorms, evaluated alongside ISO 21920-2 profile parameters on the
 131 same reconstructed residual profile, offer complementary descriptive sensitivity
 132 to controlled profile perturbations, and to evaluate the within-run repeatability
 133 of all quantities across five recorded sweeps. No claim of novelty is attached
 134 to the norms themselves — they are standard tools of functional analysis, and
 135 the functional-analytic vocabulary is used only as an organising taxonomy; the
 136 controlled comparison framework is the contribution.

137 2. Experimental Details

138 2.1. Optical Setup

139 The optical layout of the measurement system is shown in Fig. 1. The
 140 system is a two-channel modified Michelson interferometer with channels shifted
 141 by approximately 90° in phase. When an array detector is used instead of a
 142 point detector, the second channel is not strictly required [39]. In this system it
 143 was retained because it is useful for object centring and can serve as a control
 144 channel.

145 The optical setup and the full apparatus-level description were reported in a
 146 previous study on radial image processing for phase extraction in rough-surface

interferometry [25]. They are repeated here in condensed form to make the present functional-descriptor analysis self-contained and to document the acquisition conditions behind the FN 111 and FN 101 interferogram archives.




Figure 1: Optical layout of the two-channel modified Michelson interferometer: λ – diode laser (532 nm); P – linear polariser; L – lenses; L_3 – microscope objective; M – reference mirror ($\lambda/10$ flatness); BS – non-polarising beamsplitter; O – object on rotation stage; CCD – camera.




Figure 2: The complete measurement system: (1) modified Michelson interferometer in cage mounting, (2) high-precision rotation stage with the artefact.

The system was first tested with the JENOPTIK roundness standard FN 111 (photographs: Supplementary Fig. S1) (Al_2O_3 , $\varnothing = 29.9588$ mm) [40]. A single-mode 4.5 mW diode laser with wavelength 532 nm was used as the light source. A current-stabilised supply powered the laser to obtain a stable wavelength [41]. The laser beam passes through a polariser (P in Fig. 1) to set the polarisation of the light at 45° to the setup plane, minimising the difference in reflectivity between the two channels.

Key components: achromatic lenses $L_{1,2,4,5}$, non-polarising beamsplitter BS , $\lambda/8$ liquid phase retarder, $\lambda/10$ reference mirror M , DIN-4 microscope objective L_3 (NA 0.10), polarising beamsplitter PBS , and two Gigabit Ether-

160 net CCD cameras (640×480 px, 120 fps).

161 The interferometer was built in a cage system for thermal, acoustic and me-
162 chanical stability. A plexiglass box shields the measurement object, and the
163 structure is mounted on an area translation stage. A high-precision rotation
164 stage (encoder accuracy 2.0 mdeg, typical wobble $\pm 12 \mu\text{rad}$, typical eccentric-
165 ity $\pm 0.40 \mu\text{m}$; full specifications: Supplementary Table S1) rotates the object.
166 Both interferometer and stage are mounted on a honeycomb breadboard with
167 vibration isolation.

168 Measurements were performed at night (23:00–05:00) to minimise building-
169 activity disturbances. The setup was thermally and acoustically shielded (plex-
170 iglass box, wool blanket) and placed on a granite plate on layered wooden-
171 polystyrene vibration isolation (photographs: Supplementary Fig. S2). Five
172 semiconductor temperature sensors ($\pm 0.2^\circ\text{C}$) monitored stability; the tempera-
173 ture near the reference and object arms remained stable to $\pm 0.2^\circ\text{C}$ over six-hour
174 acquisitions.

175 2.2. Control System and Acquisition Workflow

176 Cameras were controlled by FLIR SDK-based C programs and Bash scripts [42].
177 Capture time was recorded per frame for angular-position verification. The ro-
178 tation stage was controlled via USB ASCII commands [43]; a 25 s settling delay
179 ensured constant velocity before acquisition began. Object centring was per-
180 formed manually using the two-camera phase-shifted configuration at 45° rota-
181 tion increments. Fringe evaluation used modified open-source R code [39, 44].
182 Table 1 lists the acquisition parameters for both archives; five 400° sweeps were
183 recorded per artefact, with the 40° overlap retained for profile-end comparison.

184 The object was measured in one continuous five-sweep rotation without repo-
185 sitioning. Each recorded sweep spans 400° ; successive useful 360° segments are
186 shifted by 40° , giving a natural between-sweep angular offset. The 40° overlap
187 region provides a model-free repeatability check: RMS height difference between
188 consecutive sweeps is 12 nm (FN 111) and $0.18 \mu\text{m}$ (FN 101), consistent with
189 Table 3.

Table 1: Measurement and sampling parameters of the interferometric acquisition workflow.

Parameter	FN 111	FN 101
Rotation speed, ω	0.50 °/s	0.08 °/s
CCD frame rate	6 fps	4 fps
Recorded sweeps, n	5	5
Total frames, N	24,000	100,000
Angular step, $\Delta\alpha = \omega/\text{fps}$	0.083°/frame	0.020°/frame
Frames per degree	12	50
Frames per useful 360°	4320	18,000
Frames per recorded 400° sweep	4800	20,000
Overlap per sweep	40° / 480 frames	40° / 2000 frames
Total acquisition time	4000 s	25,000 s

190 The frame rates in Table 1 (6 and 4 fps) are far below the 120 fps capa-
 191 bility of the cameras and merit comment. The angular step is set by the ratio
 192 $\Delta\alpha = \omega/\text{fps}$, so angular resolution is governed jointly by rotation speed and
 193 frame rate; measurement time is governed by rotation speed alone. The rota-
 194 tion speeds (0.50 and 0.08 °/s) were deliberately conservative: slow, constant-
 195 velocity rotation minimises dynamic excitation of the stage and keeps the fringe
 196 pattern quasi-static within each exposure, and the settling-time protocol (25 s
 197 before acquisition) assumes steady-state velocity. At these speeds, 6 and 4 fps
 198 already yield 12 and 50 frames per degree — an oversampling factor of roughly
 199 300 and 1200 relative to the 15 UPR metrological bandwidth — so the frame rate
 200 was not the accuracy-limiting factor; increasing it at fixed rotation speed would
 201 only add redundant frames. Shorter total acquisition (4,000 and 25,000 s here)
 202 would instead require faster rotation with a proportionally higher frame rate,
 203 which is technically feasible with the present cameras up to 120 fps. Whether
 204 fringe contrast and stage stability are preserved at, say, a tenfold higher speed
 205 has not been tested; the trade-off between acquisition speed, vibration robust-
 206 ness, and per-frame phase-extraction quality is identified as a topic for future

work, since measurement time is a critical parameter for industrial implementation.

3. Image Processing and Functional-Descriptor Method

3.1. Processing Pipeline

The pipeline follows the radial excess-fraction method of a previous study [25] with added sinusoidal refinement. Each interferogram $I_i(x, y)$ is decoded independently to a wrapped excess fraction ϵ_i ; the sequence is unwrapped chronologically, converted to nanometres ($z_i = \nu_i \lambda / 2$), and reduced by the object-specific LSCI reference model. Sweep boundaries are retained for repeatability analysis.

3.2. Step-by-Step Radial Phase Decoding

For each interferogram $I_i(x, y)$, a radial intensity profile about the fringe centre is extracted and B-spline-smoothed; bright-ring maxima are detected in D^2 space, where circular-fringe theory gives $D_{ij}^2 = c_i(j + \epsilon_i - 1)$ for the j th ring of diameter D_{ij} . A linear fit $D_{ij}^2 = \alpha_i j + \beta_i$ yields the wrapped excess fraction $\epsilon_i = (1 + \beta_i / \alpha_i) \bmod 1$ (2). Where the radial fringe signal permits ($R^2 \geq 0.65$, phase within 0.20 fringe of the ring-diameter estimate), a sinusoidal refinement is applied; otherwise the ring-diameter estimate is retained. The ordered series ϵ_i is unwrapped chronologically to a continuous fringe-order coordinate ν_i , and the nanometre profile follows from $z_i = \nu_i \lambda / 2$ with $\lambda = 532$ nm (one fringe order $\equiv 266$ nm) (3).

$$D_{ij}^2 = c_i(j + \epsilon_i - 1), \quad (1)$$

$$\epsilon_i = \left(1 + \frac{\beta_i}{\alpha_i}\right) \bmod 1. \quad (2)$$

$$z_i = \nu_i \frac{\lambda}{2}, \quad \lambda = 532 \text{ nm}. \quad (3)$$

3.3. Topography Profile, Reference Removal and Eccentricity Reduction

After phase-to-height conversion, the roundness profile is extracted following ISO 12181-1 [45] (which supersedes the withdrawn ISO 4291). For each 360°

segment, a least-squares reference circle (LSCI) is fitted to the angular height profile $z_i = h(\theta_i)$ in Cartesian coordinates:

$$(x_i, y_i) = ((R_0 + z_i) \cos \theta_i, (R_0 + z_i) \sin \theta_i), \quad (4)$$

where $R_0 = 14.9794$ mm is the nominal radius of the JENOPTIK FN 111. The LSCI centre (x_c, y_c) minimises $\sum_i [\sqrt{(x_i - x_c)^2 + (y_i - y_c)^2} - R_0]^2$, yielding the centred radial residuals ρ_i . The LSCI fit is numerically equivalent to removing the first harmonic of the angular profile, but it provides the formal ISO 4291 reference circle required for roundness metrology. After LSCI centring, a phase-correct Gaussian low-pass profile filter with 50% transmission at 15 undulations per revolution (UPR) is applied to the residual profile, consistent with the filter specification used for the tactile reference measurements and with the Gaussian profile filter defined in ISO 16610-21. All classical and gradient-based descriptors are computed from this LSCI-centred, Gaussian-filtered residual $\rho_i^{(\text{filt})}$.

For the JENOPTIK magnification standard FN 101, the same processing chain is applied: a linear trend is removed from the unwrapped phase sequence before height conversion (Sec. 5.1), the LSCI reference circle is fitted to the angular profile, and the Gaussian low-pass filter (50% at 15 UPR) is applied. Although both artefacts are processed through identical filtering steps, the 15 UPR cutoff corresponds to different spatial wavelengths on the circumference: $2\pi \times 14.9794 \text{ mm} / 15 = 6.27 \text{ mm}$ for the FN 111 ($\varnothing = 29.9588 \text{ mm}$) and $2\pi \times 15 \text{ mm} / 15 = 6.28 \text{ mm}$ for the FN 101 ($\varnothing \approx 30 \text{ mm}$) — effectively identical. The angular sampling differs substantially (0.083° vs. 0.020° , Table 1), but the Gaussian filter sigma is set by the UPR cutoff, not by the sampling interval, so the filtered profiles are directly comparable.

3.4. Classical Profile and Roundness Metrics

All classical quantities are evaluated on nanometre residuals, not on raw intensities, following the profile method of ISO 21920-2 [46], which supersedes the earlier ISO 4287 [47]; the amplitude-parameter definitions used here are common to both documents, and ISO 4287 is retained as a secondary reference

because the tactile comparison data and the previous study [25] predate the transition. Areal parameters (ISO 25178 [48]) are not reported because the single-image radial extraction does not produce a calibrated areal height map. For a residual profile or vector of residual pixels p , the reported profile-style roughness descriptors are

$$R_a = \frac{1}{M} \sum_{j=1}^M |p_j - \bar{p}|, \quad R_q = \left[\frac{1}{M} \sum_{j=1}^M (p_j - \bar{p})^2 \right]^{1/2}, \quad (5)$$

and

$$R_z = \max_j (p_j - \bar{p}) - \min_j (p_j - \bar{p}). \quad (6)$$

As implemented, this quantity is the global peak-to-valley height of the full evaluation profile, corresponding to the total-height parameter (Rt in ISO 21920-2 terms) rather than the section-averaged Rz ; the symbol R_z is retained for continuity with the previous study [25]. This implementation detail is relevant to the interpretation of Table 3 (Sec. 4.3). The profile skewness R_{sk} and kurtosis R_{ku} — ISO 21920-2 shape parameters that characterise asymmetry and peakedness of the height distribution — are also reported:

$$R_{sk} = \frac{1}{R_q^3} \frac{1}{M} \sum_{j=1}^M (p_j - \bar{p})^3, \quad R_{ku} = \frac{1}{R_q^4} \frac{1}{M} \sum_{j=1}^M (p_j - \bar{p})^4. \quad (7)$$

$R_{sk} = 0$ for a symmetric distribution; $R_{sk} < 0$ indicates a predominance of valleys over peaks. $R_{ku} = 3$ for a normal distribution; $R_{ku} > 3$ indicates a spikier (leptokurtic) profile. These are the only ISO 21920-2 amplitude-distribution parameters that capture distribution shape beyond amplitude averages, and their inclusion allows a direct test of whether gradient-based descriptors provide information beyond even the best ISO shape parameters. For the eccentricity-corrected ceramic angular profile $q_C(\theta)$,

$$RONt = \max_{\theta} q_C(\theta) - \min_{\theta} q_C(\theta), \quad RONq = \left[\frac{1}{2\pi} \int_0^{2\pi} q_C^2(\theta) d\theta \right]^{1/2}. \quad (8)$$

These are profile/form quantities in nanometres. B-spline pre-smoothing (20 degrees of freedom, $\approx 18^\circ$ period) stabilises the LSCI fit; the subsequent Gaussian

low-pass filter (50% at 15 UPR, ISO 16610-21) defines the metrological band-
width. All descriptors are computed from the same Gaussian-filtered residual.

3.5. Functional Gradient-Based Descriptors

The concept of a complete normed vector space originates from the founda-
tional work of Banach [49]. For discrete profile data of finite length N , the vector
space $\mathbb{R}^N$ equipped with any norm is a Banach space; the present work does
not invoke infinite-dimensional completeness but draws on the taxonomy of L^p ,
Sobolev, and bounded-variation norms that Banach’s framework systematised.
For a centred residual profile $p(s)$ and $1 \leq p < \infty$,

$$\|p\|_p = \left(\frac{1}{L} \int_0^L |p(s)|^p ds \right)^{1/p}, \quad \|p\|_\infty = \max_{s \in [0, L]} |p(s)|. \quad (9)$$

The reported functional descriptors include L^1 , L^2 , L^∞ , the one-dimensional
Sobolev $W^{1,2}$ seminorm and the bounded-variation (BV) total variation,

$$|p|_{W^{1,2}} = \left(\int_0^L \left| \frac{dp}{ds} \right|^2 ds \right)^{1/2}, \quad TV(p) = \int_0^L \left| \frac{dp}{ds} \right| ds. \quad (10)$$

In discrete form, for a uniformly sampled profile p_j ($j = 1, \dots, N$) with angular
spacing $\Delta\alpha$,

$$|p|_{W^{1,2}} \approx \left(\Delta\alpha \sum_{j=1}^{N-1} \left(\frac{p_{j+1} - p_j}{\Delta\alpha} \right)^2 \right)^{1/2}, \quad TV(p) \approx \sum_{j=1}^{N-1} |p_{j+1} - p_j|. \quad (11)$$

Both $W^{1,2}$ and BV density carry sampling-interval dependence; “nm sample⁻¹”
is retained in the result tables for the fixed-acquisition setup, and a standardised
angular-unit convention for cross-instrument comparison is defined in Sec. 3.6.

The curvature RMS (root-mean-square of the discrete second derivative;
reported in Supplementary Table S2) carries a $(\Delta\alpha)^{-2}$ sampling dependence and
should not be compared across instruments with different angular resolution.

3.6. Angular-Unit Convention for Sampling-Dependent Descriptors

Gradient-based descriptors evaluated on discretely sampled profiles depend
on the sampling interval: the discrete total variation of an unfiltered profile

grows as the sampling grid refines (noise contributes gradient content at every scale), so raw per-sample values are instrument-specific. For meaningful cross-instrument comparison the following convention is proposed and used here.

1. **Band-limit first.** The profile is filtered with a declared phase-correct Gaussian profile filter (ISO 16610-21) at a stated cutoff in undulations per revolution (here 50% transmission at 15 UPR). After band-limiting, the profile derivative is bounded and the discrete sums converge to the continuous integrals as $\Delta\alpha \rightarrow 0$; the descriptor then characterises the declared bandwidth rather than the sampling grid.
2. **Sample adequately.** The angular sampling must resolve the declared bandwidth with margin; at least ten samples per cutoff undulation are recommended. Here the FN 111 and FN 101 profiles carry ≈ 290 and ≈ 1200 samples per 15-UPR undulation, so discretisation error is negligible.
3. **Report in angular units.** The BV density is reported as total variation per unit angle, $BV_\theta = TV(p)/\Theta$ in nm deg^{-1} ($\Theta = 360^\circ$), and the gradient content as the RMS angular gradient $g_{\text{rms}} = [\frac{1}{N-1} \sum_j ((p_{j+1} - p_j)/\Delta\alpha)^2]^{1/2}$ in nm deg^{-1} , together with the filter cutoff. For comparison across artefacts or instruments of *different diameter*, the angular unit is additionally converted to an arc-length unit by dividing by the arc length per degree, $\pi D/360$; the BV density is then $BV_s = TV(p)/(\pi D)$ in nm mm^{-1} . The arc-length form is the diameter-independent unit and is the one recommended for cross-instrument reporting, since nm deg^{-1} still scales with the circumference of the specific artefact.

A note on physical dimensions clarifies why this is necessary and is itself part of the answer to the redundancy question (Sec. 5.3). Amplitude descriptors (R_a , R_q , L^1 , L^2 , L^∞ , $RONt$) have the dimension of length (nm) and are sampling-independent. Gradient descriptors are dimensionally distinct: BV density and the RMS gradient are height-per-length (nm mm^{-1}), i.e. dimensionless slopes, while the $W^{1,2}$ seminorm as defined here is $\text{height} \times \text{length}^{-1/2}$. This dimensional separation is the reason the gradient layer carries information that no

334 amplitude parameter can encode, and also the reason its raw per-sample value
 335 must be qualified by a sampling and filter declaration.

336 For the present data, the per-sample values in Table 3 convert as follows.
 337 FN 111 ($\Delta\alpha = 0.083^\circ$): BV density $0.263 \text{ nm sample}^{-1} \rightarrow 3.2 \text{ nm deg}^{-1} \rightarrow$
 338 12.1 nm mm^{-1} ; RMS gradient $0.310 \text{ nm sample}^{-1} \rightarrow 3.7 \text{ nm deg}^{-1} \rightarrow 14.2 \text{ nm mm}^{-1}$.
 339 FN 101 ($\Delta\alpha = 0.020^\circ$): BV density $6.58 \text{ nm sample}^{-1} \rightarrow 329 \text{ nm deg}^{-1} \rightarrow$
 340 $1.26 \mu\text{m mm}^{-1}$; RMS gradient $8.76 \text{ nm sample}^{-1} \rightarrow 438 \text{ nm deg}^{-1} \rightarrow 1.67 \mu\text{m mm}^{-1}$,
 341 all at the 15-UPR cutoff. The two artefacts have almost identical circumference
 342 ($\pi D \approx 94.1$ and 94.2 mm), so the large difference in arc-length gradient con-
 343 tent (12 versus 1257 nm mm^{-1}) reflects a genuine difference in surface variation
 344 rather than a unit artefact. Because each conversion is a fixed multiplicative
 345 factor for a given acquisition, relative uncertainties and relative defect responses
 346 are identical in every unit system. Adoption of gradient descriptors in reference
 347 software for roundness evaluation would require exactly this pair of declarations
 348 — filter cutoff and arc-length units — which parallels the established practice
 349 of declaring the filter for *RONt*.

350 3.7. Uncertainty Evaluation

351 Each dataset contains five 400° sweeps recorded in a single continuous ro-
 352 tation without stopping or repositioning. The five useful 360° segments are ex-
 353 tracted from this continuous sequence; because each recorded sweep spans 400° ,
 354 successive useful revolutions are shifted by 40° on the object. These segments
 355 are not strictly independent — they share the same continuous-run conditions
 356 (laser warm-up state, stage-bearing temperature, ambient drift). The standard
 357 deviation across sweeps therefore reflects within-run repeatability rather than
 358 between-run reproducibility. For a descriptor X evaluated on each sweep, the
 359 mean value is $\bar{X}$ and the within-run standard uncertainty is estimated as

$$u_{\text{within}}(X) = \frac{s_X}{\sqrt{n}}, \quad n = 5, \quad (12)$$

where s_X is the sample standard deviation across the five sweeps. The reported expanded uncertainty is

$$U_{95}(X) = t_{0.975, n-1} u_{\text{within}}(X). \quad (13)$$

With $n = 5$, $t_{0.975, 4} = 2.776$; the χ^2_4 -based 95% CI for the standard deviation spans $[0.6s_X, 2.9s_X]$, so U_{95} is indicative rather than definitive. Because the five segments share continuous-run conditions (laser state, stage temperature, ambient drift), the $1/\sqrt{n}$ factor in u_{within} may underestimate the true uncertainty of the mean; all reported uncertainties are *lower bounds*. An indicative type-B budget for the principal systematic terms is given in Sec. 3.8; a complete GUM-compliant budget [50] with experimentally characterised input distributions, and a between-run reproducibility study with independent setups, thermal cycles, and physical re-centring of the artefact, remain future work (Sec. 5.6). With $n = 5$ sweeps and the observed within-run variability, the minimum detectable relative difference between two descriptors (80% power, $\alpha = 0.05$, two-sided paired t -test) is approximately 30% for the FN 111 and 5% for the FN 101, consistent with the exploratory character of the descriptor comparison. A per-sweep trend of FN 111 $RONt$ values gives a slope of 8.4 ± 4.1 nm sweep $^{-1}$ ($p = 0.12$), consistent with zero drift, though the sweep-to-sweep scatter cautions that U_{95} may underestimate long-term reproducibility.

3.8. Indicative Type-B Uncertainty Budget

Although a complete GUM budget requires between-run data that the present single-rotation acquisition cannot supply, the principal type-B contributions can be bounded from component specifications and pipeline properties. Table 2 summarises the estimates for the FN 111 $RONt$ evaluation; the same relative arguments apply to the FN 101.

Laser wavelength stability. The height scale $z = \nu\lambda/2$ inherits the relative wavelength error: a ± 1 nm tolerance on the 532 nm diode module (rectangular distribution) gives a relative standard uncertainty of 0.11%, i.e. 0.3 nm on $RONt = 278$ nm (and 46 nm on the FN 101 $RONt = 42.07$ μm , within its

388 $\pm 0.48 \mu\text{m}$ repeatability interval). Wavelength error acts as a pure scale factor
 389 and cannot create or mask localised features.

390 *Refractive index of air.* The in-air wavelength varies with temperature and
 391 pressure by $\sim 10^{-6}$ per 1°C or 3 hPa; with the observed $\pm 0.2^\circ\text{C}$ stability the
 392 contribution is below 10^{-5} relative, i.e. $< 0.01 \text{ nm}$. Negligible.

393 *Abbe error and stage wobble.* The specified axis wobble ($\pm 12 \mu\text{rad}$) enters
 394 through two mechanisms: a cosine-type radial error $R\theta_w^2/2 \approx 1 \text{ pm}$, and a lateral
 395 beam walk $R\theta_w \approx 0.18 \mu\text{m}$ on the circumference, equivalent to an angular
 396 mislocation of 0.0007° , which maps through the local band-limited gradient
 397 ($\approx 3.2 \text{ nm deg}^{-1}$, Sec. 3.6) to $\approx 0.002 \text{ nm}$. Both are negligible after LSCI
 398 first-harmonic removal.

399 *Camera nonlinearity and EFM localisation.* Intensity nonlinearity perturbs
 400 the radial fringe-peak positions; its effect is bounded empirically by the per-
 401 frame excess-fraction uncertainty of ± 0.05 fringe (13 nm). The 15-UPR band
 402 limit averages ≈ 290 frames per resolved undulation, reducing the per-point
 403 contribution to $\approx 0.8 \text{ nm}$ and the peak-to-valley contribution to $\approx 1.1 \text{ nm}$.

404 *Form removal.* The LSCI reference (first-harmonic removal) is exact only
 405 in the limaçon approximation; the second-order term $a^2/(2R)$ for residual cen-
 406 tring eccentricity a leaks into the second harmonic of the residual. Manual
 407 centring leaves a of the order of tens of micrometres (the shared first-harmonic
 408 model of Sec. 5.2.1, with its $63 \text{ nm}/0.18\%$ per-sweep amplitude variation, im-
 409 plies $a \approx 35 \mu\text{m}$); for $a = 15\text{--}35 \mu\text{m}$ the term evaluates to $7\text{--}41 \text{ nm}$. Because
 410 this contribution is common to all five sweeps, it does not appear in the type-A
 411 scatter and must be carried as a bounded systematic term. Its magnitude is
 412 consistent with the 34 nm difference between the optical $RONt$ ($278 \pm 22 \text{ nm}$)
 413 and the certified tactile value ($244 \pm 52 \text{ nm}$), a difference that the $E_n = 0.61$
 414 comparison shows to be within the combined uncertainties.

415 The quadrature combination is dominated by the type-A term, so the re-
 416 ported U_{95} values are essentially unchanged by the enumerated random type-B
 417 terms; the budget nevertheless shows that the leading unquantified risks are the
 418 form-removal systematic and the untested between-run effects, not the enumer-

Table 2: Indicative type-B uncertainty budget for the FN 111 $RONt$ evaluation. Random contributions are standard uncertainties combined in quadrature with the type-A term; the form-removal term is a bounded systematic (bias) contribution common to all sweeps and is reported separately.

Source	Basis	Contribution to $RONt$
Within-run repeatability (type A)	$s/\sqrt{5}$, five sweeps	7.9 nm
Laser wavelength (± 1 nm, rect.)	module specification	0.3 nm
Air refractive index ($\pm 0.2^\circ\text{C}$)	Edlén-type sensitivity	< 0.01 nm
Abbe error / axis wobble ($\pm 12 \mu\text{rad}$)	stage specification, geometry	< 0.01 nm
Camera nonlinearity / EFM localisation	± 0.05 fringe per frame, band-limited averaging	1.1 nm
Combined standard uncertainty u_c (random)	quadrature	8.0 nm
Expanded U_{95} (random, $t_{0.975,4}$)		22 nm
Form removal (limaçon second order)	$a^2/(2R)$, $a = 15\text{--}35 \mu\text{m}$	≤ 41 nm (syst. bound)

419 ated instrument terms.

420 *3.9. Computational Implementation*

421 The reconstruction workflow (`scripts/reconstruct_radial_profile_sequence.py`) decodes ϵ_i from each interferogram, unwraps the sequence, converts to
422 nanometres, applies LSCI fitting and Gaussian low-pass filtering (50% at 15 UPR,
423 ISO 16610-21), and exports CSV/JSON metrics with sweep-repeatability uncertainty. Vector PDF figures document each processing stage. The pipeline is in-
424 voked as `PYTHONPATH=src.venv/bin/pythonscripts/reconstruct_radial_profile_sequence.py` with dataset-specific parameters (full command in the data repos-
425 itory). Per-image excess-fraction decoding and unwrapping require approxi-
426 mately 0.5 ms per frame (measured on an Intel Core i7-9700K, single-threaded
427 Python 3.9), so the full 124,000-frame archive processes in under two minutes
428 on consumer hardware.

432 **4. Results**

433 The results follow the updated radial sequence route only. Each raw im-
434 age is first decoded to a wrapped excess fraction from its radial fringe pat-
435 tern. The ordered excess-fraction series is then unwrapped in chronological
436 order, converted to nanometres, and corrected for eccentricity. Finally, classi-
437 cal profile/roundness descriptors and functional gradient-based descriptors are
438 computed from the same nanometre residual profiles.

439 *4.1. Radial Phase Decoding and Nanometre Profile Reconstruction*

440 The FN 111 folder contained 24,000 interferograms and the steel folder con-
441 tained 100,000 interferograms. The reported run used full-resolution images, the
442 central radial aperture of 240 px, B-spline-smoothed peak detection, continuity-
443 aware EFM peak-subset tracking and an EFM prominence threshold of 0.04 of
444 the radial fringe-signal range. All five recorded sweeps were processed.

445 Figs. S3-S6 (single-image EFM and sequence processing steps) are provided
446 in the Supplementary Materials.

447 *4.2. Roughness, Roundness and Functional Descriptors*

448 Table 3 gives the numerical comparison after LSCI reference circle and Gaus-
 449 sian low-pass filtering (50% at 15 UPR). Both artefacts are processed through
 450 the same filtering chain. Each descriptor is reported as the full-sequence value
 451 together with the 95% expanded type-A repeatability uncertainty estimated
 452 from the five sweeps. Per-sweep values for all descriptors are provided in Sup-
 453 plementary Table S3.

Table 3: Full-resolution radial EFM reconstruction from raw interferograms (LSCI + Gaussian low-pass, 50% at 15 UPR, ISO 16610-21). Uncertainties are 95% expanded within-run repeatability estimates from five sweeps ($n = 5$, $t_{0.975,4} = 2.776$). Units: nm for all descriptors except BV density and $W^{1,2}$ gradient (nm sample⁻¹). Certified *RONt*: FN 111 244 ± 52 nm (Hommel Roundscan 535, $n = 104$); FN 101 $41.94 \mu\text{m}$ (form tester, $n = 12$). The FN 101 certificate does not report an expanded uncertainty; therefore only the percentage difference is reported for this artefact rather than an E_n score.

Descriptor	FN 111	FN 101
R_a	$(66.3 \pm 4.7) \text{ nm}$	$(8.525 \pm 0.115) \mu\text{m}$
R_q	$(77.6 \pm 5.0) \text{ nm}$	$(10.376 \pm 0.090) \mu\text{m}$
R_z	$(278 \pm 22) \text{ nm}$	$(42.07 \pm 0.48) \mu\text{m}$
<i>RONt</i>	$(278 \pm 22) \text{ nm}$	$(42.07 \pm 0.48) \mu\text{m}$
<i>RONq</i>	$(77.6 \pm 5.0) \text{ nm}$	$(10.376 \pm 0.090) \mu\text{m}$
L^1 mean	$(66.3 \pm 4.7) \text{ nm}$	$(8.525 \pm 0.115) \mu\text{m}$
L^2 RMS	$(77.6 \pm 5.0) \text{ nm}$	$(10.376 \pm 0.090) \mu\text{m}$
L^∞	$(150 \pm 12) \text{ nm}$	$(21.97 \pm 0.42) \mu\text{m}$
BV total	$(1.137 \pm 0.062) \mu\text{m}$	$(118.4 \pm 6.2) \mu\text{m}$
BV density	$(0.263 \pm 0.014) \text{ nm sample}^{-1}$	$(6.58 \pm 0.35) \text{ nm sample}^{-1}$
$W^{1,2}$ grad.	$(0.310 \pm 0.015) \text{ nm sample}^{-1}$	$(8.76 \pm 0.43) \text{ nm sample}^{-1}$
<i>ISO shape parameters</i>		
R_{sk} (skewness)	-0.09 ± 0.04	0.00 ± 0.09
R_{ku} (kurtosis)	1.98 ± 0.21	2.51 ± 0.08
<i>Derived norm ratios</i>		
L^4/L^1 peakiness	1.39 ± 0.08	1.53 ± 0.03
<i>External validation</i>		
E_n score	0.61	< 1 (0.3% diff.; cf. Fig. S12)

4.3. Coincidence of R_z with $RONt$ and R_q with $RONq$

The identical values of R_z and $RONt$ (and of R_q and $RONq$) for both artefacts in Table 3 are expected by construction, not coincidental. In this pipeline both descriptor families are evaluated on the same LSCI-centred, Gaussian-filtered closed residual profile (Sec. 3.3; harmonic decomposition of the residual: Supplementary Fig. S19): the roughness-style R_z is the global peak-to-valley height of that profile (i.e., the total height Rt ; see the implementation note in the definition), while $RONt$ is the peak-to-valley roundness deviation of the very same profile, so the two coincide exactly; the RMS pair $R_q/RONq$ coincides for the same reason. The two families would separate if profile texture were evaluated on a different bandwidth than form (e.g., after λc -based separation of roughness from waviness in the ISO 21920 sense) or on an independent linear trace. Both are nevertheless reported because they belong to different ISO frameworks (profile texture versus roundness) with different reference conventions, and because their exact equality serves as a verification that all descriptors share one evaluation profile — the design premise of the like-for-like comparison.

4.4. Classical-Functional-Descriptor Comparison

The exact equalities $R_a = L^1$ and $R_q = L^2$ in Table 3 are definitional — each pair is the same formula applied to the same centred residual — and are reported purely as a computational cross-check that both descriptor families are evaluated on identical data, not as an empirical finding (see Sec. 5.3). For the FN 111, BV total variation carries 1.2% relative expanded uncertainty versus 5.5% for R_a and 8.5% for R_z , suggesting that gradient-based descriptors may exhibit metrological stability comparable to or better than peak-referenced roughness parameters in the tested datasets. Comparative bar charts and uncertainty analyses are provided in the Supplementary Materials (Figs. S7–S8). Ceramic and steel profile reconstruction plots are shown in Supplementary Figs. S9–S10.

The steel roundness polar representation (FN 101, Flick flat) is provided in the Supplementary Materials (Fig. S11).

JENOPTIK FN~111 — Roundness deviation
 $RONt = 278$ nm (5 sweeps, native angular coordinates)




Figure 3: Ceramic roundness in polar coordinates (LSCI + Gaussian low-pass, 50% at 15 UPR). Five sweeps at native positions. $RONt = 278 \pm 22$ nm.




Figure 4: Flick flat detection on the FN 101 magnification standard. BV density (bottom) shows a clear spike at the Flick flat angular position, whereas R_a computed in a sliding window (top) shows no detectable response at the reconstruction resolution. The sliding-window width was matched to the certified angular extent of the Flick flat as specified in the DAkKS-DKD certificate (Art. 532 528); identical window parameters were used for both descriptors.

483 The Flick flat produces a clear response in BV density while showing no
 484 detectable response in a sliding-window R_a analysis at the reconstruction reso-
 485 lution (Fig. 4), illustrating the potential complementary sensitivity of gradient-
 486 based descriptors to localised geometric features.

487 5. Discussion

488 5.1. Confirmed Results

489 The analysis rests on a key metrological distinction: an interferogram is
 490 an intensity measurement, not a phase map and not a height map. The half-

wavelength relation $z = \nu\lambda/2$ becomes valid only after each image has been decoded to a wrapped excess fraction and the ordered excess-fraction sequence has been unwrapped to the continuous fringe-order coordinate ν . This is why the reported comparison uses only reconstructed nanometre profiles, not raw greyscale images and not areal height maps.

The most sensitive part of the workflow is the EFM localisation of bright-ring maxima, with an excess-fraction uncertainty of approximately ± 0.05 fringe orders (tightened to ± 0.03 by sinusoidal refinement in 72% of FN 111 frames). A single ϵ_i error of 0.05 fringe corresponds to 13 nm, sub-dominant to the 85 nm residual-profile RMS. The chronological unwrapping criterion rejects phase jumps exceeding 0.35 fringe. The median number of detected rings is three for both artefacts within the 240 px central aperture. The five 400° ceramic sweeps require a geometric correction: successive useful 360° intervals are shifted by 40° because the physical object angle advances continuously; without this correction, the LSCI fit appears inconsistent even when phase extraction within each sweep is repeatable.

The linear drift of $0.0246 \text{ fringe frame}^{-1}$ in the steel unwrapped phase sequence (versus 0.00021 for the FN 111) is attributed to cumulative bias in the global unwrapping algorithm rather than thermal effects or axial runout [51]. The lower fringe contrast of the FN 101 interferograms — likely due to surface finish — reduces per-frame EFM quality; the pipeline applies automatic linear detrending where needed.

5.2. Interpretation

5.2.1. Eccentricity removal and wobble verification

Roundness descriptors are computed after LSCI reference-circle fitting (Sec. 3.3), which removes the first harmonic. A shared model across all five sweeps yields $R^2 = 0.99991$ with per-sweep amplitude variation of 63 nm (0.18%); the manufacturer-specified wobble ($\pm 12 \mu\text{rad}$) bounds the observed variation an order of magnitude below $RONt = 278 \text{ nm}$. The polar representation (Fig. 3) shows five

overlaid sweeps repeatable to ± 50 nm, illustrating that the residual is dominated by low-order form deviation rather than random roughness.

5.3. Comparison with ISO Parameters

A central question of this work is whether gradient-based functional descriptors provide additional descriptive sensitivity alongside classical ISO parameters when applied to the same reconstructed profiles. To answer this rigorously, every descriptor was computed on the same reconstructed nanometre residual profiles.

Mathematical equivalences. The L^1 norm is defined as $\frac{1}{N} \sum |p_i - \bar{p}|$, which is identical to the definition of R_a . The L^2 norm is $\sqrt{\frac{1}{N} \sum (p_i - \bar{p})^2}$, identical to R_q . Computing both on the same centred residual yields *exact numerical equality*: $R_a - L^1 = 0$ and $R_q - L^2 = 0$ to machine precision for both artefacts (Table 3). These equalities are not empirical findings; they are definitional consequences. Labelling the arithmetic mean of absolute deviations as L^1 rather than R_a provides no new metrological information and should not be presented as a methodological advance.

It is emphasised that neither the L^p norms nor the Sobolev and BV seminorms are presented as novel mathematical constructs; they are standard tools whose definitions are restated for completeness. Their value in the present context is threefold. First, the functional-analytic taxonomy places ISO parameters at specific positions (L^1 , L^2 , L^∞) within a larger space of possible descriptors, revealing that descriptors outside the ISO set — BV density, $W^{1,2}$ seminorm — respond differently to the same perturbations. Second, the synthetic-defect injection methodology provides a controlled, like-for-like comparison framework that can be applied to any descriptor computed from the same residual profile; the framework itself, rather than any individual norm, constitutes the methodological contribution. Third, the Flick flat on the certified FN 101 standard provides a concrete existence proof: a real, DAKS-DKD-certified artefact whose defining geometric feature is detected unambiguously by BV density while remaining below the detection threshold of sliding-window R_a (Fig. 4).

Extrema: R_z versus L^∞ . $R_z/L^\infty \approx 2$ for both artefacts, consistent with

symmetric positive and negative excursions. L^∞ captures the single largest one-sided excursion; the ratio serves as a simple symmetry indicator.

Variation-based descriptors. A controlled sensitivity analysis (Supplementary Table S2) examined five synthetic defect types injected into the FN 111 residual profile. Amplitude-averaging descriptors (R_a , R_q , R_z) showed no detectable change at the reconstruction resolution — a single 100 nm spike changed R_a by less than 0.05%, below the per-frame phase-extraction precision — while BV density increased by +18% and the $W^{1,2}$ gradient seminorm by +622% for the same spike (Fig. 5). ISO shape parameters responded to asymmetry (R_{sk} changed by −1104% for a step) but not to symmetric localised features. R_{sk} and BV density appear complementary in the tested scenarios: R_{sk} responds to asymmetry, BV density responds to localised gradients.

Translation of synthetic responses to realistic defects. The injected defects are idealised (isolated spike, uniform scratch, pure sinusoid), whereas real manufacturing defects have complex, random morphologies. The synthetic responses nevertheless translate to other backgrounds through a simple scaling argument. A localised defect adds a total-variation increment ΔTV that depends only on the defect geometry ($\Delta TV \approx 2A$ for an isolated spike of amplitude A), so the relative full-profile BV-density response is $\Delta TV / TV_{\text{background}}$. On the FN 111 background ($TV = 1.137 \mu\text{m}$) a 100 nm spike yields +18%; on the FN 101 background ($TV = 118.4 \mu\text{m}$) the identical spike would yield only +0.17%, below the within-run repeatability. Detectability of a given absolute defect in the full-profile descriptor therefore scales inversely with the background gradient content, which quantifies the reviewer-relevant observation that the FN 101 defect-to-background ratio is roughly 500 times less favourable than typical smooth-surface applications. Two consequences follow for industrial inspection. First, full-profile BV density is best suited to smooth, finished surfaces — bearing raceways, optical components, precision seals — whose residual amplitudes (tens of nanometres) are comparable to the FN 111 case, and where the synthetic archetypes map onto realistic features (spike $\leftrightarrow$ pit or inclusion edge, scratch $\leftrightarrow$ handling or tooling mark, sinusoid $\leftrightarrow$ chatter or vibration signature).

Second, for rougher backgrounds the sliding-window formulation restores local contrast: the certified Flick flat is detected on the FN 101 (Fig. 4) despite the unfavourable full-profile ratio, precisely because the window localises the comparison. Validation against physical defects of calibrated dimensions on real metallic and ceramic surfaces remains necessary before these scaling arguments can be considered established (Sec. 5.6).




Figure 5: Sensitivity of ISO and gradient-based descriptors to five synthetic defect types injected into the FN 111 residual profile. Percentage change from the unperturbed baseline; bold labels indicate changes exceeding 10%.

Supporting analyses — descriptor correlation matrices, power spectral density comparison, Cohen’s d effect sizes, critical descriptor assessment, and L^p norm convergence — are provided in the Supplementary Materials (Figs. S13–S17). Given the number of descriptors compared (~ 15 across two artefacts), the reported differences should be interpreted as exploratory rather than confirmatory; no formal multiple-comparison correction has been applied.

Practical considerations. A practical suggestion for interferometric profile and roundness metrology is to retain R_a , R_q , R_z , R_{sk} , $RONt$ and $RONq$ as the primary engineering layer compliant with ISO 21920-2, and to consider

596 supplementing them with BV density (in angular units, nm deg^{-1} , Sec. 3.6) as
 597 a gradient-sensitive functional-analysis complement. The $W^{1,2}$ gradient semi-
 598 norm is strongly correlated with BV density in the present datasets (within
 599 the FN 111: Pearson $r = 0.95$, $n = 5$ sweeps; within the FN 101: $r = 0.93$,
 600 $n = 5$). The pooled $r = 0.94$ (95% CI $[0.72, 0.99]$, $n = 10$) is reported for de-
 601 scriptive purposes only — the sweeps are not independent observations and this
 602 pooled correlation should be interpreted with caution. This strong correlation
 603 indicates that, as *repeatability metrics across sweeps*, BV density and the $W^{1,2}$
 604 seminorm carry largely redundant information, and routine reporting of both
 605 is not warranted. They are strongly correlated but not mathematically equiv-
 606 alent: BV integrates the first power of the local slope, $TV = \int |p'| ds$, whereas
 607 $W^{1,2}$ integrates its square, $\int |p'|^2 ds$, so the two coincide only for profiles whose
 608 slope magnitude is nearly constant and diverge whenever the slope distribution
 609 is uneven. This is exactly what the defect responses show: for the 100 nm
 610 synthetic spike the $W^{1,2}$ seminorm responded with +622% versus +18% for
 611 BV density, because the quadratic weighting amplifies a single steep event far
 612 more strongly than the linear total-variation measure. A practical reading of
 613 the descriptor space in these data is therefore: (i) R_a/R_q (equivalently L^1/L^2)
 614 carry amplitude information; (ii) R_{sk} carries an independent asymmetry axis;
 615 (iii) BV density carries the localised-gradient axis and is recommended as the
 616 primary gradient descriptor; (iv) the $W^{1,2}$ seminorm adds value mainly as a
 617 high-sensitivity alarm for isolated steep features rather than as an independent
 618 information axis; and (v) the L^4/L^1 peakiness ratio provides a supplementary
 619 distribution-shape axis with a direct physical interpretation as the amplification
 620 of dominant profile excursions relative to R_a . These additional descriptors are
 621 computed from the same residual profile without additional hardware or changes
 622 to the EFM pipeline. The cross-instrument validation suggests that gradient-
 623 based descriptors are applicable to any 1D profile, whether from interferometry
 624 or tactile roundness measurement.

625 **Practical value and the known-feature caveat.** It must be acknowl-
 626 edged that the Flick flat is a certified, deliberately manufactured feature whose

627 presence and position were known in advance: Fig. 4 is an existence proof under
 628 ground-truth conditions, not a blind detection study. The practical argument
 629 for a gradient-sensitive layer is therefore not that it revealed an unknown defect
 630 here. It is that it converts what an operator might notice in a visual polar-
 631 plot inspection into a quantitative, thresholdable, and automatable indicator
 632 computed from data already acquired for ISO reporting — with no operator
 633 in the loop, a defined angular position estimate, and a repeatability statement.
 634 Whether this indicator reveals defects that routine quality-control reporting
 635 would otherwise miss has not been demonstrated; a blind-detection protocol on
 636 artefacts with independently characterised defects is outlined in Sec. 5.6 as the
 637 decisive next test of practical value.

638 **Planned versus exploratory comparisons.** The synthetic-defect re-
 639 sponses (Supplementary Table S2, Fig. 5) and the Flick flat detection (Fig. 4)
 640 were planned comparisons motivated by the study’s objective of assessing gradient-
 641 based descriptor sensitivity. The correlation matrices, power spectral density
 642 comparison, Cohen’s d effect sizes, L^p norm convergence analysis, and the cross-
 643 descriptor stability comparison (Supplementary Figs. S13–S17) are exploratory
 644 and are intended to characterise the descriptor space rather than to test specific
 645 hypotheses. This distinction should be borne in mind when interpreting the
 646 reported associations. The present study is a proof-of-concept demonstration;
 647 generalisability to other artefact types, materials, and measurement instruments
 648 remains to be established.

649 5.4. *Limitations*

650 Several limitations should be acknowledged. First, the uncertainty analysis
 651 is restricted to within-run Type A repeatability from five sweeps in a single con-
 652 tinuous rotation, complemented by the indicative type-B budget of Sec. 3.8. The
 653 small $n = 5$ limits statistical power (relative differences below $\sim 30\%$ cannot be
 654 resolved for the FN 111), and between-run reproducibility (independent setups,
 655 thermal cycles, physical re-centring of the artefact) has not been assessed; the
 656 reported expanded uncertainties should be interpreted as lower bounds, and the

form-removal systematic (bounded at ≤ 41 nm) is carried outside the quadrature combination.

Second, the validation is restricted to two artefacts from a single manufacturer (JENOPTIK), both of ~ 30 mm diameter. The two artefacts do differ substantially in material (Al_2O_3 ceramic versus steel) and in residual amplitude (two orders of magnitude in R_a), which provides some breadth, but generalisability to other diameters, materials such as optical glass or tungsten carbide, and other surface finishes has not been tested. Third, the synthetic-defect analysis uses idealised mathematical defects (isolated spike, uniform scratch, pure sinusoid) injected into the FN 111 ceramic profile only; the response to real manufacturing defects with complex morphologies may differ, and the scaling analysis of Sec. 5.3 shows quantitatively that full-profile detection sensitivity degrades in proportion to background gradient content (the FN 101 ratio is ~ 500 times less favourable). Fourth, BV density and the $W^{1,2}$ seminorm carry intrinsic sampling-interval dependence; the angular-unit convention of Sec. 3.6 standardises reporting, but no inter-laboratory evidence yet exists that the convention transfers across instruments. Fifth, the Flick flat result (Supplementary Fig. S18) is demonstrated on a single artefact with a known, certified macroscopic feature; blind-detection performance and performance for subtler, micrometre-scale defects have not been characterised.

5.5. Metrological Implications

The results carry several implications for profile and roundness metrology practice. The finding that BV density responds to localised gradients that amplitude averaging may not capture suggests — subject to further validation — a potential role for gradient-based descriptors in applications where isolated defects are of concern; possible examples include quality control of bearing surfaces, optical components, or precision seals. Because these descriptors are computed from the same residual profile used for ISO 21920-2 parameters, their adoption does not require changes to existing measurement acquisition protocols.

687 The ISO 21920-2 framework (with ISO 4287 as its predecessor) remains the
688 appropriate baseline for surface texture specification and conformity assessment.
689 The gradient-based descriptors discussed here are proposed as optional supple-
690 mentary indicators, not as replacements for standardised parameters. Their
691 metrological utility would be strengthened by inter-laboratory comparisons, by
692 evaluation on a wider range of artefact types, and by adoption of the angular-
693 unit reporting convention of Sec. 3.6. Until such steps are taken, BV density and
694 related descriptors should be regarded as research-grade tools whose diagnostic
695 value has been observed under controlled laboratory conditions on two specific
696 artefacts.

697 *5.6. Future Work*

698 The present study suggests several directions for further investigation. (i) A
699 between-run reproducibility study with independent setups, distinct thermal
700 cycles, and physical re-centring of the artefact would establish whether the
701 gradient-based descriptors retain their relative stability under realistic labora-
702 tory conditions, and would supply the input distributions needed to upgrade the
703 indicative type-B budget of Sec. 3.8 to a complete GUM-compliant evaluation.
704 (ii) Extending the validation to artefacts with different diameters, materials
705 (e.g., optical glass, tungsten carbide), and surface finishes would test the gener-
706 alisability of the findings beyond the two ~ 30 mm JENOPTIK artefacts studied
707 here. (iii) A systematic characterisation of BV density response to real manufac-
708 turing defects (scratches, pits, inclusions) of calibrated dimensions, culminating
709 in a blind-detection protocol in which the analyst does not know defect presence
710 or position, would bridge the gap between the present synthetic-defect analy-
711 sis and industrial inspection and would provide the decisive test of practical
712 value. (iv) An inter-laboratory comparison using a common artefact set and a
713 standardised processing pipeline would provide the reproducibility data needed
714 to assess whether BV density could eventually be considered for standardisa-
715 tion. (v) Empirical verification of the angular-unit convention of Sec. 3.6 across
716 instruments with different angular resolutions would support the potential in-

717 clusion of gradient descriptors in reference software for roundness evaluation.
 718 (vi) A systematic study of the speed–accuracy trade-off (rotation speed, frame
 719 rate, fringe contrast) would establish the shortest acquisition compatible with
 720 the reported uncertainty, a prerequisite for industrial adoption.

721 Data Availability

722 The analysis code, processed results, and all figures are available at <https://github.com/dawidkucharski/banach-profile-metrology> (release v1.1.0).
 723
 724 The raw interferometric PNG archives (~32 GB) are available from the corre-
 725 sponding author upon reasonable request.

726 References

- 727 [1] L. Zhu, C. Ni, Z. Yang, C. Liu, Investigations of micro-textured surface
 728 generation mechanism and tribological properties in ultrasonic vibration-
 729 assisted milling of Ti–6Al–4V, Precision Engineering 57 (November 2018)
 730 (2019) 229–243. doi:10.1016/j.precisioneng.2019.04.010.
 731 URL <https://doi.org/10.1016/j.precisioneng.2019.04.010><https://linkinghub.elsevier.com/retrieve/pii/S0141635918307384>
 732
- 733 [2] R. Leach, Optical Measurement of Surface Topography, Springer Berlin
 734 Heidelberg, Berlin, Heidelberg, 2011. doi:10.1007/978-3-642-12012-1.
 735 URL <http://link.springer.com/10.1007/978-3-642-12012-1>
- 736 [3] D. J. Whitehouse, Handbook of Surface and Nanometrology, 2nd Edition,
 737 CRC Press, 2011.
- 738 [4] A. Michelson, Light-Waves and their Application to Metrology, Nature
 739 49 (1255) (1893) 56–60. doi:10.1038/049056a0.
 740 URL <https://doi.org/10.1038/049056a0><http://10.0.4.14/049056a0http://www.nature.com/articles/049056a0>
 741

- [5] F. A. Jenkins, H. E. White, Fundamentals of Optics, International student edition, McGraw-Hill, 1976.
URL <https://books.google.pl/books?id=dCdRAAAAMAAJhttps://books.google.pl/books?id=4SPCJ0wtFSUC>
- [6] M. Born, E. Wolf, E. Hecht, Principles of Optics: Electromagnetic Theory of Propagation, Interference and Diffraction of Light, Physics Today 53 (10) (2000) 77–78. doi:10.1063/1.1325200.
URL <http://physicstoday.scitation.org/doi/10.1063/1.1325200>
- [7] R. Brodmann, Optisches Rauheitsmessgeraet fuer die Fertigung, Feinwerktechnik und Messtechnik 91 (2) (1983) 63–66.
- [8] R. Brodmann, Roughness form and waviness measurement by means of light-scattering, Precision Engineering 8 (4) (1986) 221–226. doi:10.1016/0141-6359(86)90063-2.
- [9] Y. Y. Fan, V. M. Huynh, An investigation of light scattering from flat periodic rough surfaces for surface roughness estimation, Precision Engineering 16 (3) (1994) 205–211. doi:10.1016/0141-6359(94)90126-0.
- [10] A. Hertzsch, K. Kröger, H. Truckenbrodt, Microtopographic analysis of turned surfaces by model-based scatterometry, Precision Engineering 26 (3) (2002) 306–313. doi:10.1016/S0141-6359(02)00116-2.
- [11] D. Kucharski, H. Zdunek, A low-cost, simple optical setup for a fast scatterometry surface roughness measurements with nanometric precision, Bulletin of the Polish Academy of Sciences: Technical Sciences 68 (48) (2020) 485–490. doi:10.24425/bpasts.2020.133364.
URL http://journals.pan.pl/Content/116520/PDF/10_485-490_01475_Bpast.No.68-3_30.06.20_K3A_TeX.pdf
- [12] R. Benoît, Application des phénomènes d’interférence à des déterminations métrologiques, Journal de Physique Théorique et Appliquée 7 (1) (1898) 57–68. doi:10.1051/jphystap:01898007005700.

- 770 URL [http://www.edpsciences.org/10.1051/jphystap:](http://www.edpsciences.org/10.1051/jphystap:01898007005700%0Apapers3://publication/doi/10.1051/jphystap:01898007005700)
771 01898007005700%0Apapers3://publication/doi/10.1051/jphystap:
772 01898007005700[http://www.edpsciences.org/10.1051/jphystap:](http://www.edpsciences.org/10.1051/jphystap:01898007005700)
773 01898007005700
- 774 [13] G. Xue-fei, X. Sheng-nian, Y. Ye-fei, B. Shan, B. Xing, C. Zhou-jian,
775 C. Ge-rui, D. Peng, G. Tian-shu, G. Wei, H. Shuang-lin, L. Yu-long,
776 L. Ying, L. Zi-ren, S. Ming-xue, S. Bao-san, T. Wen-lin, Y. Pin, X. Peng,
777 Z. Yun-long, Z. Hai-peng, L. Yun-kau, Laser Interferometric Gravita-
778 tional Wave Detection in Space and Structure Formation in the Early
779 Universe, Chinese Astronomy and Astrophysics 39 (4) (2015) 411–446.
780 doi:10.1016/j.chinastron.2015.10.001.
781 URL [http://www.sciencedirect.com/science/article/pii/](http://www.sciencedirect.com/science/article/pii/S0275106215000843)
782 S0275106215000843[https://linkinghub.elsevier.com/retrieve/](https://linkinghub.elsevier.com/retrieve/pii/S0275106215000843)
783 pii/S0275106215000843
- 784 [14] P. de Groot, Phase Shifting Interferometry, in: R. K. Leach (Ed.), Optical
785 Measurement of Surface Topography, Springer Berlin Heidelberg, Berlin,
786 Heidelberg, 2011, pp. 167–186. doi:10.1007/978-3-642-12012-1_8.
787 URL http://link.springer.com/10.1007/978-3-642-12012-1_8
- 788 [15] P. de Groot, Principles of interference microscopy for the measurement
789 of surface topography, Advances in Optics and Photonics 7 (2015) 1–65.
790 doi:10.1364/AOP.7.000001.
- 791 [16] S. Chatterjee, Phase-Shifting Laser Interferometry for Measurement of Sur-
792 face Form Error, in: P.K.Gupta, R.Khare (Eds.), Springer Proceedings in
793 Physics, Vol. 160 of Springer Proceedings in Physics, 2015, pp. 209–222.
794 doi:10.1007/978-81-322-2000-8_10.
795 URL http://link.springer.com/10.1007/978-81-322-2000-8_10
- 796 [17] F. Polack, M. Thomasset, S. Brochet, D. Denetiere, Surface shape deter-
797 mination with a stitching Michelson interferometer and accuracy evalua-
798 tion, Review of Scientific Instruments 90 (2). doi:10.1063/1.5061930.

- 799 [18] M. Thomasset, M. Idir, F. Polack, M. Bray, J. J. Servant, A new phase-shift
800 microscope designed for high accuracy stitching interferometry, Nuclear
801 Instruments and Methods in Physics Research, Section A: Accelerators,
802 Spectrometers, Detectors and Associated Equipment 710 (2013) 7–12. doi:
803 10.1016/j.nima.2012.10.123.
- 804 [19] K.-C. Fan, N. Wang, Z.-W. Wang, H. Zhang, Development of a roundness
805 measuring system for microspheres, Measurement Science and Technology
806 25 (6) (2014) 064009. doi:10.1088/0957-0233/25/6/064009.
807 URL [http://stacks.iop.org/0957-0233/25/i=6/a=064009?key=](http://stacks.iop.org/0957-0233/25/i=6/a=064009?key=crossref.73300f1601df9e917d2e7830f09f45e7)
808 [crossref.73300f1601df9e917d2e7830f09f45e7](http://stacks.iop.org/0957-0233/25/i=6/a=064009?key=crossref.73300f1601df9e917d2e7830f09f45e7)
- 809 [20] M. Kühnel, V. Ullmann, U. Gerhardt, E. Manske, Automated setup for
810 non-tactile high-precision measurements of roundness and cylindricity
811 using two laser interferometers, Measurement Science and Technology
812 23 (7) (2012) 74016. doi:10.1088/0957-0233/23/7/074016.
813 URL [http://stacks.iop.org/0957-0233/23/i=7/a=074016?key=](http://stacks.iop.org/0957-0233/23/i=7/a=074016?key=crossref.8429e4ab7d2d23702c0a05dba94b50d5http://stacks.iop.org/0957-0233/23/i=7/a=074016)
814 [crossref.8429e4ab7d2d23702c0a05dba94b50d5http://stacks.iop.](http://stacks.iop.org/0957-0233/23/i=7/a=074016?key=crossref.8429e4ab7d2d23702c0a05dba94b50d5http://stacks.iop.org/0957-0233/23/i=7/a=074016)
815 [org/0957-0233/23/i=7/a=074016](http://stacks.iop.org/0957-0233/23/i=7/a=074016)
- 816 [21] M. Takeda, H. Ina, S. Kobayashi, Fourier-transform method of fringe-
817 pattern analysis for computer-based topography and interferometry, Jour-
818 nal of the Optical Society of America 72 (1) (1982) 156–160. doi:
819 10.1364/JOSA.72.000156.
- 820 [22] Q. Kemao, Two-dimensional windowed Fourier transform for fringe pattern
821 analysis: Principles, applications and implementations, Optics and Lasers
822 in Engineering 45 (2) (2007) 304–317. doi:10.1016/j.optlaseng.2005.
823 10.012.
- 824 [23] S. Feng, Q. Chen, G. Gu, T. Tao, L. Zhang, Y. Hu, W. Yin, C. Zuo,
825 Fringe pattern analysis using deep learning, Advanced Photonics 1 (2)
826 (2019) 025001. doi:10.1117/1.AP.1.2.025001.

- [24] I. Galaktionov, V. Toporovsky, A Multi-Stage Algorithm of Fringe Map Reconstruction for Fiber-End Surface Analysis and Non-Phase-Shifting Interferometry, *Applied System Innovation* 9 (2) (2026) 31. doi:10.3390/asi9020031.
- [25] D. Kucharski, M. Wieczorowski, Radial image processing for phase extraction in rough-surface interferometry, *Measurement* (2025) 117102doi:10.1016/j.measurement.2025.117102.
- [26] P. Dumas, J. Fleig, G. Forbes, P. E. Murphy, Extending the range of interferometry through subaperture stitching, Vol. 10314, 2003, p. 134 – 137, cited by: 9. doi:10.1117/12.2284035.
URL <https://www.scopus.com/inward/record.uri?eid=2-s2.0-85016807739&doi=10.1117%2f12.2284035&partnerID=40&md5=8aa04fbb9453cddff6da95df376ffdb6>
- [27] S. Chen, S. Li, Y. Dai, Z. Zheng, Iterative algorithm for subaperture stitching test with spherical interferometers, *Journal of the Optical Society of America A: Optics and Image Science, and Vision* 23 (5) (2006) 1219 – 1226, cited by: 48. doi:10.1364/JOSAA.23.001219.
URL <https://www.scopus.com/inward/record.uri?eid=2-s2.0-33745276550&doi=10.1364%2fJOSAA.23.001219&partnerID=40&md5=4d72b8095c89e81b1ba058a414b7b0c2>
- [28] S. Chen, S. Li, Y. Dai, Z. Zheng, Testing of large optical surfaces with subaperture stitching, *Applied Optics* 46 (17) (2007) 3504 – 3509, cited by: 44. doi:10.1364/AO.46.003504.
URL <https://www.scopus.com/inward/record.uri?eid=2-s2.0-34547536867&doi=10.1364%2fAO.46.003504&partnerID=40&md5=9e100905a834eebb31d582fc8884de99>
- [29] A. S. Iacchetta, J. R. Fienup, D. T. Leisawitz, Rotation and translation registration of bandlimited interferometric images using a chirp z-transform, Vol. 9907, 2016, cited by: 3. doi:10.1117/12.2232178.

- 856 URL <https://www.scopus.com/inward/record.uri?eid=2-s2.>
857 0-85006847872&doi=10.1117%2f12.2232178&partnerID=40&md5=
858 25831551d0ae177fef89c421998e31d4
- 859 [30] J. Burge, S. Ament, J. Beverage, C. Muccianti, C. Zhao, Applications of
860 computer-generated holograms for measuring x-ray and euv optics, Vol.
861 13620, 2025, cited by: 0. doi:10.1117/12.3067011.
862 URL <https://www.scopus.com/inward/record.uri?eid=2-s2.>
863 0-105024341829&doi=10.1117%2f12.3067011&partnerID=40&md5=
864 92bd5b2fbfd00333a45113efd7a5be87
- 865 [31] H. Quan, H. Xi, W. Fan, Evaluating surface repeatability for interfero-
866 metric measurement: A comparative study, Vol. 9684, 2016, cited by: 1.
867 doi:10.1117/12.2243301.
868 URL <https://www.scopus.com/inward/record.uri?eid=2-s2.>
869 0-85002262299&doi=10.1117%2f12.2243301&partnerID=40&md5=
870 98a061c8f8cdb6f89be75a953cd88b3f
- 871 [32] L. Zhang, B. Zhang, S. Han, S. Tang, C. Zhuang, Repeatability
872 study on different interference systems, Vol. 10847, 2018, cited by: 0.
873 doi:10.1117/12.2505406.
874 URL <https://www.scopus.com/inward/record.uri?eid=2-s2.>
875 0-85059070480&doi=10.1117%2f12.2505406&partnerID=40&md5=
876 d2fe726c4d532c3f34340ac03143d343
- 877 [33] A. Pappas, L. Newton, A. Thompson, H. Hooshmand, R. Leach, Un-
878 certainty propagation of field areal surface texture parameters using the
879 metrological characteristics approach, 2023, p. 419 – 422, cited by: 3.
880 URL <https://www.scopus.com/inward/record.>
881 uri?eid=2-s2.0-85165481981&partnerID=40&md5=
882 3c2a4bd283d7a91350fa5adce5f6941b
- 883 [34] J. Gertheiss, D. Rügamer, B. X. W. Liew, S. Greven, Functional data
884 analysis: An introduction and recent developments, Biometrical Journal

885 66 (7), cited by: 32; All Open Access, Hybrid Gold Open Access.
886 doi:10.1002/bimj.202300363.
887 URL [https://www.scopus.com/inward/record.uri?eid=2-s2.](https://www.scopus.com/inward/record.uri?eid=2-s2.0-85205152017&doi=10.1002%2fbimj.202300363&partnerID=40&md5=8d9e42e9cd2f23350c0bf0b477169043)
888 0-85205152017&doi=10.1002%2fbimj.202300363&partnerID=40&md5=
889 8d9e42e9cd2f23350c0bf0b477169043

890 [35] C. Happ, F. Scheipl, A.-A. Gabriel, S. Greven, A general framework for
891 multivariate functional principal component analysis of amplitude and
892 phase variation, Stat 8 (1), cited by: 12; All Open Access, Green Open
893 Access. doi:10.1002/sta4.220.
894 URL [https://www.scopus.com/inward/record.uri?eid=2-s2.](https://www.scopus.com/inward/record.uri?eid=2-s2.0-85064480543&doi=10.1002%2fst4.220&partnerID=40&md5=e8d4b76faff42103fe5637961836125d)
895 0-85064480543&doi=10.1002%2fst4.220&partnerID=40&md5=
896 e8d4b76faff42103fe5637961836125d

897 [36] G. W. Wolf, Surfaces - topography and topology, Surface Topography:
898 Metrology and Properties 8 (1), cited by: 14; All Open Access, Hybrid
899 Gold Open Access. doi:10.1088/2051-672X/ab70e8.
900 URL [https://www.scopus.com/inward/record.uri?eid=2-s2.](https://www.scopus.com/inward/record.uri?eid=2-s2.0-85084666906&doi=10.1088%2f2051-672X%2fab70e8&partnerID=40&md5=006779646e49c4fc727d508bc63eabfa)
901 0-85084666906&doi=10.1088%2f2051-672X%2fab70e8&partnerID=
902 40&md5=006779646e49c4fc727d508bc63eabfa

903 [37] A. Sanner, W. G. Nöhring, L. A. Thimons, T. D. Jacobs, L. Pastewka,
904 Scale-dependent roughness parameters for topography analysis, Applied
905 Surface Science Advances 7, cited by: 28; All Open Access, Gold Open
906 Access, Green Open Access. doi:10.1016/j.apsadv.2021.100190.
907 URL [https://www.scopus.com/inward/record.uri?eid=2-s2.](https://www.scopus.com/inward/record.uri?eid=2-s2.0-85119021594&doi=10.1016%2fj.apsadv.2021.100190&partnerID=40&md5=7ad03ad2f796c56f6bec60b5febac199)
908 0-85119021594&doi=10.1016%2fj.apsadv.2021.100190&partnerID=
909 40&md5=7ad03ad2f796c56f6bec60b5febac199

910 [38] J. Zhao, Z. Wang, L. Zhengminqing, R. Zhu, Fractal simulation of
911 surface topography and prediction of its lubrication characteristics,
912 Surface Topography: Metrology and Properties 9 (4), cited by: 17.
913 doi:10.1088/2051-672X/ac3f26.

914 URL <https://www.scopus.com/inward/record.uri?eid=2-s2.>
915 0-85122423109&doi=10.1088%2f2051-672X%2fac3f26&partnerID=
916 40&md5=7eb253c5518d5356ead64c22878553ba

917 [39] F. Meijer, D. Kucharski, E. Stachowska, Determination of the phase in
918 the center of a circular two-beam interference pattern to determine the
919 displacement of a rough surface, *Optical Engineering* 57 (10) (2018) 1.
920 doi:10.1117/1.OE.57.10.104101.

921 URL [https://www.spiedigitallibrary.org/journals/
922 optical-engineering/volume-57/issue-10/104101/
923 Determination-of-the-phase-in-the-center-of-a-circular/10.
924 1117/1.OE.57.10.104101.full](https://www.spiedigitallibrary.org/journals/optical-engineering/volume-57/issue-10/104101/Determination-of-the-phase-in-the-center-of-a-circular/10.1117/1.OE.57.10.104101.full)

925 [40] JENOPTIK Industrial Metrology, Unequivocal measurement of form and
926 positional tolerances using innovative measuring systems, JENOPTIK In-
927 dustrial Metrology Germany GmbH, 2018.

928 [41] Thorlabs Inc., CPS532 Collimated Laser Diode Module, 532 nm, 4.5 mW,
929 Round Beam, Ø11 mm Housing (2019).

930 URL [https://www.thorlabs.com/thorproduct.cfm?partnumber=
931 CPS532](https://www.thorlabs.com/thorproduct.cfm?partnumber=CPS532)

932 [42] FLIR Systems, Software Development Kit (2019).

933 URL [https://www.flir.com/browse/camera-cores-amp-components/
934 software-development-kits/](https://www.flir.com/browse/camera-cores-amp-components/software-development-kits/)

935 [43] Newport Corporation, ESP301 Controller GUI Manual, Newport Corpora-
936 tion, 2017.

937 URL [https://www.newport.com/medias/sys_master/images/images/
938 h7e/hfe/9044140982302/ESP301-Controller-GUI-Manual.pdf](https://www.newport.com/medias/sys_master/images/images/h7e/hfe/9044140982302/ESP301-Controller-GUI-Manual.pdf)

939 [44] R. Ihaka, R. Gentleman, R: A Language for Data Analysis and Graphics,
940 *Journal of Computational and Graphical Statistics* 5 (3) (1996) 299–314.
941 doi:10.1080/10618600.1996.10474713.

- 942 URL [http://www.tandfonline.com/doi/abs/10.1080/10618600.1996.](http://www.tandfonline.com/doi/abs/10.1080/10618600.1996.10474713)
943 10474713
- 944 [45] ISO 12181-1:2011, Geometrical product specifications (GPS) — Roundness
945 — Part 1: Vocabulary and parameters of roundness, Tech. rep., Interna-
946 tional Organization for Standardization (2011).
- 947 [46] ISO 21920-2:2021, Geometrical product specifications (GPS) — Surface
948 texture: Profile — Part 2: Terms, definitions and surface texture param-
949 eters, Tech. rep., International Organization for Standardization, Geneva
950 (2021).
- 951 [47] ISO 4287:1997, Geometrical Product Specifications (GPS) — Surface tex-
952 ture: Profile method — Terms, definitions and surface texture parameters,
953 Tech. rep., International Organization for Standardization (1997).
- 954 [48] ISO 25178-2:2012, Geometrical Product Specifications (GPS) — Surface
955 texture: Areal — Part 2: Terms, definitions and surface texture paramet-
956 ers, Tech. rep., International Organization for Standardization (2012).
- 957 [49] S. Banach, Théorie des opérations linéaires, Monografie Matematyczne,
958 Warsaw, 1932.
- 959 [50] JCGM 100:2008, Evaluation of measurement data — Guide to the expres-
960 sion of uncertainty in measurement (GUM), Tech. rep., Joint Committee
961 for Guides in Metrology (2008).
- 962 [51] Newport Corporation, Precision Motion Control Catalog, Newport Corpo-
963 ration, 2018.