Instructions for using the ECG Noise Extraction (ECG NExT) Software Code

Contact: Christopher Scully, Ph.D.

CDRH Office of Science and Engineering Laboratories

Division of Biomedical Physics

(Christopher.Scully@fda.hhs.gov)

<Catalog URL>

Updated: October 17, 2022

Disclaimer

The mention of commercial products, their sources, or their use in connection with material reported herein is not to be construed as either an actual or implied endorsement of such products by the Department of Health and Human Services.

1. General Information

This software code outputs an estimate of a noise signal collected as part of an electrocardiograph (ECG) recording. The objective is to obtain representative samples of noise and motion artifacts from ECG devices under test (DUT) that can then be used for testing the robustness of ECG analysis algorithms. The algorithm for noise estimation is described in [1]. The core operation of this software is based on removing the instances of the QRS complex from the ECG so that an estimate of the noise component of the recording remains. Required inputs are the ECG signal recorded with noise/motion artifacts and locations (indices) of the R-peaks.

2. System Requirements

The software code was written in a MATLAB® 2021b environment and consists of a core function and a shell script. The core function and its supporting functions are encapsulated in a class called "NoiseExtractClass.m" in the package. The shell script in the package is called "NoiseExtractShell.m" and is provided for demonstration purposes. While the core function can be used as a standalone code, the shell script can help with a quick start of a project. Software has been tested in MATLAB® 2021b environment without any additional toolboxes.

3. Data preparation:

The following steps are taken to prepare data as an input for processing and noise estimation by this software code:

1. Obtain an ECG signal using the device under test – *ECG_device* – at a location where noise/motion artifacts are captured in the ECG signal.

- 2. Obtain a simultaneously recorded clean ECG signal from a standard site of collection, e.g., the chest area for reference *ECG_std*.
- 3. Apply a bandpass filter to the *ECG_device* signal to obtain *ECG_device_BP*. This is to clean the data from baseline wander and high-frequency noises as would normally be performed as part of data acquisition/processing for the device under test typically in the range of 0.5 to 40 Hz. *ECG_std* can also be passed through a similar filtering scheme if needed for the purpose of better R peak detection.
- 4. Obtain the indices of the R peaks from, e. g., *ECG_std_BP* (or *ECG_device* if it provides sufficient fidelity for reliable peak detection).

A sample dataset as described in Section 6 is included in the software for demonstration purpose. The dataset includes variables *ECG_device*, *ECG_device_BP*, *ECG_std*, *fs* (sampling frequency), and *rPeaks* (R peak locations in samples).

4. Input and Output of ECG_NExT

Inputs:

- 1. *ECG_device:* 1-dimensional array of ECG samples recorded with noise at least 30 seconds length (the test data provides 60 seconds of data)
- 2. *ECG_device_BP*: 1-dimensional array Bandpass-filtered version of *ECG_device*,.
- 3. *rPeaks*: 1-dimensional array of R peak indices in samples (not in time units).
- 4. *fs*: sampling rate set to 1000 Hz. This sampling rate is hard coded in the software and will terminate with an error message if a different sampling rate is provided.
- 5. crossFadePar: This optional input is not shown in the sample shell code and can be used to specify cross fade parameters. This is a two-element array; the first element, the Blank parameter, specifies the length of data (in seconds) to be removed from both sides of each R peak. The second element, the Blend parameter, specifies the length of data (in seconds) to be copied from neighboring segments of the removed data, crossfaded and inserted in the gap created by the process of removal of the area around the R peaks. The default values for Blank and Blend parameters are 40 and 60 milliseconds, respectively. This optional input allows the user to adjust for the differences in R-Peak widths within a data set. For more details please refer to [1].

Outputs:

- 1. *noiseEstimated*: 1-dimensional array of estimated noise samples.
- 2. *tVec*: 1-dimensional array corresponding time vector for demonstration purposes, e. g., plotting original signal and the estimated noise versus time on a shared horizontal axis. The removed segments from the original signals are accounted for in this time vector.

5. Algorithm Development Test Conditions:

The test signals used during the software code development had the following characteristics:

1. Signals equivalent to *ECG_device* and *ECG_std* were 60-second-long signals sampled at 1000 Hz.

The signals representing *ECG_device_BP* and *ECG_std_BP* were obtained by applying a bandpass filter on the signals *ECG_device* and *ECG_std*, respectively. The *ECG_std_BP* was obtained for R peak detection purpose. The bandpass filter was a 5th order Butterworth design with 0.5 and 40 Hz corner frequencies corresponding to low and high cutoff frequencies, respectively.

6. Software Package Content:

The software package includes the class containing the core and supporting functions for ECG noise estimation (NoiseExtractClass.m). In addition, a sample wrapper/shell code is provided to run the function with a test set, TestData.mat. This data set consists of a clean ECG, i.e., *ECG_std*, a noisy ECG representing an ECG from DUT, i.e., *ECG_device*, the band passed version of *ECG_device*, *ECG_device_BP*, the corresponding R peak locations, *rPeaks* and the value for sampling rate, *fs*. In the provided test set, ECG_std is a simulated ECG generated with the ECGSYN waveform generator available from the PhysioNet ECG repositories [2, 3]. In obtaining ECG_std using this model the following model parameters were used: ECG sampling frequency (sfecg = 1000 Hz), internal sampling frequency (sfint = 1000 Hz) and approximate number of heartbeats (N = 60). Remaining parameters used the default values. The first 60000 samples (60 seconds) of the simulated ECG were used for the test data *ECG_std*. *ECG_device* in this case is obtained by adding random Gaussian noise (zero mean, 500 mV std) to the *ECG_std*. The signals are 60-second long and sampled at 1000 Hz.

References

- [1]. Galeotti, L. and C.G. Scully, A method to extract realistic artifacts from electrocardiogram recordings for robust algorithm testing. Journal of Electrocardiology, 2018. **51**(6, Supplement): p. S56-S60.
- [2]. Goldberger, A., Amaral, L., Glass, L., Hausdorff, J., Ivanov, P. C., Mark, R., ... & Stanley, H. E., *PhysioBank, PhysioToolkit, and PhysioNet: Components of a new research resource for complex physiologic signals.*, in *Circulation [Online]*. 2000. p. e215 e220.
- [3]. McSharry, P.E., G.D. Clifford, L. Tarassenko and L.A. Smith, *A dynamical model for generating synthetic electrocardiogram signals*. IEEE Transactions on Biomedical Engineering, 2003. p. 289-294.

Appendix I

Software Code

The shell code:

This is the primary module that runs the core software and initiates various function calls.

```
close all
clear
clc
% User data can be added here and should replace the test data below:
testDat = load('TestData.mat');
ECG device = testDat.ECG device;
ECG device BP = testDat.ECG device BP;
rPeaks = testDat.rPeaks;
fs = testDat.fs;
NEWC = NoiseExtractClass;
[noiseEstimated, tVec] = NEWC.ecgNoiseExtractor(ECG device, ECG device BP,
rPeaks, fs);
% Reduce noise record to same length as calculated noise record
NEWC.TestPlot(noiseEstimated, tVec, fs, ECG device, testDat.ECG std)
```

The main code:

% Inputs:

```
classdef NoiseExtractClass
This classdef construct includes the functions and sub modules necessary
to implement the ECG NExT algorithm that obtains representative samples
of noise and motion artifacts from ECG devices under test.
Authors:
Ahmad Suliman
Christopher Scully
Loriano Galeotti
The ECG NExT algorithm is described and published in:
Galeotti, L. and C.G. Scully, "A method to extract realistic artifacts from
electrocardiogram recordings from robust algorithm testing," Journal of
Electrocardiology, 2018. 51(6, Supplement): p. S56-S60.
If you have found this software useful, please consider citing our
publication.
Disclaimer:
This software and documentation (the "Software") were developed at the Food
and Drug Administration (FDA) by employees of the Federal Government in the
course of their official duties. Pursuant to Title 17, Section 105 of the
United States Code, this work is not subject to copyright protection and is
in the public domain. Permission is hereby granted, free of charge, to any
person obtaining a copy of the Software, to deal in the Software without
restriction, including without limitation the rights to use, copy, modify,
merge, publish, distribute, sublicense, or sell copies of the Software or
derivatives, and to permit persons to whom the Software is furnished to do
so. FDA assumes no responsibility whatsoever for use by other parties of
the Software, its source code, documentation or compiled executables, and
makes no quarantees, expressed or implied, about its quality, reliability,
or any other characteristic. Further, use of this code in no way implies
endorsement by the FDA or confers any advantage in regulatory decisions.
Although this software can be redistributed and/or modified freely, we ask
that any derivative works bear some notice that they are derived from it,
and any modified versions bear some notice that they have been modified.
    properties
        Property1
    end
    methods (Static)
        function [NOISE Calc, tVec] = ecgNoiseExtractor(ECG test, ...
                ECGtest BP, rPeaksRef, fs, varargin)
                  [NOISE Calc, tVec] = ecgNoiseExtractor(ECG test, ...
            % ECGtest BP, rPeaksRef, fs)
            % This function extracts noise from ECG recording as described
            % in Galeotti and Scully, JECG 2018.
```

% ECG test: The ECG signal obtained from the device under test.

```
% ECGtest BP: The bandpass-filtered version of ECG test.
% rPeaksRef: Accurately and reliably detected R peak indices of
             the ECG obtained.
% fs: Sampling frequency at which the ECG signals are sampled.
% Outputs:
% NOISE Calc: The calculated noise using this algorithm/code.
% tVec: Contains the time references corresponding to the noise
% instances remained after removing areas surrounding the R
% peaks. This is only needed for demo purposes when plotting the
% noise and ECG under consideration against time is desired.
% [...] = ecgNoiseExtractor(..., CROSSFADE PAR) accepts a
% two-element array of Blank and Blend parameters ([Blank Blend])
\mbox{\ensuremath{\$}} specifying the amount of data be removed from around the R
% peaks (Blank in seconds) and the amount of data to be copied
% from the neighboring sides of the removed part (Blend in
% seconds) to be cross faded and used to fill the gap
% created after removing the data from around the R peaks.
% The default values for the Blank and Blend parameters are set
% to 40 ms and 60 ms, respectively.
% Check for sampling frequency:
% Although the code should work with any sampling frequency, the
% sampling frequency used during test and development of this
% software was 1000 Hz. We, therefore, make certain that the
% users provide signals with 1000 Hz sampling frequency.
if fs ~= 1000
    error('Please change the sampling frequency to 1 kHz.')
end
sigLen = length(ECGtest BP);
% Track time:
tTemp = (0:sigLen - 1)/fs;
% Ensure consistent dimensionality of 1xn -- row vectors:
ECG test = ECG test(:)';
ECGtest BP = ECGtest BP(:)';
rPeaksRef = rPeaksRef(:)';
% Get residuals from test ECG after filtering (noise components
% outside bandwidth of bandpass filtered signal)
ECGtest residuals = ECG test - ECGtest BP;
% Obtain median beat from filtered ECG signal
[MedBeat] = ComputeMedianBeat long(ECGtest BP, rPeaksRef, fs);
% Generate synthetic signal using the median beat:
[SynECG] = GenSynECG blend(MedBeat, rPeaksRef, sigLen, fs);
% Obtain the noise component of the signal:
noiseRaw = ECGtest BP - SynECG;
% Remove areas around the R-peaks in the obtained noise as well
% as in the residuals obtained above and the time vector:
```

```
% Handle optional inputs:
   if isempty(varargin)
        NOISE Calc = RemRpkEffect(noiseRaw, rPeaksRef, fs);
        residualLessRpeaks = RemRpkEffect(ECGtest residuals, ...
          rPeaksRef, fs);
        tVec = RemRpkEffect(tTemp, rPeaksRef, fs);
        NOISE Calc = RemRpkEffect(noiseRaw, rPeaksRef, fs, ...
         varargin[2]);
        residualLessRpeaks = RemRpkEffect(ECGtest residuals, ...
          rPeaksRef, fs, varargin[2]);
        tVec = RemRpkEffect(tTemp, rPeaksRef, fs, varargin{1});
    end
    % Add noise to residual from bandpass filtered signal:
   NOISE Calc = NOISE Calc + residualLessRpeaks;
   % Remove NaN instances:
   tVec = tVec(~isnan(NOISE Calc));
   NOISE Calc = NOISE Calc(~isnan(NOISE Calc));
% End of ecgNoiseExtractor function
end
function TestPlot(noiseEstimated, tVec, fs, ECG raw, ECG ref)
   % This is a supplemental function for demo purposes in case the
   % user wants to plot the obtained noise with the ECG under
   % consideration.
   % Inputs:
   % noiseEstimated: The estimated noise by the ECG NExT algorithm.
   % tVec: the vector of time references corresponding to the
           estimated noise samples in noiseEstimated.
   % fs: sampling frequency
   % ECG raw: The ECG collected from the device under test.
   % ECG ref: The clean reference ECG (if collected). If not
           provided, the ECG raw will be used.
   % Handle optional input(s):
   if nargin < 5
       ECG ref = ECG raw;
   end
   figure
   tECG = (0:length(ECG raw)-1)/fs;
   ax(1) = subplot(3,1,1);
   plot(tECG, ECG ref)
   title('Reference ECG')
   ax(2) = subplot(3,1,2);
   plot(tECG, ECG raw)
   title('Unfiltered Device ECG')
   ax(3) = subplot(3,1,3);
   plot(tVec, noiseEstimated)
   title('Estimated Noise')
   linkaxes(ax, 'x')
   xlabel('Time (sec)')
   set(gcf, 'units', 'normal')
   set(qcf, 'position', [0.1753 0.0977 0.5064 0.7831])
```

```
% End of function
        end
   % End of method
   end
% End of classdef
end
function [sigOut, stitch] = RemRpkEffect(sigIn, Rpeaks, fs, varargin)
    % This function removes an area around R peaks and merges the segments
   % surrounding the removed area with crossfade of neighboring signal
   % segments around the removed area.
   % Inputs:
   % sigIn: This input can represent an ECG, the estimated noise or the time
   % vector in this context as a row of samples.
   % Rpeaks: R peak indices
   % fs: sampling frequency
   % Outputs:
   % sigOut: blended sigIn
   % stitch: midpoint of crossfade region (samples of sigOut).
   % Definition of blanklen and blendlen parameters:
   % The blanklen parameter is the length of signal (in milliseconds) being
   % removed from around of each R peak in final ECG noise; this is the part
   % really canceled before and after each Rpeak.
   % The blendlen parameter is the length of the neighboring signal segment
   % (in milliseconds) around the canceled part to be crossfaded when
   % stitching.
   % An "if" construct is included to make blanklen and blendlen parameters
    % optional and assign them default values.
    if isempty(varargin)
       blanklen = 40e-3;
       blendlen = 60e-3;
    else
       blanklen = varargin{1}(1);
       blendlen = varargin{1}{(2)};
    end
    % Convert to samples:
   blanklen = blanklen * fs;
   blendlen = blendlen * fs;
   % Create linear crossfade function
   W = linspace(1, 0, blendlen + 1);
    out locs = [Rpeaks(:)-blanklen Rpeaks(:)+blanklen];
   blend locs = [out locs(:,1)-blendlen out locs(:,2)+blendlen];
    % Adjust for Rpeaks too close to end of signal
    if blend locs(end,2) >= length(sigIn)
       blend locs(end,:) = [];
        out locs(end,:) = [];
        Rpeaks(end) = [];
```

```
end
```

```
% Initialize the stitch vector
    rpkLen = length(Rpeaks);
    stitch = zeros(1, rpkLen);
    % Start after 1st beat to remove errors related to location of 1st beat
    sigOut(1:blend locs(1,2)) = NaN;
    for n = 2:rpkLen
        % Copy from end of last blend region to start of new blend region
        segadd = sigIn(blend locs(n-1,2)+1:blend locs(n,1)-1);
        sigOut(length(sigOut)+1:length(sigOut)+length(segadd)) = segadd;
        % Blend cross fade region
        segadd = sigIn(blend locs(n,1):out locs(n,1)).*W + ...
            sigIn(out locs(n,2):blend locs(n,2)).*(1-W);
        % Keep track of stitch points:
        stitch(n) = length(sigOut) + floor(length(segadd)/2);
        sigOut(length(sigOut)+1:length(sigOut)+length(segadd)) = segadd;
    end
    stitch(1) = [];
    % Add data after last Rpeak
    segadd = sigIn(blend locs(n,2)+1:end);
    sigOut(length(sigOut)+1:length(sigOut)+length(segadd)) = segadd;
% End of function
end
function [MedBeat] = ComputeMedianBeat long(ECG, Rpeaks, fs)
    % Compute median beat
    % Inputs: ECG, R-peak locations
    % Output: Median beat, correction
    % Get short median, it is used to adjust X and Y offset
   MedShort = ComputeMedianBeat short(ECG, Rpeaks);
    shortl = floor(length(MedShort)/3);
    % Arbitrary area for the QRS complex to do time sync. For normal patients
    % should be <120ms, in this case let's do 50 ms each side.
    arbLen = 50e-3;
    arbLenSamp = arbLen*fs;
    qrsa = min(arbLenSamp, shortl-1);
    % Arbitrary max lag for crosscorr search in seconds.
    maxcorrlag = 20e-3;
   maxcorrlagSamp = maxcorrlag*fs;
    % Number of extra samples in each median beat (on the left double on the
    % right) should be at least as much as the cross-fade half duration
    extrabeat = 60;
    % Since sampling frequency at development time was 1000 samples/s:
    extrabeat = ceil(extrabeat*fs/1000);
```

```
% Determine max RR interval to set segment lengths
    RRint = diff(Rpeaks);
    maxRR = max(RRint);
    % Get left and right segments, chop out a bit from 1/3>
    segl = floor(maxRR*0.35) + extrabeat + 1;
    segr = 2*segl;
    % Initialize
    ECGmatxy = NaN(length(Rpeaks), segl + segr + 1);
    yoffset = NaN(length(Rpeaks),1);
    xoffset = yoffset ;
    m = 1;
    \ensuremath{\$} Catch cases where first Rpeak is too close to start of record
    while (m < length(Rpeaks)) && (Rpeaks(m) - segl - maxcorrlagSamp < 1)
       m = m + 1;
    % Main loop on each grs peak to fill ECG beat array
    while (m < length(Rpeaks)) && ...
            (Rpeaks(m) + segr + maxcorrlagSamp < length(ECG))
        % Calculate difference from short beat on Y
        shortdiff = MedShort - ECG(Rpeaks(m)-shortl:Rpeaks(m)+2*shortl);
        yoffset(m) = mean(shortdiff);
        [crosscorr, xcorrlag] = xcorr(ECG(Rpeaks(m)-qrsa:Rpeaks(m)+qrsa), ...
            MedShort(shortl-qrsa:shortl+qrsa), maxcorrlagSamp);
        [~, maxxcorri] = max(crosscorr);
        xoffset(m) = xcorrlag(maxxcorri);
        xoff = -xoffset(m);
        % Correct on X and Y.
        ECGmatxy(m,:) = (ECG(Rpeaks(m)-segl+xoff:Rpeaks(m)+2*segl+xoff))' + ...
            yoffset(m);
        % Increase loop counter
        m = m + 1;
    % End while
    end
    % Compute median from ECG beat array
    MedBeat = median(ECGmatxy, 1, 'omitnan');
% End of function
end
function [MedBeat, ECGmat] = ComputeMedianBeat short(ECG, Rpeaks)
    % Compute median beat
    % Inputs: ECG, R-peak locations
    % Output: Median beat
    % Get RR intervals
    RRint = diff(Rpeaks);
    minRR = min(RRint);
```

```
% Get left and right segments, chop out a bit from 1/3
    segl = floor(minRR*.29);
    segr = 2*segl;
    % Remove peaks too close to the edge
    if Rpeaks(1) < segl
       Rpeaks = Rpeaks(2:end);
    end
    if length(ECG) < Rpeaks(end) + segr
        Rpeaks = Rpeaks (1:end-1);
    end
    % Initialize and fill ECG beat array
    ECGmat = zeros(length(Rpeaks), segl + segr + 1);
    for m = 1:length(Rpeaks)
        ECGmat(m,:) = (ECG(Rpeaks(m)-segl:Rpeaks(m)+segr))';
    end
    % Compute median from ECG beat array
   MedBeat = median(ECGmat, 1, "omitnan");
    % Ensure it is a row
   MedBeat = MedBeat(:)';
end
function [SynECG] = GenSynECG blend(MedBeat, Rpeaks, ecqRefLen, fs)
    % Compute synthetic ECG by blending median beats.
    % Inputs: MedBeat, Rpeak location, ecgRefLen (Length of reference ECG)
    % Output: synthetic ECG
    % Cross-fade over last xx elements
    % Half of the crossfade
    crosshalfn = 50;
    % Since sampling rate at development time was 1000 sample/s:
    crosshalfn = floor(crosshalfn*fs/1000);
   medl = floor(length(MedBeat)/3);
    % Add a dummy beat at the end to fill the full space
    Rpeaks(end+1) = Rpeaks(end)+floor(mean(diff(Rpeaks)));
    % Set total out recording duration to avoid overflow
   ECGlen = max(ecgRefLen, Rpeaks(end) + 2*medl + 2);
    % Create linear crossfade function
    W = linspace(1, 0, 2*crosshalfn + 1);
    % Create empty output.
    SynECG = NaN(1, ECGlen);
    m = 2;
    % If initial Rpeak is <= medl, delete and start with second Rpeak
    if Rpeaks(1) < medl
        Rpeaks(1) = [];
    end
```

```
while m < length(Rpeaks)</pre>
        % Get segment duration for current beat (a beat is 1 segment to the
        % left and 2 segments to the right of QRS)
        segl = floor((Rpeaks(m) - Rpeaks(m-1))*1/3);
        % Merge point is 1/3 beat back.
        mergepoint = Rpeaks(m) - seql;
        % Getting pointers for extremes of crossfade region of output
        sl = mergepoint - crosshalfn;
        sr = mergepoint + crosshalfn;
        % Pointers for xfade region of median beat
        ml = medl + 1 - segl - crosshalfn;
        mr = ml + 2*crosshalfn;
        mx = length(MedBeat) - mr;
        % Check for errors:
        if ml < 1
            disp (['ERROR! median beat too short to crossfade! add some' ...
                ' margin to median beats!!! (there is a setting in ' ...
                'compute median beat function!)'])
        end
        % Crossfade on the left
        SynECG(sl:sr) = SynECG(sl:sr).*W + MedBeat(ml:mr).*(1 - W);
        % Stamp median beat outside the crossfade region
        SynECG(sr+1: sr+mx) = MedBeat(mr+1:end);
        % Increment counter
       m = m + 1;
    % End while
    end
    SynECG = SynECG(1:ecgRefLen);
% End function
end
```