



The problem is BLIND: one measurement, two unknowns

P cannot be corrected without a clean x — and x cannot be reconstructed without the right P



Why the loop works: a learned path to the clean manifold, held on the physics at every step


the prior supplies the DIRECTION (it was trained on this very path); the measured data supplies the CONSTRAINT that keeps the direction honest




Training: a flow-matching prior on the GEOMETRY BRIDGE

the prior is trained on exactly the image manifold the inference ODE walks: reconstructions under a PARTIALLY CORRECTED geometry

1. Simulate a motion-corrupted scan (CQ500, train split)



2. Geometry bridge (whole volume, no_grad — one fused FDK+tangent pass)



Inference: blind rigid-motion correction — a PnP-TV predictor-corrector loop

θ is UNKNOWN: the image and the motion bootstrap each other, one Euler step of the FM ODE at a time ($N = 50$ steps)

