Algoritmos
 Pesaje de Filtros
 FILAB-F-0105 Rev.02 (08/2022)

Código Termohigrómetro: THI-015 Código Masa Patrón: MSP-004 Tipo de Filtro: Gravimetria PM 2.5 Monitoreado

				Т						HR	Т					HR							Т				Peso			
	Peso_1		HR	ъС	H	IR	T	Peso_2		%	ъC	HR	T	Peso_3		%	ъС	HR	T	Peso_4		HR	ъС	HR	T	Promedio	Material	Factor	Factor	I
ID FILTRO	mg	Fecha_1	% Lect	t Lect	t 9	%	₽C	mg	Fecha_2	Lect	Lect	%	ъС	mg	Fecha_3	Lect	Lect	%	ъС	mg	Fecha_4	% Lect	Lect	%	ºC	Peso Final	Particulado	HR	Т	Responsable
D9110569														167.349	2023-02-01 14:21:01	37.1	22.7	38.4	23.7	167.350	2023-02-01 17:09:37	31.4	19.3	32.7	20.3	167.350	0.174	-1.3	-1	SR
D9110571														170.046	2023-02-01 14:22:54	37.1	22.7	38.4	23.7	170.059	2023-02-01 17:12:45	31.4	19.3	32.7	20.3	170.053	0.178	-1.3	-1	SR
D9110576														166.960	2023-02-01 14:24:38	37.1	22.7	38.4	23.7	166.955	2023-02-01 17:14:02	31.4	19.3	32.7	20.3	166.958	0.838	-1.3	-1	SR
D9110557														169.618	2023-02-01 14:26:20	37.1	22.7	38.4	23.7	169.610	2023-02-01 17:15:29	31.4	19.3	32.7	20.3	169.614	0.215	-1.3	-1	SR
D9110558														169.031	2023-02-01 14:27:55	37.1	22.7	38.4	23.7	169.022	2023-02-01 17:16:50	31.4	19.3	32.7	20.3	169.027	0.281	-1.3	-1	SR
D9110559														168.632	2023-02-01 14:29:28	37.1	22.7	38.4	23.7	168.655	2023-02-01 17:18:19	31.4	19.3	32.7	20.3	168.644	0.264	-1.3	-1	SR
D9110544														167.387	2023-02-01 14:31:16	37.1	22.7	38.4	23.7	167.399	2023-02-01 17:20:37	31.4	19.3	32.7	20.3	167.393	0.926	-1.3	-1	SR
D9110545														164.086	2023-02-01 14:34:04	37.1	22.7	38.4	23.7	164.088	2023-02-01 17:22:01	31.4	19.3	32.7	20.3	164.087	0.954	-1.3	-1	SR
D9110546														164.987	2023-02-01 14:36:52	37.1	22.7	38.4	23.7	164.986	2023-02-01 17:24:17	31.4	19.3	32.7	20.3	164.987	1.109	-1.3	-1	SR
D9110531														167.011	2023-02-01 14:39:01	37.1	22.7	38.4	23.7	167.005	2023-02-01 17:25:44	31.4	19.3	32.7	20.3	167.008	0.748	-1.3	-1	SR
D9110532														166.801	2023-02-01 14:42:13	37.1	22.7	38.4	23.7	166.820	2023-02-01 17:26:50	31.4	19.3	32.7	20.3	166.811	0.657	-1.3	-1	SR
D9110533														166.507	2023-02-01 14:45:25	37.1	22.7	38.4	23.7	166.504	2023-02-01 17:28:26	31.4	19.3	32.7	20.3	166.506	0.936	-1.3	-1	SR
D9110518														165.250	2023-02-01 14:48:37	37.1	22.7	38.4	23.7	165.244	2023-02-01 17:30:22	31.4	19.3	32.7	20.3	165.247	0.640	-1.3	-1	SR
D9110519														165.075	2023-02-01 14:51:02	37.1	22.7	38.4	23.7	165.095	2023-02-01 17:32:58	31.4	19.3	32.7	20.3	165.085	0.631	-1.3	-1	SR
D9110520														164.174	2023-02-01 14:52:10	37.1	22.7	38.4	23.7	164.170	2023-02-01 17:35:51	31.4	19.3	32.7	20.3	164.172	0.718	-1.3	-1	SR
QC														99.996	2023-02-01 14:54:07	31.4	19.9	32.7	20.9	99.997	2023-02-01 17:38:52	36.5	19.4	37.8	20.4	99.997	99.997	-1.3	-1	SR