Algoritmos
 Pesaje de Filtros
 FILAB-F-0105 Rev.02 (08/2022)

Código Termohigrómetro: THI-015 Código Masa Patrón: MSP-004 Tipo de Filtro: Gravimetria PM 2.5 Monitoreado

																													
				Т					HR	Т					HR	Т						Т				Peso			
	Peso_1		HR	ъС	HR	Т	Peso_2		%	ъC	HR	Т	Peso_3		%	ōС	HR	T	Peso_4		HR	ъC	HR	T	Promedio	Material	Factor	Factor	
ID FILTRO	mg	Fecha_1	% Lect	Lect	: %	ъС	mg	Fecha_2	Lect	Lect	%	ъС	mg	Fecha_3		Lect		°C	mg	Fecha_4	% Lect		%	ъС		Particulado	HR	T	Responsable
D9110053													162.477						162.486	2023-03-03 11:58:15	30	20.3	31.3	21.3	162.482	0.868	-1.3	-1	1 YZ
C9587506													125.849						125.859	2023-03-03 11:58:45	30	20.3	31.3	21.3	125.854	1.596	-1.3	-1	1 YZ
C9587704													125.415	2023-03-03 09:54:10					125.373	2023-03-03 12:00:31	30	20.3	31.3	21.3	125.394	0.549	-1.3	-1	1 YZ
C9587705													123.818						123.813	2023-03-03 12:01:35	30	20.3	31.3	21.3	123.816	0.628	-1.3	-1	1 YZ
C9587706													126.894	2023-03-03 09:57:13					126.878	2023-03-03 12:02:52	30	20.3	31.3	21.3	126.886	0.532	-1.3	-1	1 YZ
C9587707													125.601						125.594	2023-03-03 12:04:17		20.3	31.3	21.3	125.598	0.707	-1.3	-1	1 YZ
C9587708													123.684						123.670	2023-03-03 12:05:09	30	20.3	31.3	21.3	123.677	0.662	-1.3	-1	1 YZ
C9587709													123.727						123.693	2023-03-03 12:06:26	30	20.3	31.3	21.3	123.710	0.620	-1.3		1 YZ
C9587710													127.176						127.183	2023-03-03 12:07:17	30	20.3	31.3	21.3	127.180	0.543	-1.3	-1	1 YZ
C9588171													128.930						128.896	2023-03-03 12:08:22	30	20.3	31.3	21.3	128.913	0.155	-1.3	-1	1 YZ
C9588172													126.734						126.710	2023-03-03 12:13:28	30	20.3	31.3	21.3	126.722	0.319	-1.3	-1	1 YZ
C9588173													126.080	2023-03-03 10:06:24					126.066	2023-03-03 12:14:58	30	20.3	31.3	21.3	126.073	0.521	-1.3	-1	1 YZ
C9588174													128.591						128.544	2023-03-03 12:16:03	30	20.3	31.3	21.3	128.568	0.803	-1.3	-1	1 YZ
C9588175													126.757	2023-03-03 10:08:00				22.8	126.747	2023-03-03 12:16:58	30	20.3	31.3	21.3	126.752	0.520	-1.3	-1	1 YZ
C9588176													124.586					22.8	124.588	2023-03-03 12:17:48	30	20.3	31.3	21.3	124.587	0.369	-1.3		1 YZ
QC													99.997	2023-03-03 10:09:02	29.2	21.4	30.5	22.4	99.996	2023-03-03 12:18:38	36.3	20.6	37.6	21.6	99.996	99.996	-1.3	-1	1 YZ
																											-1.3	-1	L YZ
																											-1.3		1 YZ
																											-1.3	-1	1 YZ
																											-1.3		1 YZ
																											-1.3	-1	1 YZ