


The GM100 probe has two separate coaxial leads, corresponding to transmit and receive.


Lead A/B physical TX/RX identity is unknown from the supplied product photo.

Identify the leads using marked gauge sockets or a verified probe datasheet before connection.









This is an interface map only: no pulser, receiver protection, isolation or connector footprint.

Do not energize the probe from this drawing: bench characterization and electrical review are needed.

Research prototype; electrical design requires review

Sheet: /
File: probe-interface.kicad_sch

Title: Mammary Protect — GM100 split-probe connection map

Size: A4	Date:	**Rev: 0.1**
KiCad E.D.A. 10.0.8		Id: 1/1