Deep Learning Output Comparisons



Paired synthetic model/reference examples—identify reproducible failure modes.

Synthetic training material—not clinical guidance. No live Microsoft Foundry access is required.

Each case shows the same synthetic CT-like input twice: the model localization/cue is outlined with a dashed blue boundary; the reference annotation, when present, is outlined with a solid red boundary. Use the visual relationship and the quantitative comparison row to identify patterns. Do not infer a clinical diagnosis beyond the supplied reference annotation.

Case	Acquisition Note	Model Score	Reference Lesion Marked?	Overlap (IoU)	Centroid Offset	Model/reference area ratio
DL-01	Adequate contrast	0.76	Yes	0.79	2.2	0.86
DL-02	Borderline contrast	0.63	Yes	0.31	0.0	0.31
DL-03	Adequate contrast	0.71	Yes	0.01	29.5	0.78
DL-04	Motion artifact present	0.66	No reference lesion marked	0.0		



DL-01

Model localization/cue	Reference annotation
	

Observable evidence: Use as a comparison case for relatively close model/reference localization.

Evidence citation: DL-01



DL-02

Model localization/cue	Reference annotation
	

Observable evidence: The model region covers substantially less area than the reference region; note the low overlap and area ratio.

Evidence citation: DL-02



DL-03

Model localization/cue	Reference annotation
	

Observable evidence: The model cue is spatially displaced from the reference region; note the centroid offset and low overlap.

Evidence citation: DL-03

DL-04

Model localization/cue	Reference annotation
	

Observable evidence: No reference lesion is marked, but the model shows a localized cue in a study with motion artifact.

Evidence citation: DL-04