



Name der Produktfamilie: **Kronenpräparation rotierend Hartmetall**  
*Name of product family: Crown preparation rotating carbide*

Technische Dokumentation Nr: **532862**  
*Technical Documentation No:*

entsprechend der Verordnung / according to Regulation (EU) 2017/745:

Klasse/ **IIa** Regel/ Rule: **6**  
*Class:*

Basic UDI-DI: **++E22653286254; ++E226532862SSK**

SRN: **DE-MF-000006446**

EMDN Code: **Q010199 CONSERVATIVE DENTISTRY AND ENDODONTICS - OTHER**

UMDNS Code:

**16-668 Bohrer, dental, Karbid i) wiederverwendbare Instrumente**  
**Burs, dental, cabide i) reusable instruments**

Wir, Gebr. Brasseler GmbH & Co. KG, als der Hersteller erklären hiermit in alleiniger Verantwortung, dass die oben beschriebenen Medizinprodukte gemäß der Artikelliste mit Revisionsstand vom 02.03.2026 sowie die zugehörige technische Dokumentation mit allen einschlägigen Bestimmungen der Verordnung (EU) 2017/745 über Medizinprodukte übereinstimmen.

Das angewendete Konformitätsbewertungsverfahren entspricht:

Anhang IX, Kapitel I, Abschnitt 2 und 3

Benannte Stelle: TÜV Rheinland LGA Products GmbH, Tillystr. 2, 90431 Nürnberg, Kenn- Nr.: **CE 0197**

Zertifikatsnr.: HZ 1470094-1 Gültigkeitsdatum 2031-02-28

Diese Konformitätserklärung gilt in Verbindung mit dem chargenbezogenen Freigabedokument der hergestellten Medizinprodukte.

We, Gebr. Brasseler GmbH & Co. KG, as the manufacturer herewith declare under our sole responsibility that the above mentioned medical devices, as specified on the list of articles with revision status as at 02.03.2026 and the associated technical documentation correspond to all applicable provisions of the Regulation (EU) 2017/745 on medical devices. The applied conformity assessment procedure is in compliance with:

Annex IX, Chapter I, Section 2 and 3

Notified Body: TÜV Rheinland LGA Products GmbH, Tillystr. 2, 90431 Nuremberg, ident. no.: **CE 0197**

No. of Certificate: HZ 1470094-1 Expiry date: 2031-02-28

This declaration of conformity is valid in connection with the release document for the respective batch of manufactured medical devices.

**Gebr. Brasseler GmbH & Co. KG**  
**Trophagener Weg 25**  
**32657 Lemgo · Germany**

Rev. Stand: **02.03.2026**

**Unterschrift/ Signature**

Lead Product Engineering BA and TestLab

**Unterschrift/ Signature**

Head of Regulatory Affairs (PRRC)



i.V.

*Friedrich-Wilhelm Meier*  
Friedrich-Wilhelm Meier (Mar 3, 2026 15:26:56 GMT+1)

i.V.

*Petra Meier*  
Petra Meier (Mar 3, 2026 16:30:59 GMT+1)