

Møller Electronics 1998–2008

Eugene Chudakov · 4 June 2008

The electronics and DAQ was installed in 1998 and used up to 2008 with only slight modifications. The lead glass calorimeter was replaced in 2000, by a SPACAL-type calorimeter.

Detectors

The signals from both the calorimeter and the aperture counters is about 15 ns long.

Electronics

Electronics components

Step	Module						Signals					
	function	type	bandwidth	pulse	channels	comment	input			output		comment
							step	cable	max rate	pulse	max rate	
1	Linear fan-out	P/S 748	250 MHz		8		Cal 1-8	-	-	20 ns	-	
2	Linear fan-out	P/S 740	250 MHz		4		1	1 ns	-	20 ns	-	4L, 4R
3	Discriminator	Ortec TD8000	100 MHz	$\tau > 5$ ns	8	blocking mode	2	8 ns	-	8 ns	1 MHz	ΣL , ΣR
							Apert 1-2	-	-	8 ns	3 MHz	ApL, ApR
4	Delay unit	LeCroy 4518	100 MHz	$\tau > 5$ ns	16	double > 10 ns	3	2 ns	3 MHz	8 ns	3 MHz	ΣL , ΣR , ApL, ApR
5	PLU	LeCroy 2365	75 MHz	$\tau > 7$ ns	16	8 outputs	4	2 ns	3 MHz	8 ns	1 MHz	Coincidences, BCM etc
6	Scaler	LeCroy 1151E	80 MHz	$\tau > 5$ ns	16		5	2 ns	1 MHz	-	-	Coincidences, BCM etc
7	TDC	LeCroy 2277	80 MHz	1 ns bin	32	double > 20 ns	4	2 ns	3 MHz	-	-	ΣL , ΣR , ApL, ApR