

The Møller target Superconducting Magnet Ramping Rates

Power Supply IPS120-10

Current, A	Rump, A/min	Max. Field, T
0 – 30	8	1.0
30 – 60	6	2
60 – 100	4	3.33
100 – 110	2.4	3.66
110 – 115	1.2	3.83
115 – 118	0.6	3.95
118 – 120	0.25	4.0