

November 20th 2015
Target Ladder Survey

Scan #	Linear Location	Scan of	Command To Get There		Pitch	Yaw	Foil Center (mm Above Beam)	LENC	RENC	Notes
1	- Limit	Ladder		Manually	90.1	3.6		*3.055		Ladder was 28.5 mm from beam line
	- Limit	Foil 1			90.6	3.58		*3.055		
	- Limit	Foil 4			90.8	(+?) -4		*3.055		
2	+ Limit	Ladder		Manually	90.2	3.15				
	+ Limit	Foil 1			90.8	3.1				
	+ Limit	Foil 3			90.1	2.1				
	+ Limit	Foil 4			91.4	2.4				
	+ Limit	Foil 4			90.9	3.4				Smaller Sampling Area
	+ Limit	Foil 4			90.4	2.67				
	+ Limit	Foil 4			90.7	3.3				Larger Sample Area
3	Parked	Foil 1	≈2	mm off switch			43.47	3.025		NOW USING SOFTWARE
4	Foil 1	Foil 1	43.47	mm			-0.03	2.364		
5	Foil 1	Foil 1	(±?) 0.03	mm			-0.07			Wrong Direction? Backlash Software Setting?
6	Foil 2	Foil 2	37	mm			-0.25	1.809		
7	Foil 3	Foil 3	37	mm			-0.07	1.25		
8	Foil 4	Foil 4	37	mm			-0.09	0.679		
9	Foil 1	Ladder	≈.5	degrees		-7.9			4.15	Rotational Parked Position
10	Foil 1	Ladder	4	degrees		-3.9			4.353	
11	Foil 1	Ladder	3.9	degrees		-0.2			4.558	
12	Foil 1	Ladder	4	degrees		3.9			4.75	
13	Foil 1	Ladder	4	degrees		8			4.96	
14	Foil 1	Ladder	5	degrees		13.1			5.206	
15	Foil 1	Ladder	-16.9	degrees						
	Foil 1	Ladder	4.2	degrees		0.2			4.565	
16	Foil 4	Ladder	111	mm		-0.6			4.566	Note sure of command count
17	Foil 1	Ladder	-111	mm		0.2		2.369	4.566	Note sure of command count
18	Foil 4	Ladder	111	mm		-0.6		0.684	4.566	Note sure of command count
19	Foil 4	Ladder	≈.5	degrees		-8.1			4.185	Rotational Parked Position
20	Foil 4	Ladder	8.1	degrees		-0.1			4.6	
21	Foil 4	Ladder	13	degrees		13.1			5.262	
22	Foil 3	Ladder	37	mm		-0.4			4.566	
23	Foil 2	Ladder	37	mm		0			4.566	
24	Foil 2	Ladder	Bill Bump			-0.1			4.566	
25	Foil 2	Ladder	Ethan Bump			-0.7			4.566	
26	Foil 2	Ladder	go to - limit	degrees						
26	Foil 2	Ladder	3.565	RENC		-0.7			4.565	