

Beating the reaction limits of biosensor sensitivity with dynamic tracking of single binding events

Derin Sevenler^{a,b,1}, Jacob Trueb^c, and M. Selim Ünlü^{a,1}

^aDepartment of Electrical and Computer Engineering, Boston University, Boston, MA 02215; ^bCenter for Engineering in Medicine, Massachusetts General Hospital, Harvard Medical School, Boston, MA 02129; and ^cDepartment of Mechanical Engineering, Boston University, Boston, MA 02215

Edited by David R. Walt, Brigham and Women's Hospital–Harvard Medical School, Boston, MA, and accepted by Editorial Board Member John A. Rogers January 19, 2019 (received for review September 13, 2018)

The clinical need for ultrasensitive molecular analysis has motivated the development of several endpoint-assay technologies capable of single-molecule readout. These endpoint assays are now primarily limited by the affinity and specificity of the molecular-recognition agents for the analyte of interest. In contrast, a kinetic assay with single-molecule readout could distinguish between low-abundance, high-affinity (specific analyte) and high-abundance, low-affinity (nonspecific background) binding by measuring the duration of individual binding events at equilibrium. Here, we describe such a kinetic assay, in which individual binding events are detected and monitored during sample incubation. This method uses plasmonic gold nanorods and interferometric reflectance imaging to detect thousands of individual binding events across a multiplex solid-phase sensor with a large area approaching that of leading bead-based endpoint-assay technologies. A dynamic tracking procedure is used to measure the duration of each event. From this, the total rates of binding and debinding as well as the distribution of binding-event durations are determined. We observe a limit of detection of 19 fM for a proof-of-concept synthetic DNA analyte in a 12-plex assay format.

single-molecule detection | biosensor | nanotechnology | nanophotonics | molecular recognition

Several of the promises of precision medicine rely on ultrasensitive molecular diagnostic technologies. Liquid biopsies of circulating genomic, transcriptomic, or proteomic biomarkers of cancer promise earlier detection and treatment, as well as improved guidance of targeted therapies in treating minimum residual disease (1, 2). Similarly, sensitive and specific molecular diagnostic tests for infectious pathogens are vital for the identification and management of pre- or asymptomatic individuals (3, 4). Likewise, panel assays of circulating biomarkers could soon improve the accuracy of diagnosis of injuries such as acute liver failure or traumatic brain injury (5, 6).

These pressing clinical needs have motivated the development of a variety of ultrasensitive assay technologies, culminating in technologies capable of single-molecule detection. A unifying characteristic of essentially all of these assay technologies is that they employ molecular-recognition agents (or “capture probes”) such as antibodies, nanobodies, peptides, oligonucleotides, aptamers, or other agents that bind specifically to the molecule of interest. Single-molecule detection technologies commonly then use droplet emulsions (7) or microwell arrays (8) to isolate and then enumerate the precise number of analyte molecules bound to the capture probes.

In terms of signal transduction, it is clear that single-molecule detection is “as good as it gets” (9). However, transducing the amount of captured analyte is only one half of the picture—the analyte must be captured in the first place. Even with single-molecule detection, assay performance is still limited by the affinity of the capture probes. This causes sensitivity and specificity to vary widely between different probe–analyte pairs. For example, it is now relatively routine to quantify some molecular

species (e.g., genomic DNA) with single-copy sensitivity and precision (10), while the detection limits of other analytes (e.g., microRNA) are many orders of magnitude worse (11–13). Probe affinity can also vary between samples. Variations in extensive properties of the sample such as pH and ion content change the free energy of binding, and variable amounts of nonspecific background binding further complicates quantitation.

Current leading single-molecule detection technologies rely on signal-amplification reactions. These are endpoint assays: the probe molecules are incubated with the sample for a set amount of time, after which the reaction is halted so that amplification and detection can be performed. What is measured is the amount of bound analyte at the instant the incubation is halted.

In contrast to endpoint assays, kinetic assays directly measure probe–analyte interactions during course of the incubation. Kinetic assays collect more information than endpoint assays: they can measure not just concentration but also molecular affinity via rates of association and dissociation. This could allow for intersample variations in probe affinity or nonspecific binding to be identified and mitigated without additional tests. For low concentrations of analyte, kinetic assays could also be capable of distinguishing low-abundance specific binding from a larger background of nonspecific binding, or even measuring analytes below the so-called “critical concentration” at which there is fewer than one analyte molecule bound at equilibrium—a feat impossible for endpoint assays (14).

Significance

Low-abundance molecular biomarkers are a promising reservoir of diagnostic information for the early detection and precise treatment of a range of diseases. Yet, most leading assay technologies are currently limited by the affinity and specificity of the molecular-recognition agents they use. This binding “reaction limit” could be beaten by measuring the duration of individual molecular-binding events as they happen in real time. We developed a technology to measure the duration of thousands of individual binding events across a large sensor area. These duration measurements allowed specific analyte binding to be distinguished from nonspecific background binding. The sensor is large enough to accommodate 12 ultrasensitive tests at once.

Author contributions: D.S. and M.S.Ü. designed research; D.S. performed research; D.S. and J.T. contributed new reagents/analytic tools; D.S. and M.S.Ü. analyzed data; and D.S., J.T., and M.S.Ü. wrote the paper.

The authors declare no conflict of interest.

This article is a PNAS Direct Submission. D.R.W. is a guest editor invited by the Editorial Board.

Published under the PNAS license.

¹To whom correspondence may be addressed. Email: dsevenler@mgh.harvard.edu or selim@bu.edu.

This article contains supporting information online at www.pnas.org/lookup/suppl/doi:10.1073/pnas.1815329116/-DCSupplemental.

Here, we describe a kinetic assay that measures the duration of individual binding events over time on a large sensor surface with a low-magnification objective, while retaining the advantages of kinetic analysis such as discrimination between specific and nonspecific events based on duration (i.e., affinity). In this study, we used a 20 \times , 0.45 NA objective and a 1.1-inch format camera, which yielded a sensor area of 0.38 mm², comparable with that of ultrasensitive endpoint methods. (This area could be further increased several-fold with different optical instrumentation and stage scanning.)

Results

Detection of Individual Binding Events Across a Large Field of View.

We recently described the development of a “digital microarray” assay technology, which rapidly enumerates individual captured molecules across hundreds of microarray spots (21). This technology uses probe-conjugated gold nanorods as molecular labels and an interferometric reflectance imaging sensor (IRIS) to rapidly detect individual nanorods with a low-magnification (10–20 \times) objective. The large field of view enabled a similar throughput to commercial fluorescence readers while enhancing the limit of detection (LOD) and dynamic range by a factor of $\sim 10,000$. The assay reaction is based on the barcode assay developed by others (22) and is compatible with a range of different analytes.

We adapted the IRIS digital microarray platform for dynamic measurements by designing a perfusion chamber that consists of an IRIS chip, a patterned silicone gasket, and an antireflection-coated coverglass window (Fig. 1A). Two holes for the chamber inlet and outlet are drilled in the IRIS chips by wafer-scale laser micromachining. The assembly is held by a custom clamp fixture that makes fluidic connections to the inlet and outlet on the bottom of the chip (*SI Appendix, Fig. S1*).

To demonstrate dynamic detection of single molecules, a synthetic ssDNA oligonucleotide was used as a proof-of-concept analyte. The analyte was preincubated with cDNA-conjugated gold nanorods nominally 25 nm \times 70 nm for 90 min. The concentration of nanorod labels was kept constant for all experiments at 14 pM, while the concentration of the analyte varied from 10 pM to 10 fM. After preincubation, the mixture was perfused over IRIS chips with DNA microarrays of complementary and noncomplementary probes and nanorod-labeled oligos hybridized to the complementary spots (Fig. 1B).

Images were acquired every 30 s with the IRIS instrument during perfusion. Nanorods on the IRIS chip were visible as faint




Fig. 1. Dynamic measurements of single binding events across a large microarray. (A) Rendered image of IRIS chip perfusion chamber for dynamic measurements of molecular interactions. A DNA or antibody microarray is printed on the IRIS chip. Then, the chamber is formed by layering a patterned adhesive gasket and antireflection-coated coverglass viewing window. The IRIS chip has two through-holes for sample perfusion. The entire disposable costs about \$5 USD. (B) Nucleic acid assay with IRIS. DNA-conjugated gold nanorods are preincubated with the sample solution and hybridized with complementary nucleic acids. The mixture is flowed over the chip. Complementary nucleic acid strands tether nanorods to the cDNA microarray spot. (C) Schematic of dynamic detection of single nanorods with IRIS. Images are simulated. Nanorods on the chip surface are observed as diffraction-limited spots and automatically detected using purpose-built software. (D) Plots of total nanoparticle binding to six complementary (red) and six noncomplementary (green) DNA spots over time, as measured with dynamic tracking, for one representative experiment where the analyte concentration is 316 fM.

diffraction-limited blobs in the images, which were detected in each frame independently using custom software (Fig. 1C). The binding rates were then measured determined using a custom dynamic tracking algorithm described in the following section (Fig. 1D and Movie S1).

IRIS detects individual nanorods based on light scattering. Since water has a higher refractive index than air, the polarizability and scattering cross section of the nanorods were reduced compared with dry chips. Additionally, the image suffered from spherical aberrations caused by the air-coverglass interface. Although nanorods were detectable with a 10 \times , 0.3 NA objective, visibility was much improved with a 20 \times , 0.45 NA coverglass corrected objective. The resulting field of view of 0.38 mm² (725 μ m \times 530 μ m) could comfortably accommodate 12 microarray spots, each \sim 80 μ m in diameter.

Dynamic Tracking of Binding Events over Time. Under sufficiently high flow rates, the initial rate of binding of analyte is proportional to the bulk analyte concentration. One may estimate the analyte concentration by plotting the number of bound nanorods over time and measuring the initial slope (“naïve counting”; Fig. 2A). However, this approach has several problems. The first is the fundamental issue related to finite probe affinity mentioned earlier. Ultralow analyte concentrations will reach equilibrium with very few (or even fewer than one) bound analyte molecules. In those cases, the initial slope will not be measurable even with perfect error-free readout. The second issue is that some unbound nanorods are visible in each frame as they transiently diffuse through the detection volume. These transient particles result in false positives that increase the overall noise floor of the sensor.

To address these issues, we developed a postprocessing algorithm that uses the spatial positions of particles to track them individually over the course of the experiment (Fig. 2B). First,

particles are detected in each video frame independently of other frames, and their positions in the image are recorded. Second, these positions are compared with those of particles in the next frame of the video (Fig. 2C). Particles in the same position in both frames are “matched,” indicating that they are in fact the same particle. Particle matching includes a clustering algorithm that is robust to small translations between frames. For a video with N time points, this results in $N - 1$ lists of matches. Third, these lists are compiled into a single master catalog, which tracks the contiguous series of frames in which each particle was observed (Fig. 2D). This is essentially a table that lists when each particle bound, where it bound, and when (if ever) it debound from the surface. Finally, this catalog is filtered to reduce false positives and false negatives. False negatives occur when a particle is mistakenly not detected in a single frame but was detected in the same position in previous and subsequent frames. This erroneously results in two entries in the catalog. These gaps are repaired by identifying whether the binding of each particle corresponds to the exact same place as the debinding of another particle two frames prior and then merging the two catalog entries (e.g., particles 3 and 7; Fig. 2D). False positives are caused by particles visible in just one frame and are simply removed (e.g., particle 4; Fig. 2D).

This catalog can then be used to plot the cumulative number of new binding events over time. For low-analyte concentrations, most binding sites remain empty at equilibrium, and the rate of new binding events will be constant and proportional to the bulk concentration. At an analyte concentration of 316 fM, for example, the sensor reached equilibrium with about 80 bound nanorods after 1 h (naïve counting; Fig. 3A). However, the cumulative number of new binding events continued to increase linearly even after equilibrium was reached (“total binding”; Fig. 3B). Note that the total number of binding and debinding events





Fig. 2. (A) Dynamic tracking improves sensitivity when the debinding rate is nonzero. Assuming first-order reaction kinetics, equilibrium is reached when the rate of new analyte binding is balanced by the rate of debinding from the surface. Naïvely counting the instantaneous number of bound analyte provides no additional information once equilibrium has been reached. Dynamic tracking distinguishes the binding of new particles from the debinding of old ones. The cumulative number of binding events continues to increase, even at equilibrium. (B–D) Diagram of the multistep dynamic tracking algorithm, described in the text.






Fig. 4. Rate of particle binding versus analyte concentration for the same experiments analyzed using dynamic tracking (A) or naïve counting (B). A modified first-order kinetic model was fit to the data and used to determine the LOD and critical concentration (c^*). *Insets* show the same data and model on linear axes. Dynamic tracking improved the LOD by 36-fold compared with naïve counting. Notably, the dynamic tracking LOD (19 fM) is 3.6-fold lower than the critical concentration (68 fM), at which just one molecule is bound to the sensor on average.

even when the average duration of binding events was shorter than the time between them. At the critical concentration, equilibrium is reached with one bound analyte molecule on average. This dissociation rate is simply the weighted average of the two dissociation constants found earlier: $\bar{k}_{off} = (A_1 k_1 + A_2 k_2) / A_1 A_2 =$

$0.52 \text{ min}^{-1} = 31 \text{ particles per hour}$. The critical concentration was determined by taking intercept of the dynamic tracking regression line with this binding rate: $c^* = 55 \text{ fM}$.

We observed a saturation in binding rate at concentrations above 2–3 pM (Fig. 4A). We hypothesized this could be the result of using a lower concentration of nanorods than expected. We performed spectrophotometry to estimate the concentration of the conjugated nanorods as well as the nanorod stock solution from the manufacturer. We compared the absorbances to that predicted by numerical simulations and estimated that the nanorod concentration was indeed several times lower than we expected (*SI Appendix, Fig. S7*).

Discussion

A range of endpoint-assay technologies have been developed that have single-molecule readout. For these assays, the limiting factor is the affinity and specificity of the molecular-recognition agents rather than readout sensitivity. The number of capture probes (for example, the number of functionalized beads) can almost always be increased until sensitivity is limited by non-specific binding rather than insufficient numbers of probes (23), but further improvement must come through careful optimization of washes and reactions. Protocol optimization are particularly challenging for multiplexed test development, since the optimal wash conditions (duration, ion content, surfactants, pH, and so on) are often different for different probe–analyte complexes.

In response to these limitations, we have introduced a kinetic-assay technology that measures the duration of individual binding events across a large sensor area. In this work, we distinguished specific binding events from the background without even a single wash step, which could be used to further improve specificity. Notably, kinetic analysis alleviates the need for a globally optimal wash protocol and therefore makes multiplexed tests straightforward.

“Solid-phase” surface sensors are sometimes criticized for having poor mass-transport kinetics, compared with bead-based assays. We alleviated this effect by using a high flow rate, which makes the depletion layer very thin (*SI Appendix*). For longer experiments or smaller sample volumes, peristaltic pumping and recirculation may be used.

The unexpected biexponential histogram of dwell times suggested two different conformations of immobilized nanorods, each with different binding free energy: end-tethered and side-tethered. This variability complicates the probe–analyte affinity measurement and could be problematic for heterogeneous samples. This could be further tested and perhaps mitigated by preferentially functionalizing just the ends of nanorods (24, 25).

Because these proof-of-concept experiments were conducted with synthetic analytes and pure buffer solutions, direct comparisons with more mature assay technologies cannot yet be made. However, future applications of this technique are envisioned in response to current clinical needs for multiplexed quantification of microRNAs for cancer diagnostics, mRNAs for phenotypic identification of drug susceptibility, and circulating protein biomarkers of viral infection.

Materials and Methods

Additional materials and methods are available in *SI Appendix, Supplemental Materials and Methods*.

Perfusion Chamber Assembly. No. 1 coverslips, 25.4 mm × 12.7 mm, with a broadband antireflection coating on one side were purchased from Abris Technologies. Custom patterned silicone gaskets were purchased from Grace Biolabs. Silicone gaskets were 25.4 mm × 12.7 mm, 0.15-mm thickness, with pressure-sensitive adhesive on one side. In preparation, gaskets were adhered to the noncoated side of the coverglass and stored with protective tape in place.

The perfusion chamber was assembled by aligning the gasket-window assembly to the IRIS chip, loading it into the clamp fixture, removing the protective tape, and engaging the clamp to form seals between the chip and the gasket as well as with the sample inlet and outlet. The volume of the chamber was $\sim 8 \mu\text{L}$.

IRIS Digital Microarray Instrument for Dynamic Detection. The operating principle of IRIS is thoroughly described elsewhere (26). Briefly, the IRIS instrument consists of a reflectance microscope with a single high-powered LED for illumination (M660L4 LED with FB650-10 bandpass filter; Thorlabs) and a monochrome machine vision camera (Grasshopper GS3-U3-123S6M-C; Point Gray Research). The digital microarray implementation of IRIS is optimized for rapidly detecting individual gold nanorods based on their anisotropic light-scattering properties. The design, optimization, and implementation of the optical system has been described in detail elsewhere (21). For dry IRIS chips, this system can detect single gold nanorods with a $10\times$, 0.3 NA objective. For dynamic experiments the system was entirely the same except that a $20\times$, 0.45 NA coverslip-corrected air immersion objective (Nikon CFI S Plan Fluor ELWD $20\times$) was used. The higher light-collection efficiency compensated for the decreased intensity of nanorod light scattering due to immersion of the rods in water, and the collar allowed correction of spherical aberrations from the coverslip–air interface.

Assay Protocol. The assay protocol was identical for all experiments, except that the concentration of the analyte was changed. First, the DNA–nanorod conjugates and synthetic DNA oligos were premixed in a “hybridization buffer” consisting of 10 mM phosphate ($\text{pH } 7.4$), 600 mM Na^+ , 0.1% Tween-20, and 1 mM EDTA ; $100 \mu\text{L}$ of nanorods stored at 140 pM were mixed with $900 \mu\text{L}$ of hybridization buffer containing the analyte ssDNA. The final nanorod concentration was 14 pM for all experiments. The mixture was vortexed briefly and sonicated for 10 s before storing at room temperature in a microcentrifuge tube. After 90 min , the sample was aspirated with a disposable 1-mL needle-tipped syringe. The needle tip was removed, and the syringe was connected to a Luer fitting on the end of the inlet tube. The outlet waste tube was left in conical vial. The syringe was mounted in a syringe pump, and the sample was dispensed at $10 \mu\text{L}/\text{min}$ for up to 90 min . The instrument was refocused as soon as the liquid sample filled the chamber. Video acquisition began $1\text{--}3 \text{ min}$ after the sample first contacted the chip surface.

Image Acquisition. Image acquisition was automated using the Micromanager (27) microscope control software with custom scripts. Scripts have been made freely available online at <https://www.github.com/derinsevenler/IRIS-API>. Time points were taken every 30 s at each time point, and a z-stack of nine images was acquired with a step size of $2 \mu\text{m}$ (i.e., a span of $16 \mu\text{m}$). At each z-position, four images were acquired and averaged pixelwise before saving to reduce shot noise.

Image Processing and Particle Detection. The video data from each experiment consisted of an image hyperstack of $180 (t) \times 9 (z) \times 12.4 \text{ MP } (x,y)$. A $86 \mu\text{m} \times 86 \mu\text{m}$ (500×500 pixels) region of the video was cropped around each microarray spot. Nanorods in each region and timepoint were detected independently. The particle detection method described here is a refinement of methods described earlier (21, 28), and has three steps: preprocessing, key point detection, and key point filtering. First, a sparse pseudomedian filter is applied to each frame of the z-stack (made available online by others at imagejdocu.tudor.lu/doku.php?id=plugin:filter:fast_filters:start) to estimate the image background. True median filtering is effective for removing punctate features but computationally expensive for larger kernels. We found the sparse pseudomedian algorithm preferable due to its speed. Next, the normalized intensity image was calculated by pixelwise division of the original frame from the background. Finally, the normalized intensity range (NIR) image was measured by projecting the maximum difference (i.e., maximum – minimum) at each pixel of the normalized intensity stack. Although not every nanorod is visible in every normalized intensity image in the stack, each particle is clearly visible in the resulting NIR image.

Key points in the NIR images were detected by applying a global threshold to binarize the image. Blobs in the binary image (i.e., regions brighter than the threshold) were enumerated and then filtered based on size and shape. Specifically, a minimum area, maximum area, and minimum area-to-perimeter ratio were specified. The detection threshold and key point-filtering parameters were manually selected and then kept constant for all experiments.

ACKNOWLEDGMENTS. The authors are grateful to Fulya Ekiz Kanik, Joeseeph Greene, and Abdul Bhuiya for feedback on the image acquisition and analysis software and perfusion chamber design; to Celalettin Yurdakul for help designing the chip photolithography pattern; to Matthew Geib for help preparing IRIS chips and spotting; and to Ahmad Khalil for fruitful discussions about these results.

- Cohen JD, et al. (2017) Combined circulating tumor DNA and protein biomarker-based liquid biopsy for the earlier detection of pancreatic cancers. *Proc Natl Acad Sci USA* 114:10202–10207.
- Alix-Panabières C, Pantel K (2016) Clinical applications of circulating tumor cells and circulating tumor DNA as liquid biopsy. *Cancer Discov* 6:479–491.
- Shurtleff AC, Whitehouse CA, Ward MD, Cazares LH, Bavari S (2015) Pre-symptomatic diagnosis and treatment of filovirus diseases. *Front Microbiol* 6:108.
- Alsdurf H, Hill PC, Matteelli A, Getahun H, Menzies D (2016) The cascade of care in diagnosis and treatment of latent tuberculosis infection: A systematic review and meta-analysis. *Lancet Infect Dis* 16:1269–1278.
- Antoine DJ, et al. (2013) Mechanistic biomarkers provide early and sensitive detection of acetaminophen-induced acute liver injury at first presentation to hospital. *Hepatology* 58:777–787.
- Papa L, et al. (2016) Time course and diagnostic accuracy of glial and neuronal blood biomarkers GFAP and UCH-L1 in a large cohort of trauma patients with and without mild traumatic brain injury. *JAMA Neurol* 73:551–560.
- Vogelstein B, Kinzler KW (1999) Digital PCR. *Proc Natl Acad Sci USA* 96:9236–9241.
- Cohen L, Walt DR (2017) Single-molecule arrays for protein and nucleic acid analysis. *Annu Rev Anal Chem (Palo Alto Calif)* 10:345–363.
- Walt DR (2013) Optical methods for single molecule detection and analysis. *Anal Chem* 85:1258–1263.
- Schoepp NG, et al. (2017) Rapid pathogen-specific phenotypic antibiotic susceptibility testing using digital LAMP quantification in clinical samples. *Sci Transl Med* 9:eaal3693.
- Rissin DM, et al. (2017) Polymerase-free measurement of microRNA-122 with single base specificity using single molecule arrays: Detection of drug-induced liver injury. *PLoS One* 12:e0179669.
- Cohen L, Hartman MR, Amardey-Wellington A, Walt DR (2017) Digital direct detection of microRNAs using single molecule arrays. *Nucleic Acids Res* 45:e137.
- Wu D, Katilius E, Olivas E, Dumont Milutinovic M, Walt DR (2016) Incorporation of slow off-rate modified aptamers reagents in single molecule array assays for cytokine detection with ultrahigh sensitivity. *Anal Chem* 88:8385–8389.
- Squires TM, Messenger RJ, Manalis SR (2008) Making it stick: Convection, reaction and diffusion in surface-based biosensors. *Nat Biotechnol* 26:417–426.
- Baaske MD, Foreman MR, Vollmer F (2014) Single-molecule nucleic acid interactions monitored on a label-free microcavity biosensor platform. *Nat Nanotechnol* 9:933–939.
- Besteman K, Lee J-O, Wiertz FGM, Heering HA, Dekker C (2003) Enzyme-coated carbon nanotubes as single-molecule biosensors. *Nano Lett* 3:727–730.
- Sorgenfrei S, et al. (2011) Label-free single-molecule detection of DNA-hybridization kinetics with a carbon nanotube field-effect transistor. *Nat Nanotechnol* 6:126–132.
- Halpern AR, Wood JB, Wang Y, Corn RM (2014) Single-nanoparticle near-infrared surface plasmon resonance microscopy for real-time measurements of DNA hybridization adsorption. *ACS Nano* 8:1022–1030.
- Piliarik M, Sandoghdar V (2014) Direct optical sensing of single unlabelled proteins and super-resolution imaging of their binding sites. *Nat Commun* 5:4495.
- Lamichhane R, Solem A, Black W, Rueda D (2010) Single-molecule FRET of protein-nucleic acid and protein-protein complexes: Surface passivation and immobilization. *Methods* 52:192–200.
- Sevenler D, Daaboul GG, Ekiz Kanik F, Ünlü NL, Ünlü MS (2018) Digital microarrays: Single-molecule readout with interferometric detection of plasmonic nanorod labels. *ACS Nano* 12:5880–5887.
- Hill HD, Mirkin CA (2006) The bio-barcode assay for the detection of protein and nucleic acid targets using DTT-induced ligand exchange. *Nat Protoc* 1:324–336.
- Yelleswarapu VR, Jeong H-H, Yadavali S, Issadore D (2017) Ultra-high throughput detection (1 million droplets per second) of fluorescent droplets using a cell phone camera and time domain encoded optofluidics. *Lab Chip* 17:1083–1094.
- Caswell KK, Wilson JN, Bunz UHF, Murphy CJ (2003) Preferential end-to-end assembly of gold nanorods by biotin-streptavidin connectors. *J Am Chem Soc* 125:13914–13915.
- Pramod P, Joseph STS, Thomas KG (2007) Preferential end functionalization of Au nanorods through electrostatic interactions. *J Am Chem Soc* 129:6712–6713.
- Sevenler D, Avci O, Ünlü MS (2017) Quantitative interferometric reflectance imaging for the detection and measurement of biological nanoparticles. *Biomed Opt Express* 8:2976–2989.
- Edelstein AD, et al. (2014) Advanced methods of microscope control using $\mu\text{Manager}$ software. *J Biol Methods* 1:e10.
- Trueb J, Avci O, Sevenler D, Connor JH, Ünlü MS (2017) Robust visualization and discrimination of nanoparticles by interferometric imaging. *IEEE J Sel Top Quantum Electron* 23:1–10.