
A7

Anesthesia Machine

Parts


1.1 Diagrams and Tables

1.1.1 Structure



Item No.	Order Number	Part Descriptions	QTY	Remarks
1	115-036378-00	CO2 bypass assembly	1	/
2	115-025564-00	Breathing circuit unit FRU（0635）	1	The PN not contain L-shape bracket assembly
3	115-034401-00	Auxiliary gas supply assembly (American standard)	1	American standard
	115-034400-00	Auxiliary gas supply assembly (European standard)	1	European standard
4	115-017037-00	Top Deck Assembly FRU(0631)	1	for the machine without pendant adapter
	115-024379-00	Top panel assm FRU(0634/pendant version)	1	for the machine with pendant adapter
5	801-0631-00076-00	Storage Mount for Vaporizer	1	/
6	115-034451-00	Drawer Module FRU(0634)	1	/
7	801-0631-00140-00	A5 Drawer Handle	1	/
8	801-0631-00012-00	System Switch Knob	1	/
9	115-023186-00	Panel of gauge(FRU/0635/EU)	1	Refer to the machine’s configuration，please check CRM system with serial number of device to confirm gas type and standard
	115-029978-00	Panel of gauge(FRU/0635/EU/O2&AIR)	1	
	115-023187-00	Panel of gauge(FRU/0635/EU/big tank)	1	
	115-029979-00	Panel of gauge(FRU/0635/EU/O2&AIR)	1	
	115-029974-00	Panel of gauge(FRU/0632/US)	1	
	115-029976-00	Panel of gauge(FRU/0635/US/O2&AIR)	1	
	115-029975-00	Panel of gauge(FRU/0632/reverse EU)	1	
	115-029977-00	Panel of gauge(FRU/0635/Rev_EU/O2&AIR)	1	
	115-040735-00	Panel of gauge FRU(0635/E_PM)	1	


10	115-024850-00	O2 High Pressure Gauge FRU(0635)	1	/
11	115-024851-00	N2O High Pressure Gauge FRU(0635)	1	/
12	115-024852-00	Air High Pressure Gauge FRU(0635)	1	/
13	115-024848-00	N2O pipeline pressure gauge FRU(0635)	1	/
14	115-024849-00	Air pipeline pressure gauge FRU(0635)	1	/
15	115-024847-00	O2 pipeline pressure gauge FRU(0635)	1	/
16	115-022542-00	suction control assembly(CE/0635)	1	/
17	115-027534-00	Total flowmeter assembly FRU(0635)	1	/
18	801-0631-00024-00	Vaporizer Mounting Manifold	1	/



Item No.	Order Number	Part Descriptions	QTY	Remarks
19	115-026810-00	Service door maintenance kit FRU(0635)	1	/
20	801-0631-00120-00	work surface assembly(0631)FRU	1	/
21	115-017947-00	suction bottle bracket assembly	1	/
22	115-017378-00	AGSS Assembly(0631 low-flow)FRU	1	/
	115-017377-00	AGSS Assembly(0631 high-flow)FRU	1	/
23	801-0631-00074-00	AGSS Transfer Tube	1	/
24	034-000285-00	casters.6" bule and black	1	/
25	115-018131-00	Overflow Unit	1	/
26	115-034134-00	O2 Yoke Assembly(With gauge tube)FRU	1	/
	115-040983-00	O2 Yoke Assembly(W/ tube/Air driven)FRU	1	/
	115-056630-00	O2 cylinder yoke assembly(0635)	1	Only for the O2-driven AM with electronic monitoring of gas supply
	115-056628-00	O2 cylinder yoke assembly(0635/air)	1	Only for the AIR-driven AM with electronic monitoring of gas supply
27	115-034136-00	N2O Yoke Assembly(With gauge tube)FRU	1	/
28	115-034135-00	Air Yoke Assembly(With gauge tube)FRU	1	/
	115-040985-00	Air Yoke Assembly(W/ tube/Air driven)FRU	1	/
	115-056631-00	AIR cylinder yoke assembly(0635)	1	Only for the O2-driven AM with electronic monitoring of

				gas supply
	115-056629-00	AIR cylinder yoke assembly(0635/air)	1	Only for the AIR-driven AM with electronic monitoring of gas supply

1.1.2 Upper Half




Item No.	Order Number	Part Descriptions	QTY	Remarks
1	115-064350-00	gas bench receptacle assembly(PCTG)	1	/
2	115-027532-00	Gas mixer assembly FRU(0635)		/
	115-036456-00	gas mixer service kit(0635/O2&Air)		/
3	801-0631-00116-00	Flow Compensation Valve		/
4	115-033064-00	back pressure valve assembly(0635)		/
5	115-015898-00	3-way valve assembly		/

6	115-016780-01	ORC assembly(0632)	1	Need ORC calibration, this calibration needs N2O supply, if no N2O please use “ORC Maintenance Kit” from part list
7	115-018012-00	Li-ion Bat Pack (11.1V4500mAh LI23S002A)	1	Only for warranty Service, PN only contains one battery
8	115-017748-00	vacuum generator unit	1	/
9	801-0631-00086-00	N2O Regulator Assembly	1	/
10	801-0631-00087-00	Regulator Assembly for O2 or N2O FRU	1	PN only contains one regulator
11	115-015436-00	N2O supply inlet assembly(0632/NIST)	1	/
	115-014825-00	N2O origin inlet module(0632/DISS)	1	/
12	115-015437-00	Air supply inlet assembly(0632/NIST)	1	/
	115-014826-00	air origin inlet module(0632/DISS)	1	/
13	115-015435-00	O2 supply inlet assembly(0632/NIST)	1	/
	115-014827-00	oxygen origin inlet module(0632/DISS)	1	/
14	115-018132-00	Vacuum pipeline supply assembly(NIST)	1	/
	115-018133-00	Vacuum pipeline supply assembly(DISS)	1	/
15	051-001257-00	0632 Indicator Board PCBA	1	/
16	115-017877-00	system switch(0632)	1	/
17	115-034392-00	Tube of pressure gauge(0635/left)FRU	1	/

18	115-034393-00	Tube of pressure gauge(0635/middle)FRU	1	/
19	115-034394-00	Tube of pressure gauge(0635/right)FRU	1	/


1.1.3 Hardware Box



Item No.	Order Number	Part Descriptions	QTY	Remarks
1	801-0631-00039-00	0631 top lighting board PCBA	1	/
2	115-051206-00	Ventilator Control Board(DSP) FRU	1	/
3	115-040956-00	0635 main control board FRU(Non FDA/VGA)	1	Refer to the machine's configuration
	115-044211-00	Main control board(Non FDA/EPSON)	1	
4	115-018145-00	0632 Power board FRU	1	/
5	115-040955-00	0635 mother board FRU(Non FDA/VGA)	1	/
6	801-0631-00109-00	A5 Battery Interface Board	1	/

7	801-0631-00028-00	Fan	1	/
8	801-0631-00029-00	Filter Power 250VAC 10A Panel Mount	1	/
9	801-0631-00130-00	Breaker (5.0A)(FRU)	1	/
10	801-0631-00031-00	Breaker(3.0A)	1	PN only contains one Breaker
11	801-0631-00032-00	Auxiliary Output Socket	1	Refer to the machine’s configuration, please check CRM system with serial number of device to confirm the standard
	801-0631-00131-00	3Pins India Power Socket 13A 250V(FRU)	1	
	801-0631-00133-00	0631 B type socket 220V standard(FRU)	1	
	801-0631-00134-00	3Pins UK Power Socket 13A 250V(FRU)	1	
	801-0631-00135-00	Power Socket VDE (white)(FRU)	1	
	801-0631-00136-00	Power Socket Brzl 10A/250Vac(FRU)	1	
12	801-0631-00038-00	Speaker and Connecting Cable	1	/
13	801-0631-00146-00	Solenoid valve assembly(ACGO)FRU	1	/


1.1.4 Work Surface



Item No.	Order Number	Part Descriptions	QTY	Remarks
1	115-036693-00	Drive Assembly(DSP)FRU	1	/
2	115-061228-00	ACGO assembly（0634）	1	Refer to the machine's configuration
	115-061229-00	ACGO assembly（0634/AG）	1	
	115-034440-00	E_ACGO service kit(w/o AA)	1	
	115-034441-00	E_ACGO service kit(w/ AA)	1	
3	801-0616-00007-00	Pneumatic capacitor assembly	1	/
4	801-0631-00044-00	O2 Flush Assembly	1	/
5	801-0631-00068-00	Docking Station Cover	1	/
6	801-0631-00067-00	O2 Connector	1	/

7	801-0631-00114-00	Auto/Manual spring	1	/
8	801-0611-00097-00	Journey Switch	1	/
9	801-0631-00113-00	Auto/Manual pin	1	/
10	801-0631-00052-00	Touch Pad	1	/
11	115-017032-00	Track Pad FRU(0631)	1	/
12	115-034450-00	Circuit Heater FRU(0634)	1	/


1.1.5 Main Body of the Breathing Circuit



Item No.	Order Number	Part Descriptions	QTY	Remarks
1	801-0631-00054-00	Bellows Dome, A series	1	/
2	801-0601-00001-00	Folding bag assembly	1	/


3	115-048600-00	Bag arm assembly(silvery/spacing)	1	/
4	801-0631-00065-00	Bag/mechanical ventilation switch	1	/
5	115-012263-00	APL valve assembly	1	/
6	801-0631-00104-00	One-Way Valve	1	PN only contains one One-Way Valve
7	115-051819-00	Airway pressure gauge	1	/
8	115-018436-00	Upper cover FRU(0631)	1	/
9	115-022320-00	loop low cover 2 assembly FRU	1	/
10	115-021251-00	Lower cover FRU(0632/no flowsenser)	1	/
11	801-0631-00060-00	Inspiratory Flow Sensor Assembly	1	/
12	801-0631-00056-00	Expiratory Flow Sensor Assembly	1	/
13	801-0631-00058-00	Water Collection Cup, A series	1	/
14	801-0631-00057-00	Insp/Exp Connector, A series	1	PN only contains one Insp/Exp Connector
15	801-0631-00059-00	Insp/Exp Connector Rotary Cap, A series	1	PN only contains one Insp/Exp Connector Rotary Cap
16	115-021252-00	Upper cover 2 FRU(0631/no popoff)	1	/

1.1.6 Pre-pak Absorber Canister Assembly




Item No.	Order Number	Part Descriptions	QTY	Remarks
1	115-017030-00	L-shape bracket assembly FUR(0631)	1	/
2	801-0631-00092-00	CO2 Absorber Hose, A series	1	/
3	115-036378-00	CO2 Bypass Assembly FRU, A series	1	/
4	801-0631-00100-00	CO2 Absorber Base, A series	1	/
5	801-0631-00066-00	CO2 Absorbent Canister, A series	1	Only for warranty Service

1.1.7 Expiratory Check Valve and Inspiratory Check Valve




Item No.	Order Number	Part Descriptions	QTY	Remarks
1	801-0631-00061-00	Check valve dome, A series	1	/
2	801-0631-00110-00	valve cover	1	/
3	801-0631-00111-00	Check valve	1	/
4	801-0631-00104-00	One-Way Valve	1	/

1.1.8 O2 Cell Assembly



Item No.	Order Number	Part Descriptions	QTY	Remarks
1	801-0631-00102-00	O2 sensor cabel module FRU	1	/
2	0611-10-45654	Sensor Oxygen (O2 sensor) MedicelMOX-2	1	Only for warranty Service


1.1.9 Display Assembly



Item No.	Order Number	Part Descriptions	QTY	Remarks
1	115-034072-00	encoder knob assembly (balance gas0635)	1	/

2	115-034071-00	encoder knob assembly (oxygen0635)	1	/
3	051-001260-00	0632 Encoder Board PCBA	1	/
4	115-034469-00	encoder assembly service kit(0635)		PN contains encoder knob assembly (balance gas0635), encoder knob assembly (oxygen0635), 0632 Encoder Board PCBA,and Optical Encoder 24detent 5VDC Cable


1.1.10 Main Body of Display Assembly



Item No.	Order Number	Part Descriptions	QTY	Remarks
1	801-0631-00019-00	0631 Warning Light Board PCBA	1	/
2	801-0631-00014-00	Touch Screen	1	/


3	115-036373-00	Display Exchange Package（0634）	1	/
4	801-0631-00018-00	Touch Screen Control Board	1	/
5	051-001258-00	0632 Display Interface Board PCBA	1	/

1.1.11 Vaporizer Mounting Manifold Assembly




Item No.	Order Number	Part Descriptions	QTY	Remarks
1	801-0631-00106-00	Valve of Vaporizer Mounting Manifold	1	/
2	801-0631-00107-00	spring of Vaporizer Mounting Manifold	1	/

1.1.12 Auxiliary Gas Supply Assembly




Item No.	Order Number	Part Descriptions	QTY	Remarks
1	801-0631-00002-00	Flow Meter Lighting Board PCBA	1	/

1.1.13 Base Assembly




Item No.	Order Number	Part Descriptions	QTY	Remarks
1	801-0631-00072-00	Indicator Drive Plate	1	/
2	034-000285-00	casters.6" bule and black	1	/

1.1.14 Cart Assembly



Item No.	Order Number	Part Descriptions	QTY	Remarks
1	082-000423-00	Filter.0.45um,TP026AATV004AD01	1	/
2	801-0613-00033-00	AG transfer board	1	/
3	115-015948-00	assembly of ACGO drive valve	1	PN only contains one valve
4	115-002504-00	AION 03-31 Multigas Analyzer	1	/


1.1.15 EFCS Assembly



Item No.	Order Number	Part Descriptions	QTY	Remarks
1	115-044851-00	Flow sensor board PCBA FRU	1	The O2,AIR,N2O,Total flow sensor boards are the same, PN only contains one sensor
2	115-024754-00	Electronic control board FRU(Non-FDA)	1	/
3	115-018143-00	proportional valve module	1	Only used in o2 and n2o and air configuration, please check CRM system with serial number of device to confirm the configuration
	115-024628-00	proportional valve module(0635/O2/AI	1	Only used in o2 and air configuration , please check CRM system with


		R)		serial number of device to confirm the configuration

1.1.16 BFCS Assembly (Configure with O2 and Air)



Item No.	Order Number	Part Descriptions	QTY	Remarks
1	115-024630-00	O2 needle valve (0635/with basic volume)	1	Refer to the machine's configuration, please check CRM system with serial number of device to confirm configuration
	115-029990-00	O2 needle valve FRU(0635/US/w/guard)	1	
	115-029991-00	O2 needle valve FRU(0635/EU/w/guard)	1	
	115-029993-00	O2 needle valve FRU(0635/Canada/w/guard)	1	
2	115-014468-00	needle valve assembly(0632)	1	/
3	041-006773-00	O2 knob of needle valve (0632)	1	/
4	041-006775-00	AIR knob of needle valve (0632)	1	/

1.1.17 Spare Needle Valve Telescopic Mechanism



Item No.	Order Number	Part Descriptions	QTY	Remarks
1	051-001286-01	0632 BFCS Position Switch Board PCBA	1	/
2	115-018144-00	electromagnet module of BFCS	1	/
3	115-014519-01	backup valves in-out device(0632)	1	Only used in O2&N2O&Air configuration AM, please check CRM system with serial number of device to confirm configuration
	115-029847-00	retractor of backup valves(0632/OA)	1	**Only used in O2&Air configuration AM,** please check CRM system with serial number of device to confirm configuration

1.1.18 Others

Item No.	Order Number	Part Descriptions	QTY	Remarks
1	082-000519-00	7 mm x 10 mm transparent PU tube (polyether)	/	/
2	A21-000007---	3/32"X7/32"X100ft silicone	/	/
3	M6G-020014---	8 mm x 5 mm transparent soft precision PU tubing	/	/
4	M6G-020018---	20 x 25 mm silicone	/	/
5	M6G-020026---	4 x 6 transparent PU tube	/	/
6	M6G-020046---	Tubing. 4mmx2.5mm transparent PU precision hose	/	/
7	M6G-020005---	Tubing.MPF tube 2mmX3.5mm （1feet）	/	/
8	082-001198-00	PE Tubing.ID2XOD4	/	/
9	60-12140-00	ID3/32"XOD5/32",Silicone	/	/
10	9200-10-10556	Tubing.ID3/32"XOD5/32",Silicon	/	/
11	9200-10-10557	System Tubing PUR.AION,1.4/2.8mm,60-12110-00	/	/
12	082-001461-00	System Tubing PUR.AION, 2.2/4.4mm,60-12120-00	/	/
13	115-034448-00	Drive gas spring tube assemblies	/	/
14	115-034449-00	APL valve spring tube assemblies	/	/
15	049-000140-00	Valve port sealing cushion	/	For the CO2 bypass shaft
16	049-000142-00	CO2 bypass assembly washer	/	/
17	049-000143-00	Exterior washer of the absorber canister	/	/
18	049-000145-00	Interior washer of the absorber canister	/	/
19	049-000146-00	CO2 absorber hose	/	/
20	049-000240-00	Pop-off rubber pad	/	/
21	049-000243-00	Bottom washer of the bellows absorber canister	/	/
22	082-000506-00	AGSS filter	/	/

23	082-000934-00	14 x 2.65 O-ring	/	For the vaporizer mount
24	082-001499-00	30 x 2 O-ring	/	Bag arm base
25	082-001500-00	25 x 2 O-ring	/	APL valve
26	082-001501-00	27 x 1.5 O-ring	/	For the check valve dome
27	082-001502-00	18 x 2.5 O-ring	/	Breathing system base
28	082-001503-00	20 x 1.5 O-ring	/	For the check valve
29	082-001504-00	23.47 x 2.95 O-ring	/	For the water collection cup and CO2 bypass assembly
30	082-001505-00	52 x 2 O-ring	/	Auto/Manual ventilation switch
31	082-001508-00	29 x 2.62 O-ring	/	Bellows base
32	082-001514-00	6.07 x 1.78 O-ring	//	Bottom of the breathing system cover screw
33	082-001515-00	29.82 x 2.62 O-ring	/	APL valve
34	082-001520-00	40 x 2.2 O-ring		Auto/Manual ventilation switch
35	082-001525-00	8.5 x 2.0 O-ring	/	O2 cell port
36	082-000665-00	8.5 x 2 O-ring	/	For the rotating block of breathing circuit
37	082-000667-00	4.7 x 1.8 O-ring	/	For the rotating block of breathing circuit
38	082-000669-00	6 x 1 O-ring	/	For the Auto/Manual ventilation switch
39	082-000673-00	15.54 x 2.62 O-ring	/	For the bag alarm and O2 cell cover
40	082-000679-00	4.47 x 1.78 O-ring	/	For the CO2 bypass shaft
41	801-0631-00118-00	O-ring (for the Paw gauge)	/	For Paw gauge
42	M6M-010058---	16 x 2 O-ring	/	For the rotating block of breathing circuit
43	801-0631-00141-00	Maintenance kit for O rings	/	/