WATO EX-55Pro&65Pro(F1-&KY-)

Anesthesia Machine

Parts

MINDRAY CONFIDENTIAL

MINDRAY CONFIDENTIAL

Contents

1 Parts..... 4

1.1 Main Unit..... 4

1.1.1 Exploded View..... 4

1.1.2 Parts List..... 4

1.2 Gas Bench Receptacle Assembly..... 5

1.2.1 Exploded View..... 5

1.2.2 Parts List..... 5

1.3 AGSS Assembly..... 6

1.3.1 Exploded View..... 6

1.3.2 Parts List..... 6

1.4 Work Surface..... 7

1.4.1 Exploded View..... 7

1.4.2 Parts List..... 7

1.5 Work Table..... 9

1.5.1 Exploded View..... 9

1.5.2 Parts List..... 9

1.6 Hardware Box..... 11

1.6.1 Exploded View..... 11

1.6.2 Parts List..... 11

1.7 Display Assembly..... 14

1.7.1 Exploded View..... 14

1.7.2 Parts List..... 14

MINDRAY CONFIDENTIAL


1.8 Upper Half.....15
1.8.1 Exploded View.....15
1.8.2 Parts List.....15
1.9 EFCS Assembly.....18
1.9.1 Exploded View.....18
1.9.2 Parts List.....18
1.10 Gas Inlet Assembly.....19
1.10.1 Exploded View.....19
1.10.2 Parts List.....19
1.11 Main Body of the Breathing Circuit.....20
1.11.1 Exploded View.....20
1.11.2 Parts List.....21
1.12 Lifting Device Assembly.....23
1.12.1 Exploded View.....23
1.12.2 Parts List.....23
1.13 Pre-pak Absorb Canister Assembly.....24
1.13.1 Exploded View.....24
1.13.2 Parts List.....24
1.14 Others.....25

MINDRAY CONFIDENTIAL

1 Parts

1.1 Main Unit

1.1.1 Exploded View



1.1.2 Parts List


No.	Order Number	Part Description	Qty	Remarks
1	115-030766-00	Single vaporizer assembly(0625)	1	/

	115-032993-00	double vaporizer assembly(0625)		/
	115-027827-00	D-vaporizer assembly FRU(0635/plug-in)		/
2	031-000041-00	ball guided.3flod 350mm blue zinc-plated	6	Only 1 rail in each 031-000041-00
3	115-025663-00	drawer assembly(0625)	3	Only 1 drawer in each 115-025663-00
4	115-017947-00	suction bottle bracket assembly	1	/

MINDRAY CONFIDENTIAL

1.2 Gas Bench Receptacle Assembly

1.2.1 Exploded View




1.2.2 Parts List

No.	Order Number	Part Description	Qty	Remarks
1	115-016480-00	gas bench receptacle assembly	1	/
2	801-0621-00165-00	0621 Infrared Communication Board	1	/
3	024-000407-00	FAN 12V 40*40*20mm 6.3CFM 18dB typeF	1	/

MINDRAY CONFIDENTIAL

1.3 AGSS Assembly

1.3.1 Exploded View




1.3.2 Parts List

No.	Order Number	Part Description	Qty	Remarks
1	115-042912-00	AGSS 3 ways connector assembly	1	/

MINDRAY CONFIDENTIAL

1.4 Work Surface

1.4.1 Exploded View



1.4.2 Parts List

No.	Order Number	Part Description	Qty	Remarks
1	115-028423-00	suction control assembly(0625)	1	w/o neither Select switch of suction nor Switch of suction needle valve
2	115-040113-00	Pressure gauge assembly(O2/0627)	1	
3	115-040115-00	Pressure gauge	1	


		assembly(N2O/0627)		
4	115-040114-00	Pressure gauge assembly(Air/0627)	1	
5	115-040111-00	High pressure gauge assembly(N2O/0627)	1	
6	115-040110-00	High pressure gauge assembly(Air/0627)	1	
7	115-006954-00	system switch assembly(0631)	1	Only O2
8	010-000088-00	SWITCH Latching 1X90°form V	1	System Switch Knob
9	051-002581-00	0627 BFCS Key Board PCBA	1	
10	051-001934-00	0625 Indicator Board PCBA	1	
11	082-002459-00	Backup Oxygen Flowmeter 0~10L/min	1	
12	115-034687-00	Auxiliary O2 supply kit(EU)	1	W/ O2 knob label(EU)
	115-034688-00	Auxiliary O2 supply kit(US)		W/ O2 knob label(US)
13	115-040112-00	High pressure gauge assembly(O2/0627)	1	
14	115-033001-00	Switch of suction needle valve □0625□	1	/
15	115-033000-00	Select switch of suction□0625□	1	/
16	115-015247-00	suction needle valve assembly	1	/
17	115-024874-00	ction gauge assembly(0635)	1	/

MINDRAY CONFIDENTIAL

MINDRAY CONFIDENTIAL

1.5 Work Table

1.5.1 Exploded View



1.5.2 Parts List


No.	Order Number	Part Description	Qty	Remarks
1	115-032992-00	Gas reservoir assembly （0625）	1	/
2	115-001372-01	AVMA(0623)	1	/

3	115-037130-00	O2 flush module service kit(0619/0625)	1	/
4	115-037147-00	O2 flush button service kit(0619/0625)	1	/
5	115-061227-00	mechanical ACGO assembly(0625)	1	/
	115-033251-00	ACGO(AG/0625)		Only for machines with Built-in AG module
6	043-003125-00	HOLSTER_G_SMALLCLIP	1	/
7	115-002250-01	Cylinder Bracket for O2	1	W/O electronic pressure sensor
	115-056316-00	O2 cylinder yoke assembly(0627)		W/ electronic pressure sensor
8	115-002251-02	Cylinder Bracket for N2O(Φ6)	1	W/O electronic pressure sensor
	115-056317-00	N2O cylinder yoke assembly(0627)		W/ electronic pressure sensor
9	115-002728-01	Air cylinder yoke assembly	1	W/O electronic pressure sensor
	115-056318-00	Air cylinder yoke assembly(0627)		W/ electronic pressure sensor
10	115-031022-00	assembly of drive gas valve(0625)	1	/

MINDRAY CONFIDENTIAL

1.6 Hardware Box

1.6.1 Exploded View



1.6.2 Parts List

No.	Order Number	Part Description	Qty	Remarks
1	115-038712-00	Top Cover Assembly (0627)	1	/
2	051-001866-00	0625 toplight board PCBA	1	/
3	115-038711-00	Top Plate Asembly (0627)	1	/
4	115-066430-00	0625 Monitor PCBA FRU	1	/
5	115-040643-00	Main Board （VCV+PCV） FRU(0627)	1	For 65Pro. Factory configuration support VCV+PCV mode only. If you want the board support the same function with your AM, please provide Machine ID and Function configuration when make Order.
	115-040736-00	Main board kit (0627)		For 65Pro. Factory configuration support VCV mode only. If you want the board support the same function with your AM, please provide Machine ID and Function configuration when make Order.
	115-041386-00	Main Board		For 55Pro. Factory configuration support VCV+PCV mode only. If

MINDRAY CONFIDENTIAL

		（VCV+PCV/0628）		you want the board support the same function with your AM, please provide Machine ID and Function configuration when make Order.
	115-041387-00	Main board (0628)		For 55Pro. Factory configuration support VCV mode only. If you want the board support the same function with your AM, please provide Machine ID and Function configuration when make Order.
6	115-042876-00	three way valve asm(0625)	1	3 piece of valves, for machine w/o ACGO
	115-042877-00	three way valve asm(0625/3way)		4 piece of valves, for machine w/ ACGO
7	051-002637-00	0627 Mother Board PCBA	1	/
8	051-000768-00	0623 battery board PCBA	1	/
9	051-003468-00	0627Power Board PCBA	1	/
10	801-0631-00028-00	Fan	1	Back cover
11	051-001996-00	0625 Ventilater&Protect PCBA（Norgren）	1	/


MINDRAY CONFIDENTIAL

12	801-0631-00038-00	Speaker and Connecting Cable	1	/

MINDRAY CONFIDENTIAL

1.7 Display Assembly

1.7.1 Exploded View



1.7.2 Parts List


No.	Order Number	Part Description	Qty	Remarks
1	051-002623-00	0627 Key Board PCBA	1	/
2	115-040737-00	LCD FRU KIT	1	15" LCD
3	021-000287-00	Touchscreen 15" 5-wire resistive	1	/
4	801-0613-00035-00	Encoder(FRU)	1	W/O metal fixing plate
5	801-0631-00019-00	0631 warning light board PCBA	1	/
6	115-040641-00	Front Display Cover FRU	1	/

		□0627□		
7	043-000342-00	Knob	1	/
8	051-001867-00	0625 auxiliary lighting board PCBA	1	/
9	051-001260-00	0632 Encoder Board PCBA	1	/

MINDRAY CONFIDENTIAL

1.8 Upper Half

1.8.1 Exploded View



1.8.2 Parts List

No.	Order Number	Part Description	Qty	Remarks
1	115-040738-00	Module of auto-flowmeter FRU(O2&N2O&Air)	1	/
	115-040739-00	Module of auto-flowmeter FRU(O2&Air)		/
	115-040740-00	Module of auto-flowmeter		/

		FRU(O2&N2O)		
2	115-030786-01	O2 supply inlet assembly(NIST/0625)	1	W/O electronic pressure sensor, w/ pressure switch
	115-056320-00	O2 origin inlet module(0627/NIST)		W/ electronic pressure sensor, include 051-002521-00, w/o backup O2
	115-056326-00	O2 supply inlet assembly(NIST/0627)		W/ electronic pressure sensor, include 051-002521-00, w/ backup O2
	115-030789-01	O2 supply inlet assembly(DISS/0625)		W/O electronic pressure sensor
	115-056321-00	O2 origin inlet module(0627/DISS)		W/ electronic pressure sensor, include 051-002521-00,
3	115-040098-00	Air supply inlet (NIST/0627)	1	W/O electronic pressure sensor, w/o pressure switch
	115-039357-00	Air supply inlet (P switch/NIST/0627)		W/O electronic pressure sensor, w/ pressure switch
	115-056324-00	Air origin inlet module(0627/NIST)		W/ electronic pressure sensor, include 051-002521-00, w/o pressure switch
	115-040099-00	Air supply inlet (DISS/0627)		W/O electronic pressure sensor, w/o pressure switch
	115-040097-00	Air supply inlet (P		W/O electronic pressure

MINDRAY CONFIDENTIAL

		switch/DISS/0627)		sensor, w/ pressure switch
	115-056325-00	Air origin inlet module(0627/DISS)		W/ electronic pressure sensor, include 051-002521-00, w/o pressure switch
4	115-017748-00	vacuum generator unit	1	/
5	115-018356-00	Muffler Asembly	1	/
6	045-000032-00	Dust filter	1	/
7	115-018131-00	Overflow Unit	1	/
8	115-048529-00	Air filter (F92G2VNQT1)	1	Only for Air-compressor configuration.
9	115-034706-01	Backup O2 supply assembly(0615)	1	/
	115-004520-01	N2O supply inlet assembly(Big Cylinder)		NIST
	115-004559-01	N2O inlet assembly(DISS/Big Cylinder)		/
	115-004519-01	O2 input assembly(Big Cylinder)		NIST
	115-004558-01	O2 input assembly(DISS/Big Cylinder)		/
	115-018132-00	Vacuum pipeline supply		/


MINDRAY CONFIDENTIAL

		assembly(NIST)		
	115-018133-00	Vacuum pipeline supply assembly(DISS)		/
10	115-039268-00	N2O supply inlet assembly（NIST/0627）	1	W/O electronic pressure sensor, w/o pressure switch
	115-056322-00	N20 origin inlet module(0627/NIST)	1	W/ electronic pressure sensor, include 051-002521-00
	115-040096-00	N2O supply inlet assembly（DISS/0627）	1	W/O electronic pressure sensor, w/o pressure switch
	115-056323-00	N20 origin inlet module(0627/DISS)	1	W/ electronic pressure sensor, include 051-002521-00

MINDRAY CONFIDENTIAL

1.9 EFCS Assembly

1.9.1 Exploded View



1.9.2 Parts List


No.	Order Number	Part Description	Qty	Remarks
1	082-002567-00	Gas valve 2Pos. 3Port DC12V 2W 10bar	2	/
2	009-007236-00	0627 Proportional valve cable（valve）	2	/
3	009-007237-00	0627 NO valve for auto-flowmeter (w	1	/

		cable)		
4	115-040642-00	Auto-Flow Monitor Board FRU(0627)	1	/
5	009-007238-00	0627 NC valve for auto-flowmeter (w cable)	1	/

MINDRAY CONFIDENTIAL

1.10 Gas Inlet Assembly

1.10.1 Exploded View



1.10.2 Parts List


No.	Order Number	Part Description	Qty	Remarks
1	051-002521-00	VT70 air source press sensor PCBA	1	/

MINDRAY CONFIDENTIAL

1.11 Main Body of the Breathing Circuit

1.11.1 Exploded View

MINDRAY CONFIDENTIAL


①
②
③
④
⑤
⑥
⑦
⑧
⑨
⑩
⑪
⑫
⑬
⑭
⑮
⑯
⑰

1.11.2 Parts List

No.	Order Number	Part Description	Qty	Remarks
1	115-025563-00	Airway pressure gauge	1	/
2	115-048600-00	Bag arm assembly(silvery/spacing)	1	w/ position limit, refer to TMZ-17026
	801-0601-00005-00	Bag arm assembly		w/o position limit, refer to TMZ-17036
3	801-0631-00065-00	Bag/mechanical ventilation switch	1	/
4	801-0631-00061-00	Check valve dome, A series	2	w/ Seal(O-ring). Only one check valve dome in each 801-0631-00061-00
5	115-030842-00	Upper cover 2 assembly(0625)	1	For lifting absorb canister
	115-006547-00	upper cover 2 assembly(0631)		For pre-pak absorb canister
6	115-030839-00	Upper cover assembly(0625)	1	For lifting absorb canister
	115-030845-00	Upper cover assembly(gas return/0625)		For lifting absorb canister
	115-006544-00	upper cover assembly(0631)		For pre-pak absorb canister
7	115-048639-00	mid-plate assembly(heat/arm)FRU	1	For lifting absorb canister

MINDRAY CONFIDENTIAL

	115-048640-00	mid-plate assembly(0625/prepak)FRU		For pre-pak absorb canister
	115-048641-00	mid-plate assembly(heat/no arm)FRU		For lifting absorb canister
8	115-030840-00	Lower cover assembly(0625)	1	For lifting absorb canister
	115-030841-00	Lower cover assembly(bag connector/0625)		For lifting absorb canister, w/o bag arm
	115-009046-00	Lower Cover Assembly		For pre-pak absorb canister
9	801-0631-00102-00	O2 sensor cable module FRU	1	/
10	0611-10-45654	O2 sensor,MedicelMOX-2	1	/
11	801-0631-00059-00	Insp/Exp Connector Rotary Cap, A series	2	Only one cap in each 801-0631-00059-00
12	801-0631-00057-00	Insp/Exp Connector, A series	2	Only one connector in each 801-0631-00057-00
13	801-0631-00058-00	Water Collection Cup, A series	1	w/ Seal(O-ring)
14	115-030843-00	Lower cover 2 assembly(0625)	1	For lifting absorb canister
	115-006548-00	lover cover 2 assembly(0631)		For pre-pak absorb canister
15	801-0601-	Pop-off valve assembly	1	/


MINDRAY CONFIDENTIAL

	00002-00			
16	115-033004-00	Patient main unit(with bag arm/0625)	1	For lifting absorb canister
	115-033005-00	Patient main unit(no bag arm/0625)		For lifting absorb canister
	115-033006-00	Patient main unit(gas return/0625)		For lifting absorb canister
	801-0623-00011-00	patient main unit		For pre-pak absorb canister. W/O bag arm position limit, refer to TMZ-17026
	115-048843-00	patient main unit		For pre-pak absorb canister. W/ bag arm position limit, refer to TMZ-17026
17	801-0631-00054-00	Bellows Dome, A series	1	/

MINDRAY CONFIDENTIAL

1.12 Lifting Device Assembly

1.12.1 Exploded View



1.12.2 Parts List


No.	Order Number	Part Description	Qty	Remarks
1	115-030836-00	Lifting device(0625)	1	w/o bypass
	115-030837-00	Lifting device(with bypass/0625)		/
2	115-034689-00	Absorb canister shore(0625)	1	/

3	115-030838-00	Canister assembly(0625)	1	Only for warranty

MINDRAY CONFIDENTIAL

1.13 Pre-pak Absorb Canister Assembly

1.13.1 Exploded View



1.13.2 Parts List

No.	Order Number	Part Description	Qty	Remarks
1	801-0631-00066-00	CO2 Absorbent Canister, A series	1	/
2	801-0631-00100-00	CO2 Absorber Base, A series	1	/
3	115-006554-00	pre-pak assembly(0631)	1	/
4	801-0631-00092-00	CO2 Absorber Hose, A series	1	/
5	115-009490-00	L-shape bracket assembly(0623)	1	/

MINDRAY CONFIDENTIAL

1.14 Others

No.	Order Number	Part Description	Qty	Remarks
1	115-033008-00	USB flash memory （PCV-VG）	1	Need to provide machine's SN and machine ID
2	115-033011-00	USB flash memory （SIMV-VG）	1	Need to provide machine's SN and machine ID
3	115-033012-00	USB flash memory （PS）	1	Need to provide machine's SN and machine ID
4	115-033013-00	USB flash memory （CPAP/PS）	1	Need to provide machine's SN and machine ID
5	115-033014-00	USB flash memory （Loop）	1	Need to provide machine's SN and machine ID
6	115-036966-00	USB flash memory(SIMV-VC+SIMV-PC)	1	Need to provide machine's SN and machine ID
7	115-045689-00	Activation Code(USB/Recruitment)	1	Need to provide machine's SN and machine ID
8	082-001327-00	Disposable Filter.0.45μm	1	Used with liquid collection bottle
9	115-039569-00	Filler cap of ENF （Pour Fill）	1	/

MINDRAY CONFIDENTIAL

10	115-039571-00	Filler cap of ISO□Pour Fill□	1	/
11	115-039572-00	Filler cap of SEV □Pour Fill□	1	/
12	115-039573-00	Filler cap of SEV □Quik-Fil□	1	W/ medicine filler sleeve
13	115-039574-00	Filler cap of Hal□Pour Fill□	1	/
14	115-060430-00	0627 display module data cable FRU	1	/
15	115-060429-00	0627 alarm LED board cable FRU	1	/
16	010-000257-00	Optical Encoder 24detent 5VDC Cable	1	/
17	M07-00131F---	FUSE Time-lag 250V 3.15AD5X20	8	Only one fuse in each M07-00131F---
18	115-034694-00	Maintain Kit(0625,Basic,1 yr/Prepark)	1	/
19	115-034696-00	Maintain Kit(0625,Extend,1yr/Prepark)	1	/
20	115-034698-00	Maintain Kit(0625,Advanced,1	1	/

MINDRAY CONFIDENTIAL

		yr/Prepark)		
21	115-021121-00	Maintain Kit(0616/0625,Basic,1 yr)	1	/
22	115-021122-00	Maintain Kit(0616/0625,Extend, 1 yr)	1	/
23	115-021123-00	Maintain Kit(0616/0625,Advanced,1 yr)	1	/
24	034-000353-00	Drawer lock	1	/
25	049-000243-00	Gasket,Bellows Canister Base	1	/
26	115-034690-00	Check valve assembly □0625□	1	/
27	115-034693-00	Pressure gage label&cover kit(0625/US)	1	/
28	082-000527-00	Check valve AKH08-00	1	Only for Air-compressor configuration
29	115-034692-00	Pressure gage label&cover kit(0625/EU)	1	/
30	115-051204-00	Breathing system FRU(O2 drive)	1	115-066430-00 (0625 Monitor PCBA FRU) and 115-001372-01(AVMA(0623)) are included

MINDRAY CONFIDENTIAL

	115-051203-00	Breathing system FRU(Air drive)		115-066430-00 (0625 Monitor PCBA FRU) and 115-001372-01(AVMA(0623)) are included
31	803-011600-00	WATO EX-65 Pro main unit warranty	1	/
32	803-011601-00	WATO EX-55 Pro main unit warranty	1	/

MINDRAY CONFIDENTIAL