

Calibration Certificate

P - 17 F03-00

Certificate/ULR No.: CALCERT-20260702-6F5FF2		Page 1 of 1	
Customer Details		Instrument Receipt Date	1-Jul-2026
M/S. ABC Industries Pvt. Ltd. Plot 45, Sector 62, Noida, Uttar Pradesh - 201309		SRF No.	JOB-20260701-001
		Date of Calibration	1-Jul-2026
		Next Calibration Date	30-Jun-2027
		Certificate Issue Date	1-Jul-2026
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Digital Thermometer	Make / Model No.	Multi/ST-9283B
Range	-50 to 300 deg C/-58F to 572F	Instrument Condition	OK
Least Count	0.1 deg C/0.1F	Location of Calibration	At Laboratory
Serial No.	SN-784521	Location	Bisauli 154+200 LHS
Instrument ID.	INS-2026-001	Parameter	Thermal
ENVIRONMENTAL CONDITION	Temperature: 23.4 degC, Relative Humidity: 48%	Calibration Reference Standard	ITS-90
		Calibration Procedure	TNC-06-T 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
RTD Sensor with 6 1/2 Digital Multimeter	TNC/T/C/005/17	CC317223000000132F	RK labs	24/01/2024

Calibration Results

Sr.No.	UUC Value (°C)	Standard Value (°C)	Deviation (°C)	Uncertainty at k=2 at 95 % Confidence Level (±) °C
1	-30.0	-29.99	-0.01	0.117774
2	30.0	30.01	-0.01	0.117803
3	100.0	100	0	0.34506
4	200.0	200	0	0.682544
5	300.0	299.99	0.01	1.021698

Note: -

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated. Total liability of Testncal Laboratory. will be limited to the invoiced amount only.



Calibrated By

Issued & Approved By