

Calibration Certificate

P-17 F03-00

Certificate/ULR No.:	CC205523000011047F	Page 1 of 1
----------------------	--------------------	-------------

Customer Details M/S H.G. INFRA ENGINEERING LTD DMRC Project, DC-01A Casting Yard Mundka, Delhi	Instrument Receipt Date	16-Sep-2023
	SRF No.	TNC1/JC/23-24/1099
	Date of Calibration	16-Sep-2023
	Next Calibration Date	15-Sep-2024
	Certificate Issue Date	18-Sep-2023

A) Identification On UUC (Unit Under Calibration)

Instrument Name	Dig. Earth Resistance Tester	Make / Model No.	Waco/-
Range	10 Ω to 1000 Ω	Instrument Condition	OK
Least Count	As per Range	Location of Calibration	At Laboratory
Serial No.	420319	Location	-
Instrument ID.	-	Parameter	Electro Technical
ENVIRONMENTAL CONDITION	Temperature:- 24 $\pm$ 2°C ,Relative Humidity: - 55 $\pm$ 10%	Calibration Reference Standard	IS:1248-2003
		Calibration Procedure	TNC -03-WI 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
6 ½ Digit DMM	TNC/ET/C/001/15	CC202023000000237F	Anshakan India Pvt.Ltd.	02/02/2024

Resistance: Range

Sl. No	Range	UOM	Observed Value on UUC (Ω)	Observed Value on Standard	Uncertainty at k=2 $\pm$ %
1	10 Ω to 200 Ω	Ω	20	19.9989	0.00
2		Ω	100	99.9921	0.00
3		Ω	199	198.9927	0.00
4	200 to 1000 Ω	Ω	300	299.993	0.00
5		Ω	500	499.9956	0.00
6		k Ω	999	0.998972	0.00

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC205523000011049F		Page 1 of 1	
Customer Details		Instrument Receipt Date	16-Sep-2023
M/S H.G. INFRA ENGINEERING LTD DMRC Project, DC-01A Casting Yard Mundka, Delhi		SRF No.	TNC1/JC/23-24/1099
		Date of Calibration	16-Sep-2023
		Next Calibration Date	15-Sep-2024
		Certificate Issue Date	18-Sep-2023
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	RCD Tester	Make / Model No.	Metravi/1813
Range	As per Manual	Instrument Condition	OK
Least Count	As per Range	Location of Calibration	At Laboratory
Serial No.	220813564	Location	-
Instrument ID.	-	Parameter	Electro Technical
ENVIRONMENTAL CONDITION	Temperature:- 24±2 °C,Relative Humidity: - 55±10%	Calibration Reference Standard	IS:1248-2003
		Calibration Procedure	TNC -03-WI 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
6 ½ Digit DMM	TNC/ET/C/001/15	CC202023000000237F	Anshakan India Pvt.Ltd.	02/02/2024
Multifunction Calibrator	TNC/Cal/ET/015/017	CC205523000000426F	TESTNCAL	12/08/2024

Calibration Results(AC Current @ 50Hz

Sl. No	Range	UOM	Observed Value on UUC	Observed Value on Standard	Uncertainty at k=2 ±%
1	10mA	mA	10	9.99959	0.30
2	20mA	mA	20	19.9838	0.30
3	30mA	mA	30	29.9934	0.30
4	100mA	mA	100	99.983	0.58
5	300mA	mA	300	299.979	0.58
6	500mA	mA	500	499.897	0.58

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC205523000011050F		Page 1 of 1	
Customer Details		Instrument Receipt Date	16-Sep-2023
M/S H.G. INFRA ENGINEERING LTD DMRC Project, DC-01A Casting Yard Mundka, Delhi		SRF No.	TNC1/JC/23-24/1099
		Date of Calibration	16-Sep-2023
		Next Calibration Date	15-Sep-2024
		Certificate Issue Date	18-Sep-2023
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	RCD Tester	Make / Model No.	Metravi/1813
Range	As per Manual	Instrument Condition	OK
Least Count	As per Range	Location of Calibration	At Laboratory
Serial No.	220813602	Location	-
Instrument ID.	-	Parameter	Electro Technical
ENVIRONMENTAL CONDITION	Temperature:- 24±2 °C,Relative Humidity: - 55±10%	Calibration Reference Standard	IS:1248-2003
		Calibration Procedure	TNC -03-WI 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
6 ½ Digit DMM	TNC/ET/C/001/15	CC202023000000237F	Anshakan India Pvt.Ltd.	02/02/2024
Multifunction Calibrator	TNC/Cal/ET/015/017	CC205523000000426F	TESTNCAL	12/08/2024

Calibration Results(AC Current @ 50Hz

Sl. No	Range	UOM	Observed Value on UUC	Observed Value on Standard	Uncertainty at k=2 ±%
1	10mA	mA	10	9.99955	0.30
2	20mA	mA	20	19.9832	0.30
3	30mA	mA	30	29.9924	0.30
4	100mA	mA	100	99.979	0.58
5	300mA	mA	300	299.983	0.58
6	500mA	mA	500	499.917	0.58

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

Certificate/ULR No.:	CC205523000012094F	Page 1 of 1
----------------------	--------------------	-------------

Customer Details	Instrument Receipt Date	1-Oct-2023
M/S. H.G. INFRA ENGINEERING LTD.. Vill- Papad Hamjapur (Khasra No-422) Near Brahamdev Tample Teh. & P.O. Dataganj Dist.- Badaun,, Uttar Pradesh Pin Code : 243635	SRF No.	TNC1/JC/23-24/1206
	Date of Calibration	1-Oct-2023
	Next Calibration Date	30-Sep-2024
	Certificate Issue Date	3-Oct-2023

A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Test Sieve (GI 450mm Dia)	Make / Model No.	Sethi/-
Range	63 mm	Instrument Condition	OK
Least Count	-	Location of Calibration	At Laboratory
Serial No.	538	Location	208+000 LHS Dataganj
Instrument ID.	HG/TS/63mm/01	Parameter	Dimension
ENVIRONMENTAL CONDITION	Temperature:- 20±2°C, Relative Humidity: - 50±10%	Calibration Reference Standard	IS:460:1985
		Calibration Procedure	TNC -03-D 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Digital Vernier Caliper	0511201536	CC205523000000003F	Testncal Lanoratory	01/01/2024

Calibration Results

Standard value in mm	Obs 1	Obs 2	Obs 3	Obs 4	Obs 5	Average in mm	Uncertainty of Measurement at k = 2 at 95 % Confidence Level : ± µm
63	62.91	62.89	62.90	62.90	62.91	62.90	13.5

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

Certificate/ULR No.: CC205523000012095F		Page 1 of 1	
Customer Details		Instrument Receipt Date	1-Oct-2023
M/S. H.G. INFRA ENGINEERING LTD.. Vill- Papad Hamjapur (Khasra No-422) Near Brahamdev Tample Teh. & P.O. Dataganj Dist.- Badaun,, Uttar Pradesh Pin Code : 243635		SRF No.	TNC1/JC/23-24/1206
		Date of Calibration	1-Oct-2023
		Next Calibration Date	30-Sep-2024
		Certificate Issue Date	3-Oct-2023
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Test Sieve (GI 450mm Dia)	Make / Model No.	Allied Sales Corp./-
Range	45 mm	Instrument Condition	OK
Least Count	-	Location of Calibration	At Laboratory
Serial No.	-	Location	208+000 LHS Dataganj
Instrument ID.	HG/TS/45mm/01	Parameter	Dimension
ENVIRONMENTAL CONDITION	Temperature:- 20±2°C, Relative Humidity: - 50±10%	Calibration Reference Standard	IS:460:1985
		Calibration Procedure	TNC -03-D 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Digital Vernier Caliper	0511201536	CC205523000000003F	Testncal Lanoratory	01/01/2024

Calibration Results

Standard value in mm	Obs 1	Obs 2	Obs 3	Obs 4	Obs 5	Average in mm	Uncertainty of Measurement at k = 2 at 95 % Confidence Level : ± µm
45	45.12	45.12	45.11	45.12	45.13	45.12	13.5

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

Certificate/ULR No.:	CC205523000012096F	Page 1 of 1
----------------------	--------------------	-------------

Customer Details		Instrument Receipt Date	1-Oct-2023
M/S. H.G. INFRA ENGINEERING LTD.. Vill- Papad Hamjapur (Khasra No-422) Near Brahamdev Tample Teh. & P.O. Dataganj Dist.- Badaun,, Uttar Pradesh Pin Code : 243635		SRF No.	TNC1/JC/23-24/1206
		Date of Calibration	1-Oct-2023
		Next Calibration Date	30-Sep-2024
		Certificate Issue Date	3-Oct-2023

A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Test Sieve (GI 450mm Dia)	Make / Model No.	Allied Sales Corp./-
Range	37.5 mm	Instrument Condition	OK
Least Count	-	Location of Calibration	At Laboratory
Serial No.	-	Location	208+000 LHS Dataganj
Instrument ID.	HG/TS/37.5mm/01	Parameter	Dimension
ENVIRONMENTAL CONDITION	Temperature:- 20±2°C, Relative Humidity: - 50±10%	Calibration Reference Standard	IS:460:1985
		Calibration Procedure	TNC -03-D 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Digital Vernier Caliper	0511201536	CC205523000000003F	Testncal Lanoratory	01/01/2024

Calibration Results

Standard value in mm	Obs 1	Obs 2	Obs 3	Obs 4	Obs 5	Average in mm	Uncertainty of Measurement at k = 2 at 95 % Confidence Level : ± µm
37.5	37.57	37.57	37.56	37.59	37.58	37.57	13.5

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

Certificate/ULR No.:	CC205523000012098F	Page 1 of 1
----------------------	--------------------	-------------

Customer Details	Instrument Receipt Date	1-Oct-2023
M/S. H.G. INFRA ENGINEERING LTD.. Vill- Papad Hamjapur (Khasra No-422) Near Brahamdev Tample Teh. & P.O. Dataganj Dist.- Badaun,, Uttar Pradesh Pin Code : 243635	SRF No.	TNC1/JC/23-24/1206
	Date of Calibration	1-Oct-2023
	Next Calibration Date	30-Sep-2024
	Certificate Issue Date	3-Oct-2023

A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Test Sieve (GI 450mm Dia)	Make / Model No.	Allied Sales Corp./-
Range	31.5 mm	Instrument Condition	OK
Least Count	-	Location of Calibration	At Laboratory
Serial No.	-	Location	208+000 LHS Dataganj
Instrument ID.	HG/TS/31.5mm/01	Parameter	Dimension
ENVIRONMENTAL CONDITION	Temperature:- 20±2°C, Relative Humidity: - 50±10%	Calibration Reference Standard	IS:460:1985
		Calibration Procedure	TNC -03-D 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Digital Vernier Caliper	0511201536	CC205523000000003F	Testncal Lanoratory	01/01/2024

Calibration Results

Standard value in mm	Obs 1	Obs 2	Obs 3	Obs 4	Obs 5	Average in mm	Uncertainty of Measurement at k = 2 at 95 % Confidence Level : ± µm
31.5	31.58	31.59	31.58	31.59	31.57	31.58	13.5

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

Certificate/ULR No.:	CC205523000012101F	Page 1 of 1
----------------------	--------------------	-------------

Customer Details		Instrument Receipt Date	1-Oct-2023
M/S. H.G. INFRA ENGINEERING LTD.. Vill- Papad Hamjapur (Khasra No-422) Near Brahamdev Tample Teh. & P.O. Dataganj Dist.- Badaun,, Uttar Pradesh Pin Code : 243635		SRF No.	TNC1/JC/23-24/1206
		Date of Calibration	1-Oct-2023
		Next Calibration Date	30-Sep-2024
		Certificate Issue Date	3-Oct-2023

A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Test Sieve (450mm Dia)	Make / Model No.	Allied Sales Corp./-
Range	40 mm	Instrument Condition	OK
Least Count	-	Location of Calibration	At Laboratory
Serial No.	-	Location	208+000 LHS Dataganj
Instrument ID.	HG/TS/40mm/01	Parameter	Dimension
ENVIRONMENTAL CONDITION	Temperature:- 20±2°C, Relative Humidity: - 50±10%	Calibration Reference Standard	IS:460:1985
		Calibration Procedure	TNC -03-D 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Digital Vernier Caliper	0511201536	CC205523000000003F	Testncal Lanoratory	01/01/2024

Calibration Results

Standard value in mm	Obs 1	Obs 2	Obs 3	Obs 4	Obs 5	Average in mm	Uncertainty of Measurement at k = 2 at 95 % Confidence Level : ± µm
40	39.87	39.88	39.88	39.89	39.87	39.88	13.5

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

Certificate/ULR No.:	CC205523000012103F	Page 1 of 1
----------------------	--------------------	-------------

Customer Details		Instrument Receipt Date	1-Oct-2023
M/S. H.G. INFRA ENGINEERING LTD.. Vill- Papad Hamjapur (Khasra No-422) Near Brahamdev Tample Teh. & P.O. Dataganj Dist.- Badaun,, Uttar Pradesh Pin Code : 243635		SRF No.	TNC1/JC/23-24/1206
		Date of Calibration	1-Oct-2023
		Next Calibration Date	30-Sep-2024
		Certificate Issue Date	3-Oct-2023

A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Test Sieve (450mm Dia)	Make / Model No.	Allied Sales Corp./-
Range	25 mm	Instrument Condition	OK
Least Count	-	Location of Calibration	At Laboratory
Serial No.	-	Location	208+000 LHS Dataganj
Instrument ID.	HG/TS/25mm/01	Parameter	Dimension
ENVIRONMENTAL CONDITION	Temperature:- 20±2°C, Relative Humidity: - 50±10%	Calibration Reference Standard	IS:460:1985
		Calibration Procedure	TNC -03-D 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Digital Vernier Caliper	0511201536	CC205523000000003F	Testncal Lanoratory	01/01/2024

Calibration Results

Standard value in mm	Obs 1	Obs 2	Obs 3	Obs 4	Obs 5	Average in mm	Uncertainty of Measurement at k = 2 at 95 % Confidence Level : ± µm
25	24.91	24.93	24.94	24.92	24.93	24.93	13.5

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

Certificate/ULR No.:	CC205523000012104F	Page 1 of 1
----------------------	--------------------	-------------

Customer Details		Instrument Receipt Date	1-Oct-2023
M/S. H.G. INFRA ENGINEERING LTD.. Vill- Papad Hamjapur (Khasra No-422) Near Brahamdev Tample Teh. & P.O. Dataganj Dist.- Badaun,, Uttar Pradesh Pin Code : 243635		SRF No.	TNC1/JC/23-24/1206
		Date of Calibration	1-Oct-2023
		Next Calibration Date	30-Sep-2024
		Certificate Issue Date	3-Oct-2023

A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Test Sieve (450mm Dia)	Make / Model No.	Allied Sales Corp./-
Range	25 mm	Instrument Condition	OK
Least Count	-	Location of Calibration	At Laboratory
Serial No.	-	Location	208+000 LHS Dataganj
Instrument ID.	HG/TS/25mm/02	Parameter	Dimension
ENVIRONMENTAL CONDITION	Temperature:- 20±2°C, Relative Humidity: - 50±10%	Calibration Reference Standard	IS:460:1985
		Calibration Procedure	TNC -03-D 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Digital Vernier Caliper	0511201536	CC205523000000003F	Testncal Lanoratory	01/01/2024

Calibration Results

Standard value in mm	Obs 1	Obs 2	Obs 3	Obs 4	Obs 5	Average in mm	Uncertainty of Measurement at k = 2 at 95 % Confidence Level : ± µm
25	24.92	24.91	24.93	24.93	24.92	24.92	13.5

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

Certificate/ULR No.:	CC205523000012105F	Page 1 of 1
----------------------	--------------------	-------------

Customer Details		Instrument Receipt Date	1-Oct-2023
M/S. H.G. INFRA ENGINEERING LTD.. Vill- Papad Hamjapur (Khasra No-422) Near Brahamdev Tample Teh. & P.O. Dataganj Dist.- Badaun,, Uttar Pradesh Pin Code : 243635		SRF No.	TNC1/JC/23-24/1206
		Date of Calibration	1-Oct-2023
		Next Calibration Date	30-Sep-2024
		Certificate Issue Date	3-Oct-2023

A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Test Sieve (450mm Dia)	Make / Model No.	Allied Sales Corp./-
Range	19 mm	Instrument Condition	OK
Least Count	-	Location of Calibration	At Laboratory
Serial No.	-	Location	208+000 LHS Dataganj
Instrument ID.	HG/TS/19mm/01	Parameter	Dimension
ENVIRONMENTAL CONDITION	Temperature:- 20±2°C, Relative Humidity: - 50±10%	Calibration Reference Standard	IS:460:1985
		Calibration Procedure	TNC -03-D 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Digital Vernier Caliper	0511201536	CC205523000000003F	Testncal Lanoratory	01/01/2024

Calibration Results

Standard value in mm	Obs 1	Obs 2	Obs 3	Obs 4	Obs 5	Average in mm	Uncertainty of Measurement at k = 2 at 95 % Confidence Level : ± µm
19	18.92	18.93	18.91	18.92	18.91	18.92	13.5

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

Certificate/ULR No.:	CC205523000012106F	Page 1 of 1
----------------------	--------------------	-------------

Customer Details	Instrument Receipt Date	1-Oct-2023
M/S. H.G. INFRA ENGINEERING LTD.. Vill- Papad Hamjapur (Khasra No-422) Near Brahamdev Tample Teh. & P.O. Dataganj Dist.- Badaun,, Uttar Pradesh Pin Code : 243635	SRF No.	TNC1/JC/23-24/1206
	Date of Calibration	1-Oct-2023
	Next Calibration Date	30-Sep-2024
	Certificate Issue Date	3-Oct-2023

A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Test Sieve (450mm Dia)	Make / Model No.	Allied Sales Corp./-
Range	19 mm	Instrument Condition	OK
Least Count	-	Location of Calibration	At Laboratory
Serial No.	-	Location	208+000 LHS Dataganj
Instrument ID.	HG/TS/19mm/02	Parameter	Dimension
ENVIRONMENTAL CONDITION	Temperature:- 20±2°C, Relative Humidity: - 50±10%	Calibration Reference Standard	IS:460:1985
		Calibration Procedure	TNC -03-D 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Digital Vernier Caliper	0511201536	CC205523000000003F	Testncal Lanoratory	01/01/2024

Calibration Results

Standard value in mm	Obs 1	Obs 2	Obs 3	Obs 4	Obs 5	Average in mm	Uncertainty of Measurement at k = 2 at 95 % Confidence Level : ± µm
19	18.94	18.95	18.96	18.95	18.96	18.95	13.5

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

Certificate/ULR No.:	CC205523000012107F	Page 1 of 1
----------------------	--------------------	-------------

Customer Details	Instrument Receipt Date	1-Oct-2023
M/S. H.G. INFRA ENGINEERING LTD.. Vill- Papad Hamjapur (Khasra No-422) Near Brahamdev Tample Teh. & P.O. Dataganj Dist.- Badaun,, Uttar Pradesh Pin Code : 243635	SRF No.	TNC1/JC/23-24/1206
	Date of Calibration	1-Oct-2023
	Next Calibration Date	30-Sep-2024
	Certificate Issue Date	3-Oct-2023

A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Test Sieve (450mm Dia)	Make / Model No.	Allied Sales Corp./-
Range	20 mm	Instrument Condition	OK
Least Count	-	Location of Calibration	At Laboratory
Serial No.	-	Location	208+000 LHS Dataganj
Instrument ID.	HG/TS/20mm/01	Parameter	Dimension
ENVIRONMENTAL CONDITION	Temperature:- 20±2°C, Relative Humidity: - 50±10%	Calibration Reference Standard	IS:460:1985
		Calibration Procedure	TNC -03-D 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Digital Vernier Caliper	0511201536	CC205523000000003F	Testncal Lanoratory	01/01/2024

Calibration Results

Standard value in mm	Obs 1	Obs 2	Obs 3	Obs 4	Obs 5	Average in mm	Uncertainty of Measurement at k = 2 at 95 % Confidence Level : ± µm
20	20.11	20.13	20.12	20.11	20.12	20.12	13.5

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

Certificate/ULR No.:	CC205523000012108F	Page 1 of 1
----------------------	--------------------	-------------

Customer Details		Instrument Receipt Date	1-Oct-2023
M/S. H.G. INFRA ENGINEERING LTD.. Vill- Papad Hamjapur (Khasra No-422) Near Brahamdev Tample Teh. & P.O. Dataganj Dist.- Badaun,, Uttar Pradesh Pin Code : 243635		SRF No.	TNC1/JC/23-24/1206
		Date of Calibration	1-Oct-2023
		Next Calibration Date	30-Sep-2024
		Certificate Issue Date	3-Oct-2023

A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Test Sieve (450mm Dia)	Make / Model No.	Allied Sales Corp./-
Range	16 mm	Instrument Condition	OK
Least Count	-	Location of Calibration	At Laboratory
Serial No.	-	Location	208+000 LHS Dataganj
Instrument ID.	HG/TS/16mm/01	Parameter	Dimension
ENVIRONMENTAL CONDITION	Temperature:- 20±2°C, Relative Humidity: - 50±10%	Calibration Reference Standard	IS:460:1985
		Calibration Procedure	TNC -03-D 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Digital Vernier Caliper	0511201536	CC205523000000003F	Testncal Lanoratory	01/01/2024

Calibration Results

Standard value in mm	Obs 1	Obs 2	Obs 3	Obs 4	Obs 5	Average in mm	Uncertainty of Measurement at k = 2 at 95 % Confidence Level : ± µm
16	15.89	15.90	15.88	15.89	15.91	15.89	13.5

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national /International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

Certificate/ULR No.:	CC205523000012109F	Page 1 of 1
----------------------	--------------------	-------------

Customer Details		Instrument Receipt Date	1-Oct-2023
M/S. H.G. INFRA ENGINEERING LTD.. Vill- Papad Hamjapur (Khasra No-422) Near Brahamdev Tample Teh. & P.O. Dataganj Dist.- Badaun,, Uttar Pradesh Pin Code : 243635		SRF No.	TNC1/JC/23-24/1206
		Date of Calibration	1-Oct-2023
		Next Calibration Date	30-Sep-2024
		Certificate Issue Date	3-Oct-2023

A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Test Sieve (450mm Dia)	Make / Model No.	Allied Sales Corp./-
Range	13.2 mm	Instrument Condition	OK
Least Count	-	Location of Calibration	At Laboratory
Serial No.	-	Location	208+000 LHS Dataganj
Instrument ID.	HG/TS/13.2mm/01	Parameter	Dimension
ENVIRONMENTAL CONDITION	Temperature:- 20±2°C, Relative Humidity: - 50±10%	Calibration Reference Standard	IS:460:1985
		Calibration Procedure	TNC -03-D 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Digital Vernier Caliper	0511201536	CC205523000000003F	Testncal Lanoratory	01/01/2024

Calibration Results

Standard value in mm	Obs 1	Obs 2	Obs 3	Obs 4	Obs 5	Average in mm	Uncertainty of Measurement at k = 2 at 95 % Confidence Level : ± µm
13.2	13.29	13.28	13.27	13.28	13.28	13.28	13.5

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

Certificate/ULR No.:	CC205523000012112F	Page 1 of 1
----------------------	--------------------	-------------

Customer Details	Instrument Receipt Date	1-Oct-2023
M/S. H.G. INFRA ENGINEERING LTD.. Vill- Papad Hamjapur (Khasra No-422) Near Brahamdev Tample Teh. & P.O. Dataganj Dist.- Badaun,, Uttar Pradesh Pin Code : 243635	SRF No.	TNC1/JC/23-24/1206
	Date of Calibration	1-Oct-2023
	Next Calibration Date	30-Sep-2024
	Certificate Issue Date	3-Oct-2023

A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Test Sieve (450mm Dia)	Make / Model No.	Allied Sales Corp./-
Range	11.2 mm	Instrument Condition	OK
Least Count	-	Location of Calibration	At Laboratory
Serial No.	-	Location	208+000 LHS Dataganj
Instrument ID.	HG/TS/11.2mm/01	Parameter	Dimension
ENVIRONMENTAL CONDITION	Temperature:- 20±2°C, Relative Humidity: - 50±10%	Calibration Reference Standard	IS:460:1985
		Calibration Procedure	TNC -03-D 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Digital Vernier Caliper	0511201536	CC205523000000003F	Testncal Lanoratory	01/01/2024

Calibration Results

Standard value in mm	Obs 1	Obs 2	Obs 3	Obs 4	Obs 5	Average in mm	Uncertainty of Measurement at k = 2 at 95 % Confidence Level : ± µm
11.2	10.97	10.99	10.98	10.97	10.96	10.97	13.5

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

Certificate/ULR No.: CC205523000012115F Page 1 of 1

Customer Details		Instrument Receipt Date	1-Oct-2023
M/S. H.G. INFRA ENGINEERING LTD.. Vill- Papad Hamjapur (Khasra No-422) Near Brahamdev Tample Teh. & P.O. Dataganj Dist.- Badaun,, Uttar Pradesh Pin Code : 243635		SRF No.	TNC1/JC/23-24/1206
		Date of Calibration	1-Oct-2023
		Next Calibration Date	30-Sep-2024
		Certificate Issue Date	3-Oct-2023

A) Identification On UUC (Unit Under Calibration)

Instrument Name	Test Sieve (450mm Dia)	Make / Model No.	Allied Sales Corp./-
Range	9.5 mm	Instrument Condition	OK
Least Count	-	Location of Calibration	At Laboratory
Serial No.	-	Location	208+000 LHS Dataganj
Instrument ID.	HG/TS/9.5mm/01	Parameter	Dimension
ENVIRONMENTAL CONDITION	Temperature:- 20±2°C, Relative Humidity: - 50±10%	Calibration Reference Standard	IS:460:1985
		Calibration Procedure	TNC -03-D 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Digital Vernier Caliper	0511201536	CC205523000000003F	Testncal Lanoratory	01/01/2024

Calibration Results

Standard value in mm	Obs 1	Obs 2	Obs 3	Obs 4	Obs 5	Average in mm	Uncertainty of Measurement at k = 2 at 95 % Confidence Level : ± µm
9.5	9.66	9.63	9.64	9.66	9.67	9.65	13.5

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

Certificate/ULR No.:	CC205523000012118F	Page 1 of 1
----------------------	--------------------	-------------

Customer Details	Instrument Receipt Date	1-Oct-2023
M/S. H.G. INFRA ENGINEERING LTD.. Vill- Papad Hamjapur (Khasra No-422) Near Brahamdev Tample Teh. & P.O. Dataganj Dist.- Badaun,, Uttar Pradesh Pin Code : 243635	SRF No.	TNC1/JC/23-24/1206
	Date of Calibration	1-Oct-2023
	Next Calibration Date	30-Sep-2024
	Certificate Issue Date	3-Oct-2023

A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Test Sieve (450mm Dia)	Make / Model No.	Allied Sales Corp./-
Range	9.5 mm	Instrument Condition	OK
Least Count	-	Location of Calibration	At Laboratory
Serial No.	-	Location	208+000 LHS Dataganj
Instrument ID.	HG/TS/9.5mm/02	Parameter	Dimension
ENVIRONMENTAL CONDITION	Temperature:- 20±2°C, Relative Humidity: - 50±10%	Calibration Reference Standard	IS:460:1985
		Calibration Procedure	TNC -03-D 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Digital Vernier Caliper	0511201536	CC205523000000003F	Testncal Lanoratory	01/01/2024

Calibration Results

Standard value in mm	Obs 1	Obs 2	Obs 3	Obs 4	Obs 5	Average in mm	Uncertainty of Measurement at k = 2 at 95 % Confidence Level : ± µm
9.5	9.69	9.67	9.66	9.68	9.67	9.67	13.5

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

Certificate/ULR No.:	CC205523000012121F	Page 1 of 1
----------------------	--------------------	-------------

Customer Details	Instrument Receipt Date	1-Oct-2023
M/S. H.G. INFRA ENGINEERING LTD.. Vill- Papad Hamjapur (Khasra No-422) Near Brahamdev Tample Teh. & P.O. Dataganj Dist.- Badaun,, Uttar Pradesh Pin Code : 243635	SRF No.	TNC1/JC/23-24/1206
	Date of Calibration	1-Oct-2023
	Next Calibration Date	30-Sep-2024
	Certificate Issue Date	3-Oct-2023

A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Test Sieve (450mm Dia)	Make / Model No.	Allied Sales Corp./-
Range	6.3 mm	Instrument Condition	OK
Least Count	-	Location of Calibration	At Laboratory
Serial No.	-	Location	208+000 LHS Dataganj
Instrument ID.	HG/TS/6.3mm/01	Parameter	Dimension
ENVIRONMENTAL CONDITION	Temperature:- 20±2°C, Relative Humidity: - 50±10%	Calibration Reference Standard	IS:460:1985
		Calibration Procedure	TNC -03-D 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Digital Vernier Caliper	0511201536	CC205523000000003F	Testncal Lanoratory	01/01/2024

Calibration Results

Standard value in mm	Obs 1	Obs 2	Obs 3	Obs 4	Obs 5	Average in mm	Uncertainty of Measurement at k = 2 at 95 % Confidence Level : ± µm
6.3	6.21	6.23	6.24	6.22	6.23	6.23	13.5

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

Certificate/ULR No.: CC205523000012123F		Page 1 of 1	
Customer Details		Instrument Receipt Date	1-Oct-2023
M/S. H.G. INFRA ENGINEERING LTD.. Vill- Papad Hamjapur (Khasra No-422) Near Brahamdev Tample Teh. & P.O. Dataganj Dist.- Badaun,, Uttar Pradesh Pin Code : 243635		SRF No.	TNC1/JC/23-24/1206
		Date of Calibration	1-Oct-2023
		Next Calibration Date	30-Sep-2024
		Certificate Issue Date	3-Oct-2023
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Test Sieve (450mm Dia)	Make / Model No.	Allied Sales Corp./-
Range	5.6 mm	Instrument Condition	OK
Least Count	-	Location of Calibration	At Laboratory
Serial No.	-	Location	208+000 LHS Dataganj
Instrument ID.	HG/TS/5.6mm/01	Parameter	Dimension
ENVIRONMENTAL CONDITION	Temperature:- 20±2°C, Relative Humidity: - 50±10%	Calibration Reference Standard	IS:460:1985
		Calibration Procedure	TNC -03-D 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Digital Vernier Caliper	0511201536	CC205523000000003F	Testncal Lanoratory	01/01/2024

Calibration Results

Standard value in mm	Obs 1	Obs 2	Obs 3	Obs 4	Obs 5	Average in mm	Uncertainty of Measurement at k = 2 at 95 % Confidence Level : ± µm
5.6	5.49	5.51	5.48	5.49	5.50	5.49	13.5

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

Certificate/ULR No.: CC205523000012125F Page 1 of 1

Customer Details		Instrument Receipt Date	1-Oct-2023
M/S. H.G. INFRA ENGINEERING LTD.. Vill- Papad Hamjapur (Khasra No-422) Near Brahamdev Tample Teh. & P.O. Dataganj Dist.- Badaun,, Uttar Pradesh Pin Code : 243635		SRF No.	TNC1/JC/23-24/1206
		Date of Calibration	1-Oct-2023
		Next Calibration Date	30-Sep-2024
		Certificate Issue Date	3-Oct-2023

A) Identification On UUC (Unit Under Calibration)

Instrument Name	Test Sieve (450mm Dia)	Make / Model No.	Allied Sales Corp./-
Range	4.75 mm	Instrument Condition	OK
Least Count	-	Location of Calibration	At Laboratory
Serial No.	-	Location	208+000 LHS Dataganj
Instrument ID.	HG/TS/4.75mm/01	Parameter	Dimension
ENVIRONMENTAL CONDITION	Temperature:- 20±2°C, Relative Humidity: - 50±10%	Calibration Reference Standard	IS:460:1985
		Calibration Procedure	TNC -03-D 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Digital Vernier Caliper	0511201536	CC205523000000003F	Testncal Lanoratory	01/01/2024

Calibration Results

Standard value in mm	Obs 1	Obs 2	Obs 3	Obs 4	Obs 5	Average in mm	Uncertainty of Measurement at k = 2 at 95 % Confidence Level : ± µm
4.75	4.78	4.77	4.78	4.79	4.77	4.78	13.5

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P - 17 F03-00

Certificate/ULR No.: CC205523000012128F		Page 1 of 1	
Customer Details		Instrument Receipt Date	1-Oct-2023
M/S. H.G. INFRA ENGINEERING LTD.. Vill- Papad Hamjapur (Khasra No-422) Near Brahamdev Tample Teh. & P.O. Dataganj Dist.- Badaun,, Uttar Pradesh Pin Code : 243635		SRF No.	TNC1/JC/23-24/1206
		Date of Calibration	1-Oct-2023
		Next Calibration Date	30-Sep-2024
		Certificate Issue Date	3-Oct-2023
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Hot Air Oven	Make / Model No.	I-therm/AI-7781
Range	0 to 200 °C	Instrument Condition	OK
Least Count	0.1 °C	Location of Calibration	At Site
Serial No.	-	Location	208+000 LHS Dataganj
Instrument ID.	HG/Oven/01	Parameter	Thermal
ENVIRONMENTAL CONDITION	Temperature:- 25±2 °C,Relative Humidity: - 50±10%	Calibration Reference Standard	ITS-90
		Calibration Procedure	TNC-06-T 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
RTD Sensor with 6 1/2 Digital Multimeter	TNC/T/C/005/17	CC317223000000132F	RK labs	24/01/2024

Calibration Results

Sr.No.	UUC Value (°C)	Standard Value (°C)	Deviation (°C)	Uncertainty at k=2 at 95 % Confidence Level (±) °C
1	50	50.13	0.13	0.48
2	70	70.19	0.19	0.48
3	100	100.27	0.27	0.48
4	150	150.35	0.35	0.48
5	200	200.46	0.46	0.48

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

Certificate/ULR No.:	CC205523000012129F	Page 1 of 1
----------------------	--------------------	-------------

Customer Details	Instrument Receipt Date	1-Oct-2023
M/S. H.G. INFRA ENGINEERING LTD.. Vill- Papad Hamjapur (Khasra No-422) Near Brahamdev Tample Teh. & P.O. Dataganj Dist.- Badaun,, Uttar Pradesh Pin Code : 243635	SRF No.	TNC1/JC/23-24/1206
	Date of Calibration	1-Oct-2023
	Next Calibration Date	30-Sep-2024
	Certificate Issue Date	3-Oct-2023

A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Brass Sieve (200mm Dia)	Make / Model No.	Allied Sales Corp./-
Range	10 mm	Instrument Condition	OK
Least Count	-	Location of Calibration	At Laboratory
Serial No.	-	Location	208+000 LHS Dataganj
Instrument ID.	HG/TS/10mm/01	Parameter	Dimension
ENVIRONMENTAL CONDITION	Temperature:- 20±2°C, Relative Humidity: - 50±10%	Calibration Reference Standard	IS:460:1985
		Calibration Procedure	TNC -03-D 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Digital Vernier Caliper	0511201536	CC205523000000003F	Testncal Lanoratory	01/01/2024

Calibration Results

Standard value in mm	Obs 1	Obs 2	Obs 3	Obs 4	Obs 5	Average in mm	Uncertainty of Measurement at k = 2 at 95 % Confidence Level : ± µm
10	9.86	9.87	9.86	9.85	9.87	9.86	13.5

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national /International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

Certificate/ULR No.:	CC205523000012130F	Page 1 of 1
----------------------	--------------------	-------------

Customer Details		Instrument Receipt Date	1-Oct-2023
M/S. H.G. INFRA ENGINEERING LTD.. Vill- Papad Hamjapur (Khasra No-422) Near Brahamdev Tample Teh. & P.O. Dataganj Dist.- Badaun,, Uttar Pradesh Pin Code : 243635		SRF No.	TNC1/JC/23-24/1206
		Date of Calibration	1-Oct-2023
		Next Calibration Date	30-Sep-2024
		Certificate Issue Date	3-Oct-2023

A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Brass Sieve (200mm Dia)	Make / Model No.	Allied Sales Corp./-
Range	10 mm	Instrument Condition	OK
Least Count	-	Location of Calibration	At Laboratory
Serial No.	-	Location	208+000 LHS Dataganj
Instrument ID.	HG/TS/10mm/02	Parameter	Dimension
ENVIRONMENTAL CONDITION	Temperature:- 20±2°C, Relative Humidity: - 50±10%	Calibration Reference Standard	IS:460:1985
		Calibration Procedure	TNC -03-D 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Digital Vernier Caliper	0511201536	CC205523000000003F	Testncal Lanoratory	01/01/2024

Calibration Results

Standard value in mm	Obs 1	Obs 2	Obs 3	Obs 4	Obs 5	Average in mm	Uncertainty of Measurement at k = 2 at 95 % Confidence Level : ± µm
10	9.91	9.90	9.92	9.91	9.92	9.91	13.5

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

Certificate/ULR No.: CC205523000012131F Page 1 of 1

Customer Details		Instrument Receipt Date	1-Oct-2023
M/S. H.G. INFRA ENGINEERING LTD.. Vill- Papad Hamjapur (Khasra No-422) Near Brahamdev Tample Teh. & P.O. Dataganj Dist.- Badaun,, Uttar Pradesh Pin Code : 243635		SRF No.	TNC1/JC/23-24/1206
		Date of Calibration	1-Oct-2023
		Next Calibration Date	30-Sep-2024
		Certificate Issue Date	3-Oct-2023

A) Identification On UUC (Unit Under Calibration)

Instrument Name	Brass Sieve (200mm Dia)	Make / Model No.	Allied Sales Corp./-
Range	4.75 mm	Instrument Condition	OK
Least Count	-	Location of Calibration	At Laboratory
Serial No.	-	Location	208+000 LHS Dataganj
Instrument ID.	HG/TS/4.75mm/02	Parameter	Dimension
ENVIRONMENTAL CONDITION	Temperature:- 20±2°C, Relative Humidity: - 50±10%	Calibration Reference Standard	IS:460:1985
		Calibration Procedure	TNC -03-D 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Digital Vernier Caliper	0511201536	CC205523000000003F	Testncal Lanoratory	01/01/2024

Calibration Results

Standard value in mm	Obs 1	Obs 2	Obs 3	Obs 4	Obs 5	Average in mm	Uncertainty of Measurement at k = 2 at 95 % Confidence Level : ± µm
4.75	4.79	4.77	4.79	4.78	4.78	4.78	13.5

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.:	CC205523000012135F	Page 1 of 2
----------------------	--------------------	-------------

Customer Details		Instrument Receipt Date	1-Oct-2023
M/S. H.G. INFRA ENGINEERING LTD.. Vill- Papad Hamjapur (Khasra No-422) Near Brahamdev Tample Teh. & P.O. Dataganj Dist.- Badaun,, Uttar Pradesh Pin Code : 243635		SRF No.	TNC1/JC/23-24/1206
		Date of Calibration	1-Oct-2023
		Next Calibration Date	30-Sep-2024
		Certificate Issue Date	3-Oct-2023
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Electronic Balance	Make / Model No.	Apollo/-
Range	20g to 10 kg	Instrument Condition	OK
Least Count	0.5 g	Location of Calibration	At Site
Serial No.	-	Location	208+000 LHS Dataganj
Instrument ID.	HG/WB/10KG/01	Parameter	Mass
ENVIRONMENTAL CONDITION	Temperature-24±2°C Relative Humidity-55±10% RH	Calibration Reference Standard	OIML R-76-1
		Calibration Procedure	TNC-05/M/01

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Weights E2 Class (1 mg to 200 g)	TNC/M/C/002/17	FL/C/M/11012021-C001	CC2792	16/01/2024
Standard Weight (500g to 2kg)	WI/12/2017/018- F1	CC274322000001733F	WEGHTRONIX	13/03/2024
Standard Weight(5kg to 20kg)	4911/WI/01/2018/009-F2	WMCL/F/2022-01/1664		01/02/2024

Calibration Results-Weighing Performance Test

Load	Actual (g)	Indication (g)	Deviation (g)	C MAX= Maximum Correction (g)
20g	20.000056	20.0	0.0	1.0
50g	50.000105	50.0	0.0	
100g	100.00025	100.0	0.0	
200g	200.00024	200.0	0.0	
500g	500.0011	500.0	0.0	
1kg	999.9992	1000.0	0.0	
2kg	1999.9963	2000.5	0.5	
5kg	5000.012	5000.5	0.5	
10kg	10000.027	10001.0	1.0	

Eccentricity Test

1	2	3	4	5	1	5	4	3	2	1
5000.5	5000.5	5000.0	5000.5	5000.0	5000.0	5000.5	5001.0	5000.5	5000.0	5000.0

Repeatability test

CAL.POINT	R1	R2	R3	R4	R5	R6	R7	R8	R9	R10	Standard Dev. (g)	Maximum Error Between Centre and Other Points (g)	Uncertainty (g)
5000g	5000.5	5000.5	5001.0	5000.5	5000.5	5000.0	5000.5	5001.0	5000.5	5000.0	0.3	-0.1	0.7



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC205523000012135F

Page 2 of 2

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national /International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.:	CC205523000012141F	Page 1 of 1
----------------------	--------------------	-------------

Customer Details	Instrument Receipt Date	1-Oct-2023
M/S. H.G. INFRA ENGINEERING LTD.. Vill- Papad Hamjapur (Khasra No-422) Near Brahamdev Tample Teh. & P.O. Dataganj Dist.- Badaun,, Uttar Pradesh Pin Code : 243635	SRF No.	TNC1/JC/23-24/1206
	Date of Calibration	1-Oct-2023
	Next Calibration Date	30-Sep-2024
	Certificate Issue Date	3-Oct-2023

A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Electronic Balance	Make / Model No.	Apollo Scales/-
Range	10g to 30 kg	Instrument Condition	OK
Least Count	0.001 kg	Location of Calibration	At Site
Serial No.	-	Location	208+000 LHS Dataganj
Instrument ID.	HG/WB/30KG/01	Parameter	Mass
ENVIRONMENTAL CONDITION	Temperature-24±2°C Relative Humidity-55±10% RH	Calibration Reference Standard	OIML R-76-1
		Calibration Procedure	TNC-05/M/01

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Weights E2 Class (1 mg to 200 g)	TNC/M/C/002/17	FL/C/M/11012021-C001	CC2792	16/01/2024
Standard Weight (500g to 2kg)	WI/12/2017/018- F1	CC274322000001733F	WEGHTRONIX	13/03/2024
Standard Weight(5kg to 20kg)	4911/WI/01/2018/009-F2	WMCL/F/2022-01/1664		01/02/2024

Calibration Results-Weighing Performance Test

Load	Actual (g)	Indication (kg)	Deviation (kg)	C MAX= Maximum Correction (kg)
100g	100.00025	0.100	0.000	0.014
200g	200.00024	0.200	0.000	
500g	500.0011	0.500	0.000	
1kg	999.9992	1.000	0.000	
2kg	1999.9963	2.001	-0.001	
5kg	5000.012	5.003	-0.003	
10kg	10000.040	10.007	-0.007	
20kg	20000.08	20.011	-0.011	
30kg	30000.12	30.014	-0.014	

Eccentricity Test

1	2	3	4	5	1	5	4	3	2	1
10.003	10.005	10.004	10.002	10.005	10.003	10.004	10.002	10.004	10.005	10.003

Repeatability test

CAL.POIN T	R1	R2	R3	R4	R5	R6	R7	R8	R9	R10
10 kg	10.005	10.007	10.006	10.005	10.004	10.006	10.007	10.008	10.007	10.006

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .



Calibrated By

Issued & Approved By

Calibration Certificate

P -17 F03-00

Certificate/ULR No.: CC205523000012142F		Page 1 of 1	
Customer Details		Instrument Receipt Date	1-Oct-2023
M/S. H.G. INFRA ENGINEERING LTD.. Vill- Papad Hamjapur (Khasra No-422) Near Brahamdev Tample Teh. & P.O. Dataganj Dist.- Badaun,, Uttar Pradesh Pin Code : 243635		SRF No.	TNC1/JC/23-24/1206
		Date of Calibration	1-Oct-2023
		Next Calibration Date	30-Sep-2024
		Certificate Issue Date	3-Oct-2023
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Hot Plate	Make / Model No.	NA/-
Range	0 to 100 °C	Instrument Condition	OK
Least Count	5 °C	Location of Calibration	At Site
Serial No.	-	Location	208+000 LHS Dataganj
Instrument ID.	HG/HP/01	Parameter	Thermal
ENVIRONMENTAL CONDITION	Temperature:- 25±2 °C,Relative Humidity: - 50±10%	Calibration Reference Standard	ITS-90
		Calibration Procedure	TNC-06-T 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
RTD Sensor with 6 1/2 Digital Multimeter	TNC/T/C/005/17	CC317223000000132F	RK labs	24/01/2024

Calibration Results

Sr.No.	UUC Value (°C)	Avg. of Standard Observed Value (°C)	Deviation (°C)	Uncertainty at k=2 at 95 % Confidence Level (± °C)
1	20	20.13	0.13	2.90
2	40	40.19	0.19	2.90
3	60	60.24	0.24	2.90
4	80	80.31	0.31	2.90
5	100	100.38	0.38	2.90

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P -17 F03-00

Certificate/ULR No.: CC205523000012144F		Page 1 of 1	
Customer Details		Instrument Receipt Date	1-Oct-2023
M/S. H.G. INFRA ENGINEERING LTD.. Vill- Papad Hamjapur (Khasra No-422) Near Brahamdev Tample Teh. & P.O. Dataganj Dist.- Badaun,, Uttar Pradesh Pin Code : 243635		SRF No.	TNC1/JC/23-24/1206
		Date of Calibration	1-Oct-2023
		Next Calibration Date	30-Sep-2024
		Certificate Issue Date	3-Oct-2023
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Hot Plate	Make / Model No.	NA/-
Range	0 to 100 °C	Instrument Condition	OK
Least Count	5 °C	Location of Calibration	At Site
Serial No.	-	Location	208+000 LHS Dataganj
Instrument ID.	HG/HP/02	Parameter	Thermal
ENVIRONMENTAL CONDITION	Temperature:- 25±2 °C,Relative Humidity: - 50±10%	Calibration Reference Standard	ITS-90
		Calibration Procedure	TNC-06-T 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
RTD Sensor with 6 1/2 Digital Multimeter	TNC/T/C/005/17	CC317223000000132F	RK labs	24/01/2024

Calibration Results

Sr.No.	UUC Value (°C)	Avg. of Standard Observed Value (°C)	Deviation (°C)	Uncertainty at k=2 at 95 % Confidence Level (± °C)
1	20	20.11	0.11	2.90
2	40	40.17	0.17	2.90
3	60	60.26	0.26	2.90
4	80	80.33	0.33	2.90
5	100	100.41	0.41	2.90

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC205523000012145F		Page 1 of 1	
Customer Details		Instrument Receipt Date	1-Oct-2023
M/S. H.G. INFRA ENGINEERING LTD.. Vill- Papad Hamjapur (Khasra No-422) Near Brahamdev Tample Teh. & P.O. Dataganj Dist.- Badaun,, Uttar Pradesh Pin Code : 243635		SRF No.	TNC1/JC/23-24/1206
		Date of Calibration	1-Oct-2023
		Next Calibration Date	30-Sep-2024
		Certificate Issue Date	3-Oct-2023
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Electronic Balance	Make / Model No.	Swisser/-
Range	50g to 50 kg	Instrument Condition	OK
Least Count	0.001kg	Location of Calibration	At Site
Serial No.	-	Location	208+000 LHS Dataganj
Instrument ID.	HG/WB/50KG/01	Parameter	Mass
ENVIRONMENTAL CONDITION	Temperature-24±2°C Relative Humidity-55±10% RH	Calibration Reference Standard	OIML R-76-1
		Calibration Procedure	TNC-05/M/01

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Weights E2 Class (1 mg to 200 g)	TNC/M/C/002/17	FL/C/M/11012021-C001	CC2792	16/01/2024
Standard Weight (500g to 2kg)	WI/12/2017/018- F1	CC274322000001733F	WEGHTRONIX	13/03/2024
Standard Weight(5kg to 20kg)	4911/WI/01/2018/009-F2	WMCL/F/2022-01/1664		01/02/2024

Calibration Results-Weighing Performance Test

Load	Actual (g)	Indication (kg)	Deviation (kg)	C MAX= Maximum Correction (kg)
100g	100.00025	0.100	0.000	0.016
200g	200.00024	0.200	0.000	
500g	500.0011	0.500	0.000	
1kg	999.9992	1.000	0.000	
2kg	1999.9963	2.001	-0.001	
5kg	5000.012	5.002	-0.002	
10kg	10000.040	10.005	-0.005	
20kg	20000.08	20.008	-0.008	
30kg	30000.12	30.013	-0.013	
50kg	50000.14	50.016	-0.016	

Eccentricity Test

1	2	3	4	5	1	5	4	3	2	1
20.007	20.005	20.008	20.006	20.005	20.007	20.005	20.008	20.007	20.008	20.006

Repeatability test

CAL.POIN T	R1	R2	R3	R4	R5	R6	R7	R8	R9	R10
20 kg	20.007	20.005	20.008	20.006	20.005	20.007	20.006	20.007	20.008	20.006

Note:

- The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
- This report should not be reproduced except in full without our prior permission in writing.
- Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
- All our certificates are Traceable to national/International standard.
- This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC205523000012146F		Page 1 of 1	
Customer Details		Instrument Receipt Date	1-Oct-2023
M/S. H.G. INFRA ENGINEERING LTD.. Vill- Papad Hamjapur (Khasra No-422) Near Brahamdev Tample Teh. & P.O. Dataganj Dist.- Badaun,, Uttar Pradesh Pin Code : 243635		SRF No.	TNC1/JC/23-24/1206
		Date of Calibration	1-Oct-2023
		Next Calibration Date	30-Sep-2024
		Certificate Issue Date	3-Oct-2023
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Electronic Balance	Make / Model No.	Swisser/-
Range	50g to 50 kg	Instrument Condition	OK
Least Count	0.001kg	Location of Calibration	At Site
Serial No.	-	Location	208+000 LHS Dataganj
Instrument ID.	HG/WB/50KG/02	Parameter	Mass
ENVIRONMENTAL CONDITION	Temperature-24±2°C Relative Humidity-55±10% RH	Calibration Reference Standard	OIML R-76-1
		Calibration Procedure	TNC-05/M/01

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Weights E2 Class (1 mg to 200 g)	TNC/M/C/002/17	FL/C/M/11012021-C001	CC2792	16/01/2024
Standard Weight (500g to 2kg)	WI/12/2017/018- F1	CC274322000001733F	WEGHTRONIX	13/03/2024
Standard Weight(5kg to 20kg)	4911/WI/01/2018/009-F2	WMCL/F/2022-01/1664		01/02/2024

Calibration Results-Weighing Performance Test

Load	Actual (g)	Indication (kg)	Deviation (kg)	C MAX= Maximum Correction (kg)
100g	100.00025	0.100	0.000	0.012
200g	200.00024	0.200	0.000	
500g	500.0011	0.500	0.000	
1kg	999.9992	1.000	0.000	
2kg	1999.9963	2.000	0.000	
5kg	5000.012	4.999	0.001	
10kg	10000.040	9.998	0.002	
20kg	20000.08	19.995	0.005	
30kg	30000.12	29.992	0.008	
50kg	50000.14	49.988	0.012	

Eccentricity Test

1	2	3	4	5	1	5	4	3	2	1
19.991	19.989	19.992	19.991	19.989	19.990	19.989	19.988	19.991	19.990	19.989

Repeatability test

CAL.POIN T	R1	R2	R3	R4	R5	R6	R7	R8	R9	R10
20 kg	19.992	19.990	19.989	19.988	19.990	19.992	19.991	19.989	19.988	19.987

Note:

- The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
- This report should not be reproduced except in full without our prior permission in writing.
- Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
- All our certificates are Traceable to national/International standard.
- This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .



Calibrated By

Issued & Approved By

Calibration Certificate

P -17 F03-00

Certificate/ULR No.: CC205523000012147F		Page 1 of 1	
Customer Details		Instrument Receipt Date	1-Oct-2023
M/S. H.G. INFRA ENGINEERING LTD.. Vill- Papad Hamjapur (Khasra No-422) Near Brahamdev Temple Teh. & P.O. Dataganj Dist.- Badaun,, Uttar Pradesh Pin Code : 243635		SRF No.	TNC1/JC/23-24/1206
		Date of Calibration	1-Oct-2023
		Next Calibration Date	30-Sep-2024
		Certificate Issue Date	3-Oct-2023
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Compression Testing Machine	Make / Model No.	Heico/HLM592401
Range	0 to 3000 KN	Instrument Condition	OK
Least Count	0.1 KN	Location of Calibration	At Site
Serial No.	H-04157	Location	208+000 LHS Dataganj
Instrument ID.	-	Parameter	Force
ENVIRONMENTAL CONDITION	Temperature:- 24±2 °C,Relative Humidity: - 50±10	Calibration Reference Standard	IS 1828(Part1) :2022
		Calibration Procedure	TNC -01-F 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
PROVING RING 2000 KN	1821.2000KN	22011317/D1.05/C-445	CSIR	14/04/2024

Calibration Results

Force(kN)	Division at Std. Temp. 24°C	Division at room Temp. 24.4°C	Observed Reading (Division s) 0°	Observed Reading (Division s) 120°	Observed Reading (Division s) 240°	Average of Reading (Division)	Relative accuracy error (q)%	Relative Repeatability error (b)%	Relative resolution Error (a)%	Relative Zero Error (fo)%	Expende d Uncertain ty of Measure ment ± %
200	84.8	84.8	85.5	85.0	85.5	85.33	0.63	0.59	0.03	0.00	1.00
400	167.0	167.0	167.5	167.5	168.0	167.67	0.40	0.30	0.01	0.00	1.00
600	249.5	249.5	249.5	250.0	250.0	249.83	0.13	0.20	0.01	0.00	1.00
800	333.3	333.3	333.5	334.0	334.0	333.83	0.16	0.15	0.01	0.00	1.00
1000	417.6	417.6	418.0	418.5	418.5	418.33	0.17	0.12	0.01	0.00	1.00
1200	500.3	500.2	500.0	500.5	500.5	500.33	0.03	0.10	0.00	0.00	1.00
1400	583.7	583.6	583.5	584.0	584.0	583.83	0.04	0.09	0.00	0.00	1.00
1600	669.2	669.1	669.0	669.5	669.5	669.33	0.03	0.07	0.00	0.00	1.00
1800	754.5	754.4	754.0	754.5	754.0	754.17	-0.03	0.07	0.00	0.00	1.00
0	0	0	0	0	0	0	0	0	0	0	0

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
6. Measurement of uncertainty at k=2, Approximate 95% confidence level.



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC205523000012148F		Page 1 of 1	
Customer Details		Instrument Receipt Date	1-Oct-2023
M/S. H.G. INFRA ENGINEERING LTD.. Vill- Papad Hamjapur (Khasra No-422) Near Brahamdev Tample Teh. & P.O. Dataganj Dist.- Badaun,, Uttar Pradesh Pin Code : 243635		SRF No.	TNC1/JC/23-24/1206
		Date of Calibration	1-Oct-2023
		Next Calibration Date	30-Sep-2024
		Certificate Issue Date	3-Oct-2023
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	High Frequency Cube Vibrator M/c	Make / Model No.	NA/-
Range	12000±400 rpm	Instrument Condition	OK
Least Count	NA	Location of Calibration	At Site
Serial No.	-	Location	208+000 LHS Dataganj
Instrument ID.	HG/HFCVM/01	Parameter	Speed
ENVIRONMENTAL CONDITION	Temperature: 25±2°C, Relative Humidity: 50±10%	Calibration Reference Standard	IS: 12508
		Calibration Procedure	TNC -02-RPM 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Dig. Tachometer	TNC/RPM/C/002/19	CC233924000000443F	PRIMA	30/09/2024

Calibration Results

Sl. No.	UUC Value (rpm)	Std. Value (rpm)	Error (rpm)	Uncertainty at k=2 ,± %
1	12000	11986	14	0.40

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC205523000012149F		Page 1 of 1	
Customer Details		Instrument Receipt Date	1-Oct-2023
M/S. H.G. INFRA ENGINEERING LTD.. Vill- Papad Hamjapur (Khasra No-422) Near Brahamdev Tample Teh. & P.O. Dataganj Dist.- Badaun,, Uttar Pradesh Pin Code : 243635		SRF No.	TNC1/JC/23-24/1206
		Date of Calibration	1-Oct-2023
		Next Calibration Date	30-Sep-2024
		Certificate Issue Date	3-Oct-2023
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Hydrometer	Make / Model No.	Omsons/-
Range	0.800 to 0.850 g/ml	Instrument Condition	OK
Least Count	0.001 g/ml	Location of Calibration	At Laboratory
Serial No.	21441602	Location	208+000 LHS Dataganj
Instrument ID.	-	Parameter	Density
ENVIRONMENTAL CONDITION	Temperature:- 24±2°C,Relative Humidity: - 50±10%	Calibration Reference Standard	IS:3104 (Part I & II)
		Calibration Procedure	TNC-05/D/2.0

Calibration Results

S. No.	UUC. Value (g/ml)	STD. Value (g/ml)	Deviation (g/ml)	Uncertainty at k=2 ± (g/ml)
1	0.800	0.801	-0.001	0.0011
2	0.830	0.831	-0.001	0.0011
3	0.850	0.848	0.002	0.0011

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

Certificate/ULR No.:	CC205523000012150F	Page 1 of 1
----------------------	--------------------	-------------

Customer Details		Instrument Receipt Date	1-Oct-2023
M/S. H.G. INFRA ENGINEERING LTD.. Vill- Papad Hamjapur (Khasra No-422) Near Brahamdev Tample Teh. & P.O. Dataganj Dist.- Badaun,, Uttar Pradesh Pin Code : 243635		SRF No.	TNC1/JC/23-24/1206
		Date of Calibration	1-Oct-2023
		Next Calibration Date	30-Sep-2024
		Certificate Issue Date	3-Oct-2023

A) Identification On UUC (Unit Under Calibration)			
Instrument Name	GI Sieve (450mm Dia)	Make / Model No.	Allied Sales Corp./-
Range	63 mm	Instrument Condition	OK
Least Count	-	Location of Calibration	At Laboratory
Serial No.	-	Location	208+000 LHS Dataganj
Instrument ID.	HG/TS/63mm/02	Parameter	Dimension
ENVIRONMENTAL CONDITION	Temperature:- 20±2°C, Relative Humidity: - 50±10%	Calibration Reference Standard	IS:460:1985
		Calibration Procedure	TNC -03-D 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Digital Vernier Caliper	0511201536	CC205523000000003F	Testncal Lanoratory	01/01/2024

Calibration Results

Standard value in mm	Obs 1	Obs 2	Obs 3	Obs 4	Obs 5	Average in mm	Uncertainty of Measurement at k = 2 at 95 % Confidence Level : ± µm
63	62.93	62.92	62.94	62.91	62.89	62.92	13.5

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By