

Calibration Certificate

P -17 F03-00

Certificate/ULR No.: CC205524000002952F		Page 1 of 1	
Customer Details		Instrument Receipt Date	16-Jul-2024
M/S AA Electro Magnetic Test Laboratory Private Limited Plot No 174, Udyog Vihar Phase 4, Sector 18, Gurgaon,, Haryana		SRF No.	TNC/JC/24-25/158
		Date of Calibration	16-Jul-2024
		Next Calibration Date	15-Jul-2025
		Certificate Issue Date	17-Jul-2024
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Crush Test Machine	Make / Model No.	Harriss & Tarriss/-
Range	0 to 15 KN	Instrument Condition	OK
Least Count	0.01 KN	Location of Calibration	At Site
Serial No.	-	Location	-
Instrument ID.	AAEMT/ELEC/240	Parameter	Force
ENVIRONMENTAL CONDITION	Temperature:- 24±2 °C,Relative Humidity: - 50±10	Calibration Reference Standard	IS 1828 (Part1) :2022
		Calibration Procedure	TNC -01-F 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Load Cell with Indicator	TNC/F/C/001/17	CC210122000000844F	DVG Laboratory	18/07/2024

Calibration Results

Force(kN)	Division at Std. Temp. 24°C	Division at room Temp. 24.4°C	Observed Reading (Division s) 0°	Observed Reading (Division s) 120°	Observed Reading (Division s) 240°	Average of Reading (Division)	Relative accuracy error (q)%	Relative Repeatability error (b)%	Relative resolution Error (a)%	Relative Zero Error (fo)%	Expende d Uncertain ty of Measure ment ± %
2	2000	2000	2001	2012	2016	2009.67	0.48	0.75	0.50	0.00	0.90
4	3998	3998	4012	4016	4006	4011.33	0.33	0.25	0.25	0.00	0.88
5	4998	4999	5009	5014	5021	5014.67	0.33	0.24	0.20	0.00	0.88
10	9998	9999	10021	10009	10016	10015.33	0.17	0.12	0.10	0.00	0.88
13	12997	12998	13007	13015	13021	13014.33	0.13	0.11	0.08	0.00	0.88
15	14996	14998	15019	15012	15016	15015.67	0.13	0.05	0.07	0.00	0.88
0	0	0	0	0	0	0	0	0	0	0	0

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC205524000003054F		Page 1 of 1	
Customer Details		Instrument Receipt Date	16-Jul-2024
M/S AA Electro Magnetic Test Laboratory Private Limited Plot No 174, Udyog Vihar Phase 4, Sector 18, Gurgaon,, Haryana		SRF No.	TNC/JC/24-25/158
		Date of Calibration	17-Jul-2024
		Next Calibration Date	16-Jul-2025
		Certificate Issue Date	18-Jul-2024
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Computerized Battery Short Circuit	Make / Model No.	Map Electronics/-
Range	80 MΩ	Instrument Condition	OK
Least Count	As per Range	Location of Calibration	At Site
Serial No.	-	Location	-
Instrument ID.	AAEMT/ELEC/245	Parameter	Electro Technical
ENVIRONMENTAL CONDITION	Temperature:- 24±2 °C,Relative Humidity: - 55±10%	Calibration Reference Standard	IS:1248-2003
		Calibration Procedure	TNC -03-WI 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
6 ½ Digit DMM	TNC/ET/C/001/15	AIPL/300124/3-2	CC-2020 (ANSHANKAN)	31/01/2025

Resistance

Sl. No	Range	Range	UOM	Observed Value on UUC	Observed Value on Standard	Uncertainty at k=2 ±%
1	80	80MΩ	MΩ	80	79.9678	0.06

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC205524000003306F		Page 1 of 1	
Customer Details		Instrument Receipt Date	19-Jul-2024
M/S. AA ELECTRO MAGNETIC TEST LABORATORY PRIVATE LIMITED Plot No 174, Udyog Vihar-Phase 4, Sector 18, Gurgaon, Haryana		SRF No.	TNC/JC/24-25/220
		Date of Calibration	19-Jul-2024
		Next Calibration Date	18-Jul-2025
		Certificate Issue Date	20-Jul-2024
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	DC Power Supply	Make / Model No.	Shangdong/-
Range	0 to 60 V	Instrument Condition	OK
Least Count	0.01 V	Location of Calibration	At Site
Serial No.	-	Location	-
Instrument ID.	AAEMT/ELEC/225	Parameter	Electro Technical
ENVIRONMENTAL CONDITION	Temperature:- 24±2 °C,Relative Humidity: - 55±10%	Calibration Reference Standard	IS:1248-2003
		Calibration Procedure	TNC -03-WI 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
6 ½ Digit DMM	TNC/ET/C/001/15	AIPL/300124/3-2	CC-2020 (ANSHANKAN)	31/01/2025
Multifunction Calibrator	TNC/Cal/ET/015/017	CC205523000000426F	TESTNCAL	12/08/2024

AC Voltage @50Hz

Sl. No	Range	UOM	Observed Value on UUC	Observed Value on Standard	Uncertainty at k=2 ±%
1	10V	V	10.00	9.9956	0.00
2	36V	V	36.00	35.9871	0.00
3	42V	V	42.00	41.9844	0.00
4	48V	V	48.00	47.9885	0.00
5	60V	V	60.00	59.9815	0.00

Note:

- The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
- This report should not be reproduced except in full without our prior permission in writing.
- Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
- All our certificates are Traceable to national/International standard.
- This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC205524000003307F		Page 1 of 1	
Customer Details		Instrument Receipt Date	19-Jul-2024
M/S. AA ELECTRO MAGNETIC TEST LABORATORY PRIVATE LIMITED Plot No 174, Udyog Vihar-Phase 4, Sector 18, Gurgaon, Haryana		SRF No.	TNC/JC/24-25/220
		Date of Calibration	19-Jul-2024
		Next Calibration Date	18-Jul-2025
		Certificate Issue Date	20-Jul-2024
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	DC Power Supply	Make / Model No.	Shangdong/-
Range	0 to 60 V	Instrument Condition	OK
Least Count	0.01 V	Location of Calibration	At Site
Serial No.	-	Location	-
Instrument ID.	AAEMT/ELEC/224	Parameter	Electro Technical
ENVIRONMENTAL CONDITION	Temperature:- 24±2 °C,Relative Humidity: - 55±10%	Calibration Reference Standard	IS:1248-2003
		Calibration Procedure	TNC -03-WI 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
6 ½ Digit DMM	TNC/ET/C/001/15	AIPL/300124/3-2	CC-2020 (ANSHANKAN)	31/01/2025
Multifunction Calibrator	TNC/Cal/ET/015/017	CC205523000000426F	TESTNCAL	12/08/2024

AC Voltage @50Hz

Sl. No	Range	UOM	Observed Value on UUC	Observed Value on Standard	Uncertainty at k=2 ±%
1	10V	V	10.00	9.9937	0.00
2	36V	V	36.00	35.9845	0.00
3	42V	V	42.00	41.9830	0.00
4	48V	V	48.00	47.9855	0.00
5	60V	V	60.00	59.9803	0.00

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC405824000000402F		Page 1 of 1	
Customer Details		Instrument Receipt Date	2-Nov-2024
M/S. AA ELECTRO MAGNETIC TEST LABORATORY PRIVATE LIMITED Plot No 174, Udyog Vihar-Phase 4, Sector 18, Gurgaon, Haryana		SRF No.	TNC/JC/24-25/279
		Date of Calibration	2-Nov-2024
		Next Calibration Date	1-Nov-2025
		Certificate Issue Date	4-Nov-2024
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Digital Insulation Tester	Make / Model No.	Tinsley/5871P
Range	As per Specification	Instrument Condition	SATISFACTORY
Least Count	As per Range	Location of Calibration	At Laboratory
Serial No.	67114071055P	Location	Electrical Lab
Instrument ID.	AAEMT/ELEC/116	Parameter	Electro Technical
ENVIRONMENTAL CONDITION	Temperature:- 24±2°C ,Relative Humidity: - 50±10%	Calibration Reference Standard	IS:1248-2003
		Calibration Procedure	TNC -03-WI 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
HV Megohm Box (20M to 200G) ohm	TNC/ET/C/02/18	CC223124000052518F	CC-2231	13/05/2025
Standard Resistance Box	TNC/ET/C/03/18	CC243824000000230F	CC-2438	31/01/2025

Resistance: Range

Parameter	Range	UOM	Std. Value	Observed Value on UUC	Uncertainty at k=2 ±%
Resistance	@250 V	MΩ	1	1.0	0.58
		MΩ	20	19.9	0.58
		MΩ	100	99.8	0.58
		MΩ	1000	998	1.42
Resistance	@500 V	MΩ	2	1.9	0.58
		MΩ	20	19.8	0.58
		MΩ	100	99.7	0.58
		MΩ	1000	997	1.42
Resistance	@1000 V	MΩ	2	1.9	5.77
		MΩ	20	19.6	5.77
		MΩ	100	99.7	5.77
		MΩ	1000	997	5.77

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P -17 F03-00

Certificate/ULR No.: CC405824000000410F		Page 1 of 1	
Customer Details		Instrument Receipt Date	2-Nov-2024
M/S. AA ELECTRO MAGNETIC TEST LABORATORY PRIVATE LIMITED Plot No 174, Udyog Vihar-Phase 4, Sector 18, Gurgaon, Haryana		SRF No.	TNC/JC/24-25/279
		Date of Calibration	2-Nov-2024
		Next Calibration Date	1-Nov-2025
		Certificate Issue Date	4-Nov-2024
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Heating Encloser	Make / Model No.	Sunrise Scientific Instrument
Range	Amb. to 150 °C	Instrument Condition	SATISFACTORY
Least Count	0.1°C	Location of Calibration	At Site
Serial No.	-	Location	-
Instrument ID.	AAEMT/ELEC/060	Parameter	Thermal
ENVIRONMENTAL CONDITION	Temperature:- 25±2 °C,Relative Humidity: - 50±10%	Calibration Reference Standard	ITS-90
		Calibration Procedure	TNC-06-T 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
RTD Sensor with 6 1/2 Digital Multimeter	TNC/T/C/005/17	CC317224000000216F	CC-3172 (RK LABS)	23/01/2025

Calibration Results

Sr.No.	UUC Value (°C)	Avg. of Standard Observed Value (°C)	Deviation (°C)	Uncertainty at k=2 at 95 % Confidence Level (± °C)
1	25.0	25.103	-0.103	0.480
2	50.0	50.136	-0.136	0.480
3	80.0	80.172	-0.172	0.480
4	110.0	110.219	-0.219	0.480
5	150.0	150.273	-0.273	0.480

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC405824000000413F		Page 1 of 1	
Customer Details		Instrument Receipt Date	2-Nov-2024
M/S. AA ELECTRO MAGNETIC TEST LABORATORY PRIVATE LIMITED Plot No 174, Udyog Vihar-Phase 4, Sector 18, Gurgaon, Haryana		SRF No.	TNC/JC/24-25/279
		Date of Calibration	13-Nov-2024
		Next Calibration Date	13-Nov-2025
		Certificate Issue Date	13-Nov-2024
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Digital Insulation Tester	Make / Model No.	MC/LK2679D
Range	As per Specification	Instrument Condition	SATISFACTORY
Least Count	As per Range	Location of Calibration	At Laboratory
Serial No.	179D-1215-054	Location	Electrical Lab
Instrument ID.	AAEMT/ELEC/087	Parameter	Electro Technical
ENVIRONMENTAL CONDITION	Temperature:- 24±2°C ,Relative Humidity: - 50±10%	Calibration Reference Standard	IS:1248-2003
		Calibration Procedure	TNC -03-WI 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Standard Resistance Box	TNC/ET/C/03/18	CC243824000000230F	CC-2438	31/01/2025
HV Megohm Box (20M to 200G) ohm	TNC/ET/C/02/18	CC223124000052518F	CC-2231	13/05/2025

Resistance: Range

Parameter	Range	UOM	Std. Value	Observed Value on UUC	Uncertainty at k=2 ±%
Resistance	@500 V	MΩ	2	2.0	0.58
		MΩ	20	19.9	0.58
		MΩ	100	99.7	0.58
		MΩ	1000	998	1.42
Resistance	@1000 V	MΩ	2	1.9	5.77
		MΩ	20	19.7	5.77
		MΩ	100	99.4	5.77
		MΩ	1000	997	5.77

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P - 17 F03-00

Certificate/ULR No.: CC405824000000424F		Page 1 of 1	
Customer Details		Instrument Receipt Date	2-Nov-2024
M/S. AA ELECTRO MAGNETIC TEST LABORATORY PRIVATE LIMITED Plot No 174, Udyog Vihar-Phase 4, Sector 18, Gurgaon, Haryana		SRF No.	TNC/JC/24-25/279
		Date of Calibration	2-Nov-2024
		Next Calibration Date	1-Nov-2025
		Certificate Issue Date	4-Nov-2024
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Dust Test Chamber	Make / Model No.	Warmax India/WMI/DTC/10C
Range	Amb. to 50 °C, 20 to 40 % RH	Instrument Condition	SATISFACTORY
Least Count	0.1°C, 1% RH	Location of Calibration	At Site
Serial No.	-	Location	-
Instrument ID.	AAEMT/ELEC/019	Parameter	Thermal
ENVIRONMENTAL CONDITION	Temperature:- 24±2 °C,Relative Humidity: - 55±10%	Calibration Reference Standard	DKD -R5-7
		Calibration Procedure	TNC-05/T/01

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
RH & Temperature Sensor with Indicator	HTD1220240058	CC388324000000003F	CC-3883	11/04/2025

Calibration Results (Temperature)

Sr.No.	UUC Value (°C)	Avg. of Standard Observed Value (°C)	Deviation (°C)	Uncertainty at k=2 at 95 % Confidence Level (±) (°C)
1	25.0	25.06	0.06	0.48
2	30.0	30.11	0.11	0.48
3	40.0	40.17	0.17	0.48
4	50.0	50.22	0.22	0.48

Calibration Result (Humidity) @ 25°C

Sr.No.	UUC Value (%RH)	Avg. of Standard Observed Value (%RH)	Deviation (%RH)	Uncertainty at k=2 at 95 % Confidence Level (±) (%RH)
1	25	25.17	0.17	1.8
2	30	30.24	0.24	1.8

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC405825000000208F		Page 1 of 1	
Customer Details		Instrument Receipt Date	7-Jan-2025
M/S. AA ELECTRO MAGNETIC TEST LABORATORY PRIVATE LIMITED Plot No 174, Udyog Vihar-Phase 4, Sector 18, Gurgaon, Haryana		SRF No.	TNC/JC/24-25/356
		Date of Calibration	7-Jan-2025
		Next Calibration Date	6-Jan-2026
		Certificate Issue Date	8-Jan-2025
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Push Pull Gauge	Make / Model No.	Inogu/ZP-500N
Range	0 to 500 N	Instrument Condition	SATISFACTORY
Least Count	0.1 N	Location of Calibration	At Laboratory
Serial No.	-	Location	-
Instrument ID.	AAEMT/ELEC/081	Parameter	Force
ENVIRONMENTAL CONDITION	Temperature-24±2°C RH-50±10%	Calibration Reference Standard	VDI/VDE 2624 Part 2.1 (Dec
		Calibration Procedure	TNC -05/F/01

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Push Pull Gauge Calibrator	TNC/CAL/F/024-030/18	AMC/240016/06	CC-2364	07/01/2025
Push pull Gauge Calibrator	TNC/CAL/F/019-023/18	AMC/240016/05	CC-2364	07/01/2025
Push Pull Gauge Calibrator	TNC/CAL/F/008-011/18	AMC/240016/02	CC-2364	07/01/2025
Push Pull Gauge Calibrator	TNC/CAL/F/031-036/18	AMC/240016/07	CC-2364	07/01/2025
Push Pull Gauge Calibrator	TNC/CAL/F/004-007/18	AMC/240016/01	CC-2364	07/01/2025
Push Pull Gauge Calibrator	TNC/CAL/F/015-018/18	AMC/240016/04	CC-2364	07/01/2025

Calibration Results

Mode	Load (N)	Reading at Increase	Reading at Decrease	Uncertainty at k=2 ± %	Mode	Reading at Increase	Reading at Decrease	Uncertainty at k=2 ± %
Push(N)	0	0.0	0.0	-	Pull(N)	0.0	0.0	-
	100	100.0	100.0	1.00		100.0	100.0	1.00
	200	200.0	200.0	0.50		200.0	200.0	0.50
	300	299.9	300.0	0.50		299.9	300.0	0.50
	400	399.9	399.9	0.50		400.0	399.9	0.50
	500	499.8	499.9	0.50		499.8	499.8	0.50
LOAD (PUSH)		Stdev at half load (PUSH)		LOAD (PULL)		Stdev at half load (PULL)		
250 (N)		0.06		250 (N)		0.06		

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC405825000000210F		Page 1 of 1	
Customer Details		Instrument Receipt Date	7-Jan-2025
M/S. AA ELECTRO MAGNETIC TEST LABORATORY PRIVATE LIMITED Plot No 174, Udyog Vihar-Phase 4, Sector 18, Gurgaon, Haryana		SRF No.	TNC/JC/24-25/356
		Date of Calibration	7-Jan-2025
		Next Calibration Date	6-Jan-2026
		Certificate Issue Date	8-Jan-2025
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Tachometer	Make / Model No.	Luolider/UT372
Range	0 to 99999 rpm	Instrument Condition	SATISFACTORY
Least Count	1 rpm	Location of Calibration	At Site
Serial No.	-	Location	-
Instrument ID.	AAEMT/ELEC/085	Parameter	Speed
ENVIRONMENTAL CONDITION	Temperature: 25±2°C, Relative Humidity: 50±10%	Calibration Reference Standard	IS: 12508
		Calibration Procedure	TNC -02-RPM 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Dig. Tachometer	TNC/RPM/C/002/19	CC380824000000839F	CC-3808 (REALTECH)	30/09/2026

Calibration Results

Sl. No.	UUC Value (rpm)	Std. Value (rpm)	Error (rpm)	Uncertainty at k=2 ,± %
1	100.0	100.1	-0.1	4.24
2	500.0	500.2	-0.2	3.40
3	5000	5001	-1	1.62
4	15000	15003	-3	1.61
5	35000	35006	-6	1.61

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC405825000000211F		Page 1 of 1	
Customer Details		Instrument Receipt Date	7-Jan-2025
M/S. AA ELECTRO MAGNETIC TEST LABORATORY PRIVATE LIMITED Plot No 174, Udyog Vihar-Phase 4, Sector 18, Gurgaon, Haryana		SRF No.	TNC/JC/24-25/356
		Date of Calibration	7-Jan-2025
		Next Calibration Date	6-Jan-2026
		Certificate Issue Date	8-Jan-2025
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Timer (Dripping Test Equipment)	Make / Model No.	E-lin/JSS20-106-A
Range	0 to 999 sec	Instrument Condition	SATISFACTORY
Least Count	0.01 sec	Location of Calibration	At Laboratory
Serial No.	-	Location	-
Instrument ID.	AAEMT/ELEC/119	Parameter	Electro Technical
ENVIRONMENTAL CONDITION	Temperature:- 24±2°C,Relative Humidity: - 50±10%	Calibration Reference Standard	IS : 5834
		Calibration Procedure	TNC -07- ET 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Time Interval Meter	TNC/ET/C/015	CC294024000000324F	CC-2940 (Digital Calibration Technology)	02/02/2025

Calibration Results

Sl. No	UUC Value (S)	Std. Value (S)	Error (S)	Uncertainty at k=2 ,± (S)
1	10	10.002	0.002	0.61
2	100	100.009	0.009	0.66
3	200	200.014	0.014	0.69
4	300	300.029	0.029	0.73
5	500	500.063	0.063	0.87
6	999	999.18	0.18	1.32

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.:	CC405825000000212F	Page 1 of 1
----------------------	--------------------	-------------

Customer Details	Instrument Receipt Date	7-Jan-2025
M/S. AA ELECTRO MAGNETIC TEST LABORATORY PRIVATE LIMITED Plot No 174, Udyog Vihar-Phase 4, Sector 18, Gurgaon, Haryana	SRF No.	TNC/JC/24-25/356
	Date of Calibration	7-Jan-2025
	Next Calibration Date	6-Jan-2026
	Certificate Issue Date	8-Jan-2025

A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Ampere Meter (Cord Flexing)	Make / Model No.	Sunrise Scientific Instrument
Range	0 to 4 A	Instrument Condition	SATISFACTORY
Least Count	0.01 A	Location of Calibration	At Laboratory
Serial No.	-	Location	-
Instrument ID.	AAEMT/ELEC/047	Parameter	Electro Technical
ENVIRONMENTAL CONDITION	Temperature:- 24±2 °C,Relative Humidity: - 55±10%	Calibration Reference Standard	IS:1248-2003
		Calibration Procedure	TNC -03-WI 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Digital Clamp Meter	TNC/ETC/CM/01/24	CC205524000002894F	CC-2055	11/07/2025

Calibration Results(AC Current @ 50Hz

Sl. No	Range	UOM	Observed Value on UUC	Observed Value on Standard	Uncertainty at k=2 ±%
1	1A	A	1.00	1.00089	0.01
2	2A	A	2.00	2.00096	0.01
3	3A	A	3.00	3.00098	0.01
4	4A	A	4.00	4.00136	0.01

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .



Calibrated By

Issued & Approved By

Calibration Certificate

P -17 F03-00

Certificate/ULR No.: CC405825000000213F		Page 1 of 1	
Customer Details		Instrument Receipt Date	7-Jan-2025
M/S. AA ELECTRO MAGNETIC TEST LABORATORY PRIVATE LIMITED Plot No 174, Udyog Vihar-Phase 4, Sector 18, Gurgaon, Haryana		SRF No.	TNC/JC/24-25/356
		Date of Calibration	7-Jan-2025
		Next Calibration Date	6-Jan-2026
		Certificate Issue Date	8-Jan-2025
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Sulfur Dioxide Chamber	Make / Model No.	-/WEW-SO-90D
Range	Amb. to 40 °C	Instrument Condition	SATISFACTORY
Least Count	1 °C	Location of Calibration	At Site
Serial No.	-	Location	4th Floor
Instrument ID.	AAEMT/ELEC/278	Parameter	Thermal
ENVIRONMENTAL CONDITION	Temperature:- 25±2 °C,Relative Humidity: - 50±10%	Calibration Reference Standard	ITS-90
		Calibration Procedure	TNC-06-T 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
RTD Sensor with 6 1/2 Digital Multimeter	TNC/T/C/005/17	CC317224000000216F	CC-3172 (RK LABS)	23/01/2025

Calibration Results

Sr.No.	UUC Value (°C)	Avg. of Standard Observed Value (°C)	Deviation (°C)	Uncertainty at k=2 at 95 % Confidence Level (± °C)
1	10	10.172	-0.172	0.670
2	20	20.225	-0.225	0.663
3	30	30.331	-0.331	0.663
5	40	40.424	-0.424	0.662

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P -17 F03-00

Certificate/ULR No.: CC405825000000214F		Page 1 of 1	
Customer Details		Instrument Receipt Date	7-Jan-2025
M/S. AA ELECTRO MAGNETIC TEST LABORATORY PRIVATE LIMITED Plot No 174, Udyog Vihar-Phase 4, Sector 18, Gurgaon, Haryana		SRF No.	TNC/JC/24-25/356
		Date of Calibration	7-Jan-2025
		Next Calibration Date	6-Jan-2026
		Certificate Issue Date	8-Jan-2025
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Salt Spray Chamber	Make / Model No.	Sunrise Scientific Instrument
Range	0 - 50 °C, 0 to 99 % RH	Instrument Condition	SATISFACTORY
Least Count	0.1°C, 1% RH	Location of Calibration	At Laboratory
Serial No.	-	Location	4th Floor
Instrument ID.	AAEMT/ELEC/279	Parameter	Thermal
ENVIRONMENTAL CONDITION	Temperature:- 24±2 °C,Relative Humidity: - 55±10%	Calibration Reference Standard	DKD –R5-7
		Calibration Procedure	TNC-05/T/01

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
RH & Temperature Sensor with Indicator	HTD1220240058	CC388324000000003F	CC-3883	11/04/2025

Calibration Results (Temperature)

Sr.No.	UUC Value (°C)	Avg. of Standard Observed Value (°C)	Deviation (°C)	Uncertainty at k=2 at 95 % Confidence Level (±) (°C)
1	20.0	20.20	0.20	0.48
2	25.0	25.23	0.23	0.48
3	30.0	30.31	0.31	0.49
4	40.0	40.35	0.35	0.51
5	50.0	50.40	0.40	0.50

Calibration Result (Humidity) @ 25°C

Sr.No.	UUC Value (%RH)	Avg. of Standard Observed Value (%RH)	Deviation (%RH)	Uncertainty at k=2 at 95 % Confidence Level (±) (%RH)
1	10	10.04	0.04	1.8
2	30	30.15	0.15	1.8
3	50	50.22	0.22	1.8
4	70	70.28	0.28	1.8
5	95	95.38	0.38	1.8

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P -17 F03-00

Certificate/ULR No.: CC405825000000215F		Page 1 of 1	
Customer Details		Instrument Receipt Date	7-Jan-2025
M/S. AA ELECTRO MAGNETIC TEST LABORATORY PRIVATE LIMITED Plot No 174, Udyog Vihar-Phase 4, Sector 18, Gurgaon, Haryana		SRF No.	TNC/JC/24-25/356
		Date of Calibration	7-Jan-2025
		Next Calibration Date	6-Jan-2026
		Certificate Issue Date	8-Jan-2025
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Water Bath	Make / Model No.	Sunrise Scientific Instrument
Range	0 to 100 °C	Instrument Condition	SATISFACTORY
Least Count	0.1°C	Location of Calibration	At Site
Serial No.	-	Location	High Voltage Section
Instrument ID.	AAEMT/ELEC/244	Parameter	Thermal
ENVIRONMENTAL CONDITION	Temperature:- 25±2 °C,Relative Humidity: - 50±10%	Calibration Reference Standard	ITS-90
		Calibration Procedure	TNC-06-T 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
RTD Sensor with 6 1/2 Digital Multimeter	TNC/T/C/005/17	CC317224000000216F	CC-3172 (RK LABS)	23/01/2025

Calibration Results

Sr.No.	UUC Value (°C)	Standard Value (°C)	Deviation (°C)	Uncertainty at k=2 at 95 % Confidence Level (± °C)
1	20.0	20.078	0.078	0.480
2	40.0	40.174	0.174	0.480
3	60.0	60.209	0.209	0.480
4	80.0	80.256	0.256	0.480
5	100.0	100.339	0.339	0.480

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC405825000000216F		Page 1 of 1	
Customer Details		Instrument Receipt Date	7-Jan-2025
M/S. AA ELECTRO MAGNETIC TEST LABORATORY PRIVATE LIMITED Plot No 174, Udyog Vihar-Phase 4, Sector 18, Gurgaon, Haryana		SRF No.	TNC/JC/24-25/356
		Date of Calibration	7-Jan-2025
		Next Calibration Date	6-Jan-2026
		Certificate Issue Date	8-Jan-2025
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Digital Vernier Caliper	Make / Model No.	Insize/-
Range	0 to 300 mm	Instrument Condition	SATISFACTORY
Least Count	0.01mm	Location of Calibration	At Laboratory
Serial No.	2805211986	Location	-
Instrument ID.	AAEMT/ELEC/223	Parameter	Dimension
ENVIRONMENTAL CONDITION	Temperature:- 20±2°C,Relative Humidity: - 50±10%	Calibration Reference Standard	IS 16491 (Part 1 & 2)
		Calibration Procedure	TNC -05-WI 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Slip Gauge Block Set	TNC/M/C/001/15	TI/B/SGS/001/23	CC-2660 (TANSON INSTRUMENT)	02/02/2025
Slip Gauge Access.	TNC/M/C/011/16	CC266023000002231F	CC-2660 (TANSION INSTRUMENT)	17/12/2025

Calibration Results (mm)

Standard Value	Observed Value on UUC (Inner)	Deviation(In ner)	Uncertainty (± μm)	Observed Value on UUC(Outer)	Deviation(O uter)	Uncertainty (± μm)	Observed Value on UUC (Depth)	Deviation (Depth)	Uncertainty (± μm)
2.000	2.00	0.000	14.4	2.00	0.000	14.4	2.00	0.000	14.4
5.000	5.00	0.000	14.4	5.00	0.000	14.4	5.00	0.000	14.4
10.010	10.01	0.000	14.4	10.01	0.000	14.4	10.01	0.000	14.4
20.040	20.04	0.000	14.4	20.04	0.000	14.4	20.04	0.000	14.4
50.090	50.09	0.000	14.4	50.09	0.000	14.4	50.09	0.000	14.4
100.100	100.10	0.000	14.4	100.10	0.000	14.4	100.10	0.000	14.4
150.000	150.01	0.010	14.4	150.01	0.010	14.4	150.01	0.010	14.4
200.000	200.01	0.010	14.4	200.01	0.010	14.4	200.01	0.010	14.4
300.000	300.01	0.010	14.4	300.01	0.010	14.4	300.01	0.010	14.4

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P -17 F03-00

Certificate/ULR No.: CC405825000000219F		Page 1 of 1	
Customer Details		Instrument Receipt Date	7-Jan-2025
M/S. AA ELECTRO MAGNETIC TEST LABORATORY PRIVATE LIMITED Plot No 174, Udyog Vihar-Phase 4, Sector 18, Gurgaon, Haryana		SRF No.	TNC/JC/24-25/356
		Date of Calibration	7-Jan-2025
		Next Calibration Date	6-Jan-2026
		Certificate Issue Date	8-Jan-2025
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Salt Spray Chamber	Make / Model No.	Sunrise Scientific Instrument
Range	0 - 50 °C, 0 to 99 % RH	Instrument Condition	SATISFACTORY
Least Count	0.1°C, 1% RH	Location of Calibration	At Laboratory
Serial No.	-	Location	Roof Area
Instrument ID.	AAEMT/ELEC/214	Parameter	Thermal
ENVIRONMENTAL CONDITION	Temperature:- 24±2 °C,Relative Humidity: - 55±10%	Calibration Reference Standard	DKD –R5-7
		Calibration Procedure	TNC-05/T/01

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
RH & Temperature Sensor with Indicator	HTD1220240058	CC388324000000003F	CC-3883	11/04/2025

Calibration Results (Temperature)

Sr.No.	UUC Value (°C)	Avg. of Standard Observed Value (°C)	Deviation (°C)	Uncertainty at k=2 at 95 % Confidence Level (±) (°C)
1	20.0	20.21	0.21	0.48
2	25.0	25.23	0.23	0.48
3	30.0	30.31	0.31	0.49
4	40.0	40.35	0.35	0.51
5	50.0	50.42	0.42	0.55

Calibration Result (Humidity) @ 25°C

Sr.No.	UUC Value (%RH)	Avg. of Standard Observed Value (%RH)	Deviation (%RH)	Uncertainty at k=2 at 95 % Confidence Level (±) (%RH)
1	10	10.05	0.05	1.8
2	30	30.16	0.16	1.8
3	50	50.23	0.23	1.8
4	70	70.29	0.29	1.8
5	95	95.40	0.40	1.8

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC405825000000220F		Page 1 of 1	
Customer Details		Instrument Receipt Date	7-Jan-2025
M/S. AA ELECTRO MAGNETIC TEST LABORATORY PRIVATE LIMITED Plot No 174, Udyog Vihar-Phase 4, Sector 18, Gurgaon, Haryana		SRF No.	TNC/JC/24-25/356
		Date of Calibration	7-Jan-2025
		Next Calibration Date	6-Jan-2026
		Certificate Issue Date	8-Jan-2025
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Digital Tachometer	Make / Model No.	Lutron/DT-2234C
Range	5 to 99999 rpm	Instrument Condition	SATISFACTORY
Least Count	0.1, 1 rpm	Location of Calibration	At Site
Serial No.	-	Location	Electrical & Electronic Sectic
Instrument ID.	AAEMT/ELEC/199	Parameter	Speed
ENVIRONMENTAL CONDITION	Temperature: 25±2°C, Relative Humidity: 50±10%	Calibration Reference Standard	IS: 12508
		Calibration Procedure	TNC -02-RPM 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Dig. Tachometer	TNC/RPM/C/002/19	CC380824000000839F	CC-3808 (REALTECH)	30/09/2026

Calibration Results

Sl. No.	UUC Value (rpm)	Std. Value (rpm)	Error (rpm)	Uncertainty at k=2 ,± %
1	100.0	100.1	-0.1	4.24
2	500.0	500.2	-0.2	3.40
3	5000	5001	-1	1.62
4	15000	15003	-3	1.61
5	35000	35006	-6	1.61

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC405825000000221F		Page 1 of 1	
Customer Details		Instrument Receipt Date	7-Jan-2025
M/S. AA ELECTRO MAGNETIC TEST LABORATORY PRIVATE LIMITED Plot No 174, Udyog Vihar-Phase 4, Sector 18, Gurgaon, Haryana		SRF No.	TNC/JC/24-25/356
		Date of Calibration	7-Jan-2025
		Next Calibration Date	6-Jan-2026
		Certificate Issue Date	8-Jan-2025
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Standard Resistance Box	Make / Model No.	Sigma/-
Range	As per Specification	Instrument Condition	SATISFACTORY
Least Count	As per Range	Location of Calibration	At Laboratory
Serial No.	-	Location	-
Instrument ID.	AAEMT/ELEC/061	Parameter	Electro Technical
ENVIRONMENTAL CONDITION	Temperature:- 24±2°C ,Relative Humidity: - 50±10%	Calibration Reference Standard	IS:1248-2003
		Calibration Procedure	TNC -03-WI 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
HV Megohm Box (20M to 200G) ohm	TNC/ET/C/02/18	CC223124000052518F	CC-2231	13/05/2025

Resistance: Range

	UOM	Std. Value	Observed Value on UUC	Uncertainty at k=2 ±%
	MΩ	20	19.436	0.58
	MΩ	200	199.162	1.42
	MΩ	2000	1999.19	5.77

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC405825000000747F		Page 1 of 1	
Customer Details		Instrument Receipt Date	1-Feb-2025
M/S. AA ELECTRO MAGNETIC TEST LABORATORY PRIVATE LIMITED Plot No 174, Udyog Vihar-Phase 4, Sector 18, Gurgaon, Haryana		SRF No.	TNC/JC/24-25/402
		Date of Calibration	1-Feb-2025
		Next Calibration Date	31-Jan-2026
		Certificate Issue Date	3-Feb-2025
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Digital Sound Level Meter	Make / Model No.	Chauvin Arnoux/CA 832
Range	30 to 130 dB	Instrument Condition	SATISFACTORY
Least Count	0.1 dB	Location of Calibration	At Laboratory
Serial No.	-	Location	-
Instrument ID.	AAEMT/ELEC/005	Parameter	Acoustic
ENVIRONMENTAL CONDITION	Temperature-24±2°C RH-50±10%	Calibration Reference Standard	-
		Calibration Procedure	TNC -05/A/01

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Sound Level Calibrator	TNC/A/C/001/18	CC286424000001608F	ECOTECH	10/05/2025

Calibration Results

Std. Value (dB)	UUC Value (dB)	Deviation (dB)	Uncertainty at k=2 ,± dB
94	94.1	0.1	1.28
114	114.2	0.2	1.28

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC405825000000756F		Page 1 of 1	
Customer Details		Instrument Receipt Date	21-Jan-2025
M/S. AA ELECTRO MAGNETIC TEST LABORATORY PRIVATE LIMITED Plot No 174, Udyog Vihar-Phase 4, Sector 18, Gurgaon, Haryana		SRF No.	TNC/JC/24-25/406
		Date of Calibration	21-Jan-2025
		Next Calibration Date	20-Jan-2026
		Certificate Issue Date	22-Jan-2025
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	RPM Meter & Pressure Gauge	Make / Model No.	Kawili H Guru/-
Range	(Spraying & Spacing Machine) 0 to 999999 rpm	Instrument Condition	SATISFACTORY
Least Count	1 rpm	Location of Calibration	At Site
Serial No.	-	Location	IP Section
Instrument ID.	AAMET/ELEC/117	Parameter	Speed
ENVIRONMENTAL CONDITION	Temperature: 25±2°C, Relative Humidity: 50±10%	Calibration Reference Standard	IS: 12508
		Calibration Procedure	TNC -02-RPM 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Dig. Tachometer	TNC/RPM/C/002/19	CC380824000000839F	CC-3808 (REALTECH)	30/09/2026

Calibration Results

Sl. No.	UUC Value (rpm)	Std. Value (rpm)	Error (rpm)	Uncertainty at k=2 ,± %
1	100.0	100.1	-0.1	4.24
2	500.0	500.2	-0.2	3.40
3	5000	5001	-1	1.62
4	15000	15003	-3	1.61
5	35000	35006	-6	1.61

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC405825000000757F		Page 1 of 1	
Customer Details		Instrument Receipt Date	21-Jan-2025
M/S. AA ELECTRO MAGNETIC TEST LABORATORY PRIVATE LIMITED Plot No 174, Udyog Vihar-Phase 4, Sector 18, Gurgaon, Haryana		SRF No.	TNC/JC/24-25/406
		Date of Calibration	21-Jan-2025
		Next Calibration Date	20-Jan-2026
		Certificate Issue Date	22-Jan-2025
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	RPM Meter & Pressure Gauge	Make / Model No.	Kawili H Guru/-
Range	(Spraying & Spacing Machine) 0 to 7 kg/cm ²	Instrument Condition	SATISFACTORY
Least Count	0.2 kg/cm ²	Location of Calibration	At Laboratory
Serial No.	-	Location	IP Section
Instrument ID.	AAMET/ELEC/117	Parameter	Pressure
ENVIRONMENTAL CONDITION	Temperature-24±2°C Relative Humidity-55±10%	Calibration Reference Standard	DKD-R6-1
		Calibration Procedure	TNC-05/P/01

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Digital Pressure Gauge (0 to 70 bar)	62182203	CC282824000003214F	CC-2828	11/03/2025

Calibration Results

Sr.No.	UUC Value (kg/cm ²)	Average of Std. Readings (bar)	Average of Std. Readings (kg/cm ²)	Deviation (kg/cm ²)	Uncertainty at k=2 at 95 % Confidence Level (±) bar
1	0	0.000	0.000	0.000	0
2	1	0.983	1.002	-0.002	0.12
3	2	1.968	2.007	-0.007	0.12
4	3	2.955	3.013	-0.013	0.12
5	5	4.922	5.019	-0.019	0.12
6	7	6.889	7.025	-0.025	0.12

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
6. Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC405825000000758F		Page 1 of 1	
Customer Details		Instrument Receipt Date	21-Jan-2025
M/S. AA ELECTRO MAGNETIC TEST LABORATORY PRIVATE LIMITED Plot No 174, Udyog Vihar-Phase 4, Sector 18, Gurgaon, Haryana		SRF No.	TNC/JC/24-25/406
		Date of Calibration	21-Jan-2025
		Next Calibration Date	20-Jan-2026
		Certificate Issue Date	22-Jan-2025
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	DC Regulated Power Supply	Make / Model No.	Junke/JK20050K
Range	0 to 200 VDC/0 to 50ADC	Instrument Condition	SATISFACTORY
Least Count	1 VDC/0.1ADC	Location of Calibration	At Laboratory
Serial No.	-	Location	-
Instrument ID.	AAMET/ELEC/219	Parameter	Electro Technical
ENVIRONMENTAL CONDITION	Temperature:- 24±2°C ,Relative Humidity: - 50±10%	Calibration Reference Standard	IS:1248-2003
		Calibration Procedure	TNC -03-WI 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Multifunction Calibrator	TNC/Cal/ET/015/017	CC236424000000927F	CC-2364 (Acument Measurements)	24/07/2025

DC Voltage

Sl. No	Range	Range	UOM	Observed Value on UUC	Observed Value on Standard	Uncertainty at k=2 ±%
4	200 V	50V	V	50	49.997	0.00
5		100V	V	100	99.947	0.06
6		150V	V	150	498.998	0.00
7		175V	V	175	174.996	0.00
8		200V	V	200	199.993	0.00

DC Current

Sl. No	Range	Range	UOM	Observed Value on UUC	Observed Value on Standard	Uncertainty at k=2 ±%
1	50A	1A	A	1	0.4512	0.00
2		3A	A	3	2.7625	0.22
3		5A	A	5	4.9641	0.00
4		7A	A	7	6.9984	0.00
5		10A	A	10	9.9853	0.00

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC405825000000759F		Page 1 of 1	
Customer Details		Instrument Receipt Date	21-Jan-2025
M/S. AA ELECTRO MAGNETIC TEST LABORATORY PRIVATE LIMITED Plot No 174, Udyog Vihar-Phase 4, Sector 18, Gurgaon, Haryana		SRF No.	TNC/JC/24-25/406
		Date of Calibration	21-Jan-2025
		Next Calibration Date	20-Jan-2026
		Certificate Issue Date	22-Jan-2025
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	DC Regulated Power Supply	Make / Model No.	Junke/JK20050K
Range	0 to 200 VDC/0 to 50ADC	Instrument Condition	SATISFACTORY
Least Count	1 VDC/0.1ADC	Location of Calibration	At Laboratory
Serial No.	-	Location	Electrical and Selectronic Se
Instrument ID.	AAMET/ELEC/220	Parameter	Electro Technical
ENVIRONMENTAL CONDITION	Temperature:- 24±2°C ,Relative Humidity: - 50±10%	Calibration Reference Standard	IS:1248-2003
		Calibration Procedure	TNC -03-WI 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Multifunction Calibrator	TNC/Cal/ET/015/017	CC236424000000927F	CC-2364 (Acument Measurements)	24/07/2025

DC Voltage

Sl. No	Range	Range	UOM	Observed Value on UUC	Observed Value on Standard	Uncertainty at k=2 ±%
4	200 V	50V	V	50	49.996	0.00
5		100V	V	100	99.946	0.06
6		150V	V	150	498.997	0.00
7		175V	V	175	174.995	0.00
8		200V	V	200	199.993	0.00

DC Current

Sl. No	Range	Range	UOM	Observed Value on UUC	Observed Value on Standard	Uncertainty at k=2 ±%
1	50A	1A	A	1	0.4514	0.00
2		3A	A	3	2.7626	0.22
3		5A	A	5	4.9643	0.00
4		7A	A	7	6.9985	0.00
5		10A	A	10	9.9854	0.00

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC405825000000760F		Page 1 of 1	
Customer Details		Instrument Receipt Date	21-Jan-2025
M/S. AA ELECTRO MAGNETIC TEST LABORATORY PRIVATE LIMITED Plot No 174, Udyog Vihar-Phase 4, Sector 18, Gurgaon, Haryana		SRF No.	TNC/JC/24-25/406
		Date of Calibration	21-Jan-2025
		Next Calibration Date	20-Jan-2026
		Certificate Issue Date	22-Jan-2025
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	DC Electronic Load	Make / Model No.	Jaritul/JT6321A
Range	0 to 150 VDC/0 to 120ADC	Instrument Condition	SATISFACTORY
Least Count	0.001 VDC/0.001ADC	Location of Calibration	At Laboratory
Serial No.	-	Location	Electrical and Selectronic Se
Instrument ID.	AAMET/ELEC/230	Parameter	Electro Technical
ENVIRONMENTAL CONDITION	Temperature:- 24±2°C ,Relative Humidity: - 50±10%	Calibration Reference Standard	IS:1248-2003
		Calibration Procedure	TNC -03-WI 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Multifunction Calibrator	TNC/Cal/ET/015/017	CC236424000000927F	CC-2364 (Acument Measurements)	24/07/2025

DC Voltage

Sl. No	Range	Range	UOM	Observed Value on UUC	Observed Value on Standard	Uncertainty at k=2 ±%
4	150V	20V	V	20	19.996	0.00
5		50V	V	50	49.946	0.06
6		80V	V	80	79.997	0.00
7		110V	V	110	109.995	0.00
8		150V	V	150	149.993	0.00

DC Current

Sl. No	Range	Range	UOM	Observed Value on UUC	Observed Value on Standard	Uncertainty at k=2 ±%
1	120A	1A	A	1	0.4514	0.00
2		3A	A	3	2.7626	0.22
3		5A	A	5	4.9643	0.00
4		7A	A	7	6.9985	0.00
5		10A	A	10	9.9854	0.00

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC405825000000761F		Page 1 of 1	
Customer Details		Instrument Receipt Date	21-Jan-2025
M/S. AA ELECTRO MAGNETIC TEST LABORATORY PRIVATE LIMITED Plot No 174, Udyog Vihar-Phase 4, Sector 18, Gurgaon, Haryana		SRF No.	TNC/JC/24-25/406
		Date of Calibration	21-Jan-2025
		Next Calibration Date	20-Jan-2026
		Certificate Issue Date	22-Jan-2025
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	DC Electronic Load	Make / Model No.	Itech/IT 8511A+
Range	0 to 150 VDC/0 to 30ADC	Instrument Condition	SATISFACTORY
Least Count	0.001 VDC/0.001ADC	Location of Calibration	At Laboratory
Serial No.	-	Location	Electrical and Selectronic Se
Instrument ID.	AAMET/ELEC/231	Parameter	Electro Technical
ENVIRONMENTAL CONDITION	Temperature:- 24±2°C ,Relative Humidity: - 50±10%	Calibration Reference Standard	IS:1248-2003
		Calibration Procedure	TNC -03-WI 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Multifunction Calibrator	TNC/Cal/ET/015/017	CC236424000000927F	CC-2364 (Acument Measurements)	24/07/2025

DC Voltage

Sl. No	Range	Range	UOM	Observed Value on UUC	Observed Value on Standard	Uncertainty at k=2 ±%
4	150V	20V	V	20	19.997	0.00
5		50V	V	50	49.946	0.06
6		80V	V	80	79.998	0.00
7		110V	V	110	109.997	0.00
8		150V	V	150	149.994	0.00

DC Current

Sl. No	Range	Range	UOM	Observed Value on UUC	Observed Value on Standard	Uncertainty at k=2 ±%
1	30A	1A	A	1	0.4515	0.00
2		3A	A	3	2.7628	0.22
3		5A	A	5	4.9645	0.00
4		7A	A	7	6.9986	0.00
5		10A	A	10	9.9856	0.00

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC405825000000868F		Page 1 of 1	
Customer Details		Instrument Receipt Date	4-Feb-2025
M/S. AA ELECTRO MAGNETIC TEST LABORATORY PRIVATE LIMITED Plot No 174, Udyog Vihar-Phase 4, Sector 18, Gurgaon, Haryana		SRF No.	TNC/JC/24-25/424
		Date of Calibration	4-Feb-2025
		Next Calibration Date	3-Feb-2026
		Certificate Issue Date	6-Feb-2025
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Digital Sound Level Meter	Make / Model No.	Chauvin Arnoux/CA 832
Range	30 to 130 dB	Instrument Condition	SATISFACTORY
Least Count	0.1 dB	Location of Calibration	At Laboratory
Serial No.	-	Location	-
Instrument ID.	AAEMT/ELEC/005	Parameter	Acoustic
ENVIRONMENTAL CONDITION	Temperature-24±2°C RH-50±10%	Calibration Reference Standard	-
		Calibration Procedure	TNC -05/A/01

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Sound Level Calibrator	TNC/A/C/001/18	CC286424000001608F	ECOTECH	10/05/2025

Calibration Results

Std. Value (dB)	UUC Value (dB)	Deviation (dB)	Uncertainty at k=2 ,± dB
94	94.1	0.1	1.28
114	114.2	0.2	1.28

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC405825000000870F		Page 1 of 1	
Customer Details		Instrument Receipt Date	4-Feb-2025
M/S. AA ELECTRO MAGNETIC TEST LABORATORY PRIVATE LIMITED Plot No 174, Udyog Vihar-Phase 4, Sector 18, Gurgaon, Haryana		SRF No.	TNC/JC/24-25/424
		Date of Calibration	4-Feb-2025
		Next Calibration Date	3-Feb-2026
		Certificate Issue Date	6-Feb-2025
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Digital Earth Resistance Tester	Make / Model No.	Tinsley/-
Range	20Ω to 2000Ω	Instrument Condition	SATISFACTORY
Least Count	0.01 Ω	Location of Calibration	At Laboratory
Serial No.	-	Location	-
Instrument ID.	AAEMT/ELEC/003	Parameter	Electro Technical
ENVIRONMENTAL CONDITION	Temperature:- 24±2 °C,Relative Humidity: - 55±10%	Calibration Reference Standard	IS:1248-2003
		Calibration Procedure	TNC -03-WI 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
6 ½ Digit DMM	TNC/ET/C/001/15	AIPL/220125/2-1	Anshankan	25/01/2026

Resistance

Sl. No	Range	Range	UOM	Observed Value on UUC	Observed Value on Standard	Uncertainty at k=2 ±%
1	20	20Ω	Ω	20	19.99	5.77
2	100	100Ω	Ω	100	99.98	5.77
3	500	500Ω	Ω	500	499.96	5.77
4	1000	1000Ω	kΩ	1000	999.95	5.77
5	2000	2000Ω	kΩ	2000	1999.93	5.77

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.:	CC405825000000871F	Page 1 of 1
----------------------	--------------------	-------------

Customer Details	Instrument Receipt Date	4-Feb-2025
M/S. AA ELECTRO MAGNETIC TEST LABORATORY PRIVATE LIMITED Plot No 174, Udyog Vihar-Phase 4, Sector 18, Gurgaon, Haryana	SRF No.	TNC/JC/24-25/424
	Date of Calibration	4-Feb-2025
	Next Calibration Date	3-Feb-2026
	Certificate Issue Date	6-Feb-2025

A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Cord Real	Make / Model No.	Sunrise Scientific Instrument
Range	10 to 99999 rpm	Instrument Condition	SATISFACTORY
Least Count	0.1, 1 rpm	Location of Calibration	At Laboratory
Serial No.	-	Location	-
Instrument ID.	AAEMT/ELEC/063	Parameter	Speed
ENVIRONMENTAL CONDITION	Temperature: 25±2°C, Relative Humidity: 50±10%	Calibration Reference Standard	IS: 12508
		Calibration Procedure	TNC -02-RPM 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Dig. Tachometer	TNC/RPM/C/002/19	CC380824000000839F	CC-3808 (REALTECH)	30/09/2026

Calibration Results

Sl. No.	UUC Value (rpm)	Std. Value (rpm)	Error (rpm)	Uncertainty at k=2 ,± %
1	100.0	100.1	-0.1	4.24
2	500.0	500.2	-0.2	3.40
3	5000	5001	-1	1.62
4	15000	15003	-3	1.61
5	35000	35006	-6	1.61

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.:	CC405825000000872F	Page 1 of 2
----------------------	--------------------	-------------

Customer Details	Instrument Receipt Date	4-Feb-2025
M/S. AA ELECTRO MAGNETIC TEST LABORATORY PRIVATE LIMITED Plot No 174, Udyog Vihar-Phase 4, Sector 18, Gurgaon, Haryana	SRF No.	TNC/JC/24-25/424
	Date of Calibration	4-Feb-2025
	Next Calibration Date	3-Feb-2026
	Certificate Issue Date	6-Feb-2025

A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Digital Multimeter	Make / Model No.	Fluke/15B+
Range	As per Manual	Instrument Condition	SATISFACTORY
Least Count	As per Range	Location of Calibration	At Laboratory
Serial No.	-	Location	-
Instrument ID.	AAEMT/ELEC/216	Parameter	Electro Technical
ENVIRONMENTAL CONDITION	Temperature:- 24±2 °C,Relative Humidity: - 55±10%	Calibration Reference Standard	IS:1248-2003
		Calibration Procedure	TNC -03-WI 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
6 ½ Digit DMM	TNC/ET/C/001/15	AIPL/220125/2-1	Anshankan	25/01/2026

Calibration Results(AC Current @ 50Hz

Sl. No	Range	Range	UOM	Observed Value on UUC	Observed Value on Standard	Uncertainty at k=2 ±%
1	100µA	10µA	µA	10	9.9964	0.90
2		50µA	µA	50	49.9973	0.90
3		99µA	µA	99	98.9955	0.90
4	100mA	10mA	mA	10	9.99858	0.30
5		50mA	mA	50	49.9828	0.30
6		99mA	mA	99	98.9917	0.30
7	3A	1A	A	1	0.999782	0.30
8		2A	A	2	1.99972	0.30
10	10A	5A	A	5	4.99791	0.30
11		9A	A	9	8.9945	0.30

AC Voltage @50Hz

Sl. No	Range	Range	UOM	Observed Value on UUC	Observed Value on Standard	Uncertainty at k=2 ±%
1	4V	2V	V	2	1.99960	0.10
2		4V	V	4	3.99878	0.10
3	40V	20V	V	20	19.9953	0.10
4		40V	V	40	39.9885	0.10
5	400V	200V	V	200	199.934	0.10
6		400V	V	400	399.874	0.10
7	1000V	500V	V	500	499.818	0.10
8		900V	V	900	899.763	0.10

DC Current

Sl. No	Range	Range	UOM	Observed Value on UUC	Observed Value on Standard	Uncertainty at k=2 ±%
--------	-------	-------	-----	-----------------------	----------------------------	-----------------------



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC405825000000872F Page 2 of 2

10		9A	A	9	8.99934	0.26
----	--	----	---	---	---------	------

DC Voltage

Sl. No	Range	Range	UOM	Observed Value on UUC	Observed Value on Standard	Uncertainty at k=2 ±%
1	4V	2V	V	2	2.00014	0.21
2		4V	V	4	4.00024	0.21
3	40V	20V	V	20	20.0011	0.21
4		40V	V	40	40.0042	0.21
5	400V	200V	V	200	200.015	0.21
6		400V	V	400	400.047	0.21
7	1000V	500V	V	500	499.983	0.21
8		900V	V	900	899.942	0.21

Resistance

Sl. No	Range	Range	UOM	Observed Value on UUC	Observed Value on Standard	Uncertainty at k=2 ±%
1	400	400Ω	Ω	400	399.9928	0.06
2	4	4kΩ	kΩ	4	3.999830	0.00
3	40	40kΩ	kΩ	40	39.99931	0.01
4	400	400kΩ	kΩ	400	399.9925	0.06
5	4	4MΩ	MΩ	4	3.999790	0.00
6	40	40MΩ	MΩ	40	39.99873	0.01

Frequency

Sl. No	Range	Range	UOM	Observed Value on UUC	Observed Value on Standard	Uncertainty at k=2 ±%
1	20Hz	10Hz	Hz	10	10.00012	0.82
2		19Hz	Hz	19	18.00015	0.82
3	100Hz	50Hz	Hz	50	50.0012	0.82
4		99Hz	Hz	99	99.0016	0.82
5	1kHz	0.5kHz	kHz	0.5	0.5000014	0.82

Capacitance

Sl. No	Range	Range	UOM	Observed Value on UUC	Observed Value on Standard	Uncertainty at k=2 ±%
1	5μF	1μF	μF	1	0.998	4.13
2		4μF	μF	4	3.997	4.13
3	50μF	10μF	μF	10	10.02	4.13
4		49μF	μF	49	49.03	4.13
5	100μF	70μF	μF	70	69.98	4.13
6		99μF	μF	99	98.8	4.13

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC405825000000873F		Page 1 of 1	
Customer Details		Instrument Receipt Date	4-Feb-2025
M/S. AA ELECTRO MAGNETIC TEST LABORATORY PRIVATE LIMITED Plot No 174, Udyog Vihar-Phase 4, Sector 18, Gurgaon, Haryana		SRF No.	TNC/JC/24-25/424
		Date of Calibration	4-Feb-2025
		Next Calibration Date	3-Feb-2026
		Certificate Issue Date	6-Feb-2025
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Pink Noise Generator	Make / Model No.	DongJui/DF1681
Range	30 to 130 dB	Instrument Condition	SATISFACTORY
Least Count	0.1 dB	Location of Calibration	At Laboratory
Serial No.	-	Location	-
Instrument ID.	AAEMT/ELEC/079	Parameter	Acoustic
ENVIRONMENTAL CONDITION	Temperature-24±2°C RH-50±10%	Calibration Reference Standard	-
		Calibration Procedure	TNC -05/A/01

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Sound Level Calibrator	TNC/A/C/001/18	CC286424000001608F	ECOTECH	10/05/2025

Calibration Results

Std. Value (dB)	UUC Value (dB)	Deviation (dB)	Uncertainty at k=2 ,± dB
94	94.0	0.0	1.28
114	114.2	0.2	1.28

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By