

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC205524000002900F		Page 1 of 1	
Customer Details		Instrument Receipt Date	16-Jul-2024
M/S AA Electro Magnetic Test Laboratory Private Limited Plot No 174, Udyog Vihar Phase 4, Sector 18, Gurgaon,, Haryana		SRF No.	TNC/JC/24-25/158
		Date of Calibration	16-Jul-2024
		Next Calibration Date	15-Jul-2025
		Certificate Issue Date	17-Jul-2024
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Computerized Battery Force Discharge	Make / Model No.	Map Electronics/-
Range	5V	Instrument Condition	OK
Least Count	0.001V	Location of Calibration	At Laboratory
Serial No.	-	Location	Battery Section, 4th Floor
Instrument ID.	AAEMT/ELEC/247	Parameter	Electro Technical
ENVIRONMENTAL CONDITION	Temperature:- 24±2 °C,Relative Humidity: - 55±10%	Calibration Reference Standard	IS:1248-2003
		Calibration Procedure	TNC -03-WI 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
6 ½ Digit DMM	TNC/ET/C/001/15	AIPL/300124/3-2	CC-2020 (ANSHANKAN)	31/01/2025
Multifunction Calibrator	TNC/Cal/ET/015/017	CC205523000000426F	TESTNCAL	12/08/2024

AC Voltage @50Hz

Sl. No	Range	UOM	Observed Value on UUC	Observed Value on Standard	Uncertainty at k=2 ±%
1	1V	V	1.000	0.999915	0.00
2	2V	V	2.000	1.99972	0.00
3	3V	V	3.000	3.99820	0.00
4	4V	V	4.000	3.99866	0.00
5	5V	V	5.000	4.99979	0.00

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC205524000002901F		Page 1 of 1	
Customer Details		Instrument Receipt Date	16-Jul-2024
M/S AA Electro Magnetic Test Laboratory Private Limited Plot No 174, Udyog Vihar Phase 4, Sector 18, Gurgaon,, Haryana		SRF No.	TNC/JC/24-25/158
		Date of Calibration	16-Jul-2024
		Next Calibration Date	15-Jul-2025
		Certificate Issue Date	17-Jul-2024
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Computerized Battery Force	Make / Model No.	Map Electronics/-
Range	Charge/Discharge Tester	Instrument Condition	OK
Least Count	0.001V	Location of Calibration	At Laboratory
Serial No.	-	Location	Battery Section
Instrument ID.	AAEMT/ELEC/277	Parameter	Electro Technical
ENVIRONMENTAL CONDITION	Temperature:- 24±2 °C,Relative Humidity: - 55±10%	Calibration Reference Standard	IS:1248-2003
		Calibration Procedure	TNC -03-WI 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
6 ½ Digit DMM	TNC/ET/C/001/15	AIPL/300124/3-2	CC-2020 (ANSHANKAN)	31/01/2025
Multifunction Calibrator	TNC/Cal/ET/015/017	CC205523000000426F	TESTNCAL	12/08/2024

AC Voltage @50Hz

Sl. No	Range	UOM	Observed Value on UUC	Observed Value on Standard	Uncertainty at k=2 ±%
1	1V	V	1.000	0.998675	0.00
2	2V	V	2.000	1.99572	0.00
3	3V	V	3.000	3.99420	0.00
4	4V	V	4.000	3.99466	0.00
5	5V	V	5.000	4.99379	0.01

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC205524000002902F		Page 1 of 1	
Customer Details		Instrument Receipt Date	16-Jul-2024
M/S AA Electro Magnetic Test Laboratory Private Limited Plot No 174, Udyog Vihar Phase 4, Sector 18, Gurgaon,, Haryana		SRF No.	TNC/JC/24-25/158
		Date of Calibration	16-Jul-2024
		Next Calibration Date	15-Jul-2025
		Certificate Issue Date	17-Jul-2024
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Computerized Battery Force	Make / Model No.	Map Electronics/-
Range	Charge/Discharge Tester 22V	Instrument Condition	OK
Least Count	0.001V	Location of Calibration	At Laboratory
Serial No.	-	Location	Battery Section
Instrument ID.	AAEMT/ELEC/273	Parameter	Electro Technical
ENVIRONMENTAL CONDITION	Temperature:- 24±2 °C,Relative Humidity: - 55±10%	Calibration Reference Standard	IS:1248-2003
		Calibration Procedure	TNC -03-WI 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
6 ½ Digit DMM	TNC/ET/C/001/15	AIPL/300124/3-2	CC-2020 (ANSHANKAN)	31/01/2025
Multifunction Calibrator	TNC/Cal/ET/015/017	CC205523000000426F	TESTNCAL	12/08/2024

AC Voltage @50Hz

Sl. No	Range	UOM	Observed Value on UUC	Observed Value on Standard	Uncertainty at k=2 ±%
1	5V	V	5.000	4.99979	0.00
2	10V	V	10.000	9.99979	0.00
3	15V	V	15.000	14.9981	0.00
4	20V	V	20.000	19.9945	0.01
5	22V	V	22.000	21.9521	0.06

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC205524000002903F		Page 1 of 1	
Customer Details		Instrument Receipt Date	16-Jul-2024
M/S AA Electro Magnetic Test Laboratory Private Limited Plot No 174, Udyog Vihar Phase 4, Sector 18, Gurgaon,, Haryana		SRF No.	TNC/JC/24-25/158
		Date of Calibration	16-Jul-2024
		Next Calibration Date	15-Jul-2025
		Certificate Issue Date	17-Jul-2024
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Computerized Battery Force	Make / Model No.	Map Electronics/-
Range	Charge/Discharge Tester 22V	Instrument Condition	OK
Least Count	0.001V	Location of Calibration	At Laboratory
Serial No.	-	Location	Battery Section
Instrument ID.	AAEMT/ELEC/272	Parameter	Electro Technical
ENVIRONMENTAL CONDITION	Temperature:- 24±2 °C,Relative Humidity: - 55±10%	Calibration Reference Standard	IS:1248-2003
		Calibration Procedure	TNC -03-WI 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
6 ½ Digit DMM	TNC/ET/C/001/15	AIPL/300124/3-2	CC-2020 (ANSHANKAN)	31/01/2025
Multifunction Calibrator	TNC/Cal/ET/015/017	CC205523000000426F	TESTNCAL	12/08/2024

AC Voltage @50Hz

Sl. No	Range	UOM	Observed Value on UUC	Observed Value on Standard	Uncertainty at k=2 ±%
1	5V	V	5.000	4.99579	0.00
2	10V	V	10.000	9.98943	0.01
3	15V	V	15.000	14.9901	0.01
4	20V	V	20.000	19.9545	0.05
5	22V	V	22.000	21.8721	0.07

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC205524000002904F		Page 1 of 1	
Customer Details		Instrument Receipt Date	16-Jul-2024
M/S AA Electro Magnetic Test Laboratory Private Limited Plot No 174, Udyog Vihar Phase 4, Sector 18, Gurgaon,, Haryana		SRF No.	TNC/JC/24-25/158
		Date of Calibration	16-Jul-2024
		Next Calibration Date	15-Jul-2025
		Certificate Issue Date	17-Jul-2024
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Computerized Battery Force	Make / Model No.	Map Electronics/-
Range	Charge/Discharge Tester 22V	Instrument Condition	OK
Least Count	0.001V	Location of Calibration	At Laboratory
Serial No.	-	Location	Battery Section
Instrument ID.	AAEMT/ELEC/274	Parameter	Electro Technical
ENVIRONMENTAL CONDITION	Temperature:- 24±2 °C,Relative Humidity: - 55±10%	Calibration Reference Standard	IS:1248-2003
		Calibration Procedure	TNC -03-WI 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
6 ½ Digit DMM	TNC/ET/C/001/15	AIPL/300124/3-2	CC-2020 (ANSHANKAN)	31/01/2025
Multifunction Calibrator	TNC/Cal/ET/015/017	CC205523000000426F	TESTNCAL	12/08/2024

AC Voltage @50Hz

Sl. No	Range	UOM	Observed Value on UUC	Observed Value on Standard	Uncertainty at k=2 ±%
1	5V	V	5.000	4.93579	0.08
2	10V	V	10.000	9.94943	0.05
3	15V	V	15.000	14.8681	0.16
4	20V	V	20.000	19.9145	0.07
5	22V	V	22.000	21.6921	0.28

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC205524000002905F		Page 1 of 1	
Customer Details		Instrument Receipt Date	16-Jul-2024
M/S AA Electro Magnetic Test Laboratory Private Limited Plot No 174, Udyog Vihar Phase 4, Sector 18, Gurgaon,, Haryana		SRF No.	TNC/JC/24-25/158
		Date of Calibration	16-Jul-2024
		Next Calibration Date	15-Jul-2025
		Certificate Issue Date	17-Jul-2024
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Computerized Battery Force	Make / Model No.	Map Electronics/-
Range	Charge/Discharge Tester 22V	Instrument Condition	OK
Least Count	0.001V	Location of Calibration	At Laboratory
Serial No.	-	Location	Battery Section
Instrument ID.	AAEMT/ELEC/275	Parameter	Electro Technical
ENVIRONMENTAL CONDITION	Temperature:- 24±2 °C,Relative Humidity: - 55±10%	Calibration Reference Standard	IS:1248-2003
		Calibration Procedure	TNC -03-WI 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
6 ½ Digit DMM	TNC/ET/C/001/15	AIPL/300124/3-2	CC-2020 (ANSHANKAN)	31/01/2025
Multifunction Calibrator	TNC/Cal/ET/015/017	CC205523000000426F	TESTNCAL	12/08/2024

AC Voltage @50Hz

Sl. No	Range	UOM	Observed Value on UUC	Observed Value on Standard	Uncertainty at k=2 ±%
1	5V	V	5.000	4.93979	0.08
2	10V	V	10.000	9.95343	0.06
3	15V	V	15.000	14.8641	0.16
4	20V	V	20.000	19.8745	0.11
5	22V	V	22.000	21.8301	0.10

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC205524000002906F		Page 1 of 1	
Customer Details		Instrument Receipt Date	16-Jul-2024
M/S AA Electro Magnetic Test Laboratory Private Limited Plot No 174, Udyog Vihar Phase 4, Sector 18, Gurgaon,, Haryana		SRF No.	TNC/JC/24-25/158
		Date of Calibration	16-Jul-2024
		Next Calibration Date	15-Jul-2025
		Certificate Issue Date	17-Jul-2024
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Computerized Battery Force	Make / Model No.	Map Electronics/-
Range	Charge/Discharge Tester 22V	Instrument Condition	OK
Least Count	0.001V	Location of Calibration	At Site
Serial No.	-	Location	Battery Section
Instrument ID.	AAEMT/ELEC/276	Parameter	Electro Technical
ENVIRONMENTAL CONDITION	Temperature:- 24±2 °C,Relative Humidity: - 55±10%	Calibration Reference Standard	IS:1248-2003
		Calibration Procedure	TNC -03-WI 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
6 ½ Digit DMM	TNC/ET/C/001/15	AIPL/300124/3-2	CC-2020 (ANSHANKAN)	31/01/2025
Multifunction Calibrator	TNC/Cal/ET/015/017	CC205523000000426F	TESTNCAL	12/08/2024

AC Voltage @50Hz

Sl. No	Range	UOM	Observed Value on UUC	Observed Value on Standard	Uncertainty at k=2 ±%
1	5V	V	5.000	4.99979	0.00
2	10V	V	10.000	9.95743	0.05
3	15V	V	15.000	14.9281	0.13
4	20V	V	20.000	19.9345	0.08
5	22V	V	22.000	21.9501	0.05

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC205524000002907F		Page 1 of 1	
Customer Details		Instrument Receipt Date	16-Jul-2024
M/S AA Electro Magnetic Test Laboratory Private Limited Plot No 174, Udyog Vihar Phase 4, Sector 18, Gurgaon,, Haryana		SRF No.	TNC/JC/24-25/158
		Date of Calibration	16-Jul-2024
		Next Calibration Date	15-Jul-2025
		Certificate Issue Date	17-Jul-2024
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Computerized Battery Force	Make / Model No.	Map Electronics/-
Range	Charge/Discharge Tester	Instrument Condition	OK
Least Count	0.001V	Location of Calibration	At Site
Serial No.	-	Location	Battery Section
Instrument ID.	AAEMT/ELEC/243	Parameter	Electro Technical
ENVIRONMENTAL CONDITION	Temperature:- 24±2 °C,Relative Humidity: - 55±10%	Calibration Reference Standard	IS:1248-2003
		Calibration Procedure	TNC -03-WI 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
6 ½ Digit DMM	TNC/ET/C/001/15	AIPL/300124/3-2	CC-2020 (ANSHANKAN)	31/01/2025
Multifunction Calibrator	TNC/Cal/ET/015/017	CC205523000000426F	TESTNCAL	12/08/2024

AC Voltage @50Hz

Sl. No	Range	UOM	Observed Value on UUC	Observed Value on Standard	Uncertainty at k=2 ±%
1	1V	V	1.000	0.999915	0.00
2	3V	V	3.000	3.99972	0.00
3	5V	V	5.000	4.99820	0.00
4	6V	V	6.000	5.99866	0.00
5	7V	V	7.000	6.99979	0.00

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC205524000002908F		Page 1 of 1	
Customer Details		Instrument Receipt Date	16-Jul-2024
M/S AA Electro Magnetic Test Laboratory Private Limited Plot No 174, Udyog Vihar Phase 4, Sector 18, Gurgaon,, Haryana		SRF No.	TNC/JC/24-25/158
		Date of Calibration	16-Jul-2024
		Next Calibration Date	15-Jul-2025
		Certificate Issue Date	17-Jul-2024
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Banding Moment Test Apparatus	Make / Model No.	Sharadatronics/SI/BMT/163
Range	0 to 9.99 kgf	Instrument Condition	OK
Least Count	0.001 kgf	Location of Calibration	At Site
Serial No.	016	Location	-
Instrument ID.	AAEMT/ELEC/132	Parameter	Force
ENVIRONMENTAL CONDITION	Temperature-24±2°C Relative Humidity: - 50±10%	Calibration Reference Standard	VDI/VDE 2624 Part 2.1 (Dec
		Calibration Procedure	TNC -03-F 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Push pull Gauge Calibrator	TNC/CAL/F/019-023/18	AMC/240016/05	CC-2364	07/01/2025
Push Pull Gauge Calibrator	TNC/CAL/F/008-011/18	AMC/240016/02	CC-2364	07/01/2025
Push Pull Gauge Calibrator	TNC/CAL/F/031-036/18	AMC/240016/07	CC-2364	07/01/2025
Push Pull Gauge Calibrator	TNC/CAL/F/004-007/18	AMC/240016/01	CC-2364	07/01/2025
Push Pull Gauge Calibrator	TNC/CAL/F/015-018/18	AMC/240016/04	CC-2364	07/01/2025
Push Pull Gauge Calibrator	TNC/CAL/F/024-030/18	AMC/240016/06	CC-2364	07/01/2025

Calibration Results

Mode	Load (N)	Load (kgf)	Reading at Increase (kgf)	Reading at Decrease (kgf)	Uncertainty at k=2 ± %	Mode	Load (N)	Reading at Increase (kgf)	Reading at Decrease (kgf)	Uncertainty at k=2 ± %
Push	0	0.000	0.000	0.000	-	Pull	0	0.000	0.000	-
	20	2.039	2.039	2.039	0.61		20	2.039	2.039	0.61
	40	4.079	4.079	4.079	0.61		40	4.079	4.079	0.61
	60	6.118	6.119	6.119	0.61		60	6.119	6.120	0.61
	80	8.158	8.160	8.161	0.61		80	8.161	8.162	0.61
	90	9.177	9.179	9.179	0.61		90	9.178	9.179	0.61
Stdev at half load (PUSH)					Stdev at half load (PULL)					
0.06					0.06					

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncl Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P -17 F03-00

Certificate/ULR No.: CC205524000002909F		Page 1 of 1	
Customer Details		Instrument Receipt Date	16-Jul-2024
M/S AA Electro Magnetic Test Laboratory Private Limited Plot No 174, Udyog Vihar Phase 4, Sector 18, Gurgaon,, Haryana		SRF No.	TNC/JC/24-25/158
		Date of Calibration	16-Jul-2024
		Next Calibration Date	15-Jul-2025
		Certificate Issue Date	17-Jul-2024
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Environment Chamber	Make / Model No.	Warmmax/-
Range	-70 to 215 °C, 20 to 99 % RH	Instrument Condition	OK
Least Count	0.1°C, 0.1% RH	Location of Calibration	At Site
Serial No.	-	Location	-
Instrument ID.	AAEMT/ELEC/215	Parameter	Thermal
ENVIRONMENTAL CONDITION	Temperature:- 24±2 °C,Relative Humidity: - 55±10%	Calibration Reference Standard	----
		Calibration Procedure	TNC-05/T/01

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
RH & Temperature Sensor with Indicator	HTD1220240058	CC388324000000003F	CC-3883	11/04/2025
RTD Sensor with 6 1/2 Digital Multimeter	TNC/T/C/005/17	CC317224000000216F	CC-3172 (RK LABS)	23/01/2025

Calibration Results (Temperature)

Sr.No.	UUC Value (°C)	Avg. of Standard Observed Value (°C)	Deviation (°C)	Uncertainty at k=2 at 95 % Confidence Level (±) (°C)
1	-50.0	-50.15	-0.15	0.34
2	0.0	0.24	0.24	0.48
3	100.0	100.35	0.35	0.48
4	200.0	200.41	0.41	0.48
5	215.0	215.47	0.47	0.48

Calibration Result (Humidity) @ 25°C

Sr.No.	UUC Value (%RH)	Avg. of Standard Observed Value (%RH)	Deviation (%RH)	Uncertainty at k=2 at 95 % Confidence Level (±) (%RH)
1	20	20.01	0.01	1.8
2	30	30.03	0.03	1.8
3	40	40.07	0.07	1.8
4	50	50.10	0.10	1.8
5	80	80.16	0.16	1.8
6	90	90.22	0.22	1.8

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC205524000002910F		Page 1 of 1	
Customer Details		Instrument Receipt Date	16-Jul-2024
M/S AA Electro Magnetic Test Laboratory Private Limited Plot No 174, Udyog Vihar Phase 4, Sector 18, Gurgaon,, Haryana		SRF No.	TNC/JC/24-25/158
		Date of Calibration	16-Jul-2024
		Next Calibration Date	15-Jul-2025
		Certificate Issue Date	17-Jul-2024
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	S Channel CC/CV Charger	Make / Model No.	Map Electronic/-
Range	5V DC & 5A DC	Instrument Condition	OK
Least Count	0.01V DC & 0.01 A DC	Location of Calibration	At Site
Serial No.	-	Location	-
Instrument ID.	AAEMT/ELEC/268	Parameter	Electro Technical
ENVIRONMENTAL CONDITION	Temperature:- 24±2 °C,Relative Humidity: - 55±10%	Calibration Reference Standard	IS:1248-2003
		Calibration Procedure	TNC -03-WI 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
6 ½ Digit DMM	TNC/ET/C/001/15	AIPL/300124/3-2	CC-2020 (ANSHANKAN)	31/01/2025
Multifunction Calibrator	TNC/Cal/ET/015/017	CC205523000000426F	TESTNCAL	12/08/2024

DC Current

Sl. No	Range	UOM	Observed Value on UUC	Observed Value on Standard	Uncertainty at k=2 ±%
1	1A	A	1.00	1.00034	0.26
2	2A	A	2.00	2.00027	0.26
3	3A	A	3.00	3.00137	0.26
4	4A	A	4.00	4.01330	0.26
5	5A	A	5.00	5.13136	0.26

DC Voltage

Sl. No	Range	UOM	Observed Value on UUC	Observed Value on Standard	Uncertainty at k=2 ±%
1	1V	V	1.00	0.999979	0.21
2	2V	V	2.00	1.89974	0.21
3	3V	V	3.00	4.99998	0.21
4	4V	V	4.00	3.89345	0.21
5	5V	V	5.00	4.99966	0.21

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC205524000002911F		Page 1 of 1	
Customer Details		Instrument Receipt Date	16-Jul-2024
M/S AA Electro Magnetic Test Laboratory Private Limited Plot No 174, Udyog Vihar Phase 4, Sector 18, Gurgaon,, Haryana		SRF No.	TNC/JC/24-25/158
		Date of Calibration	16-Jul-2024
		Next Calibration Date	15-Jul-2025
		Certificate Issue Date	17-Jul-2024
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	S Channel CC/CV Charger	Make / Model No.	Map Electronic/-
Range	5V DC & 5A DC	Instrument Condition	OK
Least Count	0.01V DC & 0.01 A DC	Location of Calibration	At Site
Serial No.	-	Location	-
Instrument ID.	AAEMT/ELEC/270	Parameter	Electro Technical
ENVIRONMENTAL CONDITION	Temperature:- 24±2 °C,Relative Humidity: - 55±10%	Calibration Reference Standard	IS:1248-2003
		Calibration Procedure	TNC -03-WI 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
6 ½ Digit DMM	TNC/ET/C/001/15	AIPL/300124/3-2	CC-2020 (ANSHANKAN)	31/01/2025
Multifunction Calibrator	TNC/Cal/ET/015/017	CC205523000000426F	TESTNCAL	12/08/2024

DC Current

Sl. No	Range	UOM	Observed Value on UUC	Observed Value on Standard	Uncertainty at k=2 ±%
1	1A	A	1.00	1.00074	0.26
2	2A	A	2.00	2.00065	0.26
3	3A	A	3.00	3.00177	0.26
4	4A	A	4.00	4.01450	0.26
5	5A	A	5.00	5.15336	0.26

DC Voltage

Sl. No	Range	UOM	Observed Value on UUC	Observed Value on Standard	Uncertainty at k=2 ±%
1	1V	V	1.00	0.999779	0.21
2	2V	V	2.00	1.90174	0.21
3	3V	V	3.00	4.99978	0.21
4	4V	V	4.00	3.89145	0.21
5	5V	V	5.00	4.99566	0.21

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC205524000002912F		Page 1 of 1	
Customer Details		Instrument Receipt Date	16-Jul-2024
M/S AA Electro Magnetic Test Laboratory Private Limited Plot No 174, Udyog Vihar Phase 4, Sector 18, Gurgaon,, Haryana		SRF No.	TNC/JC/24-25/158
		Date of Calibration	17-Jul-2024
		Next Calibration Date	16-Jul-2025
		Certificate Issue Date	18-Jul-2024
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Coincell Internal AC Resistance Tester	Make / Model No.	Map Electronics/-
Range	10Ω	Instrument Condition	OK
Least Count	As per Range	Location of Calibration	At Laboratory
Serial No.		Location	--
Instrument ID.	AAEMT/ELEC/246	Parameter	Electro Technical
ENVIRONMENTAL CONDITION	Temperature:- 24±2 °C,Relative Humidity: - 55±10%	Calibration Reference Standard	IS:1248-2003
		Calibration Procedure	TNC -03-WI 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Digital Micro Ohm Meter	CES201AUTO/20922/01/ 18	CC243824000000232F	CC-2438 (BANSAL)	31/01/2025

Resistance

Sl. No	Range	Range	UOM	Observed Value on UUC	Observed Value on Standard	Uncertainty at k=2 ±%
1	10	10mΩ	mΩ	10	9.94	5.77
2	1	1Ω	Ω	1	0.998	0.58
3	5	5Ω	Ω	5	4.995	0.58
4	7	7Ω	Ω	7	9.991	0.58
5	10	10Ω	Ω	10	9.89	5.77

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC205524000002913F		Page 1 of 1	
Customer Details		Instrument Receipt Date	16-Jul-2024
M/S AA Electro Magnetic Test Laboratory Private Limited Plot No 174, Udyog Vihar Phase 4, Sector 18, Gurgaon,, Haryana		SRF No.	TNC/JC/24-25/158
		Date of Calibration	16-Jul-2024
		Next Calibration Date	15-Jul-2025
		Certificate Issue Date	17-Jul-2024
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	S Channel CC/CV Charger	Make / Model No.	Map Electronic/-
Range	5V DC & 5A DC	Instrument Condition	OK
Least Count	0.01V DC & 0.01 A DC	Location of Calibration	At Site
Serial No.	-	Location	-
Instrument ID.	AAEMT/ELEC/271	Parameter	Electro Technical
ENVIRONMENTAL CONDITION	Temperature:- 24±2 °C,Relative Humidity: - 55±10%	Calibration Reference Standard	IS:1248-2003
		Calibration Procedure	TNC -03-WI 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
6 ½ Digit DMM	TNC/ET/C/001/15	AIPL/300124/3-2	CC-2020 (ANSHANKAN)	31/01/2025
Multifunction Calibrator	TNC/Cal/ET/015/017	CC205523000000426F	TESTNCAL	12/08/2024

DC Current

Sl. No	Range	UOM	Observed Value on UUC	Observed Value on Standard	Uncertainty at k=2 ±%
1	1A	A	1.00	1.00154	0.26
2	2A	A	2.00	2.00185	0.26
3	3A	A	3.00	3.00237	0.26
4	4A	A	4.00	4.01670	0.26
5	5A	A	5.00	5.16536	0.26

DC Voltage

Sl. No	Range	UOM	Observed Value on UUC	Observed Value on Standard	Uncertainty at k=2 ±%
1	1V	V	1.00	0.999979	0.21
2	2V	V	2.00	1.90374	0.21
3	3V	V	3.00	4.99958	0.21
4	4V	V	4.00	3.89125	0.21
5	5V	V	5.00	4.99526	0.21

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC205524000002914F		Page 1 of 1	
Customer Details		Instrument Receipt Date	16-Jul-2024
M/S AA Electro Magnetic Test Laboratory Private Limited Plot No 174, Udyog Vihar Phase 4, Sector 18, Gurgaon,, Haryana		SRF No.	TNC/JC/24-25/158
		Date of Calibration	16-Jul-2024
		Next Calibration Date	15-Jul-2025
		Certificate Issue Date	17-Jul-2024
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	S Channel CC/CV Charger	Make / Model No.	Map Electronic/-
Range	5V DC & 5A DC	Instrument Condition	OK
Least Count	0.01V DC & 0.01 A DC	Location of Calibration	At Site
Serial No.	-	Location	-
Instrument ID.	AAEMT/ELEC/269	Parameter	Electro Technical
ENVIRONMENTAL CONDITION	Temperature:- 24±2 °C,Relative Humidity: - 55±10%	Calibration Reference Standard	IS:1248-2003
		Calibration Procedure	TNC -03-WI 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
6 ½ Digit DMM	TNC/ET/C/001/15	AIPL/300124/3-2	CC-2020 (ANSHANKAN)	31/01/2025
Multifunction Calibrator	TNC/Cal/ET/015/017	CC205523000000426F	TESTNCAL	12/08/2024

DC Current

Sl. No	Range	UOM	Observed Value on UUC	Observed Value on Standard	Uncertainty at k=2 ±%
1	1A	A	1.00	1.00554	0.26
2	2A	A	2.00	2.00585	0.26
3	3A	A	3.00	3.00637	0.26
4	4A	A	4.00	4.03670	0.26
5	5A	A	5.00	5.14936	0.26

DC Voltage

Sl. No	Range	UOM	Observed Value on UUC	Observed Value on Standard	Uncertainty at k=2 ±%
1	1V	V	1.00	0.999579	0.21
2	2V	V	2.00	1.90354	0.21
3	3V	V	3.00	4.99758	0.21
4	4V	V	4.00	3.88925	0.21
5	5V	V	5.00	4.95326	0.21

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC205524000002915F		Page 1 of 1	
Customer Details		Instrument Receipt Date	16-Jul-2024
M/S AA Electro Magnetic Test Laboratory Private Limited Plot No 174, Udyog Vihar Phase 4, Sector 18, Gurgaon,, Haryana		SRF No.	TNC/JC/24-25/158
		Date of Calibration	16-Jul-2024
		Next Calibration Date	15-Jul-2025
		Certificate Issue Date	17-Jul-2024
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	S Channel CC/CV Charger	Make / Model No.	Map Electronic/-
Range	5V DC & 5A DC	Instrument Condition	OK
Least Count	0.01V DC & 0.01 A DC	Location of Calibration	At Site
Serial No.	-	Location	-
Instrument ID.	AAEMT/ELEC/267	Parameter	Electro Technical
ENVIRONMENTAL CONDITION	Temperature:- 24±2 °C,Relative Humidity: - 55±10%	Calibration Reference Standard	IS:1248-2003
		Calibration Procedure	TNC -03-WI 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
6 ½ Digit DMM	TNC/ET/C/001/15	AIPL/300124/3-2	CC-2020 (ANSHANKAN)	31/01/2025
Multifunction Calibrator	TNC/Cal/ET/015/017	CC205523000000426F	TESTNCAL	12/08/2024

DC Current

Sl. No	Range	UOM	Observed Value on UUC	Observed Value on Standard	Uncertainty at k=2 ±%
1	1A	A	1.00	1.00754	0.26
2	2A	A	2.00	2.00665	0.26
3	3A	A	3.00	3.00717	0.26
4	4A	A	4.00	4.04270	0.26
5	5A	A	5.00	5.13136	0.26

DC Voltage

Sl. No	Range	UOM	Observed Value on UUC	Observed Value on Standard	Uncertainty at k=2 ±%
1	1V	V	1.00	0.998919	0.21
2	2V	V	2.00	1.90234	0.21
3	3V	V	3.00	4.99698	0.21
4	4V	V	4.00	3.88845	0.21
5	5V	V	5.00	4.95126	0.21

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P -17 F03-00

Certificate/ULR No.: CC205524000002918F		Page 1 of 1	
Customer Details		Instrument Receipt Date	16-Jul-2024
M/S AA Electro Magnetic Test Laboratory Private Limited Plot No 174, Udyog Vihar Phase 4, Sector 18, Gurgaon,, Haryana		SRF No.	TNC/JC/24-25/158
		Date of Calibration	16-Jul-2024
		Next Calibration Date	15-Jul-2025
		Certificate Issue Date	17-Jul-2024
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Tensile Testing Machine	Make / Model No.	Harriss & Tarriss/-
Range	2 kN	Instrument Condition	OK
Least Count	0.01 kg	Location of Calibration	At Site
Serial No.	-	Location	-
Instrument ID.	AAEMT/ELEC/248	Parameter	Force
ENVIRONMENTAL CONDITION	Temperature:- 24±2 °C,Relative Humidity: - 50±10	Calibration Reference Standard	IS 1828(Part1) :2022
		Calibration Procedure	TNC -01-F 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Load Cell With Indicator	1961/27490	CC352323000001542F	CC-3523	21/11/2026
Load Cell with Indicator	TNC/F/C/001/17	CC210122000000844F	DVG Laboratory	18/07/2024

Calibration Results (Tension Mode)

Force(N)	Division at Std. Temp. 24°C	Division at room Temp. 24.4°C	Observed Reading (Division s) 0°	Observed Reading (Division s) 120°	Observed Reading (Division s) 240°	Average of Reading (Division)	Relative accuracy error (q)%	Relative Repeatability error (b)%	Relative resolution Error (a)%	Relative Zero Error (fo)%	Expende d Uncertain ty of Measure ment ± %
100	9951	9951	10054	10055	10124	10077.67	1.27	0.69	0.01	0.00	1.02
200	19894	19893	20090	20091	20192	20124.33	1.16	0.51	0.00	0.00	1.02
400	39772	39771	40141	40142	40289	40190.67	1.06	0.37	0.00	0.00	1.02
500	49723	49722	50171	50172	50311	50218.00	1.00	0.28	0.00	0.00	1.02
1000	99470	99467	99495	99502	99512	99503.00	0.04	0.02	0.00	0.00	1.02
2000	2000	2000	2002	2001	2004	2002.33	0.12	0.15	0.00	0.00	1.02
0	0	0	0	0	0	0	0	0	0	0	0

Note:

- The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
- This report should not be reproduced except in full without our prior permission in writing.
- Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
- All our certificates are Traceable to national/International standard.
- This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P -17 F03-00

Certificate/ULR No.: CC205524000002919F		Page 1 of 1	
Customer Details		Instrument Receipt Date	16-Jul-2024
M/S AA Electro Magnetic Test Laboratory Private Limited Plot No 174, Udyog Vihar Phase 4, Sector 18, Gurgaon,, Haryana		SRF No.	TNC/JC/24-25/158
		Date of Calibration	16-Jul-2024
		Next Calibration Date	15-Jul-2025
		Certificate Issue Date	17-Jul-2024
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Stability Load Test	Make / Model No.	Sunrise Scientific Instrument
Range	3 kN	Instrument Condition	OK
Least Count	0.1 N	Location of Calibration	At Site
Serial No.	-	Location	-
Instrument ID.	AAEMT/ELEC/139	Parameter	Force
ENVIRONMENTAL CONDITION	Temperature:- 24±2 °C,Relative Humidity: - 50±10	Calibration Reference Standard	IS 1828(Part1) :2022
		Calibration Procedure	TNC -01-F 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Load Cell with Indicator	TNC/F/C/001/17	CC210122000000844F	DVG Laboratory	18/07/2024
Load Cell With Indicator	1961/27490	CC352323000001542F	CC-3523	21/11/2026

Calibration Results (Compression Mode)

Force(N)	Division at Std. Temp. 24°C	Division at room Temp. 24.4°C	Observed Reading (Division s) 0°	Observed Reading (Division s) 120°	Observed Reading (Division s) 240°	Average of Reading (Division)	Relative accuracy error (q)%	Relative Repeatability error (b)%	Relative resolution Error (a)%	Relative Zero Error (fo)%	Expanded Uncertainty of Measurement ± %
100	9931	9932	10053	10054	9989	10032.00	1.01	0.65	0.10	0.00	0.8
200	19884	19886	19893	20021	19924	19946.00	0.30	0.64	0.05	0.00	0.8
400	39791	39795	39821	39899	40012	39910.67	0.29	0.48	0.03	0.00	0.8
500	49744	49749	49789	49892	49924	49868.33	0.24	0.27	0.02	0.00	0.8
1000	99521	99532	99589	99621	99639	99616.33	0.08	0.05	0.01	0.00	0.8
3000	2999	2999	3012	3009	3007	3009.33	0.34	0.17	0.00	0.00	0.8
0	0	0	0	0	0	0	0	0	0	0	0

Note:

- The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
- This report should not be reproduced except in full without our prior permission in writing.
- Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
- All our certificates are Traceable to national/International standard.
- This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC205524000002920F		Page 1 of 1	
Customer Details		Instrument Receipt Date	16-Jul-2024
M/S AA Electro Magnetic Test Laboratory Private Limited Plot No 174, Udyog Vihar Phase 4, Sector 18, Gurgaon,, Haryana		SRF No.	TNC/JC/24-25/158
		Date of Calibration	16-Jul-2024
		Next Calibration Date	15-Jul-2025
		Certificate Issue Date	17-Jul-2024
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Dia ELeetric Strength Test	Make / Model No.	-/-
Range	1500 V	Instrument Condition	OK
Least Count	As per Range	Location of Calibration	At Site
Serial No.	-	Location	-
Instrument ID.	AAEMT/ELEC/032	Parameter	Electro Technical
ENVIRONMENTAL CONDITION	Temperature:- 24±2 °C,Relative Humidity: - 55±10%	Calibration Reference Standard	IS:1248-2003
		Calibration Procedure	TNC -03-WI 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
6 ½ Digit DMM	TNC/ET/C/001/15	AIPL/300124/3-2	CC-2020 (ANSHANKAN)	31/01/2025
Multifunction Calibrator	TNC/Cal/ET/015/017	CC205523000000426F	TESTNCAL	12/08/2024

AC Voltage @50Hz

Sl. No	Range	UOM	Observed Value on UUC	Observed Value on Standard	Uncertainty at k=2 ±%
1	100V	V	100	100.0074	0.00
2	300V	V	300	300.006	0.00
3	500V	V	500	500.011	0.00
4	700V	V	700	700.030	0.00
5	1000V	V	1000	999.045	0.00

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC205524000002921F		Page 1 of 1	
Customer Details		Instrument Receipt Date	16-Jul-2024
M/S AA Electro Magnetic Test Laboratory Private Limited Plot No 174, Udyog Vihar Phase 4, Sector 18, Gurgaon,, Haryana		SRF No.	TNC/JC/24-25/158
		Date of Calibration	16-Jul-2024
		Next Calibration Date	15-Jul-2025
		Certificate Issue Date	17-Jul-2024
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Computerized Battery Over Charge	Make / Model No.	Map Electronics/-
Range	40V	Instrument Condition	OK
Least Count	As per Range	Location of Calibration	At Site
Serial No.	-	Location	-
Instrument ID.	AAEMT/ELEC/242	Parameter	Electro Technical
ENVIRONMENTAL CONDITION	Temperature:- 24±2 °C,Relative Humidity: - 55±10%	Calibration Reference Standard	IS:1248-2003
		Calibration Procedure	TNC -03-WI 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
6 ½ Digit DMM	TNC/ET/C/001/15	AIPL/300124/3-2	CC-2020 (ANSHANKAN)	31/01/2025
Multifunction Calibrator	TNC/Cal/ET/015/017	CC205523000000426F	TESTNCAL	12/08/2024

AC Voltage @50Hz

Sl. No	Range	UOM	Observed Value on UUC	Observed Value on Standard	Uncertainty at k=2 ±%
1	10V	V	10	9.9984	0.00
2	20V	V	20	19.9739	0.04
3	30V	V	30	29.9596	0.04
4	40V	V	40	39.9950	0.01

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.:	CC205524000002922F	Page 1 of 1
----------------------	--------------------	-------------

Customer Details	Instrument Receipt Date	16-Jul-2024
M/S. AA ELECTRO MAGNETIC TEST LABORATORY PRIVATE LIMITED Plot No 174, Udyog Vihar-Phase 4, Sector 18, Gurgaon, Haryana	SRF No.	TNC/JC/24-25/159
	Date of Calibration	17-Jul-2024
	Next Calibration Date	16-Jul-2025
	Certificate Issue Date	18-Jul-2024

A) Identification On UUC (Unit Under Calibration)

Instrument Name	Cylindrical Weight	Make / Model No.	-/-
Range	5 kg	Instrument Condition	OK
Least Count	---	Location of Calibration	At Laboratory
Serial No.	-	Location	-
Instrument ID.	AAEMT/ELEC/263B	Parameter	---
ENVIRONMENTAL CONDITION	Temperature:- 24±2°C, Relative Humidity: - 50±10%	Calibration Reference Standard	OIML R-111
		Calibration Procedure	TNC-03-M 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Standard Weight(5kg to 20kg)	4911/WI/01/2018/009-F2	CC322624000000051F	CC-3226	01/02/2026

Weight

Sl. No	Nominal Load	Test weight mass (g)	Uncertainty of Measurement k = 2 at 95% confidence Level (± mg)
1	5kg	5000.083	0.42

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC205524000002923F		Page 1 of 1	
Customer Details		Instrument Receipt Date	16-Jul-2024
M/S. AA ELECTRO MAGNETIC TEST LABORATORY PRIVATE LIMITED Plot No 174, Udyog Vihar-Phase 4, Sector 18, Gurgaon, Haryana		SRF No.	TNC/JC/24-25/159
		Date of Calibration	17-Jul-2024
		Next Calibration Date	16-Jul-2025
		Certificate Issue Date	18-Jul-2024
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Touch Current Tester	Make / Model No.	-/-
Range	R1=375Ω, R2=500Ω, C1=0.22uF	Instrument Condition	OK
Least Count	As per Range	Location of Calibration	At Laboratory
Serial No.	-	Location	-
Instrument ID.	AAEMT/ELEC/167D	Parameter	Electro Technical
ENVIRONMENTAL CONDITION	Temperature:- 24±2 °C,Relative Humidity: - 55±10%	Calibration Reference Standard	IS:1248-2003
		Calibration Procedure	TNC -03-WI 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
6 ½ Digit DMM	TNC/ET/C/001/15	AIPL/300124/3-2	CC-2020 (ANSHANKAN)	31/01/2025
Multifunction Calibrator	TNC/Cal/ET/015/017	CC205523000000426F	TESTNCAL	12/08/2024

Resistance

Sl. No	Range	Range	UOM	Observed Value on UUC	Observed Value on Standard	Uncertainty at k=2 ±%
1	500	500Ω	Ω	500	499.9888	0.06

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.:	CC205524000002924F	Page 1 of 1
----------------------	--------------------	-------------

Customer Details	Instrument Receipt Date	16-Jul-2024
M/S. AA ELECTRO MAGNETIC TEST LABORATORY PRIVATE LIMITED Plot No 174, Udyog Vihar-Phase 4, Sector 18, Gurgaon, Haryana	SRF No.	TNC/JC/24-25/159
	Date of Calibration	17-Jul-2024
	Next Calibration Date	16-Jul-2025
	Certificate Issue Date	18-Jul-2024

A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Touch Current Tester	Make / Model No.	-/-
Range	R1=375Ω, R2=500Ω, C1=0.22uF	Instrument Condition	OK
Least Count	As per Range	Location of Calibration	At Laboratory
Serial No.	-	Location	-
Instrument ID.	AAEMT/ELEC/167A	Parameter	Electro Technical
ENVIRONMENTAL CONDITION	Temperature:- 24±2 °C,Relative Humidity: - 55±10%	Calibration Reference Standard	IS:1248-2003
		Calibration Procedure	TNC -03-WI 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
6 ½ Digit DMM	TNC/ET/C/001/15	AIPL/300124/3-2	CC-2020 (ANSHANKAN)	31/01/2025
Multifunction Calibrator	TNC/Cal/ET/015/017	CC205523000000426F	TESTNCAL	12/08/2024

Resistance

Sl. No	Range	Range	UOM	Observed Value on UUC	Observed Value on Standard	Uncertainty at k=2 ±%
1	500	500Ω	Ω	500	500.024	0.06

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC205524000002925F		Page 1 of 1	
Customer Details		Instrument Receipt Date	16-Jul-2024
M/S. AA ELECTRO MAGNETIC TEST LABORATORY PRIVATE LIMITED Plot No 174, Udyog Vihar-Phase 4, Sector 18, Gurgaon, Haryana		SRF No.	TNC/JC/24-25/159
		Date of Calibration	16-Jul-2024
		Next Calibration Date	15-Jul-2025
		Certificate Issue Date	17-Jul-2024
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Std. Weight	Make / Model No.	-/-
Range	15 kg	Instrument Condition	OK
Least Count	---	Location of Calibration	At Laboratory
Serial No.	-	Location	-
Instrument ID.	AAEMT/ELEC/166	Parameter	---
ENVIRONMENTAL CONDITION	Temperature:- 24±2°C, Relative Humidity: - 50±10%	Calibration Reference Standard	OIML R-111
		Calibration Procedure	TNC-03-M 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Standard Weight(5kg to 20kg)	4911/WI/01/2018/009-F2	CC322624000000051F	CC-3226	01/02/2026

Weight

Sl. No	Nominal Load	Test weight mass (g)	Uncertainty of Measurement k = 2 at 95% confidence Level (± mg)
1	15kg	15000.069	0.20

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC205524000002926F		Page 1 of 1	
Customer Details		Instrument Receipt Date	16-Jul-2024
M/S. AA ELECTRO MAGNETIC TEST LABORATORY PRIVATE LIMITED Plot No 174, Udyog Vihar-Phase 4, Sector 18, Gurgaon, Haryana		SRF No.	TNC/JC/24-25/159
		Date of Calibration	17-Jul-2024
		Next Calibration Date	16-Jul-2025
		Certificate Issue Date	18-Jul-2024
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Touch Current Tester	Make / Model No.	-/-
Range	R1=500Ω, R2=500Ω, R3=10kΩ, C1=0.22uF,	Instrument Condition	OK
Least Count	As per Range	Location of Calibration	At Laboratory
Serial No.	-	Location	-
Instrument ID.	AAEMT/ELEC/167C	Parameter	Electro Technical
ENVIRONMENTAL CONDITION	Temperature:- 24±2 °C, Relative Humidity: - 55±10%	Calibration Reference Standard	IS:1248-2003
		Calibration Procedure	TNC -03-WI 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
6 ½ Digit DMM	TNC/ET/C/001/15	AIPL/300124/3-2	CC-2020 (ANSHANKAN)	31/01/2025
Multifunction Calibrator	TNC/Cal/ET/015/017	CC205523000000426F	TESTNCAL	12/08/2024

Resistance

Sl. No	Range	Range	UOM	Observed Value on UUC	Observed Value on Standard	Uncertainty at k=2 ±%
1	500	500Ω	Ω	500	499.9838	0.06

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
6. Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC205524000002927F		Page 1 of 1	
Customer Details		Instrument Receipt Date	16-Jul-2024
M/S. AA ELECTRO MAGNETIC TEST LABORATORY PRIVATE LIMITED Plot No 174, Udyog Vihar-Phase 4, Sector 18, Gurgaon, Haryana		SRF No.	TNC/JC/24-25/159
		Date of Calibration	17-Jul-2024
		Next Calibration Date	16-Jul-2025
		Certificate Issue Date	18-Jul-2024
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Cylindrical Weight	Make / Model No.	-/-
Range	250 gm	Instrument Condition	OK
Least Count	---	Location of Calibration	At Laboratory
Serial No.	-	Location	-
Instrument ID.	AAEMT/ELEC/260	Parameter	Mass
ENVIRONMENTAL CONDITION	Temperature:- 24±2°C, Relative Humidity: - 55±10% RH	Calibration Reference Standard	OIML R-111
		Calibration Procedure	TNC-03-M 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Weights E1 Class (1 mg to 200 g)	TNC/M/C/004/24	WMCL/E/2024-01/2885	CC-2743	17/01/2026

Weight

Sl. No	Nominal Load (g)	Test weight mass (g)	Uncertainty of Measurement k = 2 at 95% confidence Level (± mg)
1	250g	250.0004	0.01

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.:	CC205524000002928F	Page 1 of 1
----------------------	--------------------	-------------

Customer Details	Instrument Receipt Date	16-Jul-2024
M/S. AA ELECTRO MAGNETIC TEST LABORATORY PRIVATE LIMITED Plot No 174, Udyog Vihar-Phase 4, Sector 18, Gurgaon, Haryana	SRF No.	TNC/JC/24-25/159
	Date of Calibration	17-Jul-2024
	Next Calibration Date	16-Jul-2025
	Certificate Issue Date	18-Jul-2024

A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Cylindrical Weight	Make / Model No.	-/-
Range	5 kg	Instrument Condition	OK
Least Count	---	Location of Calibration	At Laboratory
Serial No.	-	Location	-
Instrument ID.	AAEMT/ELEC/263A	Parameter	---
ENVIRONMENTAL CONDITION	Temperature:- 24±2°C, Relative Humidity: - 50±10%	Calibration Reference Standard	OIML R-111
		Calibration Procedure	TNC-03-M 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Standard Weight(5kg to 20kg)	4911/WI/01/2018/009-F2	CC322624000000051F	CC-3226	01/02/2026

Weight

Sl. No	Nominal Load	Test weight mass (g)	Uncertainty of Measurement k = 2 at 95% confidence Level (± mg)
1	5kg	5000.083	0.42

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.:	CC205524000002929F	Page 1 of 1
----------------------	--------------------	-------------

Customer Details	Instrument Receipt Date	16-Jul-2024
M/S. AA ELECTRO MAGNETIC TEST LABORATORY PRIVATE LIMITED Plot No 174, Udyog Vihar-Phase 4, Sector 18, Gurgaon, Haryana	SRF No.	TNC/JC/24-25/159
	Date of Calibration	16-Jul-2024
	Next Calibration Date	15-Jul-2025
	Certificate Issue Date	17-Jul-2024

A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Cylindrical Weight	Make / Model No.	-/-
Range	500 gm	Instrument Condition	OK
Least Count	---	Location of Calibration	At Laboratory
Serial No.	-	Location	-
Instrument ID.	AAEMT/ELEC/261	Parameter	---
ENVIRONMENTAL CONDITION	Temperature:- 24±2°C, Relative Humidity: - 50±10%	Calibration Reference Standard	OIML R-111
		Calibration Procedure	TNC-03-M 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Standard Weight (500g to 2kg)	WI/12/2017/018- F1	CC322624000000655F	CC-3226	18/03/2026

Weight

Sl. No	Nominal Load	Test weight mass (g)	Uncertainty of Measurement k = 2 at 95% confidence Level (± mg)
1	500g	500.0011	0.02

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC205524000002930F		Page 1 of 1	
Customer Details		Instrument Receipt Date	16-Jul-2024
M/S. AA ELECTRO MAGNETIC TEST LABORATORY PRIVATE LIMITED Plot No 174, Udyog Vihar-Phase 4, Sector 18, Gurgaon, Haryana		SRF No.	TNC/JC/24-25/159
		Date of Calibration	17-Jul-2024
		Next Calibration Date	16-Jul-2025
		Certificate Issue Date	18-Jul-2024
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Touch Current Tester	Make / Model No.	-/-
Range	2000 Ω	Instrument Condition	OK
Least Count	As per Range	Location of Calibration	At Laboratory
Serial No.		Location	-
Instrument ID.	AAEMT/ELEC/167B	Parameter	Electro Technical
ENVIRONMENTAL CONDITION	Temperature:- 24 ± 2 °C, Relative Humidity: - $55 \pm 10\%$	Calibration Reference Standard	IS:1248-2003
		Calibration Procedure	TNC -03-WI 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
6 ½ Digit DMM	TNC/ET/C/001/15	AIPL/300124/3-2	CC-2020 (ANSHANKAN)	31/01/2025
Multifunction Calibrator	TNC/Cal/ET/015/017	CC205523000000426F	TESTNCAL	12/08/2024

Resistance

Sl. No	Range	Range	UOM	Observed Value on UUC	Observed Value on Standard	Uncertainty at $k=2 \pm\%$
1	10	10 Ω	Ω	10	9.99916	0.01
2	100	100 Ω	Ω	100	99.9918	0.06
3	1	1k Ω	k Ω	1000	0.999968	0.00
4	2	2k Ω	k Ω	2000	1.99937	0.01

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with $k=2$



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.:	CC205524000002931F	Page 1 of 1
----------------------	--------------------	-------------

Customer Details	Instrument Receipt Date	16-Jul-2024
M/S. AA ELECTRO MAGNETIC TEST LABORATORY PRIVATE LIMITED Plot No 174, Udyog Vihar-Phase 4, Sector 18, Gurgaon, Haryana	SRF No.	TNC/JC/24-25/159
	Date of Calibration	17-Jul-2024
	Next Calibration Date	16-Jul-2025
	Certificate Issue Date	18-Jul-2024

A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Cylindrical Weight	Make / Model No.	-/-
Range	1.7 kg	Instrument Condition	OK
Least Count	---	Location of Calibration	At Laboratory
Serial No.	-	Location	-
Instrument ID.	AAEMT/ELEC/262	Parameter	---
ENVIRONMENTAL CONDITION	Temperature:- 24±2°C, Relative Humidity: - 50±10%	Calibration Reference Standard	OIML R-111
		Calibration Procedure	TNC-03-M 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Standard Weight (500g to 2kg)	WI/12/2017/018- F1	CC322624000000655F	CC-3226	18/03/2026

Weight

Sl. No	Nominal Load	Test weight mass (g)	Uncertainty of Measurement k = 2 at 95% confidence Level (± mg)
1	1.7kg	1700.0027	0.02

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC205524000002932F		Page 1 of 1	
Customer Details		Instrument Receipt Date	16-Jul-2024
M/S. AA ELECTRO MAGNETIC TEST LABORATORY PRIVATE LIMITED Plot No 174, Udyog Vihar-Phase 4, Sector 18, Gurgaon, Haryana		SRF No.	TNC/JC/24-25/159
		Date of Calibration	17-Jul-2024
		Next Calibration Date	16-Jul-2025
		Certificate Issue Date	18-Jul-2024
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Std. Weight	Make / Model No.	-/-
Range	5 kg	Instrument Condition	OK
Least Count	---	Location of Calibration	At Laboratory
Serial No.	-	Location	-
Instrument ID.	AAEMT/ELEC/185	Parameter	---
ENVIRONMENTAL CONDITION	Temperature:- 24±2°C, Relative Humidity: - 50±10%	Calibration Reference Standard	OIML R-111
		Calibration Procedure	TNC-03-M 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Standard Weight(5kg to 20kg)	4911/WI/01/2018/009-F2	CC322624000000051F	CC-3226	01/02/2026

Weight

Sl. No	Nominal Load	Test weight mass (g)	Uncertainty of Measurement k = 2 at 95% confidence Level (± mg)
1	5kg	5000.019	0.42

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC205524000002937F		Page 1 of 1	
Customer Details		Instrument Receipt Date	16-Jul-2024
M/S. AA ELECTRO MAGNETIC TEST LABORATORY PRIVATE LIMITED Plot No 174, Udyog Vihar-Phase 4, Sector 18, Gurgaon, Haryana		SRF No.	TNC/JC/24-25/159
		Date of Calibration	17-Jul-2024
		Next Calibration Date	16-Jul-2025
		Certificate Issue Date	18-Jul-2024
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Cylindrical Weight	Make / Model No.	-/-
Range	10 kg	Instrument Condition	OK
Least Count	---	Location of Calibration	At Laboratory
Serial No.	-	Location	-
Instrument ID.	AAEMT/ELEC/264	Parameter	---
ENVIRONMENTAL CONDITION	Temperature:- 24±2°C, Relative Humidity: - 50±10%	Calibration Reference Standard	OIML R-111
		Calibration Procedure	TNC-03-M 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Standard Weight(5kg to 20kg)	4911/WI/01/2018/009-F2	CC322624000000051F	CC-3226	01/02/2026

Weight

Sl. No	Nominal Load	Test weight mass (g)	Uncertainty of Measurement k = 2 at 95% confidence Level (± mg)
1	10kg	10000.069	0.20

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.: CC205524000002936F		Page 1 of 1	
Customer Details		Instrument Receipt Date	16-Jul-2024
M/S. AA ELECTRO MAGNETIC TEST LABORATORY PRIVATE LIMITED Plot No 174, Udyog Vihar-Phase 4, Sector 18, Gurgaon, Haryana		SRF No.	TNC/JC/24-25/159
		Date of Calibration	17-Jul-2024
		Next Calibration Date	16-Jul-2025
		Certificate Issue Date	18-Jul-2024
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Std. Weight	Make / Model No.	-/-
Range	1 kg	Instrument Condition	OK
Least Count	---	Location of Calibration	At Laboratory
Serial No.	-	Location	-
Instrument ID.	AAEMT/ELEC/196	Parameter	---
ENVIRONMENTAL CONDITION	Temperature:- 24±2°C, Relative Humidity: - 50±10%	Calibration Reference Standard	OIML R-111
		Calibration Procedure	TNC-03-M 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Standard Weight (500g to 2kg)	WI/12/2017/018- F1	CC322624000000655F	CC-3226	18/03/2026

Weight

Sl. No	Nominal Load	Test weight mass (g)	Uncertainty of Measurement k = 2 at 95% confidence Level (± mg)
1	1kg	999.9992	0.02

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.:	CC205524000002935F	Page 1 of 1
----------------------	--------------------	-------------

Customer Details	Instrument Receipt Date	16-Jul-2024
M/S. AA ELECTRO MAGNETIC TEST LABORATORY PRIVATE LIMITED Plot No 174, Udyog Vihar-Phase 4, Sector 18, Gurgaon, Haryana	SRF No.	TNC/JC/24-25/159
	Date of Calibration	17-Jul-2024
	Next Calibration Date	16-Jul-2025
	Certificate Issue Date	18-Jul-2024

A) Identification On UUC (Unit Under Calibration)

Instrument Name	Std. Weight	Make / Model No.	-/-
Range	1 kg	Instrument Condition	OK
Least Count	---	Location of Calibration	At Laboratory
Serial No.	-	Location	-
Instrument ID.	AAEMT/ELEC/182	Parameter	---
ENVIRONMENTAL CONDITION	Temperature:- 24±2°C, Relative Humidity: - 50±10%	Calibration Reference Standard	OIML R-111
		Calibration Procedure	TNC-03-M 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Standard Weight (500g to 2kg)	WI/12/2017/018- F1	CC322624000000655F	CC-3226	18/03/2026

Weight

Sl. No	Nominal Load	Test weight mass (g)	Uncertainty of Measurement k = 2 at 95% confidence Level (± mg)
1	1kg	1000.0057	0.06

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.:	CC205524000002938F	Page 1 of 1
----------------------	--------------------	-------------

Customer Details	Instrument Receipt Date	16-Jul-2024
M/S. AA ELECTRO MAGNETIC TEST LABORATORY PRIVATE LIMITED Plot No 174, Udyog Vihar-Phase 4, Sector 18, Gurgaon, Haryana	SRF No.	TNC/JC/24-25/159
	Date of Calibration	17-Jul-2024
	Next Calibration Date	16-Jul-2025
	Certificate Issue Date	18-Jul-2024

A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Std. Weight	Make / Model No.	-/-
Range	3 kg	Instrument Condition	OK
Least Count	---	Location of Calibration	At Laboratory
Serial No.	-	Location	-
Instrument ID.	AAEMT/ELEC/184	Parameter	---
ENVIRONMENTAL CONDITION	Temperature:- 24±2°C, Relative Humidity: - 50±10%	Calibration Reference Standard	OIML R-111
		Calibration Procedure	TNC-03-M 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Standard Weight (500g to 2kg)	WI/12/2017/018- F1	CC322624000000655F	CC-3226	18/03/2026

Weight

Sl. No	Nominal Load	Test weight mass (g)	Uncertainty of Measurement k = 2 at 95% confidence Level (± mg)
1	3kg	3000.0058	0.06

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.:	CC205524000002939F	Page 1 of 1
----------------------	--------------------	-------------

Customer Details	Instrument Receipt Date	16-Jul-2024
M/S. AA ELECTRO MAGNETIC TEST LABORATORY PRIVATE LIMITED Plot No 174, Udyog Vihar-Phase 4, Sector 18, Gurgaon, Haryana	SRF No.	TNC/JC/24-25/159
	Date of Calibration	17-Jul-2024
	Next Calibration Date	16-Jul-2025
	Certificate Issue Date	18-Jul-2024

A) Identification On UUC (Unit Under Calibration)

Instrument Name	Std. Weight	Make / Model No.	-/-
Range	2 kg	Instrument Condition	OK
Least Count	---	Location of Calibration	At Laboratory
Serial No.	-	Location	-
Instrument ID.	AAEMT/ELEC/183	Parameter	---
ENVIRONMENTAL CONDITION	Temperature:- 24±2°C, Relative Humidity: - 50±10%	Calibration Reference Standard	OIML R-111
		Calibration Procedure	TNC-03-M 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Standard Weight (500g to 2kg)	WI/12/2017/018- F1	CC322624000000655F	CC-3226	18/03/2026

Weight

Sl. No	Nominal Load	Test weight mass (g)	Uncertainty of Measurement k = 2 at 95% confidence Level (± mg)
1	2kg	2000.0078	0.06

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By

Calibration Certificate

P-17 F03-00

Certificate/ULR No.:	CC205524000002940F	Page 1 of 1
----------------------	--------------------	-------------

Customer Details	Instrument Receipt Date	16-Jul-2024
M/S. AA ELECTRO MAGNETIC TEST LABORATORY PRIVATE LIMITED Plot No 174, Udyog Vihar-Phase 4, Sector 18, Gurgaon, Haryana	SRF No.	TNC/JC/24-25/159
	Date of Calibration	17-Jul-2024
	Next Calibration Date	16-Jul-2025
	Certificate Issue Date	18-Jul-2024

A) Identification On UUC (Unit Under Calibration)

Instrument Name	Std. Weight	Make / Model No.	-/-
Range	500g	Instrument Condition	OK
Least Count	---	Location of Calibration	At Laboratory
Serial No.	-	Location	-
Instrument ID.	AAEMT/ELEC/168	Parameter	---
ENVIRONMENTAL CONDITION	Temperature:- 24±2°C, Relative Humidity: - 50±10%	Calibration Reference Standard	OIML R-111
		Calibration Procedure	TNC-03-M 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Certified By	Valid upto
Standard Weight (500g to 2kg)	WI/12/2017/018- F1	CC322624000000655F	CC-3226	18/03/2026

Weight

Sl. No	Nominal Load	Test weight mass (g)	Uncertainty of Measurement k = 2 at 95% confidence Level (± mg)
1	500g	500.0021	0.01

Note:

1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
4. All our certificates are Traceable to national/International standard.
5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only .
- 6.Measurement Uncertainty reported at approximate 95% confidence level with k=2



Calibrated By

Issued & Approved By